

Small-bore connectors for liquids and gases in healthcare applications — Part 20: Common test methods

Approved 2 December 2014 by
Association for the Advancement of Medical Instrumentation

Abstract: This standard specifies the TEST METHODS to support the functional requirements for SMALL-BORE CONNECTORS intended to be used for CONNECTIONS of MEDICAL DEVICES and related ACCESSORIES.

Published by

Association for the Advancement of Medical Instrumentation
4301 N. Fairfax Drive, Suite 301
Arlington, VA 22203-1633
www.aami.org

© 2014 by the Association for the Advancement of Medical Instrumentation

All Rights Reserved

This publication is subject to copyright claims of ISO, ANSI, and AAMI. No part of this publication may be reproduced or distributed in any form, including an electronic retrieval system, without the prior written permission of AAMI. All requests pertaining to this document should be submitted to AAMI. It is illegal under federal law (17 U.S.C. § 101, *et seq.*) to make copies of all or any part of this document (whether internally or externally) without the prior written permission of the Association for the Advancement of Medical Instrumentation. Violators risk legal action, including civil and criminal penalties, and damages of \$100,000 per offense. For permission regarding the use of all or any part of this document, complete the reprint request form at www.aami.org or contact AAMI, 4301 N. Fairfax Drive, Suite 301, Arlington, VA 22203-1633. Phone: +1-703-525-4890; Fax: +1-703-525-1067.

Printed in the United States of America

ISBN 1-57020-564-7

Contents

Page

| | |
|--|------------|
| Committee representation | iv |
| Background of AAMI/CN20(PS):2014 | vi |
| Foreword | vii |
| Introduction | ix |
| 1 Scope | 1 |
| 2 Normative references | 1 |
| 3 Terms and definitions | 1 |
| 4 TEST METHODS for SMALL-BORE CONNECTORS | 2 |
| Annex A (informative) Rationale and Guidance | 3 |
| Annex B (normative) * Leakage by pressure decay TEST METHOD | 6 |
| Annex C (normative) Leakage by pressure decay TEST METHOD | 9 |
| Annex D (normative) Subatmospheric-pressure air leakage TEST METHOD | 11 |
| Annex E (normative) Stress cracking TEST METHOD | 16 |
| Annex F (normative) Resistance to separation from axial load TEST METHOD | 18 |
| Annex G (normative) Resistance to separation from unscrewing TEST METHOD | 20 |
| Annex H (normative) Resistance to overriding TEST METHOD | 22 |
| Annex I (normative) Disconnection by unscrewing TEST METHOD | 24 |
| Annex J (normative) Modification of the TEST METHODS to generate variable data for statistical analysis | 26 |
| Bibliography | 29 |
| Terminology - Alphabetized index of defined terms | 30 |

Committee representation

Association for the Advancement of Medical Instrumentation

Quality Management and Corresponding General Aspects for Medical Devices Committee

The publication of AAMI/CN20(PS) as a new provisional American National Standard was initiated by the AAMI Quality Management and Corresponding General Aspects for Medical Devices Committee, which also functions as a U.S. Technical Advisory Group to the relevant work in the International Organization for Standardization (ISO). U.S. representatives from the AAMI Small Bore Connectors Committee (U.S. Sub-TAG for ISO/TC 210/JWG 04), chaired by Scott Colburn of FDA and Brad Noe of Becton Dickinson & Co. played an active part in developing the Final Draft International Standard 80369-20, upon which this Provisional American National Standard is based.

At the time this document was published, the **AAMI Quality Management and Corresponding General Aspects for Medical Devices Committee** had the following members:

Cochairs: Scott A. Colburn, MS, BSN, RN, FDA/CDRH
Charles B. Sidebottom, PE, PPO Standards LLC

Members: Jeffrey L. Eggleston, MS PE, Covidien
Mizanu Kebede, Kimberly-Clark Corporation
Ed R. Kimmelman, BME, JD, Kimmelman Consultancy
Kristi M. Kistner, RAC ASQ CBA, Amgen Inc
Dan Labelle, Nonin Medical Inc
Sherry M. Leichtweis, Abbott Laboratories
Emmanuel O. Nyakako, Zimmer Inc
David G. Osborn, Philips Electronics North America
Robert Sestrick, GE Healthcare
Mike Silvestri, Terumo Americas Corporate
Al Van Houdt, Spacelabs Medical Inc

Alternates: Ujjal Chakravartty, Kimberly-Clark Corporation
Richard Gardner, GE Healthcare
David J. Geraghty, Spacelabs Medical Inc
Robert Rabeh Hijazi, MS MHA CBE, St Louis VA Medical Center - John Cochran Division
Mike Hudon, Philips Electronics North America
Luann M. Pendy, Medtronic, Inc. WHQ Campus
Kimberly A. Trautman, FDA/CDRH

Small-bore Connectors Committee

At the time this document was published, the **AAMI Small-bore Connectors Committee** had the following members:

Cochairs: Scott A. Colburn, MS, BSN, RN, FDA/CDRH
Brad Noe, Becton Dickinson & Company

Members: Steve J. Bernard, Nestle HealthCare Nutrition Inc
Jim Brown, Colder Products Company
James Brugger, BSME MEEM, NxStage Medical Inc
Edwin L. Burnard, B Braun of America Inc
David Carr, ASQ CQA CMI, Teleflex Medical
Richard Chan, CCP, North Shore University Hospital
Conor Curtin, Fresenius Medical Care
Jim Dahl, Spacelabs Medical Inc
Michael Dennis, Abbvie
Bruce A Friedman, D.Eng, GE Healthcare
Carrie L. Gavriloff, MSN RN Ed, Akron Childrens Hospital
J. C. Gerancher, III, MD, Wake Forest University Baptist Medical Center
Cathie Gosnell, RN MS MBA, Premier Healthcare Alliance
Krisanne Graves

Peggi Guenter, PhD RN CNSN, American Society for Parenteral and Enteral Nutrition
Stephanne Hale, RN BSN MBA MHA, Novation LLC
Tom Hancock, Global Enteral Device Supplies Association
Brian K. Heston, Zimmer Inc
Trevor C. Huang, PhD MBA, Medtronic Inc WHQ Campus
Rory Jaffe, MD, California Hospital Patient Safety Organization (CHPSO)
Steven C. Jepson, Baxter Healthcare Corporation
David W. Johnson, Kimberly-Clark Corporation
Crystal Koelper, Corpak MedSystems
John J. Kropczynski, Jr., BS MBA, Abbott Laboratories
Tony Lair, NeoMed Inc
Lee Leichter, P/L Biomedical
Michael W. Maryan, Cook Inc
Ravi Narayanan, Value Plastics Inc
William O'Neill, Smiths Medical
David G. Osborn, Philips Electronics North America
Tricia Leilani Otstot, RN
Shailendra Parihar, Ph.D., Johnson & Johnson
James H. Philip, CCE, Brigham & Womens Hospital
David Quinn, Welch Allyn Inc
Glenda S. Rodgers, RNC, Heartland Regional Medical Center
Ben Rush, Hospira Worldwide Inc
Wayne Schuessler, Covidien
Debora Simmons, RN MSN CCRN, University of Texas MD Anderson Cancer Center
Courtney Smith, Blue Water Consulting
Paul R. Smith, Terumo BCT
William A. Tait, Cuffco LLC
Michael Turturro, Medline Industries Inc
Timothy W. Vanderveen, MS, PharmD, CareFusion
Torge Wall, Draeger Medical Systems Inc
Stanley W. Weitzner, MD, Duke University Medical Center

Alternates: Mary Ann Allen, Children's Hospital of Akron
Brian Cassidy, Terumo BCT
Rick Dodd, GE Healthcare
Eric R. Hudson, Medline Industries Inc
Lynn E. Jensen, Fresenius Medical Care
Kent Lurvey, Baxter Healthcare Corporation
John R. Miskovic, Hospira Worldwide Inc
Charles C. Monroe, Philips Electronics North America
John Rotella, Kimberly-Clark Corporation
Tony Sacchetti, Covidien
Pamela D. Scott, FDA/CDRH
Nityanand Shetty, Cexcelabs Medical Inc
Charles B. Stribbottom, PE, PPO Standards LLC
Roxanne M. Marszcz, RN MSN, Abbott Laboratories
Kyle Steele, Value Plastics Inc
Oliver Walikowitz, Draeger Medical Systems Inc
Natalya Evmenenko, Becton Dickinson & Company

NOTE—Participation by federal agency representatives in the development of this document does not constitute endorsement by the federal government or any of its agencies.

Background of AAMI/CN20(PS):2014

As indicated in the foreword to the main body of this document (page viii), the International Organization for Standardization (ISO) is a worldwide federation of national standards bodies. The United States is one of the ISO members that took an active role in the development of this standard, which was developed by a joint ISO and International Electrotechnical Commission (IEC) working group, ISO/TC 210-IEC/SC 62D/JWG4, Small-bore connectors.

U.S. participation in ISO/TC 210-IEC/SC 62D/JWG4 is organized through the U.S. sub-Technical Advisory Group for ISO/TC 210-IEC/SC 62D/JWG4, administered by the Association for the Advancement of Medical Instrumentation. Experts from the United States made a considerable contribution to this standard.

This is a provisional standard. It is being published because it is urgently needed in the United States to promote patient safety. The text in this document reproduced from ISO/FDIS 80369-20 is from a draft ISO standard that is subject to change without notice. The text in this document, therefore, might not reflect the final text of ISO 80369-20 upon publication.

Once the final version of ISO 80369-20 is approved by ISO, this provisional standard will be replaced by a parallel adoption of ISO 80369-20. The text of the parallel adoption shall be aligned to the text of the ISO standard during the parallel approval process.

If ISO 80369-20 is not approved, AAMI will withdraw this provisional standard.

Purchasers of this provisional standard will receive a PDF copy of ANSI/AAMI/ISO 80369-20, presuming it is approved, at no additional cost, when it is published.

As used within the context of this document, “shall” indicates requirements strictly to be followed to conform to the standard. “Should” indicates that among several possibilities, one is recommended as particularly suitable, without mentioning or excluding others, or that a certain course of action is preferred but not necessarily required, or that (in the negative form) a certain possibility or course of action should be avoided but is not prohibited. “May” is used to indicate that a course of action is permissible within the limits of the standard. “Can” is used as a statement of possibility and capability. Finally, “must” is used only to describe “unavoidable” situations, including those mandated by government regulation.

The concepts incorporated in this standard should not be considered inflexible or static. This standard, like any other, must be reviewed and updated periodically to assimilate progressive technological developments. To remain relevant, it must be modified as technological advances are made and as new data come to light.

Suggestions for improving this standard are invited. Comments and suggested revisions should be sent to Standards Department, AAMI, 4301 N. Fairfax Dr, Suite 301, Arlington, VA 22203-1633.

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 80369-20 was prepared jointly by Technical Committees ISO/TC 210, Quality management and corresponding general aspects for medical devices, and IEC/SC62D, Electromedical equipment. The draft was circulated for voting to the national bodies of both ISO and IEC.

ISO 80369 consists of the following parts, under the general title *Small-bore connectors for liquids and gases in healthcare applications*

- Part 1: General requirements
- Part 2: Connectors for breathing systems and diving gases applications
- Part 3: Connectors for enteral application
- Part 4: Connectors for urethral and urinary applications¹⁾
- Part 5: Connectors for limb circulation applications
- Part 6: Connectors for neurological applications
- Part 7: Connectors for intravascular or hypodermic applications
- Part 20: Common test methods (this standard)

In this standard, the following print types are used:

- Requirements and definitions: roman type.
- Informative material appearing outside of tables, such as notes, examples and references: in smaller type. Normative text of tables is also in a smaller type.

1) Planned but not yet begun as of the date of publication.

— TERMS DEFINED IN ISO 80369-1 AND CLAUSE 3 OF THIS STANDARD: SMALL CAPITALS.

In this standard, the conjunctive “or” is used as an “inclusive or” so a statement is true if any combination of the conditions is true.

The verbal forms used in this standard conform to usage described in Annex H of the ISO/IEC Directives, Part 2. For the purposes of this standard, the auxiliary verb:

- “shall” means that compliance with a requirement or a test is mandatory for compliance with this standard;
- “should” means that compliance with a requirement or a test is recommended but is not mandatory for compliance with this standard;
- “may” is used to describe a permissible way to achieve compliance with a requirement or test.

An asterisk (*) as the first character of a title or at the beginning of a paragraph or table title indicates that there is guidance or rationale related to that item in Annex A.

The following paragraph is directed to authorities with jurisdiction and is not intended to address clinical implementation:

The attention of Member Bodies and National Committees is drawn to the fact that equipment manufacturers and testing organizations may need a transitional period following publication of a new, amended or revised ISO or IEC publication in which to make products in accordance with the new requirements and to equip themselves for conducting new or revised tests. It is the recommendation of the committee that the content of this publication be adopted for implementation nationally not earlier than 3 years from the date of publication for equipment newly designed and not earlier than 5 years from the date of publication for equipment already in production.

Introduction

This part of ISO 80369 includes common TEST METHODS for evaluating the functional performance of the SMALL-BORE CONNECTORS of this series.

During the development of the ISO 80369 series it became evident that many of the TEST METHODS were very similar for each of the APPLICATIONS. It was therefore decided to standardize all the TEST METHODS into a separate part of the series to prevent unnecessary duplication and minor differences. It is also recognized that not all CONNECTORS can be evaluated using each TEST METHOD in this part. The TEST METHODS applicable to each CONNECTOR are specified in the respective part of the ISO 80369 series.

Currently in preview, click buy full version

Small-bore connectors for liquids and gases in healthcare applications — Part 20: Common test methods

1 * Scope

This part of ISO 80369 specifies the TEST METHODS to support the functional requirements for SMALL-BORE CONNECTORS intended to be used for CONNECTIONS of MEDICAL DEVICES and related ACCESSORIES.

This part of ISO 80369 does not specify the functional requirements for the MEDICAL DEVICES or ACCESSORIES that use these CONNECTORS. Such requirements are given in particular International Standards for specific MEDICAL DEVICES or ACCESSORIES.

2 Normative references

The following referenced documents, in whole or in part, are normatively referenced in this document and are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

NOTE 1 The way in which these referenced documents are cited in normative requirements determines the extent (in whole or in part) to which they apply.

NOTE 2 Informative references are listed in the bibliography on page 29.

ISO 80369-1:2010, *Small-bore connectors for liquids and gases in healthcare applications – Part 1: General requirements*

3 Terms and definitions

For the purposes of this document, the terms and definitions specified in ISO 80369-1:2010 and ISO 14971:2007 and the following apply. For convenience, the sources of all defined terms used in this document are given in the index at the end of this document.

3.1

TEST METHOD

definitive PROCEDURE for evaluating CONNECTORS that produces a test result

3.2

TYPE TEST

test on a representative sample with the objective of determining if the CONNECTOR, as designed and manufactured, can meet the requirements of this standard

Note 1 to entry: More than one set of representative samples can be required, e.g. testing the SMALL-BORE CONNECTORS produced in each cavity of a multi-cavity mould.

[SOURCE: IEC 60601-1:2005, definition 3.135 modified: deleted 'of the equipment' and replaced 'equipment' with 'CONNECTOR'.]