

American
National
Standard

ANSI/AAMI
PB70:2022

Liquid barrier performance
and classification of
protective apparel and drapes
intended for use in health
care facilities

Currently in preview, click buy full version

Liquid barrier performance and classification of protective apparel and drapes intended for use in health care facilities

Approved 17 October 2022 by
AAMI

Approved 5 December 2022 by
American National Standards Institute

Abstract: Establishes minimum barrier performance requirements, a classification system, and associated labeling requirements for protective apparel, surgical drapes, and drape accessories intended for use in health care facilities.

Keywords: acceptable quality level (AQL), Bacteriophage, barrier properties, Binding, blood-borne pathogen, body fluid, body fluid simulant, critical zone, critical zone component, decontamination garment, decontamination gown, Fenestration, full coverage gown, Gown, Hem, Hood, isolation gown, Laminate, laundry processes, Manufacturer, non-protective back gown, open back gown, other potentially infectious materials (OPIM), Ply, point of attachment, procedure gown, protective apparel, reinforced area, rejectable quality level (RQL), Seam, service life, simulated product, Sterile, sterile field, Sterilization, strike-through, surgical drape, surgical gown, surgical gown–E, surrogate microbe, synthetic blood, Toga, viral penetration, viscosity

AAMI Standard

This Association for the Advancement of Medical Instrumentation (AAMI) standard implies a consensus of those substantially concerned with its scope and provisions. The existence of an AAMI standard does not in any respect preclude anyone, whether they have approved the standard or not, from manufacturing, marketing, purchasing, or using products, processes, or procedures not conforming to the standard. AAMI standards are subject to periodic review, and users are cautioned to obtain the latest editions.

CAUTION NOTICE: This AAMI standard may be revised or withdrawn at any time. AAMI procedures require that action be taken to reaffirm, revise, or withdraw this standard no later than five years from the date of publication. Interested parties may obtain current information on all AAMI standards by calling or writing AAMI.

All AAMI standards, recommended practices, technical information reports, and other types of technical documents developed by AAMI are voluntary, and their application is solely within the discretion and professional judgment of the user of the document. Occasionally, voluntary technical documents are adopted by government regulatory agencies or procurement authorities, in which case the adopting agency is responsible for enforcement of its rules and regulations.

Published by

AAMI

901 N. Glebe Road, Suite 300
Arlington, VA 22203
www.aami.org

© 2023 by the Association for the Advancement of Medical Instrumentation

All Rights Reserved

This publication is subject to copyright claims of AAMI. No part of this publication may be reproduced or distributed in any form, including an electronic retrieval system, without the prior written permission of AAMI. All requests pertaining to this document should be submitted to AAMI. It is illegal under federal law (17 U.S.C. § 101, et seq.) to make copies of all or any part of this document (whether internally or externally) without the prior written permission of the Association for the Advancement of Medical Instrumentation. Violator's risk legal action, including civil and criminal penalties, and damages of \$100,000 per offense. For permission regarding the use of all or any part of this document, visit the [Copyright Clearance Center](http://www.copyright.com).

Printed in the United States of America

ISBN 978-1-57020-856-0

Contents

Page

Committee representation	v
Acknowledgement	vi
Foreword	vii
Introduction	viii
1 Scope	1
1.1 General	1
1.2 Inclusions	1
1.3 Exclusions	1
2 Normative references	2
3 Definitions	2
4 Requirements	7
4.1 Labeling requirements	7
4.1.1 Device labeling	7
4.1.2 Package labeling	7
4.1.3 Technical information	8
4.1.4 Education	8
4.2 Performance requirements	9
4.2.1 Barrier performance	9
4.2.2 Tracking mechanism for multiple-use products	10
4.2.3 Construction	10
5 Tests	12
5.1 Tests for the labeling requirements	12
5.2 Tests for the performance requirements	12
5.2.1 Barrier performance	13
5.2.2 Tracking mechanism for multiple-use products	14
5.2.3 Construction	14
Annex A (informative) Rationale for the development and provisions of this standard	15
A.1 Need for the standard, history of development, and scope	15
A.1.1 Need for the standard and history of development	15
A.1.2 Scope of the standard	15
A.2 Normative references	16
A.3 Definitions	16
A.4 Requirements	16
A.4.1 Rationale for the labeling requirements	16
A.4.2 Rationale for the performance requirements	17
Annex B (informative) Examples of barrier performance classification of surgical gowns, surgical gowns-E, aseptic gowns, other protective gowns, surgical drapes, hoods, togas, and other protective apparel	21
Annex C (informative) Example of a sampling plan	42
Annex D (informative) Rationale for changes to the 2012 edition of PB70	44
Bibliography	45

Figures

Figure B.1—Example of a surgical gown intended for surgical applications	21
--	----

Figure B.2—Example of a surgical gown—E intended for surgical applications.....	23
Figure B.3—Example of a gown intended for isolation applications.....	25
Figure B.4—Example of a surgical drape.....	27
Figure B.5a—Example of other protective gown defined as full coverage gown.....	28
Figure B.5b—Example of other protective gown defined as non-protective back gown.....	29
Figure B.5c—Example of other protective gown defined as open back gown.....	31
Figure B.6—Example of a protective hood.....	33
Figure B.7a—Example of a toga integrated with a surgical gown-E.....	34
Figure B.7b—Example of a toga integrated with a surgical gown.....	37
Figure B.7c—Example of a toga integrated with a full coverage gown.....	40
Figure C.1—OC curve for n=32, c=3 sampling plan.....	42

Tables

Table 1—Classification of barrier performance of surgical gowns, isolation gowns, other gowns, other protective apparel, surgical drapes, and drape accessories.....	10
Table B.1—Barrier performance classification of a surgical gown.....	22
Table B.2—Barrier performance classification of a surgical gown—E.....	24
Table B.3—Barrier performance classification of an isolation gown.....	26
Table B.4—Barrier performance classification of surgical drapes.....	28
Table B.5a—Barrier performance classification of other protective gown defined as full coverage gown.....	29
Table B.5b—Barrier performance classification of other protective gown defined as non-protective back gown.....	31
Table B.5c—Barrier performance classification of other protective gown defined as open back gown.....	32
Table B.6—Barrier performance classification of a protective hood.....	34
Table B.7a—Barrier performance classification of a toga integrated with a surgical gown-E.....	36
Table B.7b—Barrier performance classification of a toga integrated with a surgical gown.....	39
Table B.7c—Barrier performance classification of a toga integrated with a full coverage gown.....	41
Table C.1—Sampling plan for sample size code letter G, acceptable quality level (normal inspection).....	43

Committee representation

Association for the Advancement of Medical Instrumentation

AAMI Protective Barriers Committee

The publication of ANSI/AAMI PB70 as an American National Standard was initiated by the AAMI Protective Barriers Committee.

At the time this document was published, the **AAMI Protective Barriers Committee** had the following members:

Cochairs: Jill Holdsworth

Members: Michael Belkin, Brigham & Womens Hospital
Alex Bradley, DuPont Tyvek Medical and Pharmaceutical Protection
Jacqueline Daley, Providence St Joseph Health System
Kevin Eschen, Stryker Instruments Division
Wanda Folsom, Association of Surgical Technologists (AST)
Brian Fortier, Aorta Medical Inc
Becky Gilsdorf, Healthmark Industries Company Inc
Jill Holdsworth, Emory University Hospital Midtown
F Selcen Kilinc-Balci, CDC/NIOSH
Brian Lin, Owens & Minor
Lina Luu, Medline Industries Inc
Kathleen McMullen, Association for Professionals in Infection Control & Epidemiology
Samia Moore, Hollingsworth & Vose
Jill Morgan, Emory University School of Medicine
Mary Olivera, Quality Processing Resource Group LLC
Martin O'Lone
Bifeng Qian, FDA/CDRH
Barbara Rusynko
Anne Schuler, LexaMed Ltd
Harry Shaffer, Sterilization Consulting Services
Sarah Smit, Sotera Health LLC
Donna Swenson
Donald Tumminelli, HIGHPOWER Validation Testing & Lab Services Inc
Amber Wood, Association of Perioperative Registered Nurses (AORN)
Jon Wood, HSPA

Alternates: Parik Atre, Stryker Instruments Division
Alvin Bell, Medline Industries Inc
Shawn Engberg, M3 Furniture & Supply
Kevin Frey, Association of Surgical Technologists (AST)
Larry Green, ABBAN Development
Henri Huber, Quality Processing Resource Group LLC
Jennifer Johnson, Sotera Health LLC
Susan Klack, HSPA
Emily Kyle, Association of Perioperative Registered Nurses (AORN)
Jessica Lawrence, WuXi AppTec Inc
Catherine Miceli, Cardinal Health
Namita Mithani, Owens & Minor
Cheron Rojo, Healthmark Industries Company Inc
Gary Socola, HIGHPOWER Validation Testing & Lab Services Inc

NOTE: Participation by federal agency representatives in the development of this document does not constitute endorsement by the federal government or any of its agencies.

Acknowledgement

The AAMI Protective Barriers Committee would like to acknowledge previous cochair, Shane Woody, previously working with Aspen Surgical on his assistance and efforts in advancing the progression of the document to development.

Currently in preview, click buy full version

Foreword

The following verbal forms are used within AAMI documents to distinguish requirements from other types of provisions in the document:

- “shall” and “shall not” are used to express requirements;
- “should” and “should not” are used to express recommendations;
- “may” and “may not” are used to express permission;
- “can” and “cannot” are used as statements of possibility or capability;
- “might” and “might not” are used to express possibility;
- “must” is used for external constraints or obligations defined outside the document; “must” is not an alternative for “shall.”

Suggestions for improving this technical information report are invited. Comments and suggested revisions should be sent to Standards, AAMI, 901 N. Glebe Road, Suite 300, Arlington, VA 22203 or standards@aami.org.

Introduction

This standard was developed by the AAMI Protective Barriers Committee and establishes a classification system and the associated minimum requirements for the liquid barrier performance of protective apparel and drapes based on industry-accepted test methods. It is intended to assist manufacturers in testing and labeling their devices so health care personnel can make more informed decisions when selecting the appropriate product for the anticipated task at hand.

Protective apparel is worn by health care workers to help preserve the integrity of the sterile field and inhibit the transfer of blood, body fluids, other potentially infectious materials (OPIM), and associated microorganisms. Drapes and drape accessories are also intended to inhibit the transfer of microorganisms, body fluids, and OPIM. Drapes and drape accessories are used as protective patient coverings to isolate a site of surgical incision from microbial and other cross-contamination.

In the United States, surgical apparel, surgical drapes, and drape accessories are medical devices and, under the Food, Drug, and Cosmetic Act, as amended by the Medical Device Amendments of May 28, 1976, are subject to regulation by the U.S. Food and Drug Administration (FDA), including but not limited to FDA requirements for premarket notification (section 510(k) of the Act) and medical device reporting. Barrier efficacy has long been recognized as important in helping to prevent infections and is now mandated by Occupational Safety and Health Administration (OSHA) regulations limiting occupational exposure to bloodborne pathogens (29 CFR 1910.1030) [10]. See also the Centers for Disease Control and Prevention's (CDC's) Guideline for the prevention of surgical site infection (CDC, 1999 [11]).

Surgical gowns, other protective apparel, surgical drapes, and drape accessories are devices intended to promote infection control practices and help protect patients and health care workers. This standard is based on key barrier performance tests that are used to classify the subject products into levels of performance. Knowledge of these defined levels of performance will allow informed and consistent choices about the type of protective product necessary for the situation at hand.

This is the third edition of *Liquid barrier performance and classification of protective apparel and drapes intended for use in health care facilities*, which was first published as an American National Standard in 2003. In comparison to the second edition, the most significant revisions are the expansion of the scope to include additional types of personal protective equipment.

AAMI and ANSI procedures require that standards be reviewed every five years and, if necessary, revised to reflect technological advances that may have occurred since publication.

The concepts incorporated in this standard should not be considered inflexible or static. This standard, like any other, must be reviewed and updated periodically to assimilate progressive technological developments. To remain relevant, it must be modified as technological advances are made and as new data comes to light.

Suggestions for improving this standard are invited. Comments and suggested revisions should be sent to Standards Department, AAMI, 501 N. Glebe Road, Suite 300, Arlington, VA 22203.

NOTE This foreword does not contain provisions of the American National Standard ANSI/AAMI PB70:2022, *Liquid barrier performance and classification of protective apparel and drapes intended for use in health care facilities*, but it does provide important information about the development and intended use of the document.

Liquid barrier performance and classification of protective apparel and drapes intended for use in health care facilities

1 Scope

1.1 General

This standard establishes minimum barrier performance requirements, a classification system, and associated labeling requirements for protective apparel, surgical drapes, and drape accessories intended for use in health care facilities.

1.2 Inclusions

This standard covers surgical drapes, drape accessories, and all types of protective apparel that are labeled with liquid barrier claims or liquid borne microbial barrier claims that are not specifically excluded in 1.3 (e.g., single-use and multiple-use surgical gowns, isolation gowns, other gowns, other types of open back gowns including over the head gowns, decontamination attire, aprons, toga, protective headwear such as hats and caps, hoods, footwear covers, sleeves, and laboratory attire). Some of these devices are regulated by the U.S. Food and Drug Administration (FDA) as medical devices under 21 CFR 878.

NOTE 1 Surgical apparel is classified by the FDA under 21 CFR 878.4040, and surgical drapes and drape accessories are classified under 21 CFR 878.4370.

NOTE 2 Isolation gowns are classified by the FDA under 21 CFR 878.4040.

NOTE 3 For additional information regarding the scope of this standard, see Annex A, A.1.1 and A.1.2. Other informative annexes are also included in this standard.

1.3 Exclusions

This standard does not cover the following:

- a) protective apparel for the hands, such as surgical gloves, patient examination gloves, and other medical gloves;
- b) protective apparel for the face, and eyes, such as goggles, non-integrated face shields, surgical masks, and non-integrated respirators;
- c) other types of protective apparel worn by health care personnel that is not intended or labeled as a barrier to liquid or microorganisms;
- d) absorbent operating room (OR) towels;
- e) all of the requirements necessary to ensure the safety and effectiveness of the products within the scope of this standard;
- f) the interfaces between products, such as the gown/glove interface;
- g) all of the labeling or other information that a health care facility might deem necessary or desirable in product selection;