

American  
National  
Standard

ANSI/AAMI  
ST108:2023

Water for the processing of  
medical devices

Currently in preview, click buy full version

# Water for the processing of medical devices

Approved 30 June 2023 by  
**AAMI**

Approved 4 August 2023 by  
**American National Standards Institute, Inc.**

**Abstract:** This standard covers the selection and maintenance of effective water quality suitable for processing medical devices. It provides guidelines for selecting the water quality necessary for the processing of categories of medical devices and addresses water treatment equipment, water distribution and storage, quality control procedures for monitoring water quality, strategies for bacterial control, and environmental and personnel considerations.

**Keywords:** carbon filters, deionization, disinfection, distillation, medical devices, pasteurization, processing, reverse osmosis, rinsing, sediment filters, steam purity, sterilization, ultrafiltration, water filtration, water quality, washing, water softening, water quality, water treatment

## AAMI Standard

This Association for the Advancement of Medical Instrumentation (AAMI) standard implies a consensus of those substantially concerned with its scope and provisions. The existence of an AAMI standard does not in any respect preclude anyone, whether they have approved the standard or not, from manufacturing, marketing, purchasing, or using products, processes, or procedures not conforming to the standard. AAMI standards are subject to periodic review, and users are cautioned to obtain the latest editions.

**CAUTION NOTICE:** This AAMI standard may be revised or withdrawn at any time. AAMI procedures require that action be taken to reaffirm, revise, or withdraw this standard no later than five years from the date of publication. Interested parties can obtain current information on all AAMI documents by calling or writing AAMI.

All AAMI standards, recommended practices, technical information reports, and other types of technical documents developed by AAMI are *voluntary*, and their application is solely within the discretion and professional judgment of the user of the document. Occasionally, voluntary technical documents are adopted by government regulatory agencies or procurement authorities, in which case the adopting agency is responsible for enforcement of its rules and regulations.

*Published by*

**AAMI**

901 N. Glebe Road, Suite, 300  
Arlington, VA 22203  
[www.aami.org](http://www.aami.org)

© 2023 by the Association for the Advancement of Medical Instrumentation

All Rights Reserved

This publication is subject to copyright claims of AAMI. No part of this publication may be reproduced or distributed in any form, including an electronic retrieval system, without the prior written permission of AAMI. All requests pertaining to this document should be submitted to AAMI. It is illegal under federal law (17 U.S.C. § 101, et seq.) to make copies of all or any part of this document (whether internally or externally) without the prior written permission of the Association for the Advancement of Medical Instrumentation. Violators risk legal action, including civil and criminal penalties, and damages of \$100,000 per offense. For permission regarding the use of all or any part of this document, visit the [Copyright Clearance Center](#).

Printed in the United States of America

**ISBN 978-1-57020-873-7**

## Contents

Page

Committee representation .....	v
Foreword .....	x
Introduction .....	xi
1 Scope .....	1
2 Normative references .....	2
3 Terms and definitions .....	2
4 Roles and responsibilities .....	12
5 Risk analysis .....	15
6 Categories of water quality for medical device processing .....	17
7 Water quality selection and requirements .....	20
8 Water treatment systems installation and operation qualification .....	27
9 Water treatment systems performance qualification - evaluation of water quality .....	31
10 Water treatment systems routine monitoring .....	31
11 Continuous quality improvement .....	33
12 Water treatment systems maintenance .....	36
13 Special considerations .....	37
Annex A (informative) Guidance on the application of the normative requirements .....	40
Annex B (informative) Risk analysis .....	47
Annex C (informative) Automated Endoscope Reprocessor (AER) .....	50
Annex D (informative) Water used in cleaning and moist heat processes .....	52
Annex E (informative) Water treatment technologies .....	56
Annex F (informative) Water treatment system design .....	63
Annex G (informative) Routine monitoring of water treatment equipment & produced water quality .....	69
Annex H (informative) Maintaining microbiological quality .....	75
Annex I (informative) Typical presentation of water quality issues during the processing of medical devices .....	80
Bibliography .....	83

## Figures

Figure 1—Stages of medical device processing in which water quality is a consideration .....	22
Figure F.1—Example process flow diagram (PFD) .....	64

## Tables

Table 1—Multidisciplinary team responsibilities .....	15
Table 2—Categories and performance qualification levels of water quality for medical device processing .....	18
Table 3—Water quality for processing of devices requiring high-level disinfection .....	26
Table 4—Water quality monitoring requirements .....	33
Table 5—Frequency for water quality monitoring at water generation system .....	34

Table 6—Frequency for water quality monitoring at point-of-water-use .....	35
Table A.1—Water descriptions.....	40
Table A.2—Water quality for processing devices.....	41
Table B.1—Tables of risk factors and effects.....	47
Table B.2— Impact of water temperature on sterilizer vacuum water supply pump performance.....	48
Table B.3—Work-related risks to staff.....	49
Table C.1—Water quality for LCSPS .....	50
Table E.1—Summary of water treatment methods .....	61
Table G.1—Pretreatment equipment monitoring.....	71
Table G.2—Primary water treatment equipment monitoring .....	72
Table G.3—Storage and Distribution Equipment Monitoring.....	73
Table I.1—Examples of observed problems during medical device processing that can be caused by poor water quality.....	80

Currently in preview, click buy full version

## Committee representation

### Association for the Advancement of Medical Instrumentation

#### AAMI Water Quality for Medical Device Processing Working Group

This standard was developed by the AAMI Water Quality for Medical Device Processing Working Group under the auspices of the AAMI Sterilization Standards Committee. Approval of the standard does not necessarily mean that all working group members voted for its approval.

At the time This standard was published, the **AAMI Water Quality for Medical Device Processing Working Group** had the following members:

*Cochairs:* Terra Kremer  
Erin A. Kyle

*Members:* Anas Aljabo, SteriLabs Canada Inc.  
Lindsay Arencibia, Lindsay Arencibia (Individual Expert)  
Kristy L. Azzolin, American Dental Association (ADA)  
James Weldon Baker, NuStream Filtration Inc  
Timothy Blackwell, Timothy Blackwell Person  
Rachel Bradley, Rachel Bradley (Individual Expert)  
Lisa Shifflett Brown, UVA Health  
Deborah Lynnette Bunn, University of Maryland Medical System  
Jonathan Burdach, Nanosonics Limited  
Cathy Castro, Vested Medical LLC  
Michael Scott Castro, Pall Medical - Port Washington  
Roxanne Charles, Roxanne Charles (Individual Expert)  
Nancy Chobin, Nancy Chobin (Individual Expert)  
Horlando Cintron, Memorial Sloan Kettering Cancer Center - Monmouth  
Jeremy Criss, Jeremy Criss (Individual Expert)  
Jacqueline A. Daley, Providence St. Joseph Health System  
Frank Daniels, Virginia Commonwealth University Health System  
Nicholas Day, University Hospitals of Cleveland  
Lorraine Durigan, Lorraine Durigan (Individual Expert)  
Afif Jhoel Escheik, Catholic Healthcare  
Ernesto Rivera Fernandez, Edwards Lifesciences  
Michael Flanagan, Zimmer Biomet  
Brian Flannigan, Prigenics LLC  
Sarah Frank, Element Materials Technology (fka ALG)  
Marcia Gomez, Case Medical Inc  
Noemi Gannon, Abbott Laboratories  
Debbie Honey, AmeriWater  
Douglas F. Harbrecht, Sterility Assurance LLC  
Jeffrey N. Holland, Evoqua Water Technologies LLC  
Timothy Hurtado, Memorial Hermann Healthcare System  
David M. Jagrosse, David Jagrosse Consulting LLC  
Karissa Joosten, NAMSA  
Richard J. Kaestner Jr., Absolute Water Technologies  
Nancy E. Kaiser, STERIS Corporation | Healthcare  
Nicholas Kalanta, McLane Children's Hospital  
Sajedul Islam Khan, Apollo Hospital Dhaka  
Michal Koenner, Medtronic Inc Campus  
Margaret Konja, SpecialtyCare  
Terra Kremer, Johnson & Johnson

Kaumudi Kulkarni, Healthmark Industries Company Inc  
 Erin A. Kyle, Association of Perioperative Registered Nurses (AORN)  
 Wesley Lantz, Zimmer Biomet  
 William Leiva, 3M Health Care  
 Angela M. Lewellyn, Advantage Support Services Inc  
 Paul Littley, Integra LifeSciences Corporation  
 Anne Mattern, Anne Mattern Person  
 Michael Matthews, Association of Medical Device Service Organizations (AMDSO)  
 John L. Mazzilli, Saddleback Memorial Medical Center  
 Tom McElroy, IUVO BioScience  
 Emily Mitzel, GE Healthcare  
 Melissa G. Morgan, Moses H Cone Hospital - Cone Health  
 Derek Mortisen, Intuitive Surgical Inc  
 Mike Nolan, HIGHPOWER Validation Testing & Lab Services Inc  
 Susumu Nozawa, Siemens Healthineers  
 Kia Parker, The Ohio State Wexner Medical Center-University Hospital  
 Alpa Patel, Sotera Health LLC  
 Rebecca Peplau, Rebecca Peplau (Individual Expert)  
 Janet M. Prust, St. Croix Standards Consulting, LLC  
 Juan Migiel Ramos, Juan Migiel Ramos (Individual Expert)  
 Marc B. Rogers, Advanced Sterilization Products (ASP)  
 Don Rotter, Ecolab  
 Mandy Ryan, Stryker  
 Damien S. Berg, HSPA (fka IAHCSSM)  
 Chuck Schaefer, Charles Schaefer (Individual Expert)  
 Candace A. Schlicklin, Phoenix Children's Hospital  
 Michael J. Schoene, Bausch & Lomb Inc  
 Matthias Schöttler, Olympus America Inc  
 Richard William Schule, Cleveland Clinic  
 Linda Sue Schultz, Northside Hospital Surgical Services Atlanta  
 Cynthia Seward, Kaiser Foundation Health Plan/Hospitals  
 Donald Socha Esq., Getinge USA  
 Len Sparks, VERDA Water Quality Systems  
 Joan M. Spear, Joan M. Spear (Independent Expert)  
 Nadine Swenson, Boston Scientific Corporation  
 Janelle Trbojevich, Boston Scientific Corporation  
 Steven E. Turtill, FD, MCL RH  
 Melissa Walters, Melissa Walters (Individual Expert)  
 Dean Wellman, Healthmark Corporation  
 Jonathan A. Wilder, Quality Processing Resource Group LLC  
 Ariel Winbough, Emory Healthcare  
 David C. Winbough Jr., Veterans Administration (VA) Central Office  
 Robin Woodland, Robin Woodland (Individual Expert)  
 Carl Wager, Powder River Medical Resources  
 Stephanie L. Young, Stephanie Young (Individual Expert)  
 Roberto Zumbado, Philips

*Alternates:*

Jennifer Barnickel, Vested Medical LLC  
 Ralph J. Basile, Healthmark Industries Company Inc  
 Jhmeid B. Billingslea, Advantage Support Services Inc  
 Michael Bolcas, Integra LifeSciences Corporation  
 Marco Bommarito, 3M Health Care  
 Philip Ted Cantin, VERDA Water Quality Systems  
 Greg Crego, Bausch & Lomb Inc  
 Jeffery Daniels, Olympus America Inc

Shawn Engberg, MD Furniture & Supply  
Lorraine Floyd, Johnson & Johnson  
Alex Freeman, Intuitive Surgical Inc  
Tara Jacobson, Abbott Laboratories  
Herbert J. Kaiser, STERIS Corporation | Healthcare  
Susan G. Klacik, HSPA (fka IAHCSSM)  
Angela Le, ResMed Inc.  
Daniel Lightfoot PhD, Nanosonics Limited  
Shaun McGinley, Zimmer Biomet  
Kim Miller, HIGHPOWER Validation Testing & Lab Services Inc  
Ronald Moody, Veterans Administration (VA) Central Office  
Shane Pinkston, Getinge USA  
Michael Polozani, Case Medical Inc  
Shawn Powers, IUVO BioScience  
Kaleb Price, Northside Hospital Surgical Services Atlanta  
Nicola Revellin, GE Healthcare  
Jessica Rue, Boston Scientific Corporation  
Anne Schuler, LexaMed Ltd  
Richard Sparano, Medtronic Inc Campus  
Aaron Troutt, Williamson Medical Center  
Dustin E. Van Bree, Sotera Health LLC  
Jill A. Warren, Siemens Healthineers  
Mary Wen, FDA/CDRH  
Erika Wilson, Phigenics LLC

---

NOTE Participation by federal agency representatives in the development of this standard does not constitute endorsement by the federal government or any of its agencies.

---

At the time this document was published, the AAMI Sterilization Standards Committee had the following members:

**Cochairs:** Janet M. Prust  
Patrick B. Weixel

**Members:** Gregory Darnall Aldrich, Fairbanks Memorial Hospital  
Anas Aljabo, Steris Canada Inc.  
Eseosa Tony Amenaghawon, Eseosa Tony Amenaghawon (Individual Expert)  
Brett Anderson, Cochlear Ltd  
Jennifer Anderson, Quality, Microbiology & Sterilization Services LLC  
Richard Banoff, STERIS Corporation | Healthcare  
Marie K. Bower, UnityPoint Health - Surgery Center Cedar Rapids  
Tracie Daley Bryans, BryKor LLC  
Jonathan Burdach, Nanosonics Limited  
Charles Cogdill, Medtronic Inc Campus  
Emily Craven, Boston Scientific Corporation  
Jacqueline A. Daley, Providence St Joseph Health System  
Gordon M. Ely, LexaMed Ltd  
Afif Jhoel Escheik, Crothall Healthcare  
Deannard Esnard, Deannard Esnard (Individual Expert)  
Nicole Felderman, CIVCO Medical Solutions  
Brian J. Fortier, Aorta Medical Inc  
Daniel Fowler, WuXi AppTec Inc  
Magnus Graham, BSI Healthcare  
Shelley Hagan, Henry Ford Health

Ashley Hammer, Washington Regional Medical Center  
Joyce M. Hansen, Johnson & Johnson  
Douglas F. Harbrecht, Sterility Assurance LLC  
Deborah A. Havlik, DA Havlik Consulting  
Crystal Heishman, U of L Hospital - University of Louisville  
Ebow Holdbrook-Smith, WellSpan Health  
Mollie J. Holter, MicroBio Consulting LLC  
Stephanie Jean Homuth, Hennepin County Medical Center Warehouse  
Clark W. Houghtling, Cosmed Group Inc  
Angela G. Jensen, Rush Foundation Hospital  
Nicholas Kalanta, McLane Children's Hospital  
Susan G. Klacik, HSPA (fka IAHCSSM)  
Erin A. Kyle, Association of Perioperative Registered Nurses (AORN)  
Byron J. Lambert, Abbott Laboratories  
Marissa Jones Lewis, Veterans Administration (VA) Central Office  
Michael K. Loadman, Michael Loadman (Individual Expert)  
Emily Lorcheim, ClorDiSys Solutions, Inc  
Kelly Marcum, HCA Healthcare  
John L. Mazzilli, Saddleback Memorial Medical Center  
Patrick J. McCormick, Bausch & Lomb Inc  
Russell Mills, Zimmer Biomet  
Leslie Nichols, Mayo Clinic  
Gerry A. O'Dell, Gerry O'Dell Consulting  
James O'Reilly, St Joseph Health System - Trinity Health  
Ken Paddock, Baxter Healthcare Corporation  
Alberto Guzman Paret, Cordis US Corp  
Kia Parker, The Ohio State Wexner Medical Center - University Hospital  
Kimberly Patton, Performance Review Institute - Med Accred  
Nancy Pickens, Nancy Pickens (Individual Expert)  
Karana M Pierre, Karana M Pierre (Individual Expert)  
Janet M. Prust, St. Croix Standards Consulting, LLC  
Joan E. Rickard, Montefiore Medical Center  
Angela Salmen, Angela Salmen (Individual Expert)  
Linda Sue Schultz, Northside Hospital Surgical Services Atlanta  
Joan M. Spear, Joan M. Spear (Independent Expert)  
Sopheak Srun, Quality Tec. Services LLC  
Mark Swanson, Quality and Regulatory Expert Partners (QRX)  
Larry Talapa, 3M Health Care  
Julie TerWee, Peter Darenter Center of Excellence  
James Treharne, St Peter's Health  
Sara Vinson, University of Florida  
Adrienne C. Watson, Adrienne Watson (Individual Expert)  
Sean Weir, Sean Weir (Individual Expert)  
Patricia B. Weixel, FDA/CDRH  
James Sidney Wiggs, Legacy Health System  
Daniel Wimbush, Emory Healthcare  
Martell Winters, Sotera Health LLC  
Roberto Zumbado, Philips

Additional:

Jennifer Benolken, DuPont Tyvek Medical and Pharmaceutical Protection  
Damien S. Berg, HSPA (fka IAHCSSM)  
Stacy Bohl, Boston Scientific Corporation  
Carolyn Braithwaite-Nelson, Philips  
Richard Burgess, BSI Healthcare  
Densley Coke, Densley Coke (Individual Expert)  
Kimbrell Darnell, Becton Dickinson & Company

Aaron David DeMent, Sotera Health LLC  
April Doering, 3M Health Care  
Anna M. Grayson, Grayson Associates  
David Kelman, Sterile Edge  
Cathy Leckwart, WuXi AppTec Inc  
Paul Lorcheim, ClorDiSys Solutions, Inc  
Michelle Luebke, Baxter Healthcare Corporation  
Gerald E. McDonnell, Johnson & Johnson  
Shaun McGinley, Zimmer Biomet  
David Ford McGoldrick, Abbott Laboratories  
Shane Pinkston, Getinge USA  
Christine L. Render, Cosmed Group Inc  
Jimmy Reyes, Alcon Laboratories Inc  
Petra Richards, Kaiser Foundation Health Plan/Hospitals  
Susan Rogers, Veterans Administration (VA) Central Office  
Molly S. Swanson, Quality Tech Services LLC  
Lisa Ward, STERIS Corporation | Healthcare

---

NOTE Participation by federal agency representatives in the development of this document does not constitute endorsement by the federal government or any of its agencies.

---

## Foreword

This standard was developed by the AAMI Water Quality for Medical Device Processing Working Group under the auspices of the AAMI Sterilization Standards Committee. The objective of this standard is to set the quality requirements for the different categories of water used in processing of medical devices and provides guidance as to when and where to use water of each category. In addition, this standard provides information on how to verify that the water continues to meet those minimum requirements. This standard also provides valuable information for the performance qualification of a water treatment/delivery system and a monitoring program to ensure water quality remains within the stated specifications.

This is the first edition of this American National Standard, which revises and replaces AAMI TIR34 [1]. The goal of this standard is to specify the minimum requirements for the water quality and steam purity necessary to effectively process medical devices intended for use on a patient.

The professionals responsible for the processing of medical devices prior to use are experiencing increasing challenges including: the increasing complexity of modern medical devices with hidden, difficult to access areas (e.g., lumens and complex mechanisms) where clinical soil can become lodged; and the emerging and reemerging incidence of “superbugs” that must be removed or inactivated for patient safety but are able to survive processing in situations where soil removal is incomplete. Any limitation on or diminishment of cleaning efficacy can lead to patient morbidity or mortality and decreased device use life. Water of the appropriate quality for the processing of medical devices prior to clinical use is an important part of the solution to these problems.

A common factor in the processing of medical devices is the use of water. While medical devices cleared by the U.S. Food and Drug Administration’s (FDA) for sale into the health care market have been provided with validated processing instructions and procedures, these procedures may not be completely effective if water of specified quality is not used. Similarly, cleaning agents work better, and devices are rinsed more thoroughly if the water is of the specified quality. Each health care facility may require a specific approach to treating water for processing needs based on a variety of factors.

The following verbal forms are used within AAMI documents to distinguish requirements from other types of provisions in the document:

- “shall” and “shall not” are used to express requirements;
- “should” and “should not” are used to express recommendations;
- “may” and “may not” are used to express permission;
- “can” and “cannot” are used as statements of possibility or capability;
- “might” and “might not” are used to express possibility;
- “must” is used for external constraints or obligations defined outside the document; “must” is not an alternative for “shall.”

Suggestions for improving this technical information report are invited. Comments and suggested revisions should be sent to Standards, AAMI, 901 N. Glebe Road, Suite 300, Arlington, VA 22203 or [standards@aami.org](mailto:standards@aami.org).

---

NOTE This foreword does not contain provisions of the ANSI/AAMI ST108, *Water for the processing of medical devices* (ANSI/AAMI ST108:2023), but it does provide important information about the development and intended use of the document.

---

## Introduction

Water quality is an important consideration in all stages of medical device processing, as is the purity or chemical content of the steam that is generated for moist heat sterilization. The health care facility should have a multidisciplinary team in place that develops a strategy to confirm all aspects of water quality that impact the processing of medical devices within this standard. Appropriate water quality and steam purity in device processing requires collaboration between the personnel who process medical devices and the personnel who establish and maintain the water treatment system.

Water treatment and delivery systems can be configured in many ways to achieve the required water quality specifications. Water can be treated by a variety of methods that yield different levels of water quality. Gram-negative bacteria and nontuberculous mycobacteria can grow in water regardless of treatment process. The rate of growth and the microbial levels attained are a function of several factors related to the water (e.g., variety and concentration of organic contaminants, pH, temperature). Water systems that are closely monitored for quality reduce variability of processing conditions and effectiveness of cleaning and disinfection processes as well as reduce the potential of microbial proliferation. The importance of monitoring water quality to prevent problems with microbial proliferation cannot be overemphasized.

This standard defines multiple levels of water quality and steam purity suitable for medical device processing, and it describes the water treatment processes that can be utilized in order to produce water of the quality to meet each of these categories.



# Water for the processing of medical devices

## 1 Scope

### 1.1 General

This standard establishes minimum requirements for the water quality used at different stages in the processing of medical devices (e.g., reusable devices and single-use devices provided non-sterile requiring processing prior to use such as non-sterile implants and processed single-use devices) to make them ready for use on the next patient. These requirements are established to assist medical device processing professionals in the selection of the appropriate water quality needed for cleaning, rinsing, disinfection, and sterilization of medical devices.

### 1.2 Inclusions

This standard covers the quality of the water delivered to the point-of-use and used in medical device processing (cleaning, rinsing, disinfection and sterilization). It defines water types based on the qualities possessed and where it will be used. Included in this document are:

- a) Responsibility for the water management program;
- b) Importance of water quality;
- c) Adverse effects of water impurities on medical device processing;
- d) Categories and requirements of water quality for medical device processing;
- e) Selection of water of the appropriate quality;
- f) Purity requirements of water used to generate steam;
- g) Effective water treatment and qualification;
- h) Considerations for ongoing maintenance, monitoring and quality improvement of the water treatment system;
- i) Troubleshooting water quality issues.

This standard also provides definitions of terms and a bibliography.

### 1.3 Exclusions

This standard does not cover:

- a) Water requirements for hemodialysis applications;
- b) Water requirements for laboratory use;
- c) Steam quality requirements (i.e., physical qualities of steam such as: moisture content, noncondensable gas content and Superheat). See ANSI/AAMI ST79 [8];
- d) Water treatment performed within the medical device processing equipment (e.g., washers, washer/disinfectors, or automated endoscope reprocessors (AERs));