

# Technical Information Report

## AAMI TIR100: 2021

End-to-end microbiological  
quality and sterility  
assurance

Currently in preview, click buy full version

## End-to-end microbiological quality and sterility assurance

Approved 1 October 2021 by  
AAMI

**Abstract:** This technical information report (TIR) provides guidance for a comprehensive framework that integrates and connects Microbiological Quality & Sterility Assurance (MQ&SA) into an organization's end-to-end process for the development, validation, and routine control of sterile health care products.

**Keywords:** aseptic processing, end-to-end, microbiologically controlled processing, sterility assurance, terminal sterilization, process

## AAMI Technical Information Report

A technical information report (TIR) is a publication of the Association for the Advancement of Medical Instrumentation (AAMI) Standards Board that addresses a particular aspect of medical technology.

Although the material presented in a TIR may need further evaluation by experts, releasing the information is valuable because the industry and the professions have an immediate need for it.

A TIR differs markedly from a standard or recommended practice, and readers should understand the differences between these documents.

Standards and recommended practices are subject to a formal process of committee approval, public review, and resolution of all comments. This process of consensus is supervised by the AAMI Standards Board and, in the case of American National Standards, by the American National Standards Institute.

A TIR is not subject to the same formal approval process as a standard. However, a TIR is approved for distribution by a technical committee and the AAMI Standards Board.

Another difference is that, although both standards and TIRs are periodically reviewed, a standard must be acted on—reaffirmed, revised, or withdrawn—and the action formally approved usually every five years but at least every 10 years. A TIR must be acted on and the action formally approved usually every three years.

A TIR may be developed because it is more responsive to underlying safety or performance issues than a standard or recommended practice, or because achieving consensus is extremely difficult or unlikely. Unlike a standard, a TIR permits the inclusion of differing viewpoints on technical issues.

**CAUTION NOTICE:** This AAMI TIR may be revised or withdrawn at any time. Because it addresses a rapidly evolving field or technology, readers are cautioned to ensure that they have also considered information that may be more recent than this document.

All standards, recommended practices, technical information reports, and other types of technical documents developed by AAMI are voluntary, and their application is solely within the discretion and professional judgment of the user of the document. Occasionally, voluntary technical documents are adopted by government regulatory agencies or procurement authorities, in which case the adopting agency is responsible for enforcement of its rules and regulations.

Comments on this technical information report are invited and should be sent to AAMI, Attn: Standards Department, 901 N. Glebe Road, Suite 300, Arlington, VA 22203.

### *Published by*

AAMI  
901 N. Glebe Road, Suite 300  
Arlington, VA 22203  
[www.aami.org](http://www.aami.org)

© 2022 by the Association for the Advancement of Medical Instrumentation

All Rights Reserved

This publication is subject to copyright claims of AAMI. No part of this publication may be reproduced or distributed in any form, including an electronic retrieval system, without the prior written permission of AAMI. All requests pertaining to this document should be submitted to AAMI. It is illegal under federal law (17 U.S.C. § 101, et seq.) to make copies of all or any part of this document (whether internally or externally) without the prior written permission of the Association for the Advancement of Medical Instrumentation. Violator's risk legal action, including civil and criminal penalties, and damages of \$100,000 per offense. For permission regarding the use of all or any part of this document, visit the [Copyright Clearance Center](http://Copyright Clearance Center).

Printed in the United States of America

ISBN 978-1-57020-837-9

## Contents

Page

Committee representation .....	iv
Foreword .....	viii
Introduction .....	ix
1 Scope .....	1
2 Normative references .....	1
3 Terms and definitions .....	1
4 General .....	2
5 Research and Development (R&D) .....	3
6 Plan .....	6
7 Source .....	8
8 Make .....	10
9 Deliver .....	11
10 Customer .....	12
Bibliography .....	13

## Figures

Figure 1—End-to-end process .....	ix
Figure 2—Interactions and information flow within the end-to-end process .....	x

## Committee representation

### Association for the Advancement of Medical Instrumentation

#### AAMI Assurance of Sterility Working Group

This technical information report was developed by the AAMI Assurance of Sterility Working Group under the auspices of the AAMI Sterilization Standards Committee. Approval of the technical information report does not necessarily mean that all working group members voted for its approval.

At the time this technical information report was published, the **AAMI Assurance of Sterility Working Group** had the following members:

**Cochairs:** Christopher Anderson  
Byron Lambert

**Members:** Christopher Anderson, Johnson & Johnson  
Edward Arscott, NAMSA  
Jennifer Asleson, Quality, Microbiology & Sterilization Services LLC  
Carolyn Braithwaite-Nelson, Philips  
David Brodersen, Getinge USA  
Trabue Bryans, BryKor LLC  
Glenn Calvert, West Pharmaceutical Services, Inc.  
Jeremy Criss  
Sophia Czechowicz, Advanced Sterilization Products (ASP)  
Elaine Daniell, EDan-SA LLC  
April Doering, Cantel Inc  
Gordon Ely, LexaMed Ltd  
Dan Floyd, DuPont Tyvek Medical and Pharmaceutical Protection  
Chad Geiger, DexCom Inc  
Becky Gilsdorf, Healthmark Industries Company LLC  
Dr. Zory Glaser, Johns Hopkins University-School of Public Health  
Douglas Harbrecht, Sterility Assurance LLC  
Deborah Havlik, DA Havlik Consulting  
Jason Hedrick, Medtronic Inc Company  
Betty Howard, STERIS Corporation | Healthcare  
Henri Hubert, Quality Process and Resource Group LLC  
Timothy Hurtado  
Nupur Jain, Intuitive Surgical Inc  
Valerie Kozik, ProGanex  
Mark Krocko, Medtronic Corporation  
Byron Lambert, Abbott Laboratories  
John LeClair  
Jeffrey Martin, Sterilization and Quality System Consulting LLC  
Mauricio Martinez, Smiths Medical  
Patrick McCormick, Bausch & Lomb Inc  
Liam McLaren, Boston Scientific Corporation  
Nicole McLees, 3M Health Care  
Susan Messier, Ethide Laboratories  
Vanessa Molloy-Simard, Stryker Instruments Division  
Melissa Morgan  
Paul Nirgenau, Mesa Laboratories Biological Indicator Division- Bozeman Facility  
Susumu Nozawa, Siemens Healthineers  
Gerry O'Dell, Gerry O'Dell Consulting  
Ken Paddock, Baxter Healthcare Corporation  
David Parente, Sterwell LLC  
Kimberly Patton, Becton Dickinson & Company  
Michael Polozani, Case Medical Inc  
Claudia Romero-Waas, BSI Healthcare  
Tyrone Rouse, Owens & Minor

Jody Rupert, WL Gore & Associates Inc  
Manuel Saavedra, Avanos Medical  
Harry Shaffer, Sterilization Consulting Services  
Myra Smith, FDA/CDRH  
Sopheak Srun, Quality Tech Services LLC  
Elizabeth Steiner, IUVO BioScience  
Aiyng Sun, Verrix LLC  
Julie TerWee, Pfizer Parenteral Center of Excellence  
Donald Tumminelli, HIGHPOWER Validation Testing & Lab Services Inc  
Stephanie Volk, ConvaTec Inc QARAC Global Headquarters  
Sean Weir  
Dean Wellman, Midmark Corporation  
Beverly Whitaker, Indigo Consulting Group LLC  
Jarl Yeager, Powder River Medical Resources  
Shelley Zimmerman, WuXi AppTec Inc

*Alternates:*

Kenneth Aston, Intuitive Surgical Inc  
Ralph Basile, Healthmark Industries Company Inc  
Stacy Bohl, Boston Scientific Corporation  
Charles Cogdill, Medtronic Inc Campus  
Chris Coker, Advanced Sterilization Products (ASP)  
Kimbrell Darnell, Becton Dickinson & Company  
Mark Duro  
Shawn Engberg, MD Furniture & Supply  
Chris Haas, Getinge USA  
John Harper, Sterile Edge  
Nichole Jackson, Ecolab  
Melinda Johnson, WuXi AppTec Inc  
Satu King, Philips  
Sarith Koruprolu, Ethide Laboratories  
Jean-Luc Lemyre, Stryker Instruments Division  
Michelle (Shelly) Luebke, Baxter Healthcare Corporation  
Gerald McDonnell, Johnson & Johnson  
Gloria Narsayah, Propper Manufacturing Co Inc  
Brian Nissen, Quality Tech Services LLC  
Mike Nolan, HIGHPOWER Validation Testing & Lab Services Inc  
Janelle Pence, WL Gore & Associates Inc  
Jason Pope, Sotera Health LLC  
Victoria Rodriguez, FDA/CDRH  
Jason Rogers, STERIS Corporation | Healthcare  
Michael Schoene, Bausch & Lomb Inc  
Mark Sundt, DuPont Tyco Medical and Pharmaceutical Protection  
Molly Thompson, STERIS Health Care  
Hemant Vekaria, Abbott Laboratories  
Cynthia Vics, STS Healthcare  
Jill Warren, Biogen Healthineers  
James Whitcomb, LexaMed Ltd

---

NOTE Participation by federal agency representatives in the development of this document does not constitute endorsement by the federal government or any of its agencies.

---

At the time this technical information report was published, the **AAMI Sterilization Standards Committee** had the following members:

*Cochairs:* Janet Prust  
Patrick Weixel

*Members:* Gregory Aldrich  
Anas Aljabo, CMC Sterilization Ltd  
Brett Anderson, Cochlear Ltd  
Jennifer Asleson, Quality, Microbiology & Sterilization Services LLC  
Richard Bancroft, STERIS Corporation | Healthcare

Marie Brewer  
Trabue Bryans, BryKor LLC  
Jonathan Burdach, Nanosonics Limited  
Charles Cogdill, Medtronic Inc Campus  
Lena Cordie-Bancroft, Qualitas Professional Services LLC  
Emily Craven, Boston Scientific Corporation  
Jacqueline Daley  
Gordon Ely, LexaMed Ltd  
Lisa Foster, Adiuvio QS & SA Consulting  
Daniel Fowler, WuXi AppTec Inc  
Joel Gorski, NAMSA  
Magnus Graham, BSI Healthcare  
Ashley Hammer  
Joyce Hansen, Johnson & Johnson  
Douglas Harbrecht, Sterility Assurance LLC  
John Harper, Sterile Edge  
Deborah Havlik, DA Havlik Consulting  
Mollie Holter, MicroBio Consulting LLC  
Clark Houghtling, Cosmed Group Inc  
Susan Klacik, IAHCMM  
Erin Kyle, Association of Perioperative Registered Nurses (AORN)  
Byron Lambert, Abbott Laboratories  
Patrick McCormick, Bausch & Lomb Inc  
Leslie Nichols, Mayo Clinic  
Gerry O'Dell, Gerry O'Dell Consulting  
Tarika Onishi, DexCom Inc  
Ken Paddock, Baxter Healthcare Corporation  
Kia Parker, The Ohio State Wexner Medical Center-University Hospital  
Renee Patterson, CIVCO Medical Solutions  
Kimberly Patton, Becton Dickinson & Company  
Janet Prust, 3M Health Care  
Nancy Rakiewicz, IUVO BioScience  
Joan Rickard, Montefiore Medical Center  
Linda Schultz, Northside Hospital Surgical Services Atlanta  
Kristen Singleton, Getinge USA  
Joan Spear, Joan M. Spear (Independent Expert)  
Sopheak Srun, Quality Tech Services, LLC  
Mark Swanson, Quality and Regulatory Expert Partners (QRX)  
Julie TerWee, Pfizer Parenteral Center for Excellence  
James Treharn, St Peter's Hospital  
Marjorie Wall, Kaiser Permanente Foundation - Downey  
Sean Weir  
Patrick Weixel, FDA/CDRH  
James Wigglesworth, Leggett Health System  
Martell Winters, Sotera Health LLC  
Roberto Zurruaño, Philips

*Alternates:* Josh Abdo, CIVCO Medical Solutions  
Jennifer Benolken, DuPont Tyvek Medical and Pharmaceutical Protection  
Cecily Bohl, Boston Scientific Corporation  
Carolyn Braithwaite-Nelson, Philips  
Greg Crego, IUVO BioScience  
Kimbrell Darnell, Becton Dickinson & Company  
Aaron DeMent, Sotera Health LLC  
Chad Geiger, DexCom Inc  
Anna Grayson, Grayson Associates  
David Kelman, Sterile Edge  
Michelle (Shelly) Luebke, Baxter Healthcare Corporation  
Gerald McDonnell, Johnson & Johnson  
David McGoldrick, Abbott Laboratories  
Nicole McLees, 3M Health Care  
Christine Render, Cosmed Group Inc

Jimmy Reyes, Alcon Laboratories Inc  
Mary Sheehan, BSI Healthcare  
Molly Swanson, Quality Tech Services LLC  
Lisa Ward, STERIS Corporation | Healthcare  
Jon Wood, IAHCMM  
Shelley Zimmerman, WuXi AppTec Inc

---

NOTE Participation by federal agency representatives in the development of this technical information report does not constitute endorsement by the federal government or any of its agencies.

---

## Foreword

The following verbal forms are used within AAMI documents to distinguish requirements from other types of provisions in the document:

- “shall” and “shall not” are used to express requirements;
- “should” and “should not” are used to express recommendations;
- “may” and “may not” are used to express permission;
- “can” and “cannot” are used as statements of possibility or capability;
- “might” and “might not” are used to express possibility;
- “must” is used for external constraints or obligations defined outside the document; “must” is not an alternative for “shall.”

Suggestions for improving this document are invited. Comments and suggested revisions should be sent to Standards, AAMI, 901 N. Glebe Road, Suite 300, Arlington, VA 22203 or [standards@aami.org](mailto:standards@aami.org).

---

NOTE This foreword does not contain provisions of the AAMI TIR100, *End-to-end microbiological quality and sterility assurance*, but it does provide important information about the development and intended use of the document.

---

## Introduction


This AAMI Technical Information Report (TIR) was developed to describe a comprehensive framework that integrates and connects Microbiological Quality & Sterility Assurance (MQ&SA) into an organization's end-to-end process for the development, validation, and routine control of sterile health care products.

The Supply Chain for a health care product is an end-to-end process that begins and ends with the customer. The health care product requirements are established based on the specific customer needs, and, in the end, the customer is the ultimate user of the health care product. The end-to-end process is defined as the following: Research and Development (R&D), Plan, Source, Make, Deliver, & Customer (See Figure 1). Addressing MQ&SA throughout the end-to-end process is critical to ensure that the finished health care product meets its defined requirements.



Figure 1—End-to-end process

The interactions and information flow within the end-to-end process are interwoven and not necessarily linear as shown in Figure 1. For smaller companies, the same personnel could perform numerous activities whilst in a large company, there might be specific groups of individuals that perform each activity within the end-to-end process. Additionally, smaller companies might not have as definitive distinctions between the activities shown. Understanding the roles and responsibilities within the organization is important to facilitate collaboration. An example of the interactions and information flow between the individuals that provide these functions is shown in Figure 2.



**Figure 2—Interactions and information flow within the end-to-end process**

Interactions and information flow between personnel that support the different aspects of the end-to-end process will ensure that the product is designed to meet customer expectations. R&D works with the customer to understand their expectations and translate their needs into product requirements. R&D in collaboration with the rest of the Supply Chain (e.g., Plan, Source, Make, and Deliver) will develop the product plan for manufacturing (a.k.a. production) and delivery to the customer. Throughout the end-to-end process, a feedback loop is important between R&D, Plan, Source, Make and Deliver to ensure the product continues to meet the needs of the customer. This collaboration and information flow can be used for internally and externally manufactured products, or during the assessment and integration of products acquired through mergers or acquisitions.

MQ&SA requirements should be integrated throughout the end-to-end process rather than just at the beginning and end of the process. For example, if during the product design phase, MQ&SA is only considered once the design has been finalized, this can have significant impact on cost or product launch timelines (e.g., launch delays) due to need to modify product design to address MQ&SA requirements identified during the final design review. In addition, if the Plan for manufacturing has already been developed, it might not include considerations for critical facility design criteria, as well as microbiological controls (e.g., air, water, personnel).

This technical information report (TIR) provides guidance on a framework that integrates and connects MQ&SA during the entire end-to-end process for terminally-sterilized and aseptically-processed products. As new product innovations challenge current terminal sterilization and aseptic processing, the need for continued focus on the end-to-end process to meet the MQ&SA needs for future products becomes more critical.

# End-to-end microbiological quality and sterility assurance

## 1 Scope

This document provides guidance on integration of Microbiological Quality & Sterility Assurance (MQ&SA) into the end-to-end process. This framework is intended to connect the MQ&SA activities that should be addressed throughout the end-to-end process: from defining MQ&SA requirements during R&D through customer receipt and use of the product.

NOTE 1 These principles can be applied to other microbiologically controlled products. For example, certain hand care products, wound dressings, drainage bags, etc.

NOTE 2 While reusable devices are not included in the scope of this document, the general principles of design inputs, design process, design outputs, design verification and design validation can be applied.

## 2 Normative references

There are no normative references for this document.

## 3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

### 3.1

#### **aseptic processing**

handling of sterile product, containers, and/or the product in a controlled environment in which the air supply, materials, equipment, and personnel are regulated to maintain sterility.

[SOURCE: ISO 11139:2018, 3.14]

### 3.2

#### **clean**

visually free of soil and below specified levels of analytes

[SOURCE: ISO 11139:2018, 3.45]

### 3.3

#### **customer**

person (e.g., patient, clinician, health care provider) or organization that receives a product

### 3.4

#### **deliver**

provide ordered services or products to the customer

### 3.5

#### **disinfection**

process to inactivate viable microorganisms to a level previously specified as being appropriate for a defined purpose

[SOURCE: ISO 11139:2018, 3.84]