

Technical Information Report

AAMI TIR15: 2016/(R)2024

Physical aspects of ethylene oxide
sterilization

Currently in preview, click buy full version

Physical aspects of ethylene oxide sterilization

Approved 18 November 2016 and reaffirmed 11 September 2020 and 26 March 2024 by
AAMI

Abstract: This technical information report (TIR) provides additional guidance to augment the ANSI/AAMI/ISO 11135 series. Topics covered in this TIR are sterilization equipment, considerations for preconditioning, calculations for relative humidity, calculations of ethylene oxide concentration, and flammability.

Keywords: sterilization equipment, ethylene oxide, relative humidity, flammability, preconditioning

AAMI Technical Information Report

A technical information report (TIR) is a publication of the Association for the Advancement of Medical Instrumentation (AAMI) Standards Board that addresses a particular aspect of medical technology.

This document is not an American National Standard, and the material contained herein is not normative in nature.

Although the material presented in a TIR may need further evaluation by experts, releasing the information is valuable because the industry and the professions have an immediate need for it.

A TIR differs markedly from a standard or recommended practice, and readers should understand the differences between these documents.

Standards and recommended practices are subject to a formal process of committee approval, public review, and resolution of all comments. This process of consensus is supervised by the AAMI Standards Board and, in the case of American National Standards, by the American National Standards Institute.

A TIR is not subject to the same formal approval process as a standard. However, a TIR is approved for distribution by a technical committee and the AAMI Standards Board.

Another difference is that, although both standards and TIRs are periodically reviewed, a standard must be acted on—reaffirmed, revised, or withdrawn—and the action formally approved usually every 5 years but at least every 10 years. For a TIR, AAMI consults with a technical committee about 5 years after the publication date (and periodically thereafter) for guidance on whether the document is still useful—that is, to check that the information is relevant or of historical value. If the information is not useful, the TIR is removed from circulation.

A TIR may be developed because it is more responsive to underlying safety or performance issues than a standard or recommended practice, or because achieving consensus is extremely difficult or unlikely. Unlike a standard, a TIR permits the inclusion of differing viewpoints on technical issues.

CAUTION NOTICE: This AAMI TIR may be revised or withdrawn at any time. Because it addresses a rapidly evolving field or technology, readers are cautioned to ensure that they have also considered information that may be more recent than this document.

All standards, recommended practices, technical information reports, and other types of technical documents developed by AAMI are *voluntary*, and their application is solely within the discretion and professional judgment of the user of the document. Occasionally, voluntary technical documents are adopted by government regulatory agencies or procurement authorities, in which case the adopting agency is responsible for enforcement of its rules and regulations.

Comments on this technical information report are invited and should be sent to AAMI, Attn: Standards Department, 901 N. Glebe Road, Suite 300, Arlington, VA 22203.

Published by

AAMI
901 N. Glebe Road, Suite 300
Arlington, VA 22203
www.aami.org

© 2017 by the Association for the Advancement of Medical Instrumentation

All Rights Reserved

Publication, reproduction, photocopying, storage, or transmission, electronically or otherwise, of all or any part of this document without the prior written permission of the Association for the Advancement of Medical Instrumentation is strictly prohibited by law. It is illegal under federal law (17 U.S.C. § 101, *et seq.*) to make copies of all or any part of this document (whether internally or externally) without the prior written permission of the Association for the Advancement of Medical Instrumentation. Violators risk legal action, including civil and criminal penalties, and damages of \$100,000 per offense. For permission regarding the use of all or any part of this document, contact the Copyright Clearance Center.

Printed in the United States of America

ISBN 978-1-57020-653-5

Contents

	Page
Glossary of equivalent standards	iv
Committee representation	v
Foreword	viii
1 Scope.....	1
2 Definitions	1
3 EO sterilization processing equipment.....	2
4 Determination of minimum product temperature prior to preconditioning	8
5 Calculation of moisture content.....	9
6 Calculation of EO concentration	12
7 Calculations for use of flammability curve.....	18
8 Guide for statistical comparison to determine process equivalence	22
Bibliography	26
Figure	
1 EO Cycle Flammability Diagram.....	19
Tables	
1 List of constants and molecular weights of some common EO/diluent combinations	12
2 Gas constants (R).....	13
3 Properties of saturated steam—Commonly used temperatures	16
4 Flammability review, partial pressures, and volume percentage.....	21

Glossary of equivalent standards

International Standards adopted in the United States may include normative references to other International Standards. AAMI maintains a current list of each International Standard that has been adopted by AAMI (and ANSI). Available on the AAMI website at the address below, this list gives the corresponding U.S. designation and level of equivalency to the International Standard.

www.aami.org/standards/glossary.pdf

Currently in preview, click buy full versi

Committee representation

Association for the Advancement of Medical Instrumentation

Industrial Ethylene Oxide Sterilization Working Group

This technical information report (TIR) was developed by the AAMI Industrial Ethylene Oxide Sterilization Working Group under the auspices of the AAMI Sterilization Standards Committee. Working Group approval of the TIR does not necessarily imply that all committee members voted for its approval.

At the time this document was published, the **AAMI Industrial Ethylene Oxide Sterilization Working Group** had the following members:

Cochairs: Jonathan Bull
Jeffrey Martin

Members: Erin Armstrong, B Braun Medical
Edward Arscott, NAMSA
David Ballard, Dynatec Scientific Labs
Stacy Bohl, Boston Scientific Corporation
Anne F. Booth, MS, Booth Scientific
Carolyn Braithwaite-Nelson, Spectranetics
Jonathan Bull, Johnson & Johnson
Tim Carlson, BD Medical
Sarah Chamberlain, Accuratus Labs Services
Dennis Christensen, BS, Sterilization Validation & Consulting
Gary N. Cranston, Consulting & Technical Services/PC
Elaine Daniell, Bard Medical Division
Douglas D. Davie, Sterilization Validation Services
Darci Diage, Direct Flow Medical
David Dion, Cardinal Health
Mary Ann Drosnock, MS, Healthmark Industries
William F. FitzGerald, PE, FitzGerald & Associates Ltd.
Dan B. Floyd, RM, Dupont Protective Solutions
Lisa Foster, Adiuvo QS & SA Consulting
Matthew Freeman, Terumo BSC
Zory R. Glaser, PhD MPH CS-CM, Johns Hopkins University
Michael Groendyk, ArthroX
Douglas Harbrecht, Sterility Assurance
Arthur C. Harris, Cook
Deborah A. Havlik, Copira, a Pfizer Company
Jason Hedrick, Medtronic
Clark W. Hougliding, Cosmed Group
Krista Howard, WL Gore & Associates
Naipur Jaiswal, Inuitive Surgical
Carolyn L. Kinsley, LexaMed
Kerem A. Kowalczyk, Centurion Sterilization Services
Christine Loshbaugh, Edwards LifeSciences
Mollie Love, Smiths Medical
Jeffrey Martin, Sterilization and Quality System Consulting
Albert May, Andersen Products Inc.
Patrick McCormick, PhD, Bausch & Lomb
David Ford McGoldrick, BS, Abbott Vascular Devices
Russell D. Mills, GE Healthcare
Gerry A. O'Dell, MS, Gerry O'Dell Consulting
Dave Parente, ECOLAB Healthcare
Michelle Peterson, Stryker Instruments
Andrew Porteous, Baxter Healthcare
Nancy Rakiewicz, IUVO BioScience
Keith Reiner, Terumo Cardiovascular Systems

Beth Ridgeway, Mesa Laboratories
Manuel Saavedra Jr., Halyard Health
Harry Shaffer, Sterilization Consulting Services
Arnold Shechtman, BS, Validation Challenges Consulting
David Silor, Zimmer
Bill South, Isomedix Services
Michael G. Sprague, Ethide Laboratories
Sopheak Srun, MPH SM(NRCM), Quality Tech Services
Fenil Sutaria, Medline Industries
Mara Tafoya, WuXi AppTec
Radhakrishna S. Tirumalai, US Pharmacopeial Convention
Steven E. Turtill, FDA/CDRH
Craig A. Wallace, 3M Healthcare
P. Richard Warburton, ChemDAQ
Richard L. Weisman, Fresenius Medical Care
Beverly Whitaker, CQA RAB MBA, Indigo Consulting Group
Dennis L. Wildes, St. Jude Medical
William T. Young, Sterigenics International
Roberto Zumbado, Sr., Philips

Alternates:

August Baur, Centurion Sterilization Services
MarJean Boyter, Fresenius Medical Care
Robert Bradley, CBET, Mesa Laboratories
Trabue D. Bryans, BryCor
Greg Bush, Alcon Research
Liza Chenette, Accuratus Labs Solutions
Sean N. Colwell, WuXi Aptec
Allison Diemert, Arthrex
David Dominguez, Becton Dickinson
Brian Drumheller, Brookhaven Medical
Venice Eldred, Medline Industries
Steven Elliott, FDA/CDRH
Diane Faivre-Swiat, Cardinal Health
Mark Finley, Boston Scientific
Naomi Gamm, St. Jude Medical
Scott Giraud, Medtronic Cardiac Surgical Products
William K. Gordon, Steris Isomedix Services
David M. Hilliker, ChemDAQ
Brent Huberty, Boston Scientific
Nicole Jackson, Ecolab
Satu King, Spectranetics
Chris Kobus, GE Healthcare
Ezra Koski, Terumo BCT
James P. Kuller, BS, MS, LexaMed
Wesley Lantz, Zimmer
John Linney, Andersen Products Inc.
Paul L. Linn, BSE, Nelson Laboratories
Mauricio Martinez, Smiths Medical
Nicole M. McLees, 3M Critical & Chronic Care Solutions
Joseph Mello, Ethide Laboratories
Laurie Nawrocki, NAMSA
Koyejo Obadina, Abbott Vascular
Ken Paddock, Baxter Healthcare
Michael A. Padilla, SteriPro Labs
Tyrone S. Rouse, Halyard Health
Michael Shoene, Bausch & Lomb
Kristen Spigiel, Stryker Instruments
Elizabeth Steiner, IUVO BioScience
Brian Wallace, Intuitive Surgical
Scott Weiss, Ethicon
Daryl Woodman, Andersen Products
Casimir John Wos, PhD, FitzGerald & Associates

AAMI also acknowledges the **TIR15 Editing Task Group**, comprising the following members, for its special contribution in the development of this document:

Task group leader Greg Bush, Alcon Research

Members: Gary Cranston, Consulting & Technical Services/PCS
 William F. Fitzgerald, Fitzgerald & Associates
 Arnold Shechtman, BS, Validation Challenges Consulting
 Steven E. Turtill, FDA/CDRH

NOTE—Participation by federal agency representatives in the development of this technical information report does not constitute endorsement by the federal government or any of its agencies.

Currently in preview, click buy full version

Foreword

This document is part of a series of technical information reports (TIRs) intended for use in conjunction with ANSI/AAMI/ISO 11135:2014. The other reports in the series are:

- AAMI TIR14:2009, *Contract sterilization using ethylene oxide*;
- AAMI TIR16:2009, *Microbiological aspects of ethylene oxide sterilization*;
- AAMI TIR28:2009, *Product adoption and process equivalence for ethylene oxide sterilization*;
- AAMI TIR56:2013, *Guidance for the development, validation and routine control of an ethylene oxide sterilization process utilizing flexible bag systems for the sterilization of medical devices*.

The original TIR15, along with other AAMI TIRs, provided additional guidance to the 1994 edition of the industrial ethylene oxide sterilization standard ANSI/AAMI/ISO 11135, which was revised in 2007 and 2014 under a new designation, ANSI/AAMI/ISO 11135:2014, *Sterilization of health care products—Ethylene oxide - Requirements for development, validation, and routine control of a sterilization process for medical devices*. In 2008 and 2014, the International Organization for Standardization (ISO) published its own guidance document for the 11135 standard, ISO/TR 11135-2:2008, *Sterilization of health care products—Ethylene oxide—Part 2: Guidance on the application of ISO 11135-1*, which was based to a great extent on the earlier AAMI TIRs. Correspondingly, the AAMI Industrial Ethylene Oxide Sterilization Working Group is updating its TIRs to take into account changes to the 11135 standard and to avoid redundancy with ANSI/AAMI/ISO 11135:2014.

This TIR contains guidelines that are not intended to be absolute or to apply in all circumstances. One should use judgment in applying the information in this TIR.

As used within the context of this document, “should” indicates that among several possibilities, one is recommended as particularly suitable, without mentioning or excluding others, or that a certain course of action is preferred but not necessarily required, or that (in the negative form) a certain possibility or course of action should be avoided but is not prohibited. “May” is used to indicate that a course of action is permissible within the limits of the technical information report. “Can” is used as a statement of possibility and capability. Finally, “must” is used only to describe “unavoidable” situations, including those mandated by government regulation. See also the NOTE on Page 1.

Suggestions for improving this technical information report are invited. Comments and suggested revisions should be sent to AAMI, 901 N. Glebe Road, Suite 300, Arlington, VA 22203.

NOTE—This foreword does not contain provisions of AAMI TIR15:2016, *Physical aspects of ethylene oxide sterilization*, but it does provide important information about the development and intended use of the document.

Physical aspects of ethylene oxide sterilization

NOTE—This technical information report (TIR) is not a standard and the material contained herein is not normative in nature. In a few instances, the committee has used the term “shall” based on its knowledge of requirements contained in relevant standards, regulatory requirements, or both.

1 Scope

This AAMI TIR provides additional guidance to augment ANSI/AAMI/ISO 11135:2014. This TIR addresses some of the physical aspects of ethylene oxide (EO) sterilization and provides guidance on sterilization equipment, considerations for preconditioning, how to calculate relative humidity, EO concentration and flammability, as well as guidance on the use of statistics for process equivalence.

2 Definitions

For the purposes of this TIR, the terms and definitions in ANSI/AAMI/ISO 11135:2014 and the following apply:

2.1

dew point

temperature at which the saturation water vapor pressure is equal to the partial pressure of the water vapor in the atmosphere

NOTE Any cooling of the atmosphere below the dew point would produce water condensation.

2.2

humidity

A measure of the water vapor present in a gas. It is usually measured as absolute humidity, relative humidity, or dew point temperature.

2.3

preconditioning area

chamber or room under defined temperature and humidity conditions for the purpose of conditioning the product prior to entering the gas exposure phase of the process.

2.4

relative humidity (RH)

Ratio of water vapor partial pressure actually present to water vapor pressure required for saturation at a given temperature.

$$\%RH = \frac{\text{actual vapor pressure}}{\text{saturation vapor pressure}} \times 100$$

2.5

chamber humidification

phase of the sterilization process whereby moisture is added to the chamber to replace any environmental moisture potentially lost in the vacuum stages for air removal or inert gas dilution.

2.6

steam purity

measurement of the amount of solid, liquid, or vaporous contamination in the steam.

Note 1 to entry: Normally, steam purity is reported as the solids content. The steam purity needed for an EO sterilization system is defined by the equipment user and needs to be suitable for its intended use.