

## Product adoption and process equivalency for ethylene oxide sterilization

## Product adoption and process equivalency for ethylene oxide sterilization

Approved 24 December 2001 by  
Association for the Advancement of Medical Instrumentation

### **Abstract.**

This AAMI technical information report (TIR) provides guidance for the adoption of a new or modified product into an existing validated sterilization process and for the determination of equivalency of a sterilization process as conducted in different equipment. Its guidance is intended to augment ANSI/AAMI/ISO 11135:1994, *Medical devices—Validation and routine control of ethylene oxide sterilization*, and to expand on the areas of product adoption and process equivalency that are not addressed in ANSI/AAMI/ISO 11135:1994.

### **Keywords:**

adoption, equivalency, process equivalency, product adoption, product family

## AAMI Technical Information Report

A technical information report (TIR) is a publication of the Association for the Advancement of Medical Instrumentation (AAMI) Standards Board that addresses a particular aspect of medical technology.

Although the material presented in a TIR may need further evaluation by experts, releasing the information is valuable because the industry and the professions have an immediate need for it.

A TIR differs markedly from a standard or recommended practice, and readers should understand the differences between these documents.

Standards and recommended practices are subject to a formal process of committee approval, public review, and resolution of all comments. This process of consensus is supervised by the AAMI Standards Board and, in the case of American National Standards, by the American National Standards Institute.

A TIR is not subject to the same formal approval process as a standard. However, a TIR is approved for distribution by a technical committee and the AAMI Standards Board.

Another difference is that, although both standards and TIRs are periodically reviewed, a standard must be acted on—reaffirmed, revised, or withdrawn—and the action formally approved usually every 5 years but at least every 10 years. For a TIR, AAMI consults with a technical committee about 5 years after the publication date (and periodically thereafter) for guidance on whether the document is still useful—that is, to check that the information is relevant or of historical value. If the information is not useful, the TIR is removed from circulation.

A TIR may be developed because it is more responsive to underlying safety or performance issues than a standard or recommended practice, or because achieving consensus is extremely difficult or unlikely. Unlike a standard, a TIR permits the inclusion of differing viewpoints on technical issues.

**CAUTION NOTICE:** This AAMI TIR may be revised or withdrawn at any time. Because it addresses a rapidly evolving field or technology, readers are cautioned to ensure that they have also considered information that may be more recent than this document.

All standards, recommended practices, technical information reports, and other types of technical documents developed by AAMI are *voluntary*, and their application is solely within the discretion and professional judgment of the user of the document. Occasionally, voluntary technical documents are adopted by government regulatory agencies or procurement authorities, in which case the adopting agency is responsible for enforcement of its rules and regulations.

Comments on this technical information report are invited and should be sent to AAMI, Attn: Standards Department, 1110 N. Glebe Road, Suite 220, Arlington, VA 22201-4795.

### *Published by*

Association for the Advancement of Medical Instrumentation  
1110 N. Glebe Road, Suite 220  
Arlington, VA 22201-4795

© 2002 by the Association for the Advancement of Medical Instrumentation

All Rights Reserved

Publication, reproduction, photocopying, storage, or transmission, electronically or otherwise, of all or any part of this document without the prior written permission of the Association for the Advancement of Medical Instrumentation is strictly prohibited by law. It is illegal under federal law (17 U.S.C. § 101, *et seq.*) to make copies of all or any part of this document (whether internally or externally) without the prior written permission of the Association for the Advancement of Medical Instrumentation. Violators risk legal action, including civil and criminal penalties, and damages of \$100,000 per offense. For permission regarding the use of all or any part of this document, contact AAMI at 1110 N. Glebe Road, Suite 220, Arlington, VA 22201-4795. Phone: (703) 525-4890; Fax: (703) 525-1067.

Printed in the United States of America

ISBN 1-57020-173-0

## Contents

	Page
Glossary of equivalent standards .....	iv
Committee representation .....	vi
Acknowledgments .....	viii
Introduction.....	ix
<b>1</b> Scope .....	1
<b>2</b> References and bibliography .....	2
<b>3</b> Terms and definitions .....	2
<b>4</b> Product adoption .....	2
<b>4.1</b> Introduction.....	2
<b>4.2</b> Families .....	3
<b>4.3</b> Determination of adverse effects.....	3
<b>4.4</b> Determination of sterilization challenge.....	3
<b>4.4.1</b> Product.....	3
<b>4.4.2</b> Packaging system.....	4
<b>4.4.3</b> Cycle lethality studies .....	4
<b>4.5</b> Documentation .....	4
<b>5</b> Process equivalency .....	5
<b>5.1</b> General.....	5
<b>5.2</b> Introduction.....	5
<b>5.3</b> Determination of process equivalency .....	5
<b>5.3.1</b> Phase 1—Design and engineering evaluation.....	6
<b>5.3.2</b> Phase 2—Process analysis and evaluation.....	6
<b>5.3.3</b> Phase 3—Microbiological evaluation.....	7
<b>5.3.4</b> Results evaluation.....	7
<b>5.4</b> Process requalification and maintenance of equivalency.....	7
<b>5.5</b> Documentation .....	8
<b>Annexes</b>	
<b>A</b> Guide for evaluation of a product adoption .....	9
<b>B</b> Guide for design and engineering evaluation of candidate equipment .....	11
<b>Table</b>	
<b>1</b> Evaluation results .....	7

## Glossary of equivalent standards

International Standards adopted in the United States may include normative references to other International Standards. For each International Standard that has been adopted by AAMI (and ANSI), the table below gives the corresponding U.S. designation and level of equivalency to the International Standard.

Note—Documents are sorted by international designation.

Other normatively referenced International Standards may be under consideration for U.S. adoption by AAMI; therefore, this list should not be considered exhaustive.

International designation	U.S. designation	Equivalency
IEC 60601-1-2:2001	ANSI/AAMI/IEC 60601-1-2:2001	Identical
IEC 60601-2-21:1994 and Amendment 1:1996	ANSI/AAMI/IEC 60601-2-21 & Amendment 1:2000 (consolidated texts)	Identical
IEC 60601-2-24:1998	ANSI/AAMI ID26:1998	Major technical variations
ISO 5840:1996	ANSI/AAMI/ISO 5840:1996	Identical
ISO 7198:1998	ANSI/AAMI/ISO 7198:1998/2001	Identical
ISO 7199:1996	ANSI/AAMI/ISO 7199:1996	Identical
ISO 10993-1:1997	ANSI/AAMI/ISO 10993-1:1997	Identical
ISO 10993-2:1992	ANSI/AAMI/ISO 10993-2:1993/(R)2001	Identical
ISO 10993-3:1992	ANSI/AAMI/ISO 10993-3:1993	Identical
ISO 10993-4:1992	ANSI/AAMI/ISO 10993-4:1993	Identical
ISO 10993-5:1999	ANSI/AAMI/ISO 10993-5:1999	Identical
ISO 10993-6:1994	ANSI/AAMI/ISO 10993-6:1995/(R)2001	Identical
ISO 10993-7:1995	ANSI/AAMI/ISO 10993-7:1995/(R)2001	Identical
ISO 10993-8:2000	ANSI/AAMI/ISO 10993-8:2000	Identical
ISO 10993-9:1999	ANSI/AAMI/ISO 10993-9:1999	Identical
ISO 10993-10:1995	ANSI/AAMI/ISO 10993-10:1995	Identical
ISO 10993-11:1993	ANSI/AAMI 10993-11:1993	Minor technical variations
ISO 10993-12:1996	ANSI/AAMI/ISO/CEN 10993-12:1996	Identical
ISO 10993-13:1998	ANSI/AAMI/ISO 10993-13:1999	Identical
ISO 10993-14:2001	ANSI/AAMI/ISO 10993-14:2001	Identical
ISO 10993-15:2000	ANSI/AAMI/ISO 10993-15:2000	Identical
ISO 10993-16:1991	ANSI/AAMI/ISO 10993-16:1997	Identical
ISO 11134:1994	ANSI/AAMI/ISO 11134:1993	Identical
ISO 11135:1994	ANSI/AAMI/ISO 11135:1994	Identical
ISO 11137:1995	ANSI/AAMI/ISO 11137:1994	Identical
ISO 11138-1:1994	ANSI/AAMI ST59:1999	Major technical variations
ISO 11138-2:1994	ANSI/AAMI ST21:1999	Major technical variations
ISO 11138-3:1995	ANSI/AAMI ST19:1999	Major technical variations

International designation	U.S. designation	Equivalency
ISO 11140-1:1995 and Technical Corrigendum 1:1998	ANSI/AAMI ST60:1996	Major technical variations
ISO 11607: 200x <sup>1</sup>	ANSI/AAMI/ISO 11607:2000	Identical
ISO 11737-1:1995	ANSI/AAMI/ISO 11737-1:1995	Identical
ISO 11737-2:1998	ANSI/AAMI/ISO 11737-2:1998	Identical
ISO TR 13409:1996	AAMI/ISO TIR13409:1996	Identical
ISO 13485:1996	ANSI/AAMI/ISO 13485:1996	Identical
ISO 13488:1996	ANSI/AAMI/ISO 13488:1996	Identical
ISO 14155:1996	ANSI/AAMI/ISO 14155:1996	Identical
ISO 14160:1998	ANSI/AAMI/ISO 14160:1998	Identical
ISO 14161:2000	ANSI/AAMI/ISO 14161:2000	Identical
ISO 14937:2000	ANSI/AAMI/ISO 14937:2000	Identical
ISO 14969:1999	ANSI/AAMI/ISO 14969:1999	Identical
ISO 14971:2000	ANSI/AAMI/ISO 14971:2000	Identical
ISO 15223:2000	ANSI/AAMI/ISO 15223:2000	Identical
ISO 15223/A1:2001	ANSI/AAMI/ISO 15223:2000/A1:2001	Identical
ISO 15225:2000	ANSI/AAMI/ISO 15225:2000	Identical
ISO 15674:2001	ANSI/AAMI/ISO 15674:2001	Identical
ISO 15675:2001	ANSI/AAMI/ISO 15675:2001	Identical
ISO TS 15843:2000	ANSI/AAMI/ISO TIR15843:2000	Identical
ISO TR 15844:1998	AAMI/ISO TIR15844:1998	Identical
ISO TR 16142:1999	ANSI/AAMI/ISO TIR16142:2000	Identical

<sup>1</sup> FDIS approved; being prepared for publication.

## Committee representation

### Association for the Advancement of Medical Instrumentation

#### AAMI Sterilization Standards Committee

This technical information report was developed by the AAMI Industrial Ethylene Oxide Sterilization Working Group under the auspices of the AAMI Sterilization Standards Committee. Approval of the TIR does not necessarily imply that all working group members voted for its approval.

At the time this document was published, the **AAMI Sterilization Standards Committee** had the following members:

<i>Cochairs:</i>	Victoria Hitchins, PhD William E. Young
<i>Members:</i>	Trabue D. Bryans, AppTec Laboratory Services Virginia C. Chamberlain, PhD, Hendersonville, NC Anne M. Cofiell, CRST, IAHCSSM Loretta L. Fauerbach, MS, CIC, Association for Professionals in Infection Control and Epidemiology Dorothy M. Fogg, RN, BSN, MA, Association of Perioperative Nurses Lisa Foster, Ion Beam Applications James M. Gibson, Jr., JM Gibson Associates Barbara J. Goodman, RN, CNOR, Rising Sun, MD Joel R. Gorski, PhD, NAMSA Susan Hadfield, Canadian Standards Association Deborah A. Havlik, Abbott Laboratories Victoria Hitchins, PhD, U.S. Food and Drug Administration Clark W. Houghtling, Cosmed Group, Inc. Lois A. Jones, Cary, NC Sue Kuhnert, Sterilization Technical Services Byron J. Lambert, PhD, Guidant Corporation Sandra A. Lee, RN, STERIS Corporation Patrick J. McCormick, PhD, Bausch & Lomb, Inc. Thomas K. Moore, Getinge/Castle, Inc. Robert F. Morrisey, PhD, Johnson & Johnson David Orton, CR Bard Inc. Barry F.J. Page, Garner, NC Phil M. Schneider, 3M Healthcare Michael H. Scholla, PhD, DuPont Tyvek for Sterile Packaging Dupont Nonwovens Janet K. Schultz, MSN, RN, Kennesaw, GA Harry L. Shaffer, Titan Corp., Titan Scan Technologies Frank Sizemore, American Society for Healthcare Central Service Professionals William N. Thompson, Tyco Healthcare James L. Whitty, MA, MB, FRCP, University of Western Ontario Thelma Wilcott, Becton Dickinson & Company Stephen C. Yeardon, Alcon Laboratories, Inc. William E. Young, Baxter Healthcare Corporation
<i>Alternates:</i>	Bettye Becher, Alcon Laboratories, Inc. Louis M. Glasgow, Bausch & Lomb, Inc. Royce M. Hansen, Baxter Healthcare Corporation Susan G. Klacik, ACE, IAHCSSM Chiu Lin, PhD, U.S. Food and Drug Administration Lisa N. Macdonald, Becton Dickinson & Company Ralph Makinen, Guidant Corporation Janet Prust, 3M Healthcare James Whitbourne, Sterilization Technical Services William T. Young, Ion Beam Applications

At the time this document was published, the **AAMI Industrial Ethylene Oxide Sterilization Working Group** had the following members:

<i>Cochairs:</i>	James M. Gibson, Jr. Carolyn L. Kinsley
------------------	--

*Members:*

Kenneth W. Allen, Arrow International, Inc.  
 Lauren Andersen, Andersen Products, Inc.  
 Krisann Anderson, St. Jude Medical, Inc.  
 Leonard S. Bermann, PhD, Pall Corporation  
 Anne F. Booth, Barrington, IL  
 Lloyd Brown, TYCO Healthcare  
 Delores Bruce, Northview Biosciences, Inc.  
 Karla Byrne, Getinge/Castle, Inc.  
 Virginia C. Chamberlain, PhD, Hendersonville, NC  
 Charles Cogdill, Boston Scientific Corporation  
 Julee Comer, Ethox Corporation  
 Stephen A. Conviser, Honeywell International  
 Gary N. Cranston, CATS/PCS  
 Christine A. Czap, Fresenius Medical Care NA Dialysis Production Division  
 Roger Dabbah, PhD, U.S. Pharmacopeial Convention, Inc.  
 Douglas D. Davie, Sterilization Validation Services  
 Shawn A. Doyle, Sterilator Company, Inc.  
 Dan B. Floyd, Nelson Laboratories, Inc.  
 James M. Gibson, Jr., JM Gibson Associates  
 Zory R. Glaser, PhD, Johns Hopkins University, School of Public Health  
 Louis M. Glasgow, Bausch & Lomb, Inc.  
 Arthur C. Harris, Chicago Sterilization Services  
 Deborah A. Havlik, Abbott Laboratories  
 Clark W. Houghtling, Cosmed Group, Inc.  
 Carolyn L. Kinsley, Baxter Healthcare Corporation  
 Ralph Makinen, Guidant Corporation  
 Mary S. Mayo, CR Bard, Inc.  
 James E. McGowan, Kimberly-Clark Corporation  
 Craig A. Meadows, Medtronic, Inc.  
 Alexander Mello, Microtest Laboratories  
 Joseph M. Mello, Ethide Laboratories, Inc.  
 Russell D. Mills, Zimmer, Inc.  
 Sarah Mowitt, U.S. Food and Drug Administration  
 Gerry A. O'Dell, Wesley Chapel, FL  
 Dave Parente, NAMSA  
 Robert R. Reich, Pharmaceutical Systems, Inc.  
 Zenius V. Seliokas, Stericon, Inc.  
 Jon Seulean, Gambro BCT, Inc.  
 Paul J. Sordellini, Quality Solutions, Inc.  
 Bill South, STERIS Corporation  
 Ralph Stick, AppTec Laboratory Services  
 Wayne Swallow, Becton Dickinson & Company  
 James Whitbourne, Sterilization Technical Services  
 Casimir John Wood, PhD, Johnson & Johnson  
 Stephen C. Yeaman, Aicon Laboratories, Inc.  
 William T. Young, X-ray Beam Applications

*Alternates:*

Richard H. Beal, Zimmer, Inc.  
 Bettye Beebe, Aicon Laboratories, Inc.  
 John Brundage, NAMSA  
 Truhue D. Evans, AppTec Laboratory Services  
 Jennifer Carey, Bausch & Lomb, Inc.  
 John DiCaro, TYCO Healthcare  
 Joseph R. Durbin, Abbott Laboratories  
 Doug F. Harbrecht, Scimed-Boston Scientific Corporation  
 Donna Horner, Guidant Corporation  
 J. Michael Kuchinski, U.S. Food and Drug Administration  
 Sue Kuhnert, Sterilization Technical Services  
 Gary Mitchel, Johnson & Johnson  
 Ken Paddock, Allegiance Healthcare Corporation  
 Timothy Ramsey, Northview Biosciences  
 Christine L. Render, Cosmed Group, Inc.  
 Manny Saavedra, Kimberly-Clark Corporation  
 Mark Seybold, Baxter Healthcare Corporation  
 Theresa Snoozy, Becton Dickinson & Company  
 Richard L. Weisman, Fresenius Medical Care NA Dialysis Products Division

## Acknowledgments

The AAMI Industrial Ethylene Oxide Working Group gratefully acknowledges the work of its Industrial Product/Process Equivalency Task Group, which handled the preballot development of this TIR. The task group has the following members:

*Chairs:* Joseph M. Mello  
Gerry A. O'Dell

*Members:* Kenneth W. Allen, Arrow International, Inc.  
Judy Anderson, Smith and Nephew  
Michael M. Ball, TYCO Healthcare/U.S. Surgical Corporation  
Jerome Bell, Baxter International, Inc.  
Charles Cogdill, Boston Scientific Corporation  
Gary N. Cranston, Taunton, MA  
Anthony J. DeMarinis, Davol, Inc.  
Deborah A. Havlik, Abbott Laboratories  
James E. McGowan, Jr., Kimberly-Clark Corporation  
Joseph M. Mello, Ethide Laboratories, Inc.  
Gerry A. O'Dell, MS, Wesley Chapel, FL  
Michael A. Padilla, Ion Beam Applications  
Dave Parente, NAMSA  
Paul J. Sordellini, Quality Solutions, Inc.  
Bill South, STERIS Corporation  
Wayne Swallow, Becton Dickinson & Company  
Christine L. Sootkoos, Cosmed Group, Inc.  
Casimir J. Woss, PhD, Johnson & Johnson  
Stephen C. Yeadon, Alcon Laboratories, Inc.

---

NOTE—Participation by federal agency representatives in the development of this TIR does not constitute endorsement by the federal government or any of its agencies.

---

## Introduction

This document is part of a series of reports intended to be used in conjunction with ANSI/AAMI/ISO 11135:1994, *Medical devices—Validation and routine control of ethylene oxide sterilization*. The other reports in the series are:

- AAMI TIR14:1997, *Contract sterilization for ethylene oxide*
- AAMI TIR15:1997, *Ethylene oxide sterilization equipment, process considerations and pertinent calculations*
- AAMI TIR16:2000, *Process development and performance qualification for ethylene oxide sterilization—Microbiological aspects*, and
- AAMI TIR 20:2001, *Parametric release for ethylene oxide sterilization*.

This technical information report provides guidance for the adoption of new or modified products into an existing validated sterilization process and for the determination of equivalency of the sterilization process as conducted with different equipment. Although these areas are not specifically addressed by ANSI/AAMI/ISO 11135:1994 (AAMI, 1994), they are important industry practices that are used to reduce the expense and time associated with the validation process, and are based on accumulated process knowledge.

The adoption of a new or modified product into an existing validated sterilization process involves the determination that the product is no more of a challenge than the product that was used to validate the ethylene oxide (EO) sterilization process. Product adoption has been a longstanding practice in the industry. Although it has been addressed in individual papers (Lowery and DeRiso, 1982; Burgess and Reich, 1983) it has not been addressed in a guidance document. Therefore, this TIR will address how product can be adopted into an existing EO process.

The process equivalency section of this TIR will provide guidance on the level of validation testing required on the basis of the equivalence of the sterilization process and/or equipment. It will also provide guidance on how to determine the equivalence of the process and/or equipment.

NOTE—This TIR is considered “informative,” and the use of the terms “shall,” “should,” and so forth should be considered within the context of this TIR only. That is, if the decision is made to use a particular method presented in this TIR, then the method should be followed with adherence to the requirements (“shall”) and recommendations (“should”) as set forth in this TIR. The term “must” refers to regulatory requirements.



# Product adoption and process equivalency for ethylene oxide sterilization

## 1 Scope

This TIR addresses medical devices that are processed by ethylene oxide sterilization using conventional or parametric product release. The document applies to the following situations for the sterilization of medical devices:

- a new product is being added to the previously validated process;
- changes to validated products are being evaluated;
- a previously validated process is being moved to a different facility and/or equipment; and
- equivalency of a sterilization process is being evaluated.

Although the information presented was developed for application to medical devices, the content of this guideline may also be applied to other relevant products or materials. This document does not address the equivalency of two or more sterilization processes run in the same or different sterilization process equipment.

## 2 References and bibliography

ANSI/AAMI/ISO 10993-7:1995, *Biological evaluation of medical devices—Part 7: Ethylene oxide sterilization residuals*, 2ed.

ANSI/AAMI/ISO 11135:1994, *Medical devices—Validation and routine control of ethylene oxide sterilization*, 3ed.

ANSI/AAMI/ISO 11737-1:1995, *Sterilization of medical devices—Microbiological methods—Part 1: Estimation of population of microorganisms on products*, 1ed.

AAMI ST67:200X,<sup>2</sup> *Sterilization of medical devices—Requirements for products labeled 'STERILE.'* (In preparation.)

AAMI TIR14:1997, *Contract sterilization for ethylene oxide*, 1ed.

AAMI TIR15:1997, *Ethylene oxide sterilization equipment, process considerations, and pertinent calculations*, 1ed.

AAMI TIR16:2000, *Process development for ethylene oxide sterilization—Microbiological aspects*, 1ed.

AAMI TIR19:1998, *Guidance for ANSI/AAMI/ISO 10993-7:1995, Biological evaluation of medical devices—Part 7: Ethylene oxide sterilization residuals*, 1ed.

EN550:1994, *Sterilization of medical devices—Validation and routine control of ethylene oxide sterilization*.

Burgess DJ and Reich RR. Industrial Ethylene Oxide Sterilization. In Morrissey RF and Phillips GB, eds. *Sterilization technology: A practical guide for manufacturers and users of health care products*. New York, NY: Van Nostrand Reinhold, 1993.

Gillis J and Schmidt WC. Scanning electron microscopy of spores on inoculated product surfaces. *Medical Device and Diagnostic Industry*. June 1983, vol. 5, no. 6, pp. 46–49.

Lowery A and DeRisio R. *Adopting a device into a validated sterilization system*. Text of presentation at HIMA conference, April 1982, by A. Lowery.

West KL. Ethylene oxide sterilization: *A study of resistance relationships*. In Gaughran E and Kereluk K, eds. *Sterilization of medical products*. New Brunswick, NJ: Johnson & Johnson, 1977.

<sup>2</sup> AAMI expects to publish this standard in the second quarter of 2002 following completion of appropriate approval processes.