

# Technical Information Report

## **AAMI TIR28 2016/(R)2024**

Product adoption and process equivalence  
for ethylene oxide sterilization

Currently in preview, click buy full version

## Product adoption and process equivalence for ethylene oxide sterilization

Approved 18 November 2016 and reaffirmed 11 September 2020 and 26 March 2024 by  
AAMI

**Abstract:** This technical information report provides guidance for the adoption of new or modified products into an existing validated sterilization process and for the determination of equivalence of a sterilization process as conducted with different equipment. Guidance is intended to augment the ANSI/AAMI/ISO 11135 series in the areas of product adoption and process equivalence.

**Key words:** sterilization, ethylene oxide, product adoption, process equivalence, product family

## AAMI Technical Information Report

A technical information report (TIR) is a publication of the Association for the Advancement of Medical Instrumentation (AAMI) Standards Board that addresses a particular aspect of medical technology.

Although the material presented in a TIR may need further evaluation by experts, releasing the information is valuable because the industry and the professions have an immediate need for it.

A TIR differs markedly from a standard or recommended practice, and readers should understand the differences between these documents.

Standards and recommended practices are subject to a formal process of committee approval, public review, and resolution of all comments. This process of consensus is supervised by the AAMI Standards Board and, in the case of American National Standards, by the American National Standards Institute.

A TIR is not subject to the same formal approval process as a standard. However, a TIR is approved for distribution by a technical committee and the AAMI Standards Board.

Another difference is that, although both standards and TIRs are periodically reviewed, a standard must be acted on—reaffirmed, revised, or withdrawn—and the action formally approved usually every five years, but at least every 10 years. For a TIR, AAMI consults with a technical committee about five years after the publication date (and periodically thereafter) for guidance on whether the document is still useful—that is, to check that the information is relevant or of historical value. If the information is not useful, the TIR is removed from publication.

A TIR may be developed because it is more responsive to underlying safety or performance issues than a standard or recommended practice or because achieving consensus is extremely difficult or unlikely. Unlike a standard, a TIR permits the inclusion of differing viewpoints on technical issues.

**CAUTION NOTICE:** This AAMI TIR may be revised or withdrawn at any time. Because it addresses a rapidly evolving field or technology, readers are cautioned to ensure that they have also considered information that may be more recent than this document.

All standards, recommended practices, technical information reports, and other types of technical documents developed by AAMI are *voluntary*, and their application is solely within the discretion and professional judgment of the user of the document. Occasionally, voluntary technical documents are adopted by government regulatory agencies or procurement authorities, in which case the adopting agency is responsible for enforcement of its rules and regulations.

Comments on this technical information report are invited and should be sent to AAMI, Attn: Standards Department, AAMI, 901 N. Glebe Road, Suite 300, Arlington, VA 22203.

*Published by*

AAMI  
901 N. Glebe Road, Suite 300  
Arlington, VA 22203  
[www.aami.org](http://www.aami.org)

© 2017 by the Association for the Advancement of Medical Instrumentation

All Rights Reserved

Publication, reproduction, photocopying, storage, or transmission, electronically or otherwise, of all or any part of this document without the prior written permission of the Association for the Advancement of Medical Instrumentation is strictly prohibited by law. It is illegal under federal law (17 U.S.C. § 101, *et seq.*) to make copies of all or any part of this document (whether internally or externally) without the prior written permission of the Association for the Advancement of Medical Instrumentation. Violators risk legal action, including civil and criminal penalties, and damages of \$100,000 per offense. For permission regarding the use of all or any part of this document, contact the Copyright Clearance Center.

Printed in the United States of America

ISBN 978-1-57020-654-2

## Contents

Page

|  |     |
|--|-----|
| Glossary of equivalent standards.....  | iv  |
| Committee representation.....  | v   |
| Foreword.....  | vii |
| 1 Scope.....   | 1   |
| 2 Terms and definitions.....   | 1   |
| 3 Product adoption.....  | 2   |
| 3.1 Establishment of an EO product family.....   | 2   |
| 3.2 Establishment of EO processing category.....   | 3   |
| 3.3 Evaluation for product adoption.....   | 3   |
| 3.3.1 Determination of adverse effects.....  | 3   |
| 3.3.2 Determination of product design effects.....   | 3   |
| 3.3.3 Determination of product material and characteristics effects.....   | 3   |
| 3.3.4 Determination of sterile barrier system effects.....   | 4   |
| 3.3.5 Determination of load configuration effects.....   | 4   |
| 3.4 Conclusions of product adoption evaluation.....  | 4   |
| 3.4.1 Comparative resistance study.....  | 4   |
| 3.4.2 Temperature and humidity distribution study.....   | 5   |
| 3.4.3 Residual study.....  | 5   |
| 3.5 Maintenance of EO product adoption.....  | 5   |
| 3.6 Documentation.....   | 5   |
| 4 Process equivalence.....   | 6   |
| 4.1 General.....   | 6   |
| 4.2 Requirements of process equivalence.....   | 6   |
| 4.3 Determination of process equivalence.....  | 6   |
| 4.3.1 Process analysis and evaluation.....   | 6   |
| 4.3.2 Microbiological evaluation.....  | 8   |
| 4.3.3 Results evaluation.....  | 8   |
| 4.4 Process requalification and maintenance of equivalence.....  | 8   |
| 4.5 Documentation.....   | 8   |
| Annex A Guide for evaluation of a product for adoption into an EO product family or<br>EO processing category..... | 10  |
| Bibliography.....  | 12  |

## **Glossary of equivalent standards**

International Standards adopted in the United States may include normative references to other International Standards. AAMI maintains a current list of each International Standard that has been adopted by AAMI (and ANSI). Available on the AAMI website at the address below, this list gives the corresponding U.S. designation and level of equivalency to the International Standard.

[www.aami.org/standards/glossary.pdf](http://www.aami.org/standards/glossary.pdf)

Currently in preview, click buy full version

## Committee representation

### Association for the Advancement of Medical Instrumentation

#### Industrial Ethylene Oxide Sterilization Working Group

This technical information report (TIR) was developed by the AAMI Industrial Ethylene Oxide Sterilization Working Group under the auspices of the AAMI Sterilization Standards Committee. Working Group approval of the TIR does not necessarily imply that all committee members voted for its approval.

At the time this document was published, the **AAMI Industrial Ethylene Oxide Sterilization Working Group** had the following members:

*Cochairs:* Jonathan Bull, Johnson & Johnson  
Jeffrey Martin, Sterilization and Quality System Consulting

*Members:* Erin Armstrong, B Braun Medical  
Edward Arcscott, NAMSA  
David Ballard, Dynatec Scientific Labs  
Stacy Bohl, Boston Scientific Corporation  
Anne F. Booth, MS, Booth Scientific  
Carolyn Braithwaite-Nelson, Spectranetics  
Jonathan Bull, Johnson & Johnson  
Tim Carlson, BD Medical  
Sarah Chamberlain, Accuratus Labs Solutions  
Dennis Christensen, BS, Sterilization Validation & Consulting  
Gary N. Cranston, Consulting & Technical Services/PCS  
Elaine Daniell, Bard Medical Division  
Douglas D. Davie, Sterilization Validation Services  
Darci Diage, Direct Flow Medical  
David Dion, Cardinal Health  
Mary Ann Drosnock, MS, Healthmark Industries  
Paul Fioriti, PF Quality Consulting  
William F. FitzGerald, PE, FitzGerald & Associates Ltd.  
Dan B. Floyd, RM, Dupont Protection Solutions  
Lisa Foster, Adiuvo QS & SA Consulting  
Matthew Freeman, Terumo BCT  
Zory R. Glaser, PhD MPH CSPDM, Johns Hopkins University  
Michael Groendyk, Arthrex  
Douglas Harbrecht, SteriFit Assurance  
Arthur C. Harris, Cook  
Deborah A. Havlik, Hesperia, a Pfizer Company  
Jason Hedrick, Medtronic  
Clark W. Houghtling, Cosmed Group  
Krista Howard, WELCore & Associates  
Naipur Jain, InVivo Surgical  
Carolyn M. Kinley, LexaMed  
Karen A. Kowalczyk, Centurion Sterilization Services  
Christine Lombough, Edwards LifeSciences  
Mille Love, Smiths Medical  
Jeffrey Martin, Sterilization and Quality System Consulting  
Robert May, Andersen Products Inc.  
Patrick McCormick, PhD, Bausch & Lomb  
David Ford McGoldrick, BS, Abbott Vascular Devices  
Russell D. Mills, GE Healthcare  
Gerry A. O'Dell, MS, Gerry O'Dell Consulting  
Dave Parente, ECOLAB Healthcare  
Michelle Peterson, Stryker Instruments  
Andrew Porteous, Baxter Healthcare  
Nancy Rakiewicz, IUVO BioScience  
Keith Reiner, Terumo Cardiovascular Systems  
Beth Ridgeway, Mesa Laboratories  
Manuel Saavedra Jr., Halyard Health  
Harry Shaffer, Sterilization Consulting Services

Arnold Shechtman, BS, Validation Challenges Consulting  
David Silor, Zimmer  
Bill South, Isomedix Services  
Michael G. Sprague, Ethide Laboratories  
Sopheak Srun, MPH SM(NRCM), Quality Tech Services  
Fenil Sutaria, Medline Industries  
Mara Tafoya, WuXi AppTec  
Radhakrishna S. Tirumalai, US Pharmacopeial Convention  
Steven E. Turtill, FDA/CDRH  
Craig A. Wallace, 3M Healthcare  
P. Richard Warburton, ChemDAQ  
Richard L. Weisman, Fresenius Medical Care  
Beverly Whitaker, CQA RAB MBA, Indigo Consulting Group  
Dennis L. Wildes, St. Jude Medical  
William T. Young, Sterigenics International  
Roberto Zumbado, Sr., Philips

*Alternates:* August Baur, Centurion Sterilization Services  
MarJean Boyter, Fresenius Medical Care  
Robert Bradley, CBET, Mesa Laboratories  
Trabue D. Bryans, BryCor  
Greg Bush, Alcon Research  
Liza Chenette, Accuratus Labs Solutions  
Sean N. Colwell, WuXi Aptec  
Allison Diemert, Arthrex  
David Dominguez, Becton Dickinson  
Brian Drumheller, Brookhaven Medical  
Venice Eldred, Medline Industries  
Steven Elliott, FDA/CDRH  
Diane Faivre-Swiat, Cardinal Health  
Mark Finley, Boston Scientific  
Naomi Gamm, St. Jude Medical  
Scott Giraud, Medtronic Cardiac Surgical Products  
William K. Gordon, Steris Isomedix Services  
David M. Hilliker, ChemDaq  
Brent Huberty, Boston Scientific  
Nicole Jackson, Ecolab  
Satu King, Spectranetics  
Chris Kobus, GE Healthcare  
Ezra Koski, Terumo BCT  
James P. Kulla, BS MS, L...  
Wesley Lantz, Zimmer  
John Lindley, Andersen Products Inc.  
Paul L. Littley, BS, Nelson Laboratories  
Mauricio Martinez, Smiths Medical  
Nicole M. Mc... Smith Critical & Chronic Care Solutions  
Joseph M... Ethide Laboratories  
Laurie M... NAMSA  
Koyejo O... Abbott Vascular  
Ken Paddock, Baxter Healthcare  
Michael A. Padilla, SteriPro Labs  
Tyne S. Rouse, Halyard Health  
Michael Shoene, Bausch & Lomb  
Kristen Spigiel, Stryker Instruments  
Elizabeth Steiner, IUVO BioScience  
Brian Wallace, Intuitive Surgical  
Scott Weiss, Ethicon  
Daryl Woodman, Andersen Products  
Casimir John Wos, PhD, FitzGerald & Associates

---

NOTE—Participation by federal agency representatives in the development of this technical information report does not constitute endorsement by the federal government or any of its agencies.

---

## Foreword

This document is part of a series of technical information reports (TIRs) intended for use in conjunction with ANSI/AAMI/ISO 11135:2014. The other reports in the series are

- AAMI TIR14:2009, Contract sterilization using ethylene oxide;
- AAMI TIR15:2009, Ethylene oxide sterilization equipment, process considerations, and pertinent calculations (currently under revision);
- AAMI TIR16:2009, Process development and performance qualification for ethylene oxide sterilization— Microbiological aspects (currently under revision); and
- AAMI TIR56:2013, Guidance for the development, validation and routine control of an ethylene oxide sterilization process utilizing flexible bag systems for the sterilization of medical devices

The original TIR28, along with other AAMI TIRs, provided additional guidance to the 1994 edition of the industrial EO sterilization standard 11135, which was revised in 2007 under a new designation, ANSI/AAMI/ISO 11135-1:2007, *Sterilization of health care products - Ethylene oxide - Part 1: Requirements for the development, validation and routine control of a sterilization process for medical devices*. In 2008, ISO published its own guidance document for the 11135 standard, ISO/TR 11135-2:2008, *Sterilization of health care products - Ethylene oxide - Part 2: Guidance on the application of ISO 11135-1*, which was based to a great extent on the earlier AAMI technical information reports. Correspondingly, the AAMI Industrial EO sterilization working group is updating its TIRs to take into account changes to the 11135 standard.

This TIR provides guidance for the adoption of new or modified products into an existing validated sterilization process and for the determination of equivalence of a sterilization process conducted with different equipment. These areas are not addressed in depth by ANSI/AAMI/ISO 11135:2014, but they are important industry practices that are used to reduce the expense and time associated with the validation or requalification of an ethylene oxide sterilization process and are based on accumulated process knowledge.

The adoption of a new or modified product into an existing validated sterilization process involves the determination that the product is no more of a challenge than the product (i.e. master product or representative product) or process challenge device that was used in the performance qualification for the ethylene oxide sterilization process.

The process equivalence section of this TIR will provide guidance on how to establish equivalence between processes performed in separate equipment or different equipment and guidance on the level of microbiological performance qualification testing required. It also includes guidance on the maintenance of equivalence and requalification when equivalence has been established.

This TIR contains guidelines that are not intended to be absolute or to apply in all circumstances. One should use judgment in applying the information in this TIR.

As used within the context of this document, “should” indicates that among several possibilities, one is recommended as particularly suitable, without mentioning or excluding others, or that a certain course of action is preferred but not necessarily required, or that (in the negative form) a certain possibility or course of action should be avoided but is not prohibited. “May” is used to indicate that a course of action is permissible within the limits of the recommended practice. “Can” is used as a statement of possibility and capability. Finally, “must” is used only to describe “unavoidable” situations, including those mandated by government regulation. See also the NOTE on Page 1.

Suggestions for improving this technical information report are invited. Comments and suggested revisions should be sent to AAMI, 101 N. Glebe Road, Suite 300, Arlington, VA 22203.

---

NOTE—This foreword does not contain provisions of the AAMI TIR28:2016 titled *Product adoption and process equivalence for ethylene oxide sterilization*, but it does provide important information about the development and intended use of the document.

---

Currently in preview, click buy full version

# Product adoption and process equivalence for ethylene oxide sterilization

NOTE—This technical information report is not a standard and the material contained herein is not normative in nature. The committee has in a few places used the term "shall" based on their knowledge of requirements contained in relevant standards and/or regulatory requirements.

## 1 Scope

This TIR addresses medical devices that are processed by ethylene oxide (EO) sterilization using conventional or parametric product release. The document applies to the following situations for the sterilization of medical devices:

- a) a new product is being added to the previously validated process,
- b) changes to validated products are being evaluated,
- c) a previously validated process is being moved to a different facility or to different equipment, and
- d) equivalency of a sterilization process is being evaluated.

Although the information presented was developed for application to medical devices, the content of this guideline may also be applied to other relevant products or materials.

## 2 Terms and definitions

For the purposes of this AAMI TIR, the following terms and definitions apply:

### 2.1

#### **candidate equipment**

new or modified piece of equipment intended to deliver an existing validated process.

### 2.2

#### **candidate product**

new or modified product, including the packaging system, proposed for inclusion in the existing validated sterilization process.

### 2.3

#### **EO processing category**

collection of different product or product families that can be sterilized together (in the same EO sterilization process).

NOTE—All products within the processing category have been determined to present an equal or lesser challenge to the sterilization process than the challenge device for that category.

### 2.4

#### **EO product family**

group of products possessing characteristics that allow them to be sterilized using the same defined process conditions and determined to be similar or equivalent for validation purposes.

### 2.5

#### **load configuration**

totality of attributes defining the presentation of the product to the sterilization process. This configuration includes (a) the orientation of the product within the sterile barrier system (primary package); (b) the quantity and orientation of the primary package within the protective packaging (secondary or tertiary package); (c) the quantity, orientation, and placement of the product in the protective packaging on the sterilizer pallets (or within the carriers); and (d) the quantity and placement of the pallets (or carriers) within the vessel or area.