

Technical Information Report

AAMI TIR61: 2014/(R)2019

Generating reports for
human factors design
validation results
for external cardiac
defibrillators

Generating reports for human factors design validation results for external cardiac defibrillators

Approved 24 October 2014 and reaffirmed 19 September 2019 by
AAMI

Abstract: Provides guidance on the formatting and content of reports generated for the purpose of submitting human factors data for evaluation.

keywords: defibrillation, human factors, usability, external defibrillator, automated external defibrillator

AAMI Technical Information Report

A technical information report (TIR) is a publication of the Association for the Advancement of Medical Instrumentation (AAMI) Standards Board that addresses a particular aspect of medical technology.

Although the material presented in a TIR may need further evaluation by experts, releasing the information is valuable because the industry and the professions have an immediate need for it.

A TIR differs markedly from a standard or recommended practice, and readers should understand the differences between these documents.

Standards and recommended practices are subject to a formal process of committee approval, public review, and resolution of all comments. This process of consensus is supervised by the AAMI Standards Board and, in the case of American National Standards, by the American National Standards Institute.

A TIR is not subject to the same formal approval process as a standard. However, a TIR is approved for publication by a technical committee and the AAMI Standards Board.

Another difference is that, although both standards and TIRs are periodically reviewed, a standard must be acted on—reaffirmed, revised, or withdrawn—and the action formally approved, usually every five years but at least every ten years. For a TIR, AAMI consults with a technical committee about five years after the publication date (and periodically thereafter) for guidance on whether the document is still useful; i.e., to check that the information is relevant or of historical value. If the information is not useful, the TIR is removed from circulation.

A TIR may be developed because it is more responsive to underlying safety or performance issues than a standard or recommended practice, or because achieving consensus is extremely difficult or unlikely. Unlike a standard, a TIR permits the inclusion of differing viewpoints on technical issues.

CAUTION NOTICE: This AAMI TIR may be revised or withdrawn at any time. Because it addresses a rapidly evolving field or technology, readers are cautioned to ensure that they have also considered information that may be more recent than this document.

All standards, recommended practices, technical information reports, and other types of technical documents developed by AAMI are *voluntary*, and their application is solely within the discretion and professional judgment of the user of the document. Occasionally, voluntary technical documents are adopted by government regulatory agencies or procurement authorities, in which case the adopting agency is responsible for enforcement of its rules and regulations.

Comments on this technical information report are invited and should be sent to AAMI, Attn: Standards Department, 901 N. Glebe Road, Suite 300, Arlington, VA 22203.

AAMI acknowledges the contributions of the AAMI Human Factors group in the development of this TIR.

Published by

AAMI
901 N. Glebe Road, Suite 300
Arlington, VA 22203
www.aami.org

© 2014 by the Association for the Advancement of Medical Instrumentation

All Rights Reserved

Publication, reproduction, photocopying, storage, or transmission, electronically or otherwise, of all or any part of this document without the prior written permission of the Association for the Advancement of Medical Instrumentation is strictly prohibited by law. It is illegal under federal law (17 U.S.C. § 101, *et seq.*) to make copies of all or any part of this document (whether internally or externally) without the prior written permission of the Association for the Advancement of Medical Instrumentation. Violators risk legal action, including civil and criminal penalties, and damages of \$100,000 per offense. For permission regarding the use of all or any part of this document, complete the request request form at www.aami.org or contact AAMI at 901 N. Glebe Road, Suite 300, Arlington, VA 22203. Phone: +1-703-525-4890; Fax: +1-703-525-1067.

Printed in the United States of America

ISBN 1-57020-568-x

Contents

Page

| | |
|--|----|
| Committee representation..... | iv |
| Foreword..... | v |
| Introduction..... | vi |
| 1 Scope..... | 1 |
| 2 Object..... | 1 |
| 3 Definitions..... | 1 |
| 3.1 Human factors (HF) engineering..... | 1 |
| 3.2 Usability..... | 1 |
| 3.3 User-centered design (UCD)..... | 1 |
| 3.4 User interface (UI)..... | 1 |
| 3.5 Human Factors Validation testing..... | 2 |
| 4 Human Factors Validation study..... | 2 |
| 5 Contents of a validation report..... | 2 |
| 5.1 Format..... | 2 |
| 5.2 Executive summary..... | 3 |
| 5.3 Purpose..... | 3 |
| 5.4 Methods..... | 3 |
| 5.5 Results..... | 6 |
| 5.6 Deviations and exclusions..... | 7 |
| 5.7 Discussion..... | 7 |
| 5.8 Conclusions..... | 7 |
| 6 References..... | 8 |
| Annex: Design Considerations and Best Practice..... | 9 |