

Technical Information Report

AAMI TIR75: 2019

Factors to consider when
multi-vendor devices
interact via an electronic
interface: Practical
applications and examples

Factors to consider when multi-vendor devices interact via an electronic interface: Practical applications and examples

Approved 24 May 2019 by
AAMI

Abstract: Guidance on factors that manufacturers should consider when designing, testing, and monitoring interoperable medical devices.

Keywords: interoperability, connectivity, risk management

AAMI Technical Information Report

A technical information report (TIR) is a publication of the Association for the Advancement of Medical Instrumentation (AAMI) Standards Board that addresses a particular aspect of medical technology.

Although the material presented in a TIR may need further evaluation by experts, releasing the information is valuable because the industry and the professions have an immediate need for it.

A TIR differs markedly from a standard or recommended practice, and readers should understand the differences between these documents.

Standards and recommended practices are subject to a formal process of committee approval, public review, and resolution of all comments. This process of consensus is supervised by the AAMI Standards Board and, in the case of American National Standards, by the American National Standards Institute.

A TIR is not subject to the same formal approval process as a standard. However, a TIR is approved for distribution by a technical committee and the AAMI Standards Board.

Another difference is that, although both standards and TIRs are periodically reviewed, a standard must be acted on—reaffirmed, revised, or withdrawn—and the action formally approved usually every five years but at least every 10 years. A TIR must be acted on and the action formally approved usually every three years.

A TIR may be developed because it is more responsive to underlying safety or performance issues than a standard or recommended practice, or because achieving consensus is extremely difficult or unlikely. Unlike a standard, a TIR permits the inclusion of differing viewpoints on technical issues.

CAUTION NOTICE: This AAMI TIR may be revised or withdrawn at any time. Because it addresses a rapidly evolving field or technology, readers are cautioned to ensure that they have also considered information that may be more recent than this document.

All standards, recommended practices, technical information reports, and other types of technical documents developed by AAMI are *voluntary*, and their application is solely within the discretion and professional judgment of the user of the document. Occasionally, voluntary technical documents are adopted by government regulatory agencies or procurement authorities, in which case the adopting agency is responsible for enforcement of its rules and regulations.

Comments on this technical information report are invited and should be sent to AAMI, Attn: Standards Department, 901 N. Glebe Road, Suite 300, Arlington, VA 22203.

Published by

AAMI
901 N. Glebe Road, Suite 300
Arlington, VA 22203
www.aami.org

© 2019 by the Association for the Advancement of Medical Instrumentation

All Rights Reserved

This publication is subject to copyright claims of AAMI. No part of this publication may be reproduced or distributed in any form, including an electronic retrieval system, without the prior written permission of AAMI. All requests pertaining to this document should be submitted to AAMI. It is illegal under federal law (17 U.S.C. § 101, et seq.) to make copies of all or any part of this document (whether internally or externally) without the prior written permission of the Association for the Advancement of Medical Instrumentation. Violators risk legal action, including civil and criminal penalties, and damages of \$100,000 per offense. For permission regarding the use of all or any part of this document, complete the permission request form at www.aami.org or contact AAMI, 901 N. Glebe Road, Suite 300, Arlington, VA 22203. Phone: +1-703-525-4890; Fax: +1-703-276-0793.

Printed in the United States of America

ISBN 978-1-57020-719-8

Contents

Page

| | |
|--|-----|
| Committee representation..... | iv |
| Foreword..... | vi |
| Introduction..... | vii |
| 1 Scope..... | 1 |
| 2 References..... | 1 |
| 3 Factors to consider..... | 1 |
| Annex A Controls—Example analysis..... | 9 |
| Annex B Alarms—Example analysis..... | 11 |
| Annex C Data transfer—Example analysis..... | 15 |
| Annex D Remote viewing—Example analysis..... | 18 |
| Annex E Maintenance activities—Example analysis..... | 20 |
| Annex F Special considerations: Software only devices, clouds, HIT, Wi-Fi..... | 22 |
| Annex G Special considerations: Mobile devices..... | 24 |
| Annex H Special considerations: Distributed platforms, computation, and information..... | 25 |

Currently in preview, click buy full version

Committee representation

Association for the Advancement of Medical Instrumentation Interoperability Working Group

The publication of AAMI TIR75 as a new American Technical Report was initiated by the AAMI Interoperability Working Group. At the time this document was published, the **AAMI Interoperability Working Group** had the following members:

Cochairs: Julian Goldman
Sandy Weininger

Members: Heather Agler, FDA/CDRH
Robert Boyer, Medtronic - MITG
Ryan Burke, AJW Technology Consultants Inc
Dave Deaven, GE Healthcare
Sherman Eagles, SoftwareCPR
Scott Eaton, Mindray DS USA Inc
Kurt Eliason, Smiths Medical
Plamena Entcheva-Dimitrov, Preferred Regulatory Consulting Inc
Salvador Fagasten, Becton Dickinson & Company
Anura Fernando, UL LLC
Mike Flis, Genentech Inc
Ken Fuchs, Draeger Inc
Rich Gardner, Baxter Healthcare Corporation
Stuart Gardner, SG & A Consulting Inc
Tim Gee, Canon Communications
Julian Goldman, Partners Healthcare
Aaron Goldmuntz, West Wireless Health Institute
George Gray, Ivenix Inc
Kerry Griffin, Stryker Instruments Division
Rick Hampton, Partners Healthcare
David Hirschorn, Staten Island University Hospital
Sam Jacques, Pen State Health
Mike Jaffe, Cardiorespiratory Consulting LLC
Michelle Jump, Nova Leah LLC
Joshua Kim, Hill-Rom Holdings Inc
Dennis Krabbe, ICU Medical Inc
Yimin Li, Abbott Laboratories
Joern Lubadel, B Braun of America Inc
Sherman McMurray, Clarian Health Partners
Julia Merritt
Geoff Pascoe
Steven F. Packer, Software Quality Consulting
Geetha Rao, Springborne Life Sciences
Tracy Raue, DocBox Inc
John Rhoads, Philips
Jeff Rinda, Hospira, a Pfizer company
Michael Robkin, Anakena Solutions, Inc.
Elliot Sloane, Center for Healthcare Information Research and Policy
Donna-Bea Tillman, Biologics Consulting Group
Hung Trinh, Department of Defense Veterans Affairs
Fei Wang, Fresenius Medical Care
Roy Wang, Micro Systems Engineering Inc
Sandy Weininger, FDA/CDRH
Victory Yin, Amgen Inc
Daidi Zhong, Chongqing University

Alternates: Steve Anthony, Ivenix Inc
Pat Baird, Philips
Justin Bushko, AJW Technology Consultants Inc
Andrea Carrillo, Baxter Healthcare Corporation
Marty Crnkovich, Fresenius Medical Care
Bruce Friedman, GE Healthcare
Kevin Kopp, Smiths Medical
Patty Krantz-Zuppan, Medtronic Inc Campus
Michael Myers, ICU Medical
Andrew Northup, Medical Imaging & Technology Alliance a Division of NEMA
Rick Schrenker, Massachusetts General Hospital
Robert Schulz, Amgen Inc
Xianyu Shea, Stryker Instruments Division
Eugene Vasseman, Kansas State University

The AAMI Interoperability working group would like to acknowledge the efforts of the Team Leaders for AAMI TIR75, Michelle Jump and Sandy Weinginger, who helped maintain the momentum amongst the other Interoperability Working Group activity.

NOTE—Participation by federal agency representatives in the development of this technical information report does not constitute endorsement by the federal government or any of its agencies.

Foreword

This technical information report (TIR) was developed by the AAMI Interoperability Working Group under the auspices of the AAMI Software Standards Committee.

Manufacturers of medical devices that are intended to interoperate have the responsibility to manage the risks of this functionality. In many cases, there are unique factors that manufacturers should consider when designing, testing, and monitoring interoperable medical devices.

The objective of this TIR is to provide guidance to medical device manufacturers on assessing and managing the risks associated with interoperable medical devices.

NOTE—This foreword does not contain provisions of AAMI TIR75, *Factors to consider when multi-vendor devices interact via an electronic interface: Practical applications and examples* (AAMI TIR75), but it does provide important information about the development and intended use of the document.

Introduction

A number of factors impact the management of risks in a multi-vendor system of interoperating medical and non-medical devices. It would be useful if individual manufacturers could consider them in a common and consistent manner when developing or adding connectivity-based functionality to their products. Some factors are unique to connected products. This document provides a discussion of the context for important factors and questions that manufacturers should consider when designing, testing, installing, and maintaining connected medical devices. Consistent consideration of a structured set of factors can facilitate exchange of interoperability-related information, especially those factors impacting risks, among the stakeholders involved in development, deployment, and use of connected devices.

Failure can occur in the connection and connected functionality that impact the risk profile of a device in many ways, including issues of data quality, availability, authenticity, and integrity, as well as problems such as data lockup, backlogged data, and dropped data. Functionality that depends on connections includes command and control, remote access, remote software updates, and other functions controlled via connections. The factors described in this document are intended to shed light on how a manufacturer may better understand and plan for these risks. Evaluation of the impact of failure of connected functionality needs to include the potential to impact overall device functionality, not just the failure of connected functionality.

This technical information report is primarily intended to aid manufacturers as they design and develop products intended to interact with other devices via an electronic interface. It may also be useful as a reference when integrating devices from multiple vendors into a system via an electronic interface; maintaining a multi-vendor, interoperable system; auditing, inspecting, and evaluating the factors involved in an interoperable system; and selecting devices or systems for procurement or implementation.

Factors to consider when multi-vendor devices interact via an electronic interface: Practical applications and examples

1 Scope

1.1 General

This document is intended to assist stakeholders in considering risks associated with connectivity when designing, testing, installing, and maintaining devices that interact via an electronic interface. It identifies specific factors that should be considered as part of risk management activities. It also provides examples where these factors are used to identify causes, hazards, and hazardous situations related to interoperability. These factors can be relevant throughout the lifecycle of connected products.

This list of factors is not intended to be comprehensive. Instead, it is meant to direct design teams to think about common elements requiring dedicated attention when assessing the risk of the connected functionality of the medical device. These elements are referred to as, “Factors to Consider.”

This TIR applies to the following:

- devices with added or modified connectivity; and
- newly developed devices that are intended to connect via an electronic interface for the purposes of interoperability.

1.2 Exclusions

This TIR does not specifically consider the following:

- devices not connected by an electronic interface; and
- basic issues of medical device safety, not directly related to data interoperability, e.g., electrical safety.

2 References

The following referenced documents are useful for the application of this document. For undated references, the latest edition of the referenced document (including any amendments) applies. Members of IEC and ISO maintain registers of currently valid International Standards.

ISO 14971, *Medical devices— Application of risk management to medical devices*

Design Considerations and Pre-market Submission Recommendations for Interoperable Medical Devices—Guidance for Industry and Food and Drug Administration Staff. U.S. Food and Drug Administration. September 2017.

IEC 60601-1-8, *Medical electrical equipment—Part 1-8: General requirements for safety—Collateral standard: General requirements, test and guidance for alarm systems in medical electrical equipment and medical electrical systems*

3 Factors to consider

The interconnection of devices and systems in a healthcare setting presents the potential for significant innovation in healthcare, ranging from simple tracking of inventory to medical devices that interact autonomously to control functions that protect patients' lives. However, incorporating an electronic interface into a medical device also poses risks that are typically not posed by an unconnected device.