

# American National Standard

*for Ophthalmics –  
Laser Systems  
for Corneal Reshaping*



Currently in preview, click buy full version

**ANSI®**  
**Z80.11-2012 (R2017)**  
Reaffirmation of  
ANSI Z80.11-2012

American National Standard  
for Ophthalmics –

**Laser Systems for  
Corneal Reshaping**

Sponsor

**The Vision Council**

Approved August 9, 2012  
Reaffirmed December 11, 2017

**American National Standards Institute, Inc.**

# American National Standard

Approval of an American National Standard requires review by ANSI that the requirements for due process, consensus, and other criteria for approval have been met by the standards developer.

Consensus is established when, in the judgement of the ANSI Board of Standards Review, substantial agreement has been reached by directly and materially affected interests. Substantial agreement means much more than a simple majority, but not necessarily unanimity. Consensus requires that all views and objections be considered, and that a concerted effort be made towards their resolution.

The use of American National Standards is completely voluntary; their existence does not in any respect preclude anyone, whether he has approved the standards or not, from manufacturing, marketing, purchasing, or using products, processes, or procedures not conforming to the standards.

The American National Standards Institute does not develop standards and will in no circumstances give an interpretation of any American National Standard. Moreover, no person shall have the right or authority to issue an interpretation of an American National Standard in the name of the American National Standards Institute. Requests for interpretations should be addressed to the secretariat or sponsor whose name appears on the title page of this standard.

**CAUTION NOTICE:** This American National Standard may be revised or withdrawn at any time. The procedures of the American National Standards Institute require that action be taken periodically to reaffirm, revise, or withdraw this standard. Purchasers of American National Standards may receive current information on all standards by calling or writing the American National Standards Institute.

## *Developed by*

The Accredited Committee Z80 for Ophthalmic Standards -

The Vision Council  
Z80 Secretariat  
225 Reinekers Lane, Suite 700  
Alexandria, VA 22314

## *Published by*

The Vision Council  
225 Reinekers Lane, Suite 700  
Alexandria, VA 22314

Copyright © 2017 by The Vision Council  
All rights reserved.

No part of this publication may be reproduced in any form, in an electronic retrieval system or otherwise, without prior written permission of the publisher.

Printed in the United States of America

## Contents

|  | Page |
|--|------|
| Foreword.....  | iv   |
| <b>1</b> Scope and purpose .....                                   | 1    |
| <b>2</b> Normative references.....                                 | 1    |
| <b>3</b> Definitions .....   | 3    |
| <b>4</b> Mechanical, thermal, and environmental requirements ..... | 5    |
| 4.1 Combination of different devices.....                          | 5    |
| 4.2 Materials .....  | 5    |
| 4.3 Resistance to transport and storage conditions.....            | 5    |
| <b>5</b> Safety requirements.....                                  | 6    |
| 5.1 Protection against contaminants.....                           | 6    |
| 5.2 Protection against toxins and allergens .....                  | 6    |
| 5.3 Photobiological hazards.....                                   | 6    |
| 5.4 Thermal hazards.....   | 6    |
| 5.5 Mechanical hazards.....  | 6    |
| 5.6 Electrical safety.....   | 7    |
| 5.7 Radiation safety .....   | 7    |
| 5.7.1 Light hazards .....  | 7    |
| 5.8 Gas safety (for gas lasers).....                               | 7    |
| 5.9 Safety in use .....  | 7    |
| 5.10 System hazard analysis.....                                   | 8    |
| <b>6</b> Optical requirements.....                                 | 8    |
| 6.1 Alignment system.....  | 8    |
| 6.2 Laser safety monitoring.....                                   | 8    |
| 6.3 System calibration.....  | 8    |
| <b>7</b> System control and performance .....                      | 9    |
| 7.1 Software.....  | 9    |
| <b>8</b> Clinical evaluation .....                                 | 9    |
| 8.1 Clinical investigation plan.....                               | 9    |
| 8.2 Surgical procedure.....  | 9    |
| 8.3 Reporting periods and evaluations .....                        | 9    |
| 8.4 Adverse events .....   | 10   |
| <b>9</b> Test methods .....  | 10   |
| 9.1 Verification of alignment system accuracy.....                 | 10   |
| 9.1.1 Materials .....  | 10   |
| 9.1.2 Procedure .....  | 10   |

|                | Page  |
|----------------|---|
| 9.2            | Verification of cylinder axis alignment..... 10   |
| 9.2.1          | Materials ..... 10  |
| 9.2.2          | Procedure ..... 11  |
| <b>10</b>      | Accompanying documents..... 11  |
| <b>11</b>      | Marking ..... 11  |
| <b>Annexes</b> |   |
| <b>A</b>       | Spectral weighting function for ultraviolet radiation hazard analysis . 12  |
| <b>B</b>       | Methods for system calibration ..... 14   |
| B.1            | Plastic plate ablation and measurement..... 14  |
| B1.1           | Materials ..... 14  |
| B.1.2          | Procedure ..... 14  |
| B.2            | Laminated calibration plate method..... 14  |
| B.2.1          | Materials ..... 14  |
| B.2.2          | Procedure ..... 15  |
| <b>C</b>       | Characterization of laser ablation beams and treatment patterns ..... 16  |
| C.1            | Ablation characteristics of the beam..... 16  |
| C.2            | Mathematical models and simulations ..... 17  |
| C.3            | Validation of ablation algorithm software ..... 17  |
| <b>D</b>       | Guidance on clinical study design of refractive procedures<br>that use laser systems for corneal reshaping ..... 18 |
| D.1            | General ..... 18  |
| D.2            | Study objectives..... 18  |
| D.3            | Design of clinical study ..... 19   |
| D.4            | Study duration..... 19  |
| D.5            | Enrollment of subjects ..... 19   |
| D.6            | Inclusion and exclusion for subject selection..... 20   |
| D.6.1          | Inclusion criteria..... 20  |
| D.6.2          | Exclusion criteria..... 20  |
| D.7            | Examination schedule..... 21  |
| D.8            | Evaluations and methodology..... 23   |
| D.8.1          | Visual acuity and manifest refraction ..... 23  |
| D.8.2          | Measurement of intraocular pressure ..... 24  |
| D.8.3          | Subject questionnaire ..... 25  |
| D.8.4          | Mesopic pupil size ..... 25   |
| D.8.5          | Contrast sensitivity..... 25  |
| D.8.5.1        | Grating contrast sensitivity testing ..... 25   |

|               | Page  |
|---------------|---|
| D.8.6         | Low contrast letter acuity testing..... 26  |
| D.8.7         | Specular microscopy..... 27   |
| D.9           | Adverse device effects / Adverse events ..... 29  |
| <b>E</b>      | Statistical sample size considerations..... 30  |
| E.1           | Statistical symbols and definitions ..... 30  |
| E.2           | Calculation of necessary sample sizes ..... 31  |
| E.2.1         | Sample size based on safety estimates..... 31   |
| E.2.2         | Sample size based on effectiveness estimates using<br>noninferiority hypothesis testing..... 32 |
| E.3           | Clinical substudies ..... 34  |
| E.3.1         | Sample size for a contrast sensitivity study ..... 34   |
| E.3.2         | Sample size for endothelial cell density study..... 34  |
| <b>F</b>      | Presentation of results of clinical studies ..... 36  |
| F.1           | General ..... 36  |
| F.2           | Accountability of subjects..... 36  |
| F.3           | Refractive stability ..... 38   |
| F.4           | Safety ..... 39   |
| F.5           | Effectiveness..... 39   |
| F.6           | Retreatment ..... 40  |
| <b>G</b>      | Bibliography ..... 41   |
| <b>Tables</b> |   |
| <b>D.1</b>    | Recommended postoperative examination schedule ..... 22   |
| <b>E.1</b>    | Symbolic definitions ..... 30   |
| <b>E.2</b>    | Normal quantiles to use in equations ..... 31   |
| <b>F.1</b>    | Accountability by post-operative visit ..... 37   |

**Foreword** (This foreword is not part of American National Standard ANSI Z80.11-2012 (R2017).)

This American National Standard was developed to address the expressed needs of those members of the ophthalmic community who correct the refractive errors of the human eye using laser refractive correction procedures, those who manufacture the lasers systems for corneal reshaping utilized to perform these procedures, and those who protect the public interest by ensuring that such systems are made in such a way so that they reliably perform their function in a safe and effective manner when used correctly by skilled operators.

It must be realized that correcting the refractive error of the human eye with laser corrective surgery is a medical procedure involving not only a laser system for corneal reshaping but also (1) the assessment of the refractive state of the eye prior to surgery, (2) decisions on the best approach to take for treatment that involve not only the judgment of the physician but the desires of the patient, (3) other devices used during surgery, (4) the postsurgical care, and (5) the effects of healing, known and unknown. This standard only addresses the laser system for corneal reshaping and makes no attempt to standardize the procedure itself.

However, in response to a perceived need, informative annexes have been included in the standard to give guidance on types of clinical testing deemed to be adequate to ensure that the entire procedure is safe and effective. It was felt that a service would be performed for those in the field if this information were to be placed in a public document where it would be readily available for reference.

While it is true that the outcome of a laser refractive procedure will not prove acceptable if the laser system for corneal reshaping used to perform it is not adequate for the task, it cannot be assumed that a laser system is inadequate if outcomes are not acceptable, as this may be the result of deficiencies in other important parts of the overall medical procedure. Thus, no claim is made such that, if a laser system for corneal reshaping complies with this standard for the tasks it is designed to perform, that surgical procedures performed with the laser will have acceptable outcomes.

This standard was created by a special working group created by the Z80 Subcommittee on Medical Ophthalmic Devices and included experts in the field of laser refractive correction from the clinical, manufacturing and academic areas of the ophthalmic community and by experts from the regulatory agency given oversight in this field in the United States of America.

This standard contains several annexes. Annexes A and B are normative and are considered part of the standard. Annexes C through G are informative and are not considered part of this standard.

Suggestions for improvement of this standard are welcome. They should be sent to The Vision Council, 225 Reinekers Lane, Suite 700, Alexandria, VA 22314.

This standard was processed and approved for submittal to ANSI by the Accredited Standards Committee on Ophthalmic Optics, Z80. Committee approval of this standard does not necessarily imply that all committee members voted for its approval. At the time it approved this standard, the Z80 Committee had the following members:

Thomas C. White, M.D., Chairman  
Quido Cappelli, Vice-Chairman  
Robert Rosenberg, O.D., Secretary  
Jeff Endres, Secretariat

| <i>Organization Represented</i>                                       | <i>Name of Representative</i>  |
|---|--|
| Advance Medical Technologies Association.....                         | Paul Ludington<br>Richard Courtney (Alt.)<br>Glenn Davies (Alt.)<br>Bernie Liebler (Alt.)                        |
| American Academy of Ophthalmology.....                                | Tom White<br>Carl Tubbs (Alt.)<br>Pradeep Ramulu (Alt.)<br>Shannon Curtis (Alt.)                                 |
| American Academy of Optometry.....                                    | David Loshin   |
| American Ceramic Society.....   | Lyle Rubin   |
| American Glaucoma Society.....  | Herbert Hoover (Alt.)<br>Steven Gedde  |
| American Optometric Association.....                                  | Douglass Rhee (Alt.)<br>Karl Citek<br>Robert Rosenberg (Alt.)<br>William Benjamin (Alt.)<br>William Brown (Alt.) |
| American Society of Cataract and Refractive Ophthalmology ....        | Stephen Klyce<br>Jack Holladay (Alt.)  |
| Association for Research in Vision & Ophthalmology (ARVO) ...         | Jan Bergmanson<br>Gerald McGwin (Alt.)<br>Elmer Y. Tu (Alt.)   |
| Contact Lens Institute.....   | Stan Rogowski<br>Peter Mattingly (Alt.)  |
| Contact Lens Manufacturers Association.....                           | Quinto Cappelli<br>Troy Miller (Alt.)  |
| Department of Veterans Affairs.....                                   | John Townsend<br>Michael White (Alt.)  |
| Federated Cornea Societies/ASCRS.....                                 | Michael Belin<br>David Glasser (Alt.)  |
| Food & Drug Administration CDRH/<br>Division of Federal Agencies..... | Don Calogero<br>Robert James (Alt.)<br>Dexiu Shi (Alt.)  |
| National Association of Optometrists & Opticians.....                 | Nick Mileti<br>Doug Pelkey (Alt.)<br>Franklin Rozak (Alt.)   |
| Optical Laboratories Association.....                                 | Richard Tinson<br>Virginia Leshner (Alt.)<br>Jonathan Schwartz (Alt.)  |
| Opticians Association of America.....                                 | Tom Hicks<br>Chris Allen (Alt.)  |
| Sunglass Association of America.....                                  | Kenneth Frederick<br>Scott MacGuffie (Alt.)<br>Rick Van Arnem (Alt.)   |
| The Vision Council.....   | Jeff Endres<br>Neil Roche (Alt.)<br>Dick Whitney (Alt.)<br>Mike Vitale (Alt.)                                    |
| U.S. Assistant TAG Leader to ISO TC 172/SC7J.....                     | Jeff Endres  |

The working group for Laser Systems for Corneal Reshaping, which falls under the Medical Ophthalmic Devices Subcommittee, had the following members who worked on the writing of this standard:

Carl Tubbs MD, WG Chair  
Charles E. Campbell, Subcommittee Chair

Michael W. Belin MD  
Karl Citek  
John Townsend  
Thomas C. White, MD

Currently in preview, click buy full version

---

American National Standard  
for Ophthalmics –

# Laser Systems for Corneal Reshaping

## 1 Scope and purpose

This standard applies to any laser system whose primary intended use is to alter the shape of the cornea through the removal of corneal tissue, resulting in the improvement of visual performance.

This standard addresses the vocabulary, performance requirements, labeling, and clinical investigations necessary for this type of device.

## 2 Normative references

The following standards contain provisions that, through reference in this text, constitute provisions of this American National Standard. All standards are subject to revision, and parties to agreements based on this American National Standard are encouraged to investigate the possibility of applying the most recent editions of the standards indicated below. Members of the IEC and ISO maintain registers of currently valid International Standards.

ANSI Z136.1-2000, *Safe Use of Lasers*

ANSI Z136.3-2005, *Safe Use of Lasers in Health Care Facilities*

IEC 60601-1:2001, *Medical Electrical Equipment – Part 1: General Requirements for Safety*

IEC 60601-1-1:2000, *Medical Electrical Equipment – Part 1-1: General Requirements for Safety – Collateral Standard: Safety Requirements for Medical Electrical Systems*

IEC 60601-1-2:1993, *Medical Electrical Equipment – Part 1: General Requirements for Safety – 2. Collateral Standard: Electromagnetic Compatibility – Requirements and Tests*

IEC 60601-1-4:2000, *Medical Electrical Equipment – Part 1-4: General Requirements for Safety – Collateral Standard: Programmable Electrical Medical Systems*

IEC 60601-2-22:1995, *Medical Electrical Equipment – Part 2: Particular Requirements for the Safety of Diagnostic and Therapeutic Laser Equipment*

IEC 60825-1:2001, *Safety of Laser Products – Part 1: Equipment Classification, Requirements and User's Guide*

ISO 8598:1996, *Optics and optical instruments – Focimeters*