

AMERICAN NATIONAL STANDARD



*for Ophthalmics –
Focimeters*

Currently in preview, click buy full version

Currently in preview, click buy full version

ANSI®
Z80.17-2001

American National Standard
for Ophthalmics –
Focimeters

Secretariat
Optical Laboratories Association

Approved December 11, 2001
American National Standards Institute, Inc.

American National Standard

Approval of an American National Standard requires review by ANSI that the requirements for due process, consensus, and other criteria for approval have been met by the standards developer.

Consensus is established when, in the judgement of the ANSI Board of Standards Review, substantial agreement has been reached by directly and materially affected interests. Substantial agreement means much more than a simple majority, but not necessarily unanimity. Consensus requires that all views and objections be considered, and that a concerted effort be made towards their resolution.

The use of American National Standards is completely voluntary; their existence does not in any respect preclude anyone, whether he has approved the standards or not, from manufacturing, marketing, purchasing, or using products, processes, or procedures not conforming to the standards.

The American National Standards Institute does not develop standards and will in no circumstances give an interpretation of any American National Standard. Moreover, no person shall have the right or authority to issue an interpretation of an American National Standard in the name of the American National Standards Institute. Requests for interpretations should be addressed to the secretariat or sponsor whose name appears on the title page of this standard.

CAUTION NOTICE: This American National Standard may be revised or withdrawn at any time. The procedures of the American National Standards Institute require that action be taken periodically to reaffirm, revise, or withdraw this standard. Purchasers of American National Standards may receive current information on all standards by calling or writing the American National Standards Institute.

Developed by

The Accredited Committee Z80 for Ophthalmic Standards -
Optical Laboratories Association
Z80 Secretariat
P. O. Box 2000
Merrifield, VA 22116-2000

Published by

Optical Laboratories Association
P. O. Box 2000
Merrifield, VA 22116-2000

Copyright © 2002 by Optical Laboratories Association
All rights reserved.

No part of this publication may be reproduced in any form, in an electronic retrieval system or otherwise, without prior written permission of the publisher.

Printed in the United States of America

Contents

	Page
Foreword.....	ii
1 Scope	1
2 Normative References.....	1
3 Definitions.....	1
4 Design requirements and recommendations for general purpose focimeters.....	2
5 Accuracy requirements.....	4
6 Testing.....	6
7 Design requirements and recommendations for test lenses	8
Tables	
1 Tolerances of measured vertex power for continuously indicating instruments.....	5
2 Tolerances of measured prismatic power for continuously indicating instruments.....	5
3 Permissible deviations of measured vertex power reading from the nominal value of the test lenses for digitally rounding instruments	6
4 Permissible deviations of measured prismatic power reading from the nominal value of the test lenses for digitally rounding instruments	6
5 Design range for the standard test lenses.....	8
6 Tolerances for spherical test lenses.....	9
7 Tolerances for prismatic test lenses.....	10
Figures	
1 Permissible movement of the adjusting rail.....	3
2 Example of lens support for spectacle lenses in cross section	4
3 Cylindrical test lens.....	10
Annexes	
A Manufacture of test lenses for focimeters.....	11

Foreword (This foreword is not part of American National Standard ANSI Z80.17-2001.)

ANSI Z80.17-2001 was developed by a group of experts under the direction of the ANSI Instrument subcommittee chair David S. Loshin, O.D., Ph.D. This standard defines the requirement for devices that measure the sphere and cylindrical vertex power, prismatic power and axis for spectacle and contact lenses. Accuracy requirements for tolerances or deviation of readings are also defined.

This American National Standard is in harmony with the ISO 5898 Standard: Optics and Optical Instruments: Focimeters. However, this American National Standard has two major differences. First, since in the United States there is only one reference wavelength (helium d-line or 587.5618 nm), this standard eliminates the option of two reference wavelengths and only the helium d-line is allowable. Second, there is a significant improvement in defining the manufacture of test lenses for focimeters as found in Annex A. Also in Annex A is a method for determining back surface power and error tolerances as well as an example of a calculation for the expected error.

Suggestions for improvement of this standard will be welcome. They should be sent to the Optical Laboratories Association, P.O. Box 200, Merrifield, VA 22116-2000.

This standard was processed and approved for submittal to ANSI by the Accredited Standards Committee on Ophthalmic Optics, Z80. Committee approval of this standard does not necessarily imply that all committee members voted for its approval. At the time it approved this standard, the Z80 Committee had the following members:

Thomas White, M.D., Chairman
F. Dow Smith, Ph.D., Vice-Chairman
Robert Rosenberg, O.D., Secretary
Daniel Torgersen, Secretariat

<i>Organization Represented</i>	<i>Name of Representative</i>
Advance Medical Technologies Association	Douglas J. Fortunato Carolyn Jones (Alt.) Glenn Davies (Alt.) Stanley J. Rogaski (Alt.)
American Academy of Ophthalmology	Thomas C. White Gerhard Cibis (Alt.) Keith Baratz (Alt.) Paul F. Vinger (Alt.)
American Academy of Ophthalmology	David S. Loshin
American Ceramic Society	David J. Kerko Herbert Hoover (Alt.) Jackson S. Stroud (Alt.) Ron Klimek (Alt.)
American Optometric Association	Donald Pitts Gregory Stephens (Alt.) Robert Rosenberg (Alt.) William J. Benjamin (Alt.) Jeffrey Weaver (Alt.)
American Society of Cataract and Refractive Surgery	John J. Alpar Stephen H. Johnson (Alt.) Jack T. Holladay (Alt.)
Contact Lens Institute	Ed Schilling
Contact Lens Manufacturers Association	Quido Cappelli
Food & Drug Administration	David Whipple Donald Calogero (Alt.) Robert Landry (Alt.)
Industrial Safety Equipment Association	John Salce Janice C. Bradley (Alt.)

<i>Organization Represented</i>	<i>Name of Representative</i>
National Association of Optometrists & Opticians	Arthur Newman
National Academy of Opticianry	Jeffrey C. Snodgrass
	Floyd Holmgrain (Alt.)
Optical Laboratories Association	Daniel Torgersen (Alt.)
	Jeffrey S. Kosh
	Henry A. Hart (Alt.)
Optical Society of America	F. Dow Smith
	Richard A. Phillips (Alt.)
Opticians Association of America	Mike Robey
Prevent Blindness America	Tod Turriff
Sunglass Association of America	Kenneth L. Frederick
	David Elliott (Alt.)
	James Pritts (Alt.)
US Leader to ISO TC 172/SC7	Charles E. Campbell
Veterans Administration	John Townsend
	Sharon R. Atkin (Alt.)
Vision Council of America	William C. Thomas
	Darryl Meister (Alt.)
	Dick Whitney (Alt.)
	Kenneth O. Wood (Alt.)
	Richard P. White (Alt.)
	Steve Drake (Alt.)

The subcommittee on Ophthalmic Instruments, which developed this standard, had the following members:

David S. Loshin, O.D., Ph.D., Chairman	John J. Alpar
	Charles E. Campbell
	Robert Landry
	Robert Rosenberg
	Daniel Torgersen
	Tom White

Currently in preview, click buy full version

American National Standard for Ophthalmics –

Focimeters

1 Scope

This standard specifies requirements for continuously indicating and digitally rounding focimeters with which the vertex powers and prismatic powers of spherical and astigmatic lenses, including lenses mounted in frames, can be measured and with which lenses can be oriented and marked.

2 Normative References

The following standards contain provisions which, through reference in this text, constitute provisions of this American National Standard. At the time of publication, the editions indicated were valid. All standards are subject to revision, and parties to agreements based on this American National Standard are encouraged to investigate the possibility of applying the most recent editions of the standards indicated below.

ISO 8429: 1986, *Optics and Optical Instruments – Ophthalmics – Graduated Dial Scale*

3 Definitions

For the purposes of this standard, the following definitions apply.

3.1 adjusting rail: Movable rail or bar used as the reference axis for spectacles during measurement, which is aligned perpendicularly to the optical axis of the focimeter and parallel to the axis direction of 0° to 180°. This is also called the lens table or frame rest.

NOTE - The focimeter measures the vertex power relative to the surface placed against the lens support.

3.2 astigmatic power lens: Lens bringing a paraxial pencil of parallel rays to two separate line foci mutually at right angles and hence, unlike a spherical lens, having two principal powers.

One of these powers may be zero, with the corresponding focal line at infinity. Lenses referred to as toric lenses, spherocylindrical lenses and cylinder lenses are all astigmatic lenses.

3.3 centration error of the instrument: Residual prismatic error of the instrument with no lens in place.

3.4 cylindrical test lens: Lenses with cylindrical faces that are used to calibrate the axis marker and axis indicator with respect to the adjustment orientation of the rail. These lenses are usually specially designed and marked.

3.5 focimeter: An instrument that is used to measure vertex powers and prismatic effects of spectacle and contact lenses, to orient and mark uncut lenses, and to verify the correct mounting of lenses in spectacle frames.

3.5.1 continuously indicating focimeter: Focimeter with a continuous scale.