

American National Standard

*for Ophthalmics –
Instruments –
General-Purpose Clinical
Visual Acuity Charts*



Currently in preview, click buy full version

ANSI®
Z80.21-2010 (R2015)
Reaffirmation of
ANSI Z80.21-2010

American National Standard
for Ophthalmics –
Instruments –
General-Purpose Clinical
Visual Acuity Charts

Sponsor

The Vision Council

Approved May 27, 2010
Reaffirmed April 23, 2015

American National Standards Institute, Inc.

American National Standard

Approval of an American National Standard requires review by ANSI that the requirements for due process, consensus, and other criteria for approval have been met by the standards developer.

Consensus is established when, in the judgement of the ANSI Board of Standards Review, substantial agreement has been reached by directly and materially affected interests. Substantial agreement means much more than a simple majority, but not necessarily unanimity. Consensus requires that all views and objections be considered, and that a concerted effort be made towards their resolution.

The use of American National Standards is completely voluntary; their existence does not in any respect preclude anyone, whether he has approved the standards or not, from manufacturing, marketing, purchasing, or using products, processes, or procedures not conforming to the standards.

The American National Standards Institute does not develop standards and will in no circumstances give an interpretation of any American National Standard. Moreover, no person shall have the right or authority to issue an interpretation of an American National Standard in the name of the American National Standards Institute. Requests for interpretations should be addressed to the secretariat or sponsor whose name appears on the title page of this standard.

CAUTION NOTICE: This American National Standard may be revised or withdrawn at any time. The procedures of the American National Standards Institute require that action be taken periodically to reaffirm, revise, or withdraw this standard. Purchasers of American National Standards may receive current information on all standards by calling or writing the American National Standards Institute.

Developed by

The Accredited Committee Z80 for Ophthalmic Standards

The Vision Council
Z80 Secretariat
225 Reinekers Lane
Suite 700
Alexandria, VA 22314

Published by

The Vision Council
225 Reinekers Lane
Suite 700
Alexandria, VA 22314

Copyright © 2015 by The Vision Council
All rights reserved.

No part of this publication may be reproduced in any form, in an electronic retrieval system or otherwise, without prior written permission of the publisher.

Printed in the United States of America

Contents

	Page
Foreword	ii
1 Scope and purpose	1
2 Normative references.....	1
3 Clinical optotype calibration	2
4 System requirements	2
5 Application of standard	2
6 Marking	2
Table	
1 Progression of optotypes, in terms of gap size of the equivalent Landolt ring.....	5
Annexes	
A Methods of generating correlating optotypes.....	6
B Bibliography	10

Foreword (This foreword is not part of American National Standard ANSI Z80.21-2010 (R2015).)

This American National Standard addresses the optical design of optotypes used in clinical visual acuity measurement systems that use recognition of high-contrast optotypes and that are designed for general use.

ANSI Z80.21-2010 (R2015) was adapted by a group of experts within the ANSI Ophthalmic Instruments Subcommittee under the chairmanship of William L. Brown, O.D., Ph.D. It is a performance standard.

There are no substantive changes in this document compared to the previous 2010 edition.

This standard contains two annexes. Annex A, which is normative, is considered part of this standard. Annex B, which is informative, is not considered part of this standard.

Suggestions for improvement of this standard will be welcome. They should be sent to the Vision Council, 225 Reinekers Lane, Suite 700, Alexandria, VA 22304.

This standard was processed and approved for submittal to ANSI by the Accredited Standards Committee on Ophthalmic Optics, Z80. Committee approval of this standard does not necessarily imply that all committee members voted for its approval. At the time of approval of this standard, the Z80 Committee consisted of the following members:

Thomas White, M.D., Chairman
Quido Cappelli, Vice-Chairman
William Benjamin, O.D., Secretary
Jeff Endres, Secretariat

<i>Organization Represented</i>	<i>Name of Representative</i>
Abbott Medical Optics.....	Leonard Borrmann
Advance Medical Technologies Association.....	Michael Pflieger
American Academy of Ophthalmology.....	Dr. Thomas White
American Academy of Optometry.....	Dr. David Loshin
American Ceramic Society.....	Lyle Rubin
American Glaucoma Society.....	Dr. Steven Gedde
American Optometric Association.....	Dr. Karl Citek
American Society of Contact and Refractive Ophthalmology.....	Dr. Stephen Klyce
Contact Lens Institute.....	Stan Rogaski
Contact Lens Manufacturers Association.....	Quido Cappelli
Department of Veterans Affairs.....	Dr. John Townsend
Federated Colleague Societies/ASCRS.....	Dr. Michael Belin
Food & Drug Administration.....	Don Calogero
National Association of Optometrists & Opticians.....	Nick Mileti
Optical Laboratory Association.....	Michael Vitale
Opticians Association of America.....	Tom Hicks
Sunglass Association of America.....	Frederick Grethel
The Vision Council.....	Jeffrey Endres
U.S. TAG Leader to ISO TC 172/SC7.....	Jeffrey Endres

Members of the Subcommittee on Ophthalmic Instruments who contributed to this standard are:

William Brown, Chair

Charles Campbell

Bruce Drum

Robert Rosenberg

Currently in preview, click buy full version

Currently in preview, click buy full version

American National Standard for Ophthalmics –

Instruments – General-Purpose Clinical Visual Acuity Charts

1 Scope and purpose

1.1 Scope

This standard applies to displays of optotypes for all clinical visual acuity measurement systems that use recognition of high-contrast optotypes and that are designed for general use, including optotypes printed on opaque media, those intended for transillumination, electronically generated or projected displays. It does not apply to special testing of visual acuity, e.g., low vision or low-contrast charts.

1.2 Purpose

The principles of standardized visual acuity testing are presented in standards adopted by the National Academy of Sciences and the Consilium Ophthalmologicum Universal as referenced below. Due to practical design considerations and physical limitations of most general-purpose clinical visual acuity measurement systems, the chart design features specified in these reference standards can only be met for a limited range of acuity presentations. Specialized charts are often required for special clinic visual acuity measurements such as for low-vision patients. The purpose of this standard is to provide for standardized visual acuity charts for general measurement, which will enable the measurement of visual acuity over a limited, but clinically useful, range of acuities according to the principles contained in the reference standards.

2 Normative references

None