

American National Standard

*for Ophthalmics
Aqueous Shunts for
Glaucoma Application*



Currently in preview, click buy full version

ANSI®
Z80.27-2001 (R2011)

American National Standard
for Ophthalmics –

Aqueous Shunts for
Glaucoma Application

Secretariat
The Vision Council

Approved May 11, 2001
Reaffirmed June 28, 2011

American National Standards Institute, Inc.

American National Standard

Approval of an American National Standard requires review by ANSI that the requirements for due process, consensus, and other criteria for approval have been met by the standards developer.

Consensus is established when, in the judgement of the ANSI Board of Standards Review, substantial agreement has been reached by directly and materially affected interests. Substantial agreement means much more than a simple majority, but not necessarily unanimity. Consensus requires that all views and objections be considered, and that a concerted effort be made towards their resolution.

The use of American National Standards is completely voluntary; their existence does not in any respect preclude anyone, whether he has approved the standards or not, from manufacturing, marketing, purchasing, or using products, processes, or procedures not conforming to the standards.

The American National Standards Institute does not develop standards and will in no circumstances give an interpretation of any American National Standard. Moreover, no person shall have the right or authority to issue an interpretation of an American National Standard in the name of the American National Standards Institute. Requests for interpretations should be addressed to the secretariat or sponsor whose name appears on the title page of this standard.

CAUTION NOTICE: This American National Standard may be revised or withdrawn at any time. The procedures of the American National Standards Institute require that action be taken periodically to reaffirm, revise, or withdraw this standard. Purchasers of American National Standards may receive current information on all standards by calling or writing the American National Standards Institute.

Developed by

The Accredited Committee Z80 for Ophthalmic Standards -

The Vision Council
225 Reinkers Lane, Suite 700
Alexandria, VA 22314

Published by

The Vision Council
225 Reinkers Lane, Suite 700
Alexandria, VA 22314

Copyright © 2011 by The Vision Council
All rights reserved.

No part of this publication may be reproduced in any form, in an electronic retrieval system or otherwise, without prior written permission of the publisher.

Printed in the United States of America

Contents

	Page
Foreword	ii
1 Scope and purpose	1
2 Normative references	1
3 Definitions	2
4 Design attributes and evaluation	2
5 Packaging and labeling	2
Annexes	
A Examples of practices for <i>in-vitro</i> flow characterization of aqueous shunts	6
B Example of practices for biocompatibility testing of aqueous shunts	10
C Biocompatibility tests	13
D Example of practice for the clinical evaluation of aqueous shunts	19
E Bibliography	22

Foreword (This foreword is not part of American National Standard ANSI Z80.27-2001 (R2011).)

ANSI Z80.27-2001 was developed by a group of experts under the direct chairmanship of Prof. Dale Heuer. This standard on aqueous shunts for glaucoma application applies to the physical and mechanical properties and performance of finished aqueous shunts, their biocompatibility properties, and describes elements of a clinical protocol that may be used to assess the clinical performance of these devices for treatment of glaucoma. This is the first such standard developed. The committee, which consisted of scientists, industrialists, government and clinicians (several of them developers and/or manufacturers of shunts), did an excellent job in opening up this new area of a medical device standard.

Z80.27 is a subcommittee on Medical Ophthalmic Devices of Accredited Standards Committee Z80.

In 1982, the Optical Laboratories Association (OLA) assumed the responsibility of the Secretariat; and in 1985, the Z80 Committee became an accredited standards committee. The scope of the Z80 committee is for the establishment of standards that shall apply to ophthalmic lenses and to equipment, instruments and processes used in the final fabrication level that affect their performance; to ophthalmic frames, sun glasses and fashion eyewear; to contact lenses and accessories for their use; to intraocular implant lenses both monofocal and multifocal, as well as optical devices used to change the refractive capability of the eye (such as phakic intraocular lenses, intraocular contact lenses, corneal rings), as well as laser equipment to reshape the cornea for refractive purposes; to low vision aids and ophthalmic contact devices in addition to contact lenses; and to optical instrumentation used in ophthalmic procedures and vision evaluation; to computers to allow the use of such computers in conjunction with ophthalmic equipment, etc. The Medical Ophthalmic Devices subcommittee deals with the different intraocular lenses, whether phakic or aphakic; different devices to change the refractive power of the eye, including lasers, viscoelastic devices, ophthalmic irrigating solutions and glaucoma shunts.

This standard contains five annexes. Annex C is normative and is considered part of the standard. Annexes A, B, D, and E are informative and are not considered part of the standard.

Suggestions for improvement of this standard are welcome. They should be sent to the Vision Council, 225 Reinhardt Lane, Suite 700, Alexandria, VA 22314.

This standard was processed and approved for submittal to ANSI by the Accredited Standards Committee on Ophthalmic Optics, Z80. Committee approval of this standard does not necessarily imply that all committee members voted for its approval. At the time it approved this standard, the Z80 Committee had the following members:

Thomas C. White, M.D., Chairman
F. Dow Smith, Ph.D., Vice-Chairman
Robert Rubenbergl, O.D., Secretary

<i>Organization Represented</i>	<i>Name of Representative</i>
Advanced Medical Technologies Association	Douglas J. Fortunato Carolyn Jones (Alt.) Glenn Davies (Alt.) Stanley J. Rogaski (Alt.)
American Academy of Ophthalmology	Thomas C. White Gerhard Cibis (Alt.) Keith Baratz (Alt.) Paul F. Vinger (Alt.)

<i>Organization Represented</i>	<i>Name of Representative</i>
American Academy of Optometry	David S. Loshin
American Ceramic Society	David J. Kerko Herbert Hoover (Alt.) Jackson S. Stroud (Alt.) Ron Klimek (Alt.)
American Optometric Association	Donald Pitts Gregory Stephens (Alt.) Robert Rosenberg (Alt.) William J. Benjamin (Alt.) Jeffrey Weaver (Alt.)
American Society of Cataract and Refractive Surgery	John J. Alpar (Alt.) Stephen H. Johnson (Alt.) Jack T. Holladay (Alt.)
Contact Lens Institute	Ed Schilling
Contact Lens Manufacturers Association	Quido Cappelli
Food and Drug Administration	David Whipple Donald Calogero (Alt.) Robert Landry (Alt.)
Industrial Safety Equipment Association	John Salce Janice C. Bradley (Alt.)
National Academy of Opticianry	Jeffrey C. Smith Floyd Holm, Jr. (Alt.)
National Association of Optometrists and Opticians	Arthur L. Brown
Optical Laboratories Association	Jeffrey S. Losh Daniel Jorgensen (Alt.) Henry A. Hart (Alt.)
Optical Society of America	F. Dow Smith
Opticians Association of America	Richard A. Phillips (Alt.) Richard Sanders Joyce M. Otto (Alt.)
Prevent Blindness America	Tod Turriff
Sunglass Association of America	Kenneth L. Frederick David Elliott (Alt.) James Pritts (Alt.)
Veterans Administration	John Townsend Sharon R. Atkin (Alt.)
Vision Council of America	William C. Thomas Darryl Meister (Alt.) Dick Whitney (Alt.) Kenneth O. Wood (Alt.) Richard P. Waido (Alt.)
US Leader to ISO TC 172/107	Charles E. Campbell

The Subcommittee on Medical Ophthalmic Devices, which developed this standard, had the following members:

John J. Alpar, M.D., Chairman

A. Mateen Ahmed
P. Juhani Airaksinen
Robert Allen
George Baerveldt
Stephen Bartels
Keith Barton
Shelley Beadle
David A. Belyea
Michael Berlin
D.W. Blok
James D. Brandt
Reay Brown
Don Calogero
Carl Camras
Kevin J. Chismire
Michael Coote
Dennis Creedon
Hans Denffer
J. William Doyle
David Eifrig
Ron Fellman
James Ferguson, Jr.
Davis Fernquist, Jr.
Jeffrey Freedman

David George
Orna Geyer
Vijay Gulati
Claudine Hamlin-Krawczyk
Dale Heuer
Rick Hill
Jean T. Jacob-LaBarre
Neil H. Joseph
Frederick M. Kapetansky
Michael A. Kass
Changwon Kee
Robert A. Kenz
Peng T. Khaw
Yoshiaki Kitazawa
Naomi Kynin
Paul P. Lee
Richard Lippman
Mary Ann Lloyd
Gerardo Lugo
Jason Malecka
Anil K. Mandal
Lew Marten
Richard P. Mills
Donald S. Minckler

Anthony Molteno
Alok Nigam
C. Eric Olsen
David Osborne
Richard K. Parrish
Jeffrey Porter
John L. Ruiz
Mark Sherwood
Paul A. Sidoti
M. Fran Smith
Alfred Solish
Arieh S. Solomon
Joerg P. Stuermer
Albert Ting
Anja Tuulonen
Juha Valimaki
Marrie van der Mooren
Kyosti Vihanninjoki
Aharon Wegner
Richard E. Wertheimer
Thomas C. White
Erich Ziebler
Steve Ziemann

American National Standard for Ophthalmics –

Aqueous Shunts for Glaucoma Application

1 Scope and purpose

This standard applies to the physical and mechanical properties and performance of finished aqueous shunts, their biocompatibility properties, and describes elements of a clinical protocol that may be used to assess the clinical performance of these devices for treatment of glaucoma.

2 Normative references

The following standards contain provisions that, through reference in this text, constitute provisions of this American National Standard. At the time of publication, the editions indicated were valid. All standards are subject to revision, and parties to agreements based on this American National Standard are encouraged to investigate the possibility of applying the most recent editions of the standards indicated below. Members of the IEC and ISO maintain registers of currently valid International Standards.

ASTM D4169-99, *Standard Practice for Performance Testing of Shipping Containers and Systems*¹⁾

ASTM D4332-89(1994)e1, *Standard Practice for Conditioning Containers, Packages or Packaging Components for Testing*¹⁾

ISO 2248, *Packaging - Complete, Filled Transport Packages - Vertical Impact Test by Dropping*

ISO 8318, *Packaging - Complete Filled Transport Packages - Vibration Tests with a Sinusoidal Variable Frequency*

ISO 8597, *Optics and Optical Instruments - Visual Acuity Testing - Method of Correlating Optotypes*

ANSI/AAMI/ISO 10993-1-1997, *Biological Evaluation of Medical Devices - Part 1: Evaluation and Testing*

ANSI/AAMI/ISO 10993-5-1992, *Biological Evaluation of Medical Devices - Part 5: Tests for Cytotoxicity: In vitro methods*

ANSI/AAMI/ISO 11607-2000, *Sterilization for Health Care Products - Packaging for Terminally Sterilized Medical Devices*

ANSI/AAMI/ISO 14155-2000, *Clinical Investigation of Medical Devices*

ISO 14630, *Non-Active Surgical Implants - General Requirements*

ANSI Z 07.1-1994, *Ophthalmics - Intraocular Lenses*

ANSI/AAMI/ISO 11134-1993, *Sterilization of Health Care Products - Requirements for Validation and Routine Control - Industrial Moist Heat Sterilization*

ANSI/AAMI/ISO 11135-1994, *Medical Devices - Validation and Routine Control of Ethylene Oxide Sterilization*

ANSI/AAMI/ISO 11137-1994, *Sterilization of Health Care Products - Requirements for Validation and Routine Control - Radiation Sterilization*

ISO/FDIS 11979-3, *Ophthalmic Implants - Intraocular lenses - Part 3: Mechanical Properties and Test Methods*

Organization for Economic Cooperation and Development (OECD) Test No. 471 (7/97) - *Bacterial Reverse Mutation Test*²⁾

Recommended Practice for Determining Residual Ethylene Oxide in Medical Devices, ST29-1988³⁾

U.S. Pharmacopeia, Ed. XXII, 1990

¹⁾ Available from ASTM, 100 Barr Harbor Drive, West Conshohocken, PA 19428-2959.

²⁾ Available from the Office of Pollution, Prevention, and Toxics, US-EPA (7403), 401 M Street, SW, Washington, DC 20460, phone: (202) 260-1240, fax: (202) 260-1236.

³⁾ Available from the Association for Advancement of Medical Instrumentation, 1110 North Glebe Road, Suite 220, Arlington, VA 22201.