

ANSI Z80.27-2014 (R2019)

American National Standard

*for Ophthalmics –
Implantable Glaucoma Devices*

ANSI Z80.27-2014 (R2019)



Currently in preview, click buy full version

ANSI®
Z80.27-2014 (R2019)
(reaffirmation of
ANSI Z80.27-2014)

American National Standard
for Ophthalmics –
Implantable Glaucoma Devices

Secretariat
The Vision Council

Approved January 27, 2014
Reaffirmed December 9, 2019

American National Standards Institute, Inc.

American National Standard

Approval of an American National Standard requires review by ANSI that the requirements for due process, consensus, and other criteria for approval have been met by the standards developer.

Consensus is established when, in the judgement of the ANSI Board of Standards Review, substantial agreement has been reached by directly and materially affected interests. Substantial agreement means much more than a simple majority, but not necessarily unanimity. Consensus requires that all views and objections be considered, and that a concerted effort be made towards their resolution.

The use of American National Standards is completely voluntary; their existence does not in any respect preclude anyone, whether he has approved the standards or not, from manufacturing, marketing, purchasing, or using products, processes, or procedures not conforming to the standards.

The American National Standards Institute does not develop standards and will in no circumstances give an interpretation of any American National Standard. Moreover, no person shall have the right or authority to issue an interpretation of an American National Standard in the name of the American National Standards Institute. Requests for interpretations should be addressed to the secretariat or sponsor whose name appears on the title page of this standard.

CAUTION NOTICE: This American National Standard may be revised or withdrawn at any time. The procedures of the American National Standards Institute require that action be taken periodically to reaffirm, revise, or withdraw this standard. Purchasers of American National Standards may receive current information on all standards by calling or writing the American National Standards Institute.

Developed by

The Accredited Committee Z80 for Ophthalmic Standards

The Vision Council
Z80 Secretariat
225 Reinekers Lane
Alexandria, VA 22314

Published by
The Vision Council
Z80 Secretariat
225 Reinekers Lane
Alexandria, VA 22314

Copyright © 2019 by The Vision Council
All rights reserved.

No part of this publication may be reproduced in any form, in an electronic retrieval system or otherwise, without prior written permission of the publisher.

Printed in the United States of America

Contents

	Page
Foreword	ii
1 Scope and purpose.....	1
2 Normative references.....	1
3 Definitions	2
4 Design attributes and evaluation.....	2
4.1 Scope.....	2
4.2 General guidelines.....	2
5 Physical and mechanical requirements	3
5.1 Scope.....	3
5.2 Surface quality	3
5.3 Edge quality	3
5.4 Dimensions	3
5.5 Physical stability.....	3
5.6 Pressure/flow characteristics	4
5.7 Structural integrity	4
6 Biocompatibility requirements.....	4
6.1 Scope.....	4
6.2 General guidelines.....	4
6.3 Biological requirements.....	4
6.3.1 Texts on extractables.....	5
6.3.2 Cytotoxicity.....	5
6.3.3 Genotoxicity.....	5
6.3.4 Sensitization test.....	5
6.3.5 Immunotoxicity testing	5
6.3.6 Ocular implantation test	6
6.3.7 Rabbit pyrogen test.....	6
6.4 Physicochemical test requirements	6
6.4.1 Extractables by exhaustive extraction.....	6
6.4.2 Test for leachables.....	6
6.4.3 Hydrolysis testing.....	6
7 Sterility/package integrity requirements	7
7.1 Scope.....	7
7.2 General guidelines	7
7.3 Requirements.....	7
7.4 Ethylene oxide sterilent residues	7
7.5 Bacterial endotoxin	7
8 Shelf-life and transport stability requirements.....	7
8.1 Scope.....	7
8.2 Requirements.....	8

	Page
9	Additional requirements 8
9.1	Insertion method 8
9.2	Surface coatings 8
10	Clinical evaluation 8
10.1	Scope 8
10.2	Clinical investigational plan 8
11	Labeling 9
Annexes	
A	Examples of practices for in-Vitro flow characteristics of aqueous shunt type of implantable glaucoma devices 10
B	Ocular implantation test 14
C	Guidance on clinical study design for implantable glaucoma devices with refractory glaucoma indications 17
D	Guidance on clinical study design for implantable glaucoma devices with nonrefractory glaucoma indications 21
E	Evaluations, methodology and adverse events 27
F	Statistical sample size considerations for nonrefractory glaucoma devices 33
G	Recommended analysis of data from the clinical investigation 35
H	Labeling for implantable glaucoma devices 38
I	Bibliography 40
Tables	
C.1	Recommended examination schedule - Refractory glaucoma indication devices 20
D.1	Recommended minimum endothelial cell density 24
D.2	Recommended examination schedule - nonrefractory glaucoma indication devices 26
F.1	Symbol definitions 33
G.1	Accountability at each visit enrolled (N) = _____ 35

Foreword (This foreword is not part of American National Standard ANSI Z80.27-2014 (R2019).)

ANSI Z80.27-2014 was developed by a group of experts under the direct chairmanship of Dr. Carl Tubbs, with detailed editing by the group secretary, Jane Ellen Giamporcaro. The current standard represents a revision of the original and first 2001 document chaired by Dr. Dale Heuer, and pertains to the physical, biocompatibility, and mechanical properties, as well as to performance properties of implantable medical devices that are designed to lower IOP (intraocular pressure). The revision was completed by an active group of basic scientists, industrialists, regulatory agencies, clinicians, and clinician researchers, and involved spirited discussion on several fronts. Z80.27 is a subcommittee on Medical Ophthalmic Devices of ANSI (Accredited Standards Committee) Z80.

Z80 Committee was established in 1956, and is now a U.S. standards developer and accredited by ANSI. The Vision Council became Secretariat for Z80 in January of 2009. The Z80 committee, made up of 19 voting organizations and more than 200 participants, meets at least twice a year in order to regularly create and to rewrite draft standards. There are eight subcommittees that operate under the Z80 parent committee. The Medical Ophthalmic Devices (SC4) subcommittee deals with intraocular lenses (whether phakic or aphakic, toric or accommodative), devices to change the refractive power of the eye including lasers; viscoelastic devices; ophthalmic irrigating solutions; and finally, implantable glaucoma devices.

The Z80.27 standard contains nine annexes, A through I. The annexes are informative and supply additional guidance, but are not considered part of the standard.

Suggestions for improvement of this standard are welcome. They should be sent to the Vision Council, 225 Reinekers Lane, Suite 700, Alexandria, VA 22314.

This standard was processed and approved for submittal to ANSI by the Accredited Standards Committee on Ophthalmic Optics, Z80. Committee approval of this standard indicates general consensus but in no way implies that all committee members voted for approval. At the time it approved this standard, the Z80 Committee had the following members:

Thomas C. White, M.D., Chairman
Quido Cappelli, Vice-Chairman
Dr. Robert Rosenberg, O.D., Secretary
Jeff Endres, Secretariat Z80

<i>Organization Represented</i>	<i>Name of Representative</i>
Advance Medical Technologies	Michael Pflieger Richard Courtney (Alt.)
American Academy of Ophthalmology	Thomas White Carl Tubbs (Alt.) Pradeep Ramulu (Alt.) Shannon Curtis (Alt.)
American Academy of Optometry	David Loshin
American Ceramic Society	Lyle Rubin Herbert Hoover (Alt.)
American Glaucoma Society	Steven Gedde Douglas Rhee (Alt.)
American Optometric Association	Karl Citek Robert Rosenberg (Alt.) William Benjamin (Alt.)
American Society of Cataract and Refractive Surgery (ASCRS)	Steven Klyce Jack Holladay (Alt.)
Contact Lens Institute	Stan Rogaski Peter Mathers (Alt.)

<i>Organization Represented</i>	<i>Name of Representative</i>
Contact Lens Manufacturers Association	Quido Cappelli Troy Miller (Alt.)
Department of Veterans Affairs	John Townsend Michael White (Alt.)
Federated Cornea Societies	Michael Belin David Glasser (Alt.) Kathy Colby (Alt.) Elmer Tu (Alt.)
Food & Drug Administration	Don Calogero Robert James (Alt.) Dexiu Shi (Alt.)
National Association of Optometrists & Opticians	Nick Mileti Doug Pelkey (Alt.) Franklin Rozak (Alt.)
Optical Laboratories Association	Richard Tinson Jonathan Schwartz (Alt.)
Opticians Association of America	Tom Hicks Chris Allen (Alt.)
Sunglass Association of America	Frederick Grethel Nick Brown (Alt.) Rick Van Arnam (Alt.)
The Vision Council.....	Jeff Endres Neil Roche (Alt.) Dick Whitner (Alt.) Mike Vitale (Alt.)
U.S. ISO TC 172/SC7.....	Jeff Endres

American National Standard for Ophthalmics –

Implantable Glaucoma Devices

1 Scope and purpose

The scope of this standard applies to devices that are implanted in the eye to treat glaucoma by facilitating aqueous outflow. The standard excludes glaucoma devices whose effect depends upon metabolic and/or pharmacologic mechanisms.

The purpose of the standard is to describe the physical, mechanical and biocompatibility properties, as well as the elements of clinical protocols that may be useful in assessing the clinical performance of these devices.

2 Normative references

The following standards contain provisions that, through reference in this text, constitute provisions of this American National Standard. At the time of publication, the editions indicated were valid. All standards are subject to revision, and parties to agreements based on this American National Standard are encouraged to investigate the possibility of applying the most recent editions of the standards indicated below. Members of the IEC and ISO maintain registers of currently valid International Standards.

ANSI/AAMI ST72:2002, *Bacterial endotoxins – Test methodologies, routine monitoring, and alternatives to batch testing*

ISO 10993-1:2009, *Biological evaluation of medical devices – Part 1: Evaluation and testing within a risk management process*

ISO 10993-2:2006, *Biological evaluation of medical devices – Part 2: Animal welfare requirements*

ISO 10993-3:2003, *Biological evaluation of medical devices – Part 3: Tests for genotoxicity, carcinogenicity and reproductive toxicity*

ISO 10993-5:2009, *Biological evaluation of medical devices – Part 5: Tests for in vitro cytotoxicity*

ISO 10993-10:2002/Amd 1:2006, *Biological evaluation of medical devices – Part 10: Tests for irritation and delayed type hypersensitivity*

ISO 14155-1:2003, *Clinical investigation of medical devices for human subjects – Part 1: General requirements*

ISO 14155-2:2003, *Clinical investigation of medical devices for human subjects – Part 2: Clinical investigation plans*

ISO 14630:2008, *Non-active surgical implants – General requirements*

U.S. Code of Federal Regulations 21, Part 58: Good Laboratory Practices