

American National Standard

*for Ophthalmics –
Light Hazard from Operation Microscopes
Used in Ocular Surgery*



Currently in preview, click buy full version

ANSI®
Z80.38-2017

American National Standard
for Ophthalmics –

**Light Hazard from Operation Microscopes
Used in Ocular Surgery**

Secretariat
The Vision Council

Approved July 28, 2017
Published September 14, 2017

American National Standards Institute, Inc.

American National Standard

Approval of an American National Standard requires review by ANSI that the requirements for due process, consensus, and other criteria for approval have been met by the standards developer.

Consensus is established when, in the judgement of the ANSI Board of Standards Review, substantial agreement has been reached by directly and materially affected interests. Substantial agreement means much more than a simple majority, but not necessarily unanimity. Consensus requires that all views and objections be considered, and that a concerted effort be made towards their resolution.

The use of American National Standards is completely voluntary; their existence does not in any respect preclude anyone, whether he has approved the standards or not, from manufacturing, marketing, purchasing, or using products, processes, or procedures not conforming to the standards.

The American National Standards Institute does not develop standards and will in no circumstances give an interpretation of any American National Standard. Moreover, no person shall have the right or authority to issue an interpretation of an American National Standard in the name of the American National Standards Institute. Requests for interpretations should be addressed to the secretariat or sponsor whose name appears on the title page of this standard.

CAUTION NOTICE: This American National Standard may be revised or withdrawn at any time. The procedures of the American National Standards Institute require that action be taken periodically to reaffirm, revise, or withdraw this standard. Purchasers of American National Standards may receive current information on all standards by calling or writing the American National Standards Institute.

Developed by

The Accredited Committee Z80 for Ophthalmic Standards -

The Vision Council
Z80 Secretariat
225 Reinekers Lane
Suite 700
Alexandria, VA 22314

Published by

The Vision Council
225 Reinekers Lane
Suite 700
Alexandria, VA 22314

Copyright © 2017 by The Vision Council
All rights reserved.

No part of this publication may be reproduced in any form, in an electronic retrieval system or otherwise, without prior written permission of the publisher.

Printed in the United States of America

Contents

	Page
Foreword	ii
1 Scope.....	1
2 Normative references.....	1
3 Terms and definitions.....	1
4 Requirements for optical radiation hazard	2
4.1 General	2
4.2 Determination of classification group.....	2
4.3 Requirements for Group 1 instruments.....	2
4.3.1 General	2
4.3.2 Emission limits	2
4.4 Requirements for Group 2 instruments.....	2
4.4.1 General	2
4.4.2 Emission limits and guideline values	2
4.4.3 Retinal protection device.....	2
4.4.4 Stability of light intensity.....	3
5 Test methods	3
6 Additional information to be supplied by the manufacturer of Group 2 instruments	3
6.1 Time to reach recommended maximum exposure for photochemical aphakic weighted retinal radiant exposure - Continuous wave light sources	3
6.2 Number of pulses to reach recommended maximum exposure for photochemical aphakic retinal radiant exposure - Pulse light sources.....	3
7 Marking.....	3
Annex	
A Example of information regarding time to reach recommended maximum retinal exposure to be provided to the user during surgery	4

Foreword (This foreword is not part of American National Standard ANSI Z80.38-2017.)

This American National Standard specifies requirements and test methods for optical radiation hazards from operation microscopes that are used during ocular surgery.

ANSI Z80.36-2017 was adapted by a group of experts within the ANSI ASC Z80 Instruments and Low Vision Devices Subcommittee under the chair of William L. Brown, O.D., Ph.D. It is a performance standard.

This document was developed in 2016 when changes were made to ISO 15504-2, *Light hazard protection*, that resulted in recommendations for levels of radiant exposure for retinal photochemical hazard that were unacceptable to the US delegation to the ISO. Since ISO 10936-2, *Light hazard from operation microscopes used in ocular surgery*, refers to ISO 15004-2 for radiation levels, it was decided that an American National Standard was needed.

Suggestions for improvement of this standard are welcome. They should be sent to the Vision Council, 225 Reinekers Lane, Suite 700, Alexandria, VA 22314.

This standard was processed and approved for submittal to ANSI by the Accredited Standards Committee on Ophthalmic Optics, Z80. Committee approval of this standard does not necessarily imply that all committee members voted for its approval. At the time of approval of this standard, the Z80 Committee consisted of the following members:

Thomas C. White, M.D., Chair
Neil Roche, Vice-Chair
William Benjamin, O.D., Secretary
Michael Vitale, Secretariat

<i>Organization Represented</i>	<i>Name of Representative</i>
Advance Medical Technologies Association	Michael Pflieger
American Academy of Ophthalmology	Thomas White
American Academy of Optometry	David Loshin
American Ceramic Society	Lyle Rubin
American Glaucoma Society	Steven Gedde
American Optometric Association	Karl Citek
American Society of Cataract and Refractive Surgery	Stephen Klyce
Contact Lens Institute	Stan Rogaski
Contact Lens Manufacturers Association	Martin Dalsing
Cornea Society	Michael Belin
Department of Veterans Affairs	John Townsend
Food & Drug Administration, CDRH/Division	Don Calogero
Individual	Ralph Stone
Johnson & Johnson Vision	Leonard Borrmann
National Association of Optometrists & Opticians	Nick Mileti
Optical Laboratory Association	Steve Sutherlin
Opticians Association of America	Tom Hicks
Sunglass Association of America	Tibor Gross
The Vision Council	Michael Vitale
U.S. I/O TC 172/SC7	Michael Vitale

The Subcommittee on Instruments and Low Vision Devices, which developed this American National Standard, had the following members who contributed to this document:

William L. Brown, O.D., Ph.D., Chair

Charles Campbell
Bruce Drum
Sharon Miller
Robert Rosenberg
Dexiu Shi
David Sliney
Thomas White

Currently in preview, click buy full version

Currently in preview, click buy full version

American National Standard
for Ophthalmics –

Light Hazard from Operation Microscopes Used in Ocular Surgery

1 Scope

ANSI Z80.38 specifies requirements and test methods for optical radiation hazards from operation microscopes that are used during ocular surgery.

NOTE – General requirements for operation microscopes and test methods for these requirements are specified in ISO 10936-1.

2 Normative references

The following standards contain provisions which, through reference in this text, constitute provisions of this American National Standard. All standards are subject to revision and parties to agreements based on this American National Standard are encouraged to investigate the possibility of applying the most recent editions of the standards indicated below.

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ANSI Z80.36, *Light hazard protection for ophthalmic instruments*

ISO 10936-1, *Optics and optical instruments – Operation microscopes – Part 1: Requirements and test methods*

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

3.1

Group 1 instrument

ophthalmic instrument for which no potential light hazard exists and that can be shown to fulfill the requirements of 5.2 in ANSI Z80.36

3.2

Group 2 instrument

ophthalmic instrument for which a potential light hazard exists and that does not fulfill the requirements of 5.2 in ANSI Z80.36

3.3

Retinal protection device

mechanism placed in the path of illumination to increase the time to reach the recommended maximum retinal exposure limit