

AS 1169—1982

Australian Standard[®]

**Minimizing of combustion hazards
arising from medical use of
flammable anaesthetic agents**

This Australian standard was prepared by Committee MD/4, Medical Gases and Pipeline Services. It was approved on behalf of the Council of the Standards Association of Australia on 3 August 1982 and published on 11 October 1982.

The following interests were represented on the committee responsible for the preparation of this standard:

Association of Consulting Engineers Australia
Australian Society of Anaesthetists
Confederation of Australian Industry
Department of Health
Department of Housing and Construction
Department of Public Works, New South Wales
Department of Public Works, Western Australia
Electricity Trust of South Australia
Health Commission of New South Wales
Hospital architects
Institute of Hospital Engineers
Metal Trades Industry Association of Australia
Public Buildings Department, South Australia
Royal Australasian College of Surgeons
South Australian Health Commission
State Electricity Commission of Victoria

Review of Australian Standards. To keep abreast of progress in industry, Australian Standards are subject to periodic review and are kept up to date by the issue of amendments or new editions as necessary. It is important therefore that Standards users ensure that they are in possession of the latest edition, and any amendments thereto.

Full details of all Australian Standards and related publications will be found in the Standards Australia Catalogue of Publications; this information is supplemented each month by the magazine 'The Australian Standard', which subscribing members receive, and which gives details of new publications, new editions and amendments, and of withdrawn Standards.

Suggestions for improvements to Australian Standards, addressed to the head office of Standards Australia, are welcomed. Notification of any inaccuracy or ambiguity found in an Australian Standard should be made without delay in order that the matter may be investigated and appropriate action taken.

Australian Standard[®]

**Minimizing of combustion hazards
arising from medical use of
flammable anaesthetic agents**

First published as (AS CZ9)	1956
Revised and issued as AS 1169	1973
Second edition	1982

PREFACE

This edition of this standard was prepared by the Association's Committee on Medical Gases and Pipeline Services under the direction of the Medical Materials and Equipment Standards Committee. It supersedes Part 1 of AS 1169-1973, but does not supersede part 2 of AS 1169-1973. Part 2, which at present applies to the installation and testing of medical gas supply systems, is under revision and will be re-issued with a different Australian standard number.

This edition takes account of changes in medical practice that have occurred since publication of the previous edition, particularly the frequent and widespread use of surgical diathermy and the increasing use of electromedical equipment for monitoring patients. In addition, the committee recognized that in the majority of operating theatres, in most medical rooms in industrial premises, in dental surgeries and the like, flammable anaesthetic agents are no longer used. The use of non-flammable agents for anaesthesia eliminates the need for extensive precautions for minimizing fire and explosion hazards.

The precautionary requirements specified herein for the use of flammable agents are essentially the same as those prescribed in the 1973 edition. Likewise this new edition highlights as possible sources of ignition, incendive sparks of electrical, electrostatic and percussive origin, naked flames and hot materials, and inadvertent mixing of flammable anaesthetic agents. It includes appropriate precautions covering the design, construction and furnishing of buildings, administration of hospitals, equipment and practices to be adopted by personnel.

In regard to intrinsically safe equipment, in general, requirements are only meaningful in relation to low-energy-consuming circuits such as those used with endoscopic instruments incorporating sensitive relays, thermionic valves or semiconductor devices. Attention is drawn to the fact that incandescent lamps may achieve temperatures which may be unsafe within hazardous locations. Cold light sources, e.g. fibreoptic, are regarded as satisfactory provided that the generator is located outside the hazardous area.

© Copyright — STANDARDS AUSTRALIA

Users of Standards are reminded that copyright subsists in all Standards Australia publications and software. Except where the Copyright Act allows and except where provided for below no publications or software produced by Standards Australia may be reproduced, stored in a retrieval system in any form or transmitted by any means without prior permission in writing from Standards Australia. Permission may be conditional on an appropriate royalty payment. Requests for permission and information on commercial software royalties should be directed to the head office of Standards Australia.

Standards Australia will permit up to 10 percent of the technical content pages of a Standard to be copied for use exclusively in-house by purchasers of the Standard without payment of a royalty or advice to Standards Australia.

Standards Australia will also permit the inclusion of its copyright material in computer software programs for no royalty payment provided such programs are used exclusively in-house by the creators of the programs.

Care should be taken to ensure that material used is from the current edition of the Standard and that it is updated whenever the Standard is amended or revised. The number and date of the Standard should therefore be clearly identified.

The use of material in print form or in computer software programs to be used commercially, with or without payment, or in commercial contracts is subject to the payment of a royalty. This policy may be varied by Standards Australia at any time.

CONTENTS

	<i>Page</i>
SECTION 1. SCOPE AND GENERAL	
1.1 Scope	5
1.2 Application	5
1.3 Referenced Documents	5
1.4 Definitions	5
1.5 Compliance with Regulations of Statutory Authorities and Other Codes ..	5
1.6 Area Designation	5
SECTION 2. CONSTRUCTIONAL REQUIREMENTS	
2.1 Application of Section	8
2.2 Antistatic Flooring	8
2.3 Electrical Installation	8
2.4 Ventilation and Air Conditioning	8
2.5 Ventilating Equipment	10
2.6 Fire Extinguishers	10
2.7 Storage Rooms	10
2.8 Storage Cupboards	10
SECTION 3. EQUIPMENT FOR USE IN HAZARDOUS ANAESTHETIZING LOCATIONS	
3.1 Application of Section	11
3.2 Conductivity	11
3.3 Temperature Limits	11
3.4 Belt Drives	11
3.5 Rubber and Plastics Devices	11
3.6 Sheet Materials	11
3.7 Operating Tables and Trolleys	11
3.8 Anaesthetic Apparatus	11
3.9 Containers	11
3.10 Electrical Equipment	11
3.11 Electrical Tests	12
SECTION 4. HOSPITAL MANAGEMENT	
4.1 Advisory Committee	13
4.2 Availability of this Standard	13
4.3 Prescribed Procedures	13
4.4 Signs	13
4.5 Stimulation of Interest	13
4.6 Frequency of Testing	13
SECTION 5. STAFF PROCEDURES	
5.1 Application of Section	15
5.2 Clothing and Footwear	15
5.3 Smoking and Naked Flames	15
5.4 Equipment	15
5.5 Safe Handling of Flammable Anaesthetic Agents	15
5.6 Cleaning of Antistatic Floors	16
SECTION 6. MEDICAL COMPRESSED OXIDIZING GASES	
6.1 General	17
6.2 Storage	17
6.3 Oil and Grease	17
6.4 Decanting and Mixing	17

	<i>Page</i>
APPENDICES	
A Flammability of Inhalational Anaesthetics	18
B Nature of Hazards	19
C Installation of Antistatic Flooring	21
D Summary of Precautions Required to be Taken in the Preparation of a Theatre or Area for the Use of a Flammable Anaesthetic Agent	22
ANNEX. LIST OF REFERENCED DOCUMENTS	23

Currently in preview, click buy full version.

STANDARDS ASSOCIATION OF AUSTRALIA

Australian Standard

for

MINIMIZING OF COMBUSTION HAZARDS ARISING FROM THE MEDICAL
USE OF FLAMMABLE ANAESTHETIC AGENTS

SECTION 1. SCOPE AND GENERAL

1.1 SCOPE. This standard defines anaesthetizing locations and sets out requirements for minimizing the risk of fire and explosion arising from the medical use of flammable anaesthetic agents.

NOTES:

1. The prevention of fire and explosion depends not only on physical safeguards incorporated in the building structure and equipment, but also on a constant awareness of the hazards, and the adoption of specific precautions, by all persons concerned. Accordingly, information on the flammability of inhalational anaesthetics (see Appendix A), and on the nature of hazards (see Appendix B), has been included so that medical staff may appreciate the intent of this standard, and thus be better equipped to implement its requirements.
2. It is recognized that flammable liquids and gases other than for anaesthetic purposes may be used in theatres for skin preparation, aerosol spray applications or other clinical purposes. In such cases possible sources of ignition must be excluded from the area of application and the immediate surroundings. This standard, however, does not cover such situations, and in these instances reference should be made to Section 9 of AS 3000.

1.2 APPLICATION. The standard is relevant to all hospitals, and to rooms in industrial premises, dental surgeries, and the like, where anaesthetic agents are used.

In the application of the standard, it is essential that each operating theatre or other anaesthetizing location be designated either as not suitable for use with flammable anaesthetic agents or suitable (see Clause 1.6). If a decision is made to allow the use of flammable anaesthetic agents in a theatre or other location, then this standard which sets out the full precautions to be taken should be strictly implemented by those in authority (see Section 4).

1.3 REFERENCED DOCUMENTS. A list with titles of the standards referred to in this standard is given in the Annex.

1.4 DEFINITIONS. For the purpose of this standard, the following definitions apply:

1.4.1 Flammable anaesthetic agent - an anaesthetic agent, compressed gas or volatile liquid, which burns or explodes when mixed with air, oxygen or nitrous oxide in the concentration range used in medical practice under normal ambient conditions. (See Appendix A.)

1.4.2 Anaesthetizing location - any area in which it is intended to apply or administer a medical anaesthetic agent (flammable or non-flammable) in the normal course of examination or treatment.

1.4.3 Hazardous anaesthetic locations or areas -

- (a) Rooms used for bulk storage of flammable anaesthetic agents.
- (b) Anaesthetizing locations where a flammable anaesthetic agent is to be used.

1.4.4 Hazardous anaesthetic zone - the region immediately around discharge points or potential discharge points of a flammable mixture of anaesthetic gases and vapours. The zone includes the region from the floor to 0.3 m above it, a region 0.3 m above and beyond the patient, the operating table and the anaesthetic system and the region below to the floor. (See Fig. 1.)

1.4.5 Communicating location - an adjoining area within 1 m of a doorway communicating directly with a hazardous anaesthetic location as defined in Clause 1.4.3(b).

NOTE: Such areas are not considered to be hazardous anaesthetic locations, but are required to have conductive flooring to remove electrostatic charges from persons or objects before they enter a hazardous anaesthetic location.

1.4.6 Anaesthetic system - that part on the anaesthetic apparatus which may contain the anaesthetic mixture dispensed to the patient. This extends in sequence from the flowmeter manifold to the common gas outlet, breathing tubes, valves, absorber canisters, mask or endotracheal tube, scavenging tube and all interconnecting conduits.

1.5 COMPLIANCE WITH REGULATIONS OF STATUTORY AUTHORITIES AND OTHER CODES. Attention is drawn to the necessity of compliance with any relevant statutory regulations and codes, e.g. those applying to the design and construction of bulk storage rooms, and to the storage and safe handling of liquid oxygen and liquefied petroleum gases. Such regulations and codes shall be deemed to form part of this standard, being additional to any specific requirements contained herein.

1.6 AREA DESIGNATION.

1.6.1 General. Anaesthetic locations shall be designated as either suitable or not suitable for the use of flammable anaesthetic agents.