

Australian Standard 1410—1981

PRE-VACUUM PRESSURE STEAM STERILIZERS (WITH MECHANICAL AIR REMOVAL)

[Title allocated by Detective Cataloguing Authority:
STERILIZER, SURGICAL INSTRUMENT AND DRESSING (PRE-VACUUM PRESSURE
TYPE WITH MECHANICAL AIR REMOVAL)]



STANDARDS ASSOCIATION OF AUSTRALIA
Incorporated by Royal Charter



THE FOLLOWING SCIENTIFIC, INDUSTRIAL AND GOVERNMENTAL ORGANIZATIONS and departments were officially represented on the committee entrusted with the preparation of this standard:

Allied Building Trades Federation of Australia
Australian Capital Territory Health Commission
Australian Medical Association
Confederation of Australian Industry
Department of Health, Queensland
Department of Housing and Construction
Department of Industrial Relations, N.S.W.
Department of Public Works, N.S.W.
Health Commission of New South Wales
Health Commission of Victoria
Institute of Hospital Engineers (Australia)
Metal Trades Industry Association of Australia
National Biological Standards Laboratory
Public Health Department, W.A.
Royal College of Pathologists of Australia
South Australian Health Commission
Sterilizing Research and Advisory Council
Sterilizer Manufacturers Association of Australia
Sterilization & Disinfection Society (Vic.)
Sub-contractors Association of New South Wales

This standard, prepared by Committee ME/3, Sterilizing Equipment, was approved on behalf of the Council of the Standards Association of Australia on 2 March 1981, and was published on 15 July 1981.

To keep abreast of progress in industry, Australian standards are subject to continuous review and are kept up-to-date by the issue of amendments or new editions as necessary. It is important therefore that standards users ensure that their standards are up-to-date. Full details of all SAA publications will be found in the Annual List of Australian Standards; these details are supplemented by listings in the SAA monthly journal 'The Australian Standard'. Information on the Annual List and 'The Australian Standard' may be obtained from any sales office of the Association, where details are also available of the current status of individual standards. Suggestions for improvements to published standards, addressed to the head office of the Association, are welcomed.

This standard was issued in draft form for public review as DR 78157.

AUSTRALIAN STANDARD

**PRE-VACUUM PRESSURE
STEAM STERILIZERS
(WITH MECHANICAL AIR
REMOVAL)**

AS 1410—1981

First published	1973
Second edition	1981

**PUBLISHED BY THE STANDARDS ASSOCIATION OF AUSTRALIA
STANDARDS HOUSE, 80 ARTHUR ST, NORTH SYDNEY, N.S.W.**

ISBN 0 7262 2244 5



21 JUL 1981

PREFACE

This edition of this standard was prepared by the Association's Committee on Sterilizing Equipment to supersede the first edition published in 1973. This edition updates Australian requirements for sterilizers for porous loads and in particular the requirements relating to testing.

It should be noted that the standard relates only to equipment supplied with steam from an independent source. It does not relate to sterilizers of a type suitable for installation in a zone of risk associated with the use of flammable anaesthetics.

The methods of testing the sterilizers and the test requirements contained in Appendix C have been taken in their entirety from the British Department of Health and Social Security document 'Health Technical Memorandum 10—Sterilizers'. Acknowledgement is made to the Controller, Her Majesty's Stationery Office, for permission to use this material, which is covered by Crown Copyright.

Other Australian standards dealing with sterilizing and sanitizing equipment are as follows:

- AS 1714 Ethylene Oxide Sterilizers (Using Ethylene Oxide/Dichlorodifluoromethane 12%/88% *m/m* Sterilizing Gas Mixture)
- AS 1862 Aeration Cabinets (for Use with Ethylene Oxide Sterilizers)
- AS 2182 Portable Electrically Heated Steam Sterilizers (Downward Displacement Pressure Steam Type)
- AS 2192 Horizontal Sterilizers (Downward Displacement Pressure Steam Type)
- AS 2437 Flusher/Sanitizers for Bed Pans and Urin

Standards are in course of preparation for drying cabinets, dry heat sterilizers, (hot air type), washer/disinfectors for respiratory apparatus, washer/sanitizers for hospital utensils, and ultrasonic cleaners.

This standard may require reference to the following standards:

- AS 1135 SAA Non-ferrous Pressure Piping Code
- AS 1169 SAA Medical Agents and Gas Safety Code
- AS 1192 Electroplated Coatings of Nickel and Chromium
- AS 1210 SAA Unfired Pressure Vessels Code
- AS 1250 SAA Steel Structures Code
- AS 1271 Valves, Water Gauges and Other Fittings for Boilers and Unfired Pressure Vessels
- AS 1324 Air Filters for Use in Air Conditioning and General Ventilation
- AS 1349 Bourdon Tube Pressure and Vacuum Gauges
- AS 1449 Stainless Steel Heat-resisting Steel Plate, Sheet and Strip (Coils and Cut Lengths)
- AS 1650 Galvanized Coatings
- AS 1722 Pipe Threads of Whitworth Form
Part 1—Sealing Pipe Threads
- AS CB18 SAA Pressure Piping Code
Part 1—Ferrous Piping
- AS C100 Approval and Test Specification for Definitions and General Requirements for Electrical Materials and Equipment
- AS 170 Wrought Steels in the Form of Blooms, Billets, Bars and Forgings
Part 4—Stainless, Heat Resisting and Valve Steels
- AS 1041 Code for Temperature Measurement
- BS 1449 Steel Plate, Sheet and Strip
Part 2—Stainless and Heat Resisting Steel Plate, Sheet and Strip
- BS 1781 Linen Textiles for use by Government Departments, Hospitals and Local Authorities
- BS 3281 Rectangular Metal Boxes for Use in High Vacuum Steam Sterilizers
- BS 3970 Steam Sterilizers
- ANSI/ASTM Standard Recommended Practice for Cleaning and Descaling
A 380-80 Stainless Steel Parts, Equipment, and Systems
- Health Technical Memorandum 10—Sterilizers, 1980 (British Department of Health and Social Security).

CONTENTS

	<i>Page</i>
SECTION 1. SCOPE AND GENERAL	
1.1 Scope....	5
1.2 Application	5
1.3 Expression of Pressure Values	5
SECTION 2. MATERIALS OF CONSTRUCTION	
2.1 General	6
2.2 Chamber and Jacket, Door Ring and Door....	6
2.3 Materials for Connections Welded to the Chamber	6
2.4 Materials for Connections Welded to the Jacket	6
2.5 Pipework and Pipe Fittings	6
2.6 Welding Consumables	6
2.7 Framework	6
2.8 External Panels	6
SECTION 3. DESIGN	
3.1 Design Requirements....	7
3.2 Internal Dimensions of the Chamber	7
3.3 Working Height....	7
3.4 Mounting	7
3.5 Doors	7
3.6 Controls	8
3.7 Thermal Insulation	8
3.8 Connections and Fittings	9
3.9 Loading Equipment	10
3.10 Air Removal	10
3.11 Water supply	10
3.12 Vibration Isolation	10
3.13 Frame	10
3.14 External Panels	10
3.15 Ventilation....	10
SECTION 4. CONSTRUCTION	
4.1 General	11
4.2 Welding	11
4.3 Jacket	11
4.4 Finish	11
4.5 Frame	11
4.6 External Panels	11
SECTION 5. REQUIREMENTS FOR ELECTRICALLY CONTROLLED EQUIPMENT	
5.1 General	12
5.2 Controls	12
5.3 Timers	12
5.4 Cycle Indicator	12
SECTION 6. TESTING	
6.1 Tests	13
6.2 Methods of Test	13
6.3 Test Requirements	13
SECTION 7. MARKING, DATA AND INFORMATION	
7.1 Marking	14
7.2 Data to be Provided by the Manufacturer	14

APPENDICES		<i>Page</i>
A	Additional Detail Required	15
B	Method of Introducing Thermocouples into Sterilizer Chamber	16
C	Methods of Test for Performance and Functioning.....	19
D	Quality of Water	26

©Copyright — STANDARDS ASSOCIATION OF AUSTRALIA 1981

Users of standards are reminded that copyright subsists in all SAA publications. No part of this publication may be reproduced, stored in a retrieval system in any form or transmitted by any means without prior permission in writing of the Standards Association of Australia.

STANDARDS ASSOCIATION OF AUSTRALIA**Australian Standard**

for

PRE-VACUUM PRESSURE STEAM STERILIZERS (WITH MECHANICAL AIR REMOVAL)**SECTION 1. SCOPE AND GENERAL**

1.1 SCOPE. This standard specifies requirements for pre-vacuum steam sterilizers intended for use in hospitals and other non-industrial applications for the sterilization of porous loads and/or instruments and utensils wrapped in porous material. The sterilizers are fitted with mechanical means for the removal of air and are intended for use with steam supplied from a source heated independently of the sterilizer jacket and chamber.

NOTE: Certain details need to be discussed with the manufacturer at the time of enquiry or order. Appendix A provides information in this regard.

1.2 APPLICATION. This standard applies to the design, manufacture and testing of fully automatic jacketed pre-vacuum sterilizers, which utilize saturated steam, operating at a sterilizing temperature of 134°C (see Clause 3.8.10.4).

NOTE: Sterilizers made to this standard may not be installed in zones of risk associated with the use of flammable anaesthetics or volatile liquids (see AS 1169) nor used for the sterilization of fluids (see Clause 7.1(d)).

1.3 EXPRESSION OF PRESSURE VALUES. Pressure values are gauge pressure at sea level, unless otherwise specified.

SECTION 2. MATERIALS OF CONSTRUCTION

2.1 GENERAL. Materials used in the construction of the vessels shall comply with AS 1210.

2.2 CHAMBER AND JACKET, DOOR RING AND DOOR. One of the combinations of materials given in Table 2.1 should be used.

Where dissimilar materials, which may cause galvanic action, are used in the manufacture of pressure and non-pressure retaining parts, the surfaces of such materials shall be protected by electroplating or other forms of protection.

NOTE: The purchaser should specify the materials required at the time of enquiry or order (see Appendix A).

2.3 MATERIALS FOR CONNECTIONS WELDED TO THE CHAMBER. Materials for connections welded to the chamber should be as listed in Table 2.2.

2.4 MATERIALS FOR CONNECTIONS WELDED TO THE JACKET. The material for connections welded to the jacket shall be compatible with the jacket material.

2.5 PIPEWORK AND PIPE FITTINGS. Pipe-work and pipe fittings shall comply with AS 1135 or AS CB18, Part 1.

2.6 WELDING CONSUMABLES. Electrodes and filler rods used in welding and brazing processes and other welding consumables, shall comply with the relevant Australian or equivalent standards.

2.7 FRAMEWORK. The material for the framework of sterilizers shall be as follows:

- (a) *Frame*—stainless steel or hot-dip galvanized carbon steel.
- (b) *Adjustable brackets, feet or levelling bolts*—stainless steel.

2.8 EXTERNAL PANELS. For free-standing cabinet models, all external panels should be fabricated from satin finish stainless steel having a thickness not less than 0.9 mm.

NOTE: Where the purchaser requires material other than the above, he should specify the type of material required and its thickness, at the time of enquiry or order (see Appendix A).

TABLE 2.1
MATERIALS FOR CHAMBER, JACKET, DOOR RING AND DOOR

Combination No	Chamber	Jacket	Door ring	Door
1	Stainless steel	Stainless steel Carbon steel	Stainless steel Carbon steel	Stainless steel Nickel-clad or
2	Stainless steel-clad (on internal face) carbon steel	Carbon steel	Stainless steel Carbon steel	Stainless steel-clad (on internal face) carbon steel
3	Nickel-clad carbon steel	Carbon steel	Stainless steel Copper-nickel alloy Carbon steel	Carbon steel Copper-nickel alloy Gunmetal Aluminium
4	Carbon steel	Carbon steel	Carbon Steel	Carbon steel
5	Copper-nickel alloy	Copper-nickel alloy	Copper-nickel alloy Gunmetal	Stainless steel Copper-nickel alloy Gunmetal
6	Copper	Copper	Gunmetal	Gunmetal
7	Aluminium	Aluminium	Aluminium	Stainless steel Aluminium

TABLE 2.2
MATERIALS FOR CONNECTIONS WELDED TO THE CHAMBER

Chamber (See Table 2.1)	Connections
Stainless steel	Stainless steel of equivalent type to the material of the chamber
Stainless steel-clad carbon steel	Stainless steel of equivalent type to the cladding material of the chamber
Nickel-clad carbon steel	Nickel or copper-nickel alloy
Carbon steel	Carbon steel of equivalent type to the material of the chamber
Copper-nickel alloy	Copper-nickel alloy of equivalent type to the material of the chamber
Copper	Gunmetal
Aluminium	Aluminium