

Australian Standard™

**Sterile hypodermic needles
for single use**

Currently in preview, click buy full version

This Australian Standard was prepared by Committee HE-009, Hypodermic Equipment—General Medical. It was approved on behalf of the Council of Standards Australia on 26 July 2002 and published on 27 September 2002.

The following are represented on Committee HE-009:

Auckland Healthcare, New Zealand
Australian Chamber of Commerce and Industry
Australian College of Operating Room Nurses
Australian Medical Association
Certification Bodies (Australia)
Commonwealth Department of Health and Ageing
Federation of Sterilizing Research and Advisory Councils of Australia
Medical Industry Association of Australia
New Zealand Chemical Industry Council
Queensland Health
Royal Australian College of General Practitioners
Royal College of Pathologists of Australasia

Keeping Standards up-to-date

Standards are living documents which reflect progress in science, technology and systems. To maintain their currency, all Standards are periodically reviewed, and new editions are published. Between editions, amendments may be issued. Standards may also be withdrawn. It is important that readers assure themselves they are using a current Standard, which should include any amendments which may have been published since the Standard was purchased.

Detailed information about Standards can be found by visiting the Standards Australia website at www.standards.com.au and looking up the relevant Standard in the on-line catalogue.

Alternatively, the printed Catalogue provides information current at 1 January each year and the monthly magazine, *The Australian Standard*, has a full listing of revisions and amendments published each month.

We also welcome suggestions for improvement in our Standards, and especially encourage readers to notify us immediately of any apparent inaccuracies or ambiguities. Contact us via email at mail@standards.com.au, or write to the Chief Executive, Standards Australia International Ltd, GPO Box 5420, Sydney, NSW 2001.

This Standard was issued in draft form for comment as DR 01257.

Australian Standard™

**Sterile hypodermic needles
for single use**

Revised as AS T48—1970.
Previous edition AS 1946—1984.
Third edition 2002.

COPYRIGHT

© Standards Australia International

All rights are reserved. No part of this work may be reproduced or copied in any form or by any means, electronic or mechanical, including photocopying, without the written permission of the publisher.

Published by Standards Australia International Ltd
GPO Box 5420, Sydney, NSW 2001, Australia

ISBN 0 7337 4788 4

PREFACE

This Standard was prepared by the Australian members of Joint Standards Australia/Standards New Zealand Committee HE-009, Hypodermic Equipment—General Medical, to supersede AS 1946—1984, *Hypodermic equipment—Single-use needles (sterile) for general medical use*. After consultation with stakeholders in both countries, Standards Australia and Standards New Zealand decided to develop this Standard as an Australian Standard rather than an Australian/New Zealand Standard.

This Standard has been reproduced from, and is identical to, ISO 7864:1993, *Sterile hypodermic needles for single use*.

As this Standard is reproduced from an International Standard, the following modifications apply.

- Its number does not appear on each page of text and its identity is shown only on the cover and title page.
- In the source text ‘this International Standard’ should read ‘this Australian Standard’.
- A full point should be substituted for a comma when referring to a decimal marker.

The terms ‘normative’ and ‘informative’ have been used to define the application of the annex to which they apply. A ‘normative’ annex is an integral part of a Standard whereas an ‘informative’ annex is only for information and guidance.

References to International Standards should be replaced by references to the following Australian or Australian/New Zealand Standards:

<i>Reference to International Standard</i>		<i>Australian/New Zealand Standard</i>	
ISO		AS	
594	Conical fittings with a 6% (Luer) taper for syringes, needles and certain other medical equipment	1600	Medical equipment—Conical fittings with a 6% (Luer) taper for syringes, needles and certain other medical equipment
594-1	Part 1: General requirements	1600.1	Part 1: General requirements
594-2	Part 2: Lock fitting	—	
3696	Water for analytical laboratory use—Specification and test methods	—	
6009	Hypodermic needles for single use—Colour coding for identification	—	
7886	Sterile hypodermic syringes for single use	1094	Medical equipment—Single-use syringes (sterile) for general medical use
7886-1	Part 1: Syringes for manual use		
		AS/NZS	
8601	Data elements and interchange formats—Information interchange—Representation of dates and times	3802	Data elements and interchange formats—Information interchange—Representation of dates and times
		AS	
9626	Stainless steel needle tubing for the manufacture of medical devices	2145	Stainless steel needle tubing for manufacture of medical devices

CONTENTS

	<i>Page</i>
	-
1 Scope	1
2 Normative references	1
3 Nomenclature	1
4 Cleanliness	1
5 Limits for acidity or alkalinity	1
6 Limits for extractable metals	1
7 Size designation	2
8 Colour coding	2
9 Needle hub	2
10 Sheath	2
11 Needle tube	4
12 Needle point	4
13 Performance	4
14 Packaging	5
15 Labelling	5
Annexes	
A Method for preparation of extracts	7
B Example of method for determination of fragment production from rubber closures	8
C Symbol for "do not re-use"	10
D Bibliography	11

Currently in preview, click buy full version.

AUSTRALIAN STANDARD

Sterile hypodermic needles for single use**1 Scope**

This International Standard specifies requirements for sterile hypodermic needles for single use of nominal outside diameters 0,3 mm and 1,2 mm.

It does not apply to dental needles.

2 Normative references

The following standards contain provisions which, through reference in this text, constitute provisions of this International Standard. At the time of publication, the editions indicated were valid. All standards are subject to revision, and parties to agreements based on this International Standard are encouraged to investigate the possibility of applying the most recent editions of the standards indicated below. Members of IEC and ISO maintain registers of currently valid International Standards.

ISO 594-1:1986, *Conical fittings with a 6 % (Luer) taper for syringes, needles and certain other medical equipment — Part 1: General requirements.*

ISO 594-2:1991, *Conical fittings with a 6 % (Luer) taper for syringes, needles and certain other medical equipment — Part 2: Lock fittings.*

ISO 3696:1987, *Water for analytical laboratory use — Specification and test methods.*

ISO 409:1992, *Hypodermic needles for single use — Colour coding for identification.*

ISO 7886-1:—¹⁾, *Sterile hypodermic syringes for single use — Part 1: Syringes for manual use.*

1) To be published.

ISO 8601:1988, *Data elements and interchange formats — Information interchange — Representation of dates and times.*

ISO 9626:1991, *Stainless steel needle tubing for the manufacture of medical devices.*

3 Nomenclature

The nomenclature for components of hypodermic needles for single use is shown in figure 1 together with the designation for length l ; nomenclature for needle points is shown in figure 2.

4 Cleanliness

When inspected by normal or corrected-to-normal vision without magnification under an illuminance of 300 lx to 700 lx, the surface of the hypodermic needle tube shall appear free from particles and extraneous matter.

When examined under $\times 2,5$ magnification, the hub socket shall appear free from particles and extraneous matter.

5 Limits for acidity or alkalinity

When determined with a laboratory pH meter and using a general purpose electrode, the pH value of an extract prepared in accordance with annex A shall be within one unit of pH of that of the control fluid.

6 Limits for extractable metals

When tested by a recognized microanalytical method, for example by an atomic absorption