

Australian Standard[®]

**Metallic materials—Rockwell
superficial hardness test—N and T
scales**

This Australian Standard was prepared by Committee MT/6, Mechanical Testing of Metals. It was approved on behalf of the Council of Standards Australia on 25 July 1991 and published on 7 October 1991.

The following interests are represented on Committee MT/6:

Aluminium Development Council
Bureau of Steel Manufacturers of Australia
Confederation of Australian Industry
Department of Defence
Division of Applied Physics, CSIRO
Metal Trades Industry Association of Australia
National Association of Testing Authorities, Australia
Railways of Australia Committee
Society of Automotive Engineers, Australasia
University of Sydney
University of Wollongong

Additional interests participating in preparation of Standard:

Calibrating organizations
Metal manufacturing industries
Testing and research organizations

Review of Australian Standards. To keep abreast of progress in industry, Australian Standards are subject to periodic review and are kept up to date by the issue of amendments or new editions as necessary. It is important therefore that Standards users ensure that they are in possession of the latest edition, and any amendments thereto.

Full details of Australian Standards and related publications will be found in the Standards Australia Catalogue of Publications; this information is supplemented each month by the magazine 'The Australian Standard', which subscribing members receive, and which gives details of new publications, new editions and amendments, and of withdrawn Standards.

Suggestions for improvements to Australian Standards, addressed to the head office of Standards Australia, are welcomed. Notification of any inaccuracy or ambiguity found in an Australian Standard should be made without delay in order that the matter may be investigated and appropriate action taken.

This Standard was issued in draft form for comment as DR 87214.

Australian Standard[®]

**Metallic materials—Rockwell
superficial hardness test—N and T
scales**

First published as AS 2025.1 — 1977.
Revised and redesignated AS 2025— 1991.

PREFACE

This Standard was prepared under the direction of the Standards Australia Committee on Mechanical Testing of Metals to supersede AS 2025, *Method for Rockwell superficial hardness test, Part 1—1977, Testing of metals, N and T scales* and to include calibration requirements for the testing machine.

The hardness scales in the Standard are limited to the N and T scales, since these are used extensively in Australia.

This edition was prepared at the request of the National Association of Testing Authorities, Australia, to alleviate problems associated with calibration in relation to availability and cost of certified diamond indenters, and cost, adequacy and range of hardness blocks available for user checks.

During 1989 the International Organization for Standardization published the following three Standards:

- ISO 1024 *Metallic materials—Hardness test—Rockwell superficial test (scales 15N, 30N, 45N, 15T, 30T and 45T)*
- ISO 1079 *Metallic materials—Hardness test—Verification of Rockwell superficial hardness testing machines (scales 15N, 30N, 45N, 15T, 30T and 45T)*
- ISO 1355 *Metallic materials—Hardness test—Calibration of standardized blocks to be used for Rockwell superficial hardness testing machines (scales 15N, 30N, 45N, 15T, 30T and 45T)*

This Australian Standard is technically equivalent to these ISO Standards but covers all the subjects in the one volume.

© Copyright — STANDARDS AUSTRALIA

Users of Standards are reminded that copyright subsists in all Standards Australia publications and software. Except where the Copyright Act allows and except where provided for below no publications or software produced by Standards Australia may be reproduced, stored in a retrieval system in any form or transmitted by any means without prior permission in writing from Standards Australia. Permission may be conditional on an appropriate royalty payment. Requests for permission and information on commercial software royalties should be directed to the head office of Standards Australia.

Standards Australia will permit up to 10 percent of the technical content pages of a Standard to be copied for use exclusively in-house by purchasers of the Standard without payment of a royalty or advice to Standards Australia.

Standards Australia will also permit the inclusion of its copyright material in computer software programs for no royalty payment provided such programs are used exclusively in-house by the creators of the programs.

Care should be taken to ensure that material used is from the current edition of the Standard and that it is updated whenever the Standard is amended or revised. The number and date of the Standard should therefore be clearly identified.

The use of material in print form or in computer software programs to be used commercially, with or without payment, or in commercial contracts is subject to the payment of a royalty. This policy may be varied by Standards Australia at any time.

CONTENTS

	<i>Page</i>
SECTION 1 SCOPE AND GENERAL	
1.1 SCOPE	4
1.2 REFERENCED DOCUMENTS	4
1.3 DEFINITIONS	4
1.4 ROCKWELL TEST PRINCIPLE	5
1.5 DESIGNATION OF ROCKWELL SUPERFICIAL HARDNESS	5
1.6 HARDNESS CONVERSIONS	6
SECTION 2 TESTING OF METALS	
2.1 GENERAL	7
2.2 TESTING MACHINE	7
2.3 INDENTERS	7
2.4 ANVIL	7
2.5 MAINTENANCE OF THE TESTING MACHINE BY THE USER	7
2.6 TEST PIECE	8
2.7 TEST PROCEDURE	8
2.8 RECORD OF TEST RESULTS	9
2.9 TEST REPORT	10
SECTION 3 COMPLETE CALIBRATION OF THE TESTING MACHINE	
3.1 GENERAL	11
3.2 PRELIMINARY INSPECTION	11
3.3 CALIBRATION OF APPLIED FORCES	11
3.4 CALIBRATION OF MEASURING APPARATUS	12
3.5 VISUAL EXAMINATION OF INDENTERS	12
3.6 PERFORMANCE TEST	12
3.7 RECORD OF RESULTS	12
3.8 REPORT OF COMPLETE CALIBRATION	13
SECTION 4 PARTIAL CALIBRATION OF THE TESTING MACHINE	
4.1 GENERAL	14
4.2 PRELIMINARY INSPECTION	14
4.3 INDENTING PROCEDURE	14
4.4 ASSESSMENT OF PERFORMANCE	14
4.5 RECORD OF RESULTS	16
4.6 REPORT OF PARTIAL CALIBRATION	16
SECTION 5 CALIBRATION OF STANDARD HARDNESS BLOCKS	
5.1 GENERAL	17
5.2 REQUIREMENTS FOR STANDARD HARDNESS BLOCKS	17
5.3 CALIBRATION OF BLOCKS	17
5.4 STANDARDING MACHINE	17
SECTION 6 CALIBRATION OF STANDARD DIAMOND CONE INDENTERS AND STEEL BALLS	
6.1 GENERAL	19
6.2 DIAMOND CONE INDENTER	19
6.3 STEEL BALL INDENTER	19
APPENDIX A TESTS ON CURVED SURFACES	20

STANDARDS AUSTRALIA

Australian Standard

Metallic materials—Rockwell superficial hardness test—N and T scales

SECTION 1 SCOPE AND GENERAL

1.1 SCOPE This Standard sets out methods for measuring the hardness of metals and metal products using the N and T scales of Rockwell superficial hardness. It includes requirements for the installation and calibration of the testing machine and recommendations for its maintenance by the user.

This Standard does not preclude the use of portable hardness testing machines which meet the appropriate requirements of Section 2. This Standard does not meet the requirements of Rockwell scales A to K, as these are dealt with in AS 1815, nor does it cover the testing of materials other than metals.

1.2 REFERENCED DOCUMENTS The following documents are referred to in this Standard:

AS	
1199	Sampling procedures and tables for inspection by attributes
1815	Metallic materials—Rockwell hardness test
1817	Metallic materials—Vickers hardness test
2193	Methods for calibration and grading of force-measuring systems of testing machines
ISO	
4964	Steel—Hardness conversions

1.3 DEFINITIONS For the purpose of this Standard, the definitions below apply:

1.3.1 Arithmetic mean deviation (R_a)—the arithmetic average of the deviation of a profile above and below the reference curve or line. This parameter has the dimension of length and is used for surface roughness measurement.

1.3.2 Calibration—all the operations for determining the values of the errors of a Rockwell superficial hardness testing machine (and, if necessary, to determine other metrological properties).

1.3.3 Calibrating authority—any approved person or organization qualified and equipped to perform the tests set out in Section 3 or 4 or both.

NOTE: The National Association of Testing Authorities, Australia, registers laboratories for the performance of these calibrations.

1.3.4 Hardness test blocks—metal blocks, for the purpose of maintaining surveillance of the performance of hardness testing machines between periodic calibrations.

1.3.5 Indenter, reference—a diamond cone indenter which is maintained by a national standardizing authority and whose performance has been compared with reference indenters maintained by other national standardizing authorities.

1.3.6 Indenter, standard—a diamond cone indenter which is maintained by a calibrating authority for assessing the performance of hardness testing machines in the field, in the event of non-complying results being obtained when using the indenters assigned to the machine.

1.3.7 Standard hardness blocks—metal blocks which have a calibration traceable to a national standard scale of hardness which is recognized by the Australian standardizing authority, for the purpose of verifying the performance of hardness testing machines.

1.3.8 Standardizing authority—an authority which maintains the standard of Rockwell superficial hardness.

NOTE: In Australia the standardizing authority is the National Measurement Laboratory, Division of Applied Physics, CSIRO.

1.3.9 Test piece—a piece prepared for testing, made from a test specimen by some mechanical operation (see Note to Clause 1.3.10).

1.3.10 Test specimen—a portion of material or a single item taken from the test sample for applying a particular test.

NOTE: Rockwell superficial hardness tests can be made on test pieces or test specimens, the latter often being in the form of finished products or components.

1.3.11 Normative—an appendix which is essential to the understanding or implementation of the Standard.