

Australian Standard<sup>®</sup>

---

**Density hydrometers**

---

This Australian Standard was prepared by Committee CH/1, Laboratory Glassware and Related Apparatus. It was approved on behalf of the Council of Standards Australia on 11 March 1994 and published on 14 June 1994.

---

The following interests are represented on Committee CH/1:

Australian Chamber of Commerce and Industry  
Australian Government Analytical Laboratories  
National Association of Testing Authorities, Australia  
National Standards Commission  
N.S.W. Agriculture  
Royal Australian Chemical Institute  
Royal College of Pathologists of Australasia  
Royal Melbourne Institute of Technology  
University of Sydney

---

*Review of Australian Standards.* To keep abreast of progress in industry, Australian Standards are subject to periodic review and are kept up to date by the issue of amendments or new editions as necessary. It is important therefore that Standards users ensure that they are in possession of the latest edition, and any amendments thereto.

Full details of all Australian Standards and related publications will be found in the Standards Australia Catalogue of Publications; this information is supplemented each month by the magazine 'The Australian Standard', which subscribing members receive, and which gives details of new publications, new editions and amendments, and of withdrawn Standards.

Suggestions for improvements to Australian Standards, addressed to the head office of Standards Australia, are welcomed. Notification of any inaccuracy or ambiguity found in an Australian Standard should be made without delay in order that the matter may be investigated and appropriate action taken.

Australian Standard<sup>®</sup>

---

**Density hydrometers**

---

First published as AS K90—1946 (being endorsement of BS 718—1936 without amendment).  
Redesignated as AS R8—1947.  
Second edition 1956 (being endorsement of BS 718:1953 subject to Australian amendment).  
Third edition 1961 (being endorsement of BS 718:1960 subject to Australian amendment).  
Revised and redesignated AS 2026—1977.  
Second edition 1994.

PUBLISHED BY STANDARDS AUSTRALIA  
(STANDARDS ASSOCIATION OF AUSTRALIA)  
1 THE CRESCENT, HOMEBUSH, NSW 2140

ISBN 0 7262 8867 5

## PREFACE

This Standard was prepared by the Standards Australia Committee on Laboratory Glassware and Related Apparatus as a revision of AS 2026—1977 which it supersedes.

The objective of this Standard is to provide a document giving requirements for density hydrometers.

The main provisions of this Standard follow closely those of ISO/649—1981, *Laboratory glassware—Density hydrometers*, Part 1: *Specification*, and Part 2: *Test methods and use*, with minor differences in some technical details and in the units which are used. In ISO/649 densities are expressed in grams per cubic centimetre and surface tensions in dynes per centimetre, whereas in this Standard the corresponding units are kilogram per cubic metre and millinewtons per metre (mN/m).

NOTE: 1 millinewton per metre is equal to 1 dyne per centimetre.

Five series of hydrometers are specified, and dimensions were chosen that would lead to convenience in use and economy in manufacture.

To assist users who may, for special purposes, require a hydrometer which is not included in any of the five series specified, appendices are provided giving notes on the design and adjustment of hydrometers, a method for the determination of density, the measurement of liquid in bulk and for suitable vessels for use with hydrometers.

The Standard does not make provision for the high surface tension category for hydrometers.

The term 'informative' has been used in this Standard to define the application of the appendix to which it applies. An 'informative' appendix is only for information and guidance.

### © Copyright — STANDARDS AUSTRALIA

Users of Standards are reminded that copyright subsists in all Standards Australia publications and software. Except where the Copyright Act allows and except where provided for below no publications or software produced by Standards Australia may be reproduced, stored in a retrieval system in any form or transmitted by any means without prior permission in writing from Standards Australia. Permission may be conditional on an appropriate royalty payment. Requests for permission and information on commercial software royalties should be directed to the head office of Standards Australia.

Standards Australia will permit up to 10 percent of the technical content pages of a Standard to be copied for use exclusively in-house by purchasers of the Standard without payment of a royalty or advice to Standards Australia.

Standards Australia will also permit the inclusion of its copyright material in computer software programs for no royalty payment provided such programs are used exclusively in-house by the creators of the programs.

Care should be taken to ensure that material used is from the current edition of the Standard and that it is updated whenever the Standard is amended or revised. The number and date of the Standard should therefore be clearly identified.

The use of material in print form or in computer software programs to be used commercially, with or without payment, or in commercial contracts is subject to the payment of a royalty. This policy may be varied by Standards Australia at any time.

## CONTENTS

	<i>Page</i>
1 SCOPE .....	4
2 REFERENCED DOCUMENTS .....	4
3 DESIGNATION .....	4
4 RANGES .....	4
5 BASIS OF SCALE .....	4
6 REFERENCE TEMPERATURE .....	5
7 MATERIAL AND WORKMANSHIP .....	5
8 FORM .....	5
9 PRINCIPAL DIMENSIONS .....	5
10 SCALE .....	5
11 GRADUATION LINES .....	5
12 ACCURACY .....	6
13 CATEGORIES OF SURFACE TENSION FOR HYDROMETERS .....	6
14 MARKING .....	6
 APPENDICES	
A DETERMINATION OF DENSITY BY MEANS OF A HYDROMETER .....	12
B THE DESIGN OF HYDROMETERS .....	20
C VESSELS FOR HYDROMETRIC OBSERVATIONS .....	22
D THE MEASUREMENT OF LIQUIDS IN BULBS .....	24

## STANDARDS AUSTRALIA

**Australian Standard**  
**Density hydrometers**

**1 SCOPE** This specification sets out the requirements for five main series and three sub-series of glass hydrometers of constant mass which are graduated to indicate the density (mass per unit volume) of a liquid. The specification does not include hydrometers for use with liquefied petroleum gas or lead-acid batteries.

**NOTES:**

- 1 Appendix A provides information on the use of hydrometers.
- 2 Appendix B provides information on the design of hydrometers.
- 3 Appendix C provides information on vessels which may be used when making hydrometer measurements.
- 4 Appendix D provides information on the determination of hydrometer values for liquids in bulk.

**2 REFERENCED DOCUMENTS** The following documents are referred to in this Standard.

**AS**

- 2163 Graduated measuring cylinders  
2849 Density of water—Numerical value

**ISO**

- 649-1 Laboratory glassware—Density hydrometers for general purposes  
Part 1: Specification  
3675 Crude petroleum and liquid petroleum products—Laboratory determination of density or relative density—Hydrometer method

**3 DESIGNATION** The five series of hydrometers are designated by the symbols L20, L50, M50, M100, and S50, where L, M, and S indicate long, medium, and short, respectively, and 20, 50, and 100 indicate the range of the scale on each hydrometer in the particular series.

The L50, M50 and S50 series contain the special sub-series designated by the suffix SP. These hydrometers are used mainly in petroleum technology.

**4 RANGES**

**4.1 Main series** Each main series of hydrometer shall cover the range of 600 kg/m<sup>3</sup> to 2000 kg/m<sup>3</sup> or 0.600 g/mL to 2.00 g/mL according to the scale marking.

**4.2 Sub-series** Each sub-series of hydrometer shall cover the range of 600 kg/m<sup>3</sup> to 1100 kg/m<sup>3</sup> or 0.600 g/mL to 1.100 g/mL according to the scale marking.

**5 BASIS OF SCALE** The basis of scale shall be density (mass per unit volume) in kilograms per cubic metre (kg/m<sup>3</sup>). The use of g/cm<sup>3</sup>, for which the symbol g/mL may be used, is acceptable.

**NOTES:**

- 1 The term millilitre (mL) is commonly used as a special name for cubic centimetre (cm<sup>3</sup>), in accordance with a decision of the 12th conference Generale des Poids et Mesures. The term millilitre, is acceptable, in general, for reference to capacities of volumetric glassware and is used in the present text.
- 2 If water is used for calibration, AS 2849 provides density values.