

Australian Standard[®]

Medical suction equipment

**Part 3: Suction equipment powered
from a vacuum or pressure source**

This Australian Standard was prepared by Committee HT/4, Medical Gases and Pipeline Services. It was approved on behalf of the Council of Standards Australia on 22 July 1992 and published on 16 November 1992.

The following interests are represented on Committee HT/4:

Association of Consulting Engineers, Australia
Australian Society of Anaesthetists
Confederation of Australian Industry
Department of Health, Housing and Community Services (Commonwealth)
Department of Health, N.S.W.
Health Department, W.A.
Institute of Hospital Engineering, Australia
Medical Industry Association of Australia
Metal Trades Industry Association of Australia
Public Works Department, N.S.W.
Royal Australasian College of Surgeons
South Australian Health Commission

Review of Australian Standards. To keep abreast of progress in industry, Australian Standards are subject to periodic review and are kept up to date by the issue of amendments or new editions as necessary. It is important therefore that Standard users ensure that they are in possession of the latest edition, and any amendments thereto.

Full details of all Australian Standards and related publications will be found in the Standards Australia Catalogue of Publications, the information is supplemented each month by the magazine 'The Australian Standard', which subscribing members receive, and which gives details of new publications, new editions and amendments, and of withdrawn Standards.

Suggestions for improvements to Australian Standards, addressed to the head office of Standards Australia, are welcomed. Notification of any inaccuracy or ambiguity found in an Australian Standard should be made without delay so that the matter may be investigated and appropriate action taken.

© Copyright — STANDARDS AUSTRALIA

Users of Standards are reminded that copyright subsists in all Standards Australia publications and software. Except where the Copyright Act allows and except where provided for below no publications or software produced by Standards Australia may be reproduced, stored in a retrieval system in any form or transmitted by any means without prior permission in writing from Standards Australia. Permission may be conditional on an appropriate royalty payment. Requests for permission and information on commercial software royalties should be directed to the head office of Standards Australia.

Standards Australia will permit up to 10 percent of the technical content pages of a Standard to be copied for use exclusively in-house by purchasers of the Standard without payment of a royalty or advice to Standards Australia.

Standards Australia will also permit the inclusion of its copyright material in computer software programs for no royalty payment provided such programs are used exclusively in-house by the creators of the programs.

Care should be taken to ensure that material used is from the current edition of the Standard and that it is updated whenever the Standard is amended or revised. The number and date of the Standard should therefore be clearly identified.

The use of material in print form or in computer software programs to be used commercially, with or without payment, or in commercial contracts is subject to the payment of a royalty. This policy may be varied by Standards Australia at any time.

STANDARDS AUSTRALIA

RECONFIRMATION

OF

AS 2120.3—1992

Medical suction equipment

Part 3: Suction equipment powered from a vacuum or pressure source

RECONFIRMATION NOTICE

Technical Committee HE-017 has reviewed the content of this publication and in accordance with Standards Australia procedures for reconfirmation, it has been determined that the publication is still valid and does not require change.

Certain documents referenced in the publication may have been amended since the original date of publication. Users are advised to ensure that they are using the latest versions of such documents as appropriate, unless advised otherwise in this Reconfirmation Notice.

Approved for reconfirmation in accordance with Standards Australia procedures for reconfirmation on 22 November 2016.

The following are represented on Technical Committee HE-017:

Australia New Zealand Industrial Gas Association
Australian and New Zealand College of Anaesthetists
Australian Chamber of Commerce and Industry
Australian Industry Group
Australian Society of Anaesthetists
Engineers Australia
Master Plumbers Association of QLD
Master Plumbers Australia
Ministry of Consumer Affairs (NZ)

NOTES

Currently in preview, click buy full vers.

Australian Standard[®]

Medical suction equipment

**Part 3: Suction equipment powered
from a vacuum or pressure source**

First published as part of AS 2120—1977.
Revised and redesignated in part as AS 2120.3—1992.

PUBLISHED BY STANDARDS AUSTRALIA
(STANDARDS ASSOCIATION OF AUSTRALIA)
1 THE CRESCENT, HOMEBUSH, NSW 2140

ISBN 0 7262 7740 1

PREFACE

This Standard was prepared by the Standards Australia Committee on Medical Gases and Pipeline Services, under the direction of the Multitechnics Standards Policy Board, to supersede, in part, AS 2120—1977, *Suction systems for medical use in hospitals*.

This Standard is the third in a series of Standards for medical suction equipment, and deals only with suction equipment powered from a vacuum or pressure source. Part 1 deals with safety requirements for electrically-powered suction equipment, and Part 2 deals with manually-powered suction equipment.

This Standard has been prepared in response to a need for a safety and performance standard for suction systems. Suction is used to clear the airway and remove unwanted material from body cavities. Suction is also used to assist drainage and decompress body cavities. Suction and vacuum systems are used widely in health care facilities such as hospitals, for domiciliary care of patients who are nursed at home, and in emergency situations both outside hospitals in field conditions and during transport in ambulances.

As far as possible, this Standard has been written specifying performance requirements for effective and safe treatment of the patient.

Some devices specified in this Standard are intended to be operated in conjunction with a pipeline complying with AS 2896—1991, *Medical gas systems—Installation and testing of non-flammable medical gas pipeline systems*.

The clauses of this Standard supplement or modify the corresponding clauses of ISO 10079-3, *Medical suction equipment — Part 3: Suction equipment powered from a vacuum or pressure source*. Although this Standard closely follows ISO 10079-3 in format and technical content, some of the requirements of that publication have been modified to take account of local conditions.

Where this Standard deviates technically from ISO 10079-3 by way of additional or different requirements, the deviation is indicated by a rule in the margin against the clause, or part thereof, affected. Minor changes are not indicated. An annex to the Standard lists the variations from ISO 10079-3.

Appendix A is normative; it gives test methods to be used to verify compliance with the requirements given in this Standard.

CONTENTS

	<i>Page</i>
1 SCOPE	3
2 REFERENCED DOCUMENTS	3
3 DEFINITIONS	3
4 CLEANING AND STERILIZATION	5
5 DESIGN REQUIREMENTS	5
6 OPERATIONAL REQUIREMENTS	6
7 PHYSICAL REQUIREMENTS	7
8 PERFORMANCE REQUIREMENTS FOR VACUUM AND FLOW	8
9 GAS SUPPLY	8
10 VACUUM REGULATOR	8
11 RESISTANCE TO HARSH ENVIRONMENTS	8
12 MARKING	8
13 INFORMATION TO BE SUPPLIED BY THE MANUFACTURER	9
A. APPENDIX A TEST METHODS	10
ANNEX	16

STANDARDS AUSTRALIA

Australian Standard
Medical suction equipment

Part 3: Suction equipment powered from a vacuum or pressure source

1 SCOPE This Standard specifies safety and performance requirements for medical suction equipment powered from a vacuum or pressure source (see Figure 1). In particular it applies to connections to the terminal unit of vacuum systems.

NOTE: Suction equipment with electronic timing, controlled by electrical means, may also need to comply with AS 3200.1. This Standard does not apply to transportable electrically-powered suction equipment, whether mains or battery-powered, which is dealt with in AS 2120.1 nor to manually-powered suction equipment which is dealt with in AS 2120.2 nor to the following:

- (a) Central power source (by vacuum/compressed air generation), piping systems of vehicles and buildings, and wall connectors.
- (b) Catheter tubes, drains, curettes and suction tips.
- (c) Syringes.
- (d) Dental suction equipment complying with AS 2686.2.
- (e) Waste gas scavenging systems.
- (f) Laboratory suction.
- (g) Autotransfusion systems.
- (h) Passive urinary drainage.
- (i) Closed systems for wound drainage.
- (j) Gravity gastric drainage.
- (k) Orally-operated mucous extractors.
- (l) Suction equipment where the collection container is downstream of the vacuum pump.
- (m) Equipment marked as suction unit for permanent tracheostomy.
- (n) Ventouse (obstetric) equipment.
- (o) Neonatal mucous extractors.
- (p) Breast pumps.
- (q) Liposuction.
- (r) Uterine aspiration.

2 REFERENCED DOCUMENTS The following documents are referred to in this Standard:

AS	
2496	Breathing attachments for anaesthetic purposes for human use
2686	Dental equipment—Suction systems
2686.2	Part 2: Mobile systems
2700	Colour standards for general purposes
2896	Medical gas systems—Installation and testing of non-flammable medical gas pipeline systems
2902	Medical gas systems—Low pressure flexible connecting assemblies (hose assemblies)
3200	Approval and test specification—Electromedical equipment—General requirements
3200.1	Part 1: General requirements for safety
IS 2	
107	Small medical gas cylinders—Yoke-type valve connections

3 DEFINITIONS For the purposes of this Standard, the definitions below apply.

- 3.1 Collection container**—container in which liquids or solid particles are collected.
- 3.2 Drainage**—removal of fluids from a body cavity or wound, assisted by vacuum.
- 3.3 End piece**—that part of the suction equipment applied to the patient. The end piece starts at the site where material is drawn in and ends at the first detachable connection.
- 3.4 Exhaust opening**—port or ports through which exhaust is discharged.
- 3.5 Filter**—device for separation of particulate matter.