

Australian Standard™

**Sterilizers—Steam—
Downward-displacement**

Currently in preview, click buy full version

This Australian Standard was prepared by Committee ME-003, Sterilizing Equipment. It was approved on behalf of the Council of Standards Australia on 21 June 2002 and published on 25 July 2002.

The following are represented on Committee ME-003:

Australian Chamber of Commerce and Industry
Australian Dental Association
Australian Dental Industry Association
Australian Health Industry
Australian Industry Group
Commonwealth Department of Health and Ageing
Department of Human Services (Victoria)
Health Department of Western Australia
Institute of Hospital Engineering Australia
Medical Industry Association of Australia
New South Wales Health Department
Queensland Health
South Australian Health Commission
Sterilizing Research and Advisory Council of Australia (N.S.W. Branch)

Keeping Standards up-to-date

Standards are living documents which reflect progress in science, technology and systems. To maintain their currency, all Standards are periodically reviewed, and new editions are published. Between editions, amendments may be issued. Standards may also be withdrawn. It is important that readers assure themselves they are using a current Standard, which should include any amendments which may have been published since the Standard was purchased.

Detailed information about Standards can be found by visiting the Standards Australia website at www.standards.com.au and looking up the relevant Standard in the on-line catalogue.

Alternatively, the printed Catalogue provides information current at 1 January each year and the monthly magazine, *The Australian Standard*, has a full listing of revisions and amendments published each month.

We also welcome suggestions for improvement in our Standards, and especially encourage readers to notify us immediately of any apparent inaccuracies or ambiguities. Contact us via email at mail@standards.com.au, or write to the Chief Executive, Standards Australia International Ltd, GPO Box 5420, Sydney, NSW 2001.

Australian Standard™

**Sterilizers—Steam—
Downward-displacement**

Formulated as part of AS T54—1971 and AS T55—1971.
Previous edition AS 2192—1991.
Third edition 2002.

COPYRIGHT

© Standards Australia International

All rights are reserved. No part of this work may be reproduced or copied in any form or by any means, electronic or mechanical, including photocopying, without the written permission of the publisher.

Published by Standards Australia International Ltd
GPO Box 5420, Sydney, NSW 2001, Australia

ISBN 0 7337 4662 4

PREFACE

This Standard was prepared by the Standards Australia Committee ME-003, Sterilizing Equipment, to supersede AS 2192—1991.

The objective of this Standard is to provide users with equipment that will perform safely as intended.

Technical changes from the previous edition include the following:

- (a) Clause 1.4 includes some definitions.
- (b) Clause 2.8 includes amended requirements for wheels of carriages.
- (c) Clause 3.6 amends the requirements for surface finish.
- (d) Clause 3.18 includes requirements to prohibit the introduction of superheated steam and amendments to requirements for steam separation.
- (e) Clause 3.23.1 for temperature indicators has been amended.
- (f) Clause 3.25.2 includes amended requirements for colours of stage indicators.
- (g) Section 6 on testing has been amended.
- (h) Paragraphs H1 and H2, which prescribe standard fabric and instrument test packs, have been amended.
- (i) Requirements and advice regarding the sterilization of fluids are included in Clauses 3.12.1, 3.24.3, Table 4 and Appendix E.

The terms 'normative' and 'informative' have been used in this Standard to define the application of the appendix to which they apply. A 'normative' appendix is an integral part of a Standard, whereas an 'informative' appendix is only for information and guidance.

CONTENTS

	<i>Page</i>
SECTION 1 SCOPE AND GENERAL	
1.1 SCOPE.....	5
1.2 INNOVATION	5
1.3 REFERENCED DOCUMENTS.....	5
1.4 DEFINITIONS.....	5
1.5 EXPRESSION OF PRESSURE VALUES	8
SECTION 2 MATERIALS	
2.1 WELDING AND BRAZING CONSUMABLES	9
2.2 MATERIALS FOR CONNECTIONS WELDED TO THE CHAMBER OR JACKET	9
2.3 CLADDING.....	9
2.4 GALVANIZING.....	9
2.5 CHAMBERS, JACKETS, CLOSURE HOLDING ELEMENTS AND DOORS	9
2.6 FRAMEWORK.....	10
2.7 EXTERNAL PANELS.....	10
2.8 LOADING EQUIPMENT.....	11
SECTION 3 DESIGN AND CONSTRUCTION	
3.1 GENERAL.....	12
3.2 WELDING.....	12
3.3 CORROSION	12
3.4 NOISE	12
3.5 THERMAL INSULATION.....	12
3.6 EXTERNAL PANELS.....	12
3.7 VENTILATION.....	13
3.8 SUPPORT FRAME	13
3.9 PRESSURIZED COMPONENTS.....	13
3.10 JACKET	13
3.11 CHAMBERS	13
3.12 DOORS.....	14
3.13 LOADING EQUIPMENT.....	15
3.14 PIPEWORK AND PIPE FITTINGS	16
3.15 VALVES	16
3.16 MECHANICAL AIR REMOVAL	16
3.17 AIR FILTERS.....	16
3.18 STEAM.....	17
3.19 COMPRESSED AIR.....	17
3.20 DRAINS	17
3.21 EQUIPMENT ON THE FRONT PANEL	18
3.22 PRESSURE INDICATORS	18
3.23 TEMPERATURE INDICATORS	19
3.24 AUTOMATIC PROCESS CONTROL	20
3.25 STAGE INDICATORS.....	21
3.26 PROCESS RECORDER	21
3.27 ELECTRICAL.....	22
3.28 ELECTROMAGNETIC DISTURBANCE.....	22

SECTION 4 OPERATING CYCLE

4.1	AUTOMATIC CONTROL	23
4.2	MANDATORY STAGES	23
4.3	OPTIONS	24

SECTION 5 MARKING AND INFORMATION

5.1	LANGUAGE	26
5.2	MARKING	26
5.3	TECHNICAL INFORMATION.....	26

SECTION 6 TESTING

6.1	PERFORMANCE TESTING	27
6.2	COMMISSIONING TEST.....	28
6.3	DEMONSTRATION OF COMPLIANCE	29

APPENDICES

A	INFORMATION THAT SHOULD BE SUPPLIED WITH ENQUIRIES AND ORDERS	30
B	QUALITY OF WATER.....	32
C	STEAM SUPPLIES	34
D	STANDARDS FOR MACHINES THAT WASH, STEERLIZE OR DRY MEDICAL EQUIPMENT IN HEALTH CARE FACILITIES.....	43
E	REFERENCED DOCUMENTS.....	44
F	INTRODUCING THERMOCOUPLES INTO A STERILIZER CHAMBER	46
G	STERILIZATION OF BOTTLED FLUIDS.....	49
H	INDICATING MATERIAL.....	50

STANDARDS AUSTRALIA

Australian Standard Sterilizers—Steam— Downward-displacement

SECTION 1 SCOPE AND GENERAL

1.1 SCOPE

This Standard specifies requirements for horizontal downward-displacement steam sterilizers (including jacketed types and non-jacketed types) that are intended for use in hospitals and other non-industrial applications.

NOTES:

- 1 Downward-displacement-type steam sterilizers should not be used to sterilize any instrument or equipment that is lumened or blind lumened, as specified in AS 4187 (Clause 4.2.2 in the 2002 edition).
- 2 Steam sterilizers require an independently generated saturated steam supply that has a dryness value of not less than 0.97 and an adequate steam quality. Information on steam quality is provided by HTM 2031.
- 3 Steam sterilizers should not be installed in any location where flammable anaesthetics or volatile liquids may be used, unless the location and the sterilizer comply with the relevant requirements of AS 1169, AS 2380.2 and AS 2380.7.
- 4 Guidance on information that should be supplied with enquiries and orders is given in Appendix A.
- 5 Information on quality of water is given in Appendix B.
- 6 Information on steam supplies is given in Appendix C.
- 7 Standards for machines that wash, sterilize or dry medical equipment in health care facilities are listed in Appendix D.

1.2 INNOVATION

Any use of new or unusual materials or methods to achieve the intent of this Standard is encouraged. Manufacturers employing innovative materials or methods shall show documented third party proof that the innovation is not inferior to that of the intent of this Standard.

1.3 REFERENCED DOCUMENTS

The documents referred to in this Standard are listed in Appendix E.

1.4 DEFINITIONS

For the purpose of this Standard, the definitions in AS 4187 and those below apply:

1.4.1 Air removal

Removal of air from the sterilizer chamber and sterilizer load, sufficient to facilitate steam penetration.

1.4.2 Autoclave

Colloquial term for steam-under-pressure sterilizer.