

Dup

Under Revision See DR 96111

SUPERSEDED BY: AS/NZS 2243.1:1997

AS 2243.1—1990

Australian Standard®

Safety in laboratories

Part 1: General



STANDARDS AUSTRALIA



F

This Australian Standard was prepared by Committee CH/26, Safety in Laboratories. It was approved on behalf of the Council of Standards Australia on 11 June 1990 and published on 15 October 1990.

The following interests are represented on Committee CH/26:

Australian Government Analytical Laboratories
Australian Institute of Petroleum
National Association of Testing Authorities, Australia

Additional interests participating in preparation of Standard:

Australian Nuclear Science and Technology Organisation
Department of Agriculture and Rural Affairs
Prince Henry Hospital
University of Melbourne

Review of Australian Standards. To keep abreast of progress in industry, Australian Standards are subject to periodic review and are kept up-to-date by the issue of amendments or new editions as necessary. It is important, therefore, that Standards users ensure that they are in possession of the latest edition, and any amendments thereto.

Full details of all Australian Standards and related publications will be found in the Standards Australia Catalogue of Publications; this information is supplemented each month by the magazine 'The Australian Standard', which subscribing members receive, and which gives details of new publications, new editions and amendments, and of withdrawn Standards.

Suggestions for improvements to Australian Standards, addressed to the head office of Standards Australia, are welcomed. Notification of any inaccuracy or ambiguity found in an Australian Standard should be made without delay in order that the matter may be investigated and appropriate action taken.

STANDARDS AUSTRALIA

Amendment No 1
to
AS 2243.1—1990
Safety in laboratories
Part 1: General

REVISED TEXT



The 1990 edition of AS 2243.1 is amended as follows; the amendments should be inserted in the appropriate place.

SUMMARY: This Amendment applies to Clauses 1.4.18, 1.4.19, 2.7 and 3.1.1.

Published on 9 August 1991.

Page 6 New Clauses 1.4.18 and 1.4.19

Add new clauses as follows:

1.4.18 Flammable liquid—a liquid which has a flashpoint below 61 °C. (See AS 1940)

1.4.19 Laboratory compartment—a fire-isolated compartment in a laboratory, being either the whole laboratory or a subdivision of the laboratory, complying with the limitations given in AS 2982. Laboratory compartments may be of four types according to their flammability hazard rating of high, medium, low, or very low in accordance with AS 2982.

NOTE: A laboratory compartment may be divided into subcompartments, such as work benches, stores or offices, or may be an open-plan area.

Page 10 Clause 2.7

Delete existing clause and substitute the following:

2.7 STORAGE OF HAZARDOUS SUBSTANCES

2.7.1 General Special provisions shall be made for the safe handling and storage of hazardous substances. For information on the design and construction of installations for the storage and handling of flammable and combustible liquids, see AS 1940 and AS 2982. Storage cabinets for flammable liquids shall comply with AS 2982. Additional advice can be obtained from local and state regulatory authorities.

2.7.2 Flammable liquids Flammable liquids shall not be stored in laboratory compartments in amounts exceeding the quantities shown in AS 2982 and AS 1940 or in relevant government regulations.

2.7.3 Change in laboratory functions Before a change is made in the function of a laboratory compartment which involves the storage or use of chemicals, that compartment shall be assessed for its suitability for the new function.

Page 11 Clause 3.1.1(i)

Delete existing Clause 3.1.1(i) and substitute the following:

- (i) Updating safe procedures when new chemical compounds or methods are introduced, or when the function of the laboratory compartment is changed.

Australian Standard®

Safety in laboratories

Part 1: General

First published as AS 2243.1—1979.
Second edition 1982.
Third edition 1990.

PREFACE

This Standard was prepared by the Standards Australia Committee on Safety in Laboratories, under the direction of the Chemical Standards Board, to supersede the 1982 edition. This edition includes a major revision to Section 2 in light of the publication of AS 2982, *Laboratory construction*, and to Section 3 to elaborate on the responsibilities for laboratory safety.

The Standard comprises Part 1 of a nine-part series designed to provide basic coverage of all important aspects of the safety function in laboratories. It deals with the general aspects of safety common to all kinds of laboratories and is intended to be used in conjunction with other Parts of the series, which relate to particular aspects of laboratory operations and to particular kinds of hazard. It stresses the importance of preventive measures and sets out the organization of safety techniques, emergency procedures, and first aid.

The other Parts in the series are as follows:

- Part 2: Chemical aspects
- Part 3: Microbiology
- Part 4: Ionizing radiations
- Part 5: Non-ionizing radiations
- Part 6: Mechanical aspects
- Part 7: Electrical aspects
- Part 8: Fume cupboards
- Part 9: Recirculating fume cabinets

© Copyright — STANDARDS AUSTRALIA

Users of Standards are reminded that copyright subsists in all Standards Australia publications and software. Except where the Copyright Act allows and except where provided for below no publications or software produced by Standards Australia may be reproduced, stored in a retrieval system in any form or transmitted by any means without prior permission in writing from Standards Australia. Permission may be conditional on an appropriate royalty payment. Requests for permission and information on commercial software royalties should be directed to the Head Office of Standards Australia.

Standards Australia will permit up to 10 percent of the technical content pages of a Standard to be copied for use exclusively in-house by purchasers of the Standard without payment of a royalty or advice to Standards Australia.

Standards Australia will also permit the inclusion of its copyright material in computer software programs for no royalty payment provided such programs are used exclusively in-house by the creators of the programs.

Care should be taken to ensure that material used is from the current edition of the Standard and that it is updated whenever the Standard is amended or revised. The number and date of the Standard should therefore be clearly identified.

The use of material in print form or in computer software programs to be used commercially, with or without payment, or in commercial contracts is subject to the payment of a royalty. This policy may be varied by Standards Australia at any time.

CONTENTS

	<i>Page</i>
FOREWORD	4
SECTION 1 SCOPE AND GENERAL	
1.1 SCOPE	5
1.2 APPLICATION	5
1.3 REFERENCED AND RELATED DOCUMENTS	5
1.4 DEFINITIONS	5
1.5 RECOMMENDATIONS FOR SAFE CONDUCT	6
SECTION 2 LABORATORY DESIGN FOR SAFETY	
2.1 CONSTRUCTION AND DESIGN	7
2.2 FIRE DETECTION AND EXPLOSION PROTECTION (AUTOMATIC SYSTEMS)	8
2.3 EMERGENCY ALARM SYSTEMS AND ASSOCIATED OPERATIONS	8
2.4 INFORMATION ON HAZARDOUS MATERIALS	9
2.5 SAFETY INSPECTIONS	9
2.6 SAFETY EQUIPMENT	9
2.7 STORAGE OF HAZARDOUS SUBSTANCES	10
2.8 VENTILATION	10
2.9 ELECTRICAL INSTALLATIONS	10
SECTION 3 RESPONSIBILITY FOR LABORATORY SAFETY AND EMERGENCY PROCEDURES	
3.1 RESPONSIBILITY FOR SAFETY PROCEDURES	11
3.2 FIRE, EMERGENCY AND RESCUE PROCEDURES	11
3.3 FIRST-AID	12
SECTION 4 GENERAL SAFETY PROCEDURES	
4.1 USE OF PERSONAL PROTECTIVE EQUIPMENT	13
4.2 CONTAINERS FOR HAZARDOUS MATERIALS	13
4.3 USE OF FUME CUPBOARDS	14
4.4 GENERAL LABORATORY PRACTICES	15
4.5 OPERATION OF INSTRUMENTS	16
4.6 OPERATIONS UNDER VACUUM	16
4.7 OPERATIONS UNDER INTERNAL PRESSURE	17
4.8 HANDLING OF LABORATORY WASTES	18
APPENDICES	
A LIST OF REFERENCED AND RELATED DOCUMENTS	19
B HAZARDOUS AREAS NEEDING AUTOMATIC FIRE DETECTION OR PROTECTION DEVICES	21
C TYPICAL WORK INJURY REPORT	22

FOREWORD

Safety in laboratories is an individual and personal responsibility, as well as being a management responsibility. Staff training is generally directed towards making safety considerations an attitude of mind, and an integral part of all laboratory procedures so that a constant purposeful control of the laboratory environment will result.

No document can ensure safe practice within the laboratory. The best that it can do is to impart the necessary awareness and information required. The final responsibility for safe conduct always rests with the individual. To help in maintaining optimum safety in a laboratory, however, each laboratory section should become the responsibility of one individual.

Responsibility for the enforcement of safety rules, and for the promulgation of the necessary information, devolves upon the senior members of the laboratory. It is their duty to supervise their less experienced colleagues and to acquaint them with the potential hazards involved in every reaction and operation, so that, from the very beginning, new staff become safety-conscious and learn the techniques of safe working. The rules of safety should be so well ingrained in every employee that their observance becomes a habitual attitude of mind. It cannot be over-emphasized that personal attitude is just as important to safety as the learning of relevant facts. New staff, and particularly young students or graduates, are likely to be over-enthusiastic, and may take risks to save time. Firm, but diplomatic, personal supervision by senior staff at an early stage will help establish a team spirit, and foster a respect for laboratory safety—factors of foremost importance in the safe and smooth functioning of a laboratory.

STANDARDS AUSTRALIA

Australian Standard Safety in laboratories

Part 1: General

SECTION 1 SCOPE AND GENERAL

1.1 SCOPE This Standard provides information and sets out recommendations, general procedures and precautions designed to promote safety of persons and property in laboratory operations. The safety aspects described in this Standard apply to personnel who use the laboratory facilities including students, visitors, maintenance staff, cleaners, laboratory staff, contractors, security staff and other authorized personnel. This Standard deals specifically with safe practices in laboratories and does not cover the design and construction of laboratories, which is the subject of AS 2982.

1.2 APPLICATION This Standard should be used in conjunction with the appropriate part(s) of AS 2243 that are relevant to the type of work being carried out in the laboratory. Where the requirements of any part of this Standard are covered by state or Commonwealth regulations, then these should also be read in conjunction with this Standard and, where any conflict exists between the two, the provisions of the statutory regulations shall apply.

1.3 REFERENCED AND RELATED DOCUMENTS A list of references and related documents is given in Appendix A.

1.4 DEFINITIONS For the purpose of this Standard, the definitions below apply.

1.4.1 Accident—any occurrence which results in personal injury, disease or death, or property damage.

1.4.2 Auto-ignition temperature—the lowest temperature at which a vapour will spontaneously catch fire in air.

NOTE: If vapour from a liquid at room temperature is carried by draught into contact with a surface above the ignition temperature, the vapour may ignite and burn back to ignite the liquid.

1.4.3 Corrosive—having the characteristic of damaging or destroying by direct chemical action; this includes the effect of caustic substances.

1.4.4 Exposure standard—an airborne concentration of a particular substance in the worker's breathing zone, exposure to which, according to current knowledge, should not cause adverse health effects nor cause undue discomfort to nearly all workers. The exposure standard can be of three forms: time-weighted average, peak, or short-term exposure limit.

(a) *time-weighted average (TWA)*—the average airborne concentration of a particular substance when calculated over a normal eight-hour work day, for a five-day working week.

(b) *peak*—a maximum or peak airborne concentration of a particular substance determined over the shortest analytically practicable period of time, which does not exceed 15 min.

(c) *short-term exposure limit (STEL)*—a 15 min TWA exposure which should not be exceeded at any time during a work day even if the eight-hour TWA average is within the TWA exposure standard. Exposure at the STEL should not be longer than 15 min and should not be repeated more than four times per day. There should be at least 60 min between successive exposures at the STEL.

1.4.5 Flammable—capable of being readily ignited and of burning in air.

1.4.6 Flashpoint—the lowest temperature at which application of a small flame causes the vapour above a flammable liquid to produce a momentary flash when it is heated under standardized conditions specified in an appropriate test method, e.g. AS 2106.

1.4.7 Harmful processes—processes in which energy or harmful substances are used or produced and which, upon uncontrolled release, may cause injury to persons or damage property.

1.4.8 Hazardous materials—materials which, alone or in combination with others, are or could become toxic, irritant, explosive, flammable, corrosive, or noxious, and for which special precautions are required.

NOTE: Included in this definition are all materials that are classified as 'dangerous goods' by the Australian Code for the Transport of Dangerous Goods by Road and Rail.

1.4.9 Head officer—the person appointed by the organization to be accountable for all functions of the laboratory area.