

AS 2243.3

Australian National Council on AIDS. Lab safety
guidelines. Bulletin No.3 (held in SIC) 344-9208369792
Superseded by AS/ASZS 2243.3:1995

AS 2243.3—1991

Under Reconsideration

Amendment 1 - February 1993

DR94215

Australian Standard®

Safety in laboratories

Part 3: Microbiology



STANDARDS AUSTRALIA



This Australian Standard was prepared by Committee CH/26, Safety in Laboratories. It was approved on behalf of the Council of Standards Australia on 12 December 1990 and published on 28 March 1991.

The following interests are represented on Committee CH/26:

Australian Government Analytical Laboratories
Australian Institute of Petroleum
Australian Nuclear Science and Technology Organization
Bureau of Steel Manufacturers
Chemical Confederation of Australia
Confederation of Australian Industry
CSIRO, Division of Energy Technology
Department of Agricultural Rural Affairs, Vic.
Department of Defence
National Association of Testing Authorities, Australia
Royal Australian Chemical Institute
University of Melbourne

Additional interests participating in preparation of Standard:

Australian Animal Health Laboratory, Geelong, Vic.
Australian Society for Microbiology
Department of Community Services and Health
Fairfield Hospital, Fairfield, Vic.
Institute of Chemical Pathology and Medical Research
Monash University
Repatriation General Hospital, Concord, NSW
Royal Melbourne Hospital
Royal North Shore Hospital, St Leonards, NSW
Sydney College of TAFE
University of NSW

Review of Australian Standards. To keep abreast of progress in industry, Australian Standards are subject to periodic review and are kept up-to-date by the issue of amendments or new editions as necessary. It is important therefore that Standards users ensure that they are in possession of the latest edition, and any amendments thereto.

Full details of all Australian Standards and related publications will be found in the Standards Australia Catalogue of Publications; this information is supplemented each month by the magazine 'The Australian Standard', which subscribing members receive, and which gives details of new publications, new editions and amendments, and of withdrawn Standards.

Suggestions for improvements to Australian Standards, addressed to the head office of Standards Australia, are welcomed. Notification of any inaccuracy or ambiguity found in an Australian Standard should be made without delay in order that the matter may be investigated and appropriate action taken.

AS 2243.3—1991

Australian Standard®

Safety in laboratories

Part 3: Microbiology

First published as AS 2243.3—1979.
Second edition 1985.
Third edition 1991.

PUBLISHED BY STANDARDS AUSTRALIA
(STANDARDS ASSOCIATION OF AUSTRALIA)
STANDARDS HOUSE, 80 ARTHUR ST, NORTH SYDNEY NSW
ISBN 0 7262 6715 5

PREFACE

This Standard was prepared by the Standards Australia Committee on Safety in Laboratories, under the direction of the Chemical Standards Board, to supersede AS 2243.3—1985. This Standard is the third in the AS 2243 series aimed at promoting safety in laboratories. This Standard deals with microbiological aspects of laboratory work and is intended to be used in conjunction with Parts 1 and 2 in the series, relating to general and chemical safety aspects of laboratory work. This edition of the Standard contains a complete revision of the requirements for laboratories dealing with infectious diseases. It classifies microorganisms into four risk groups and contains a description of four levels of containment for microbiology laboratories.

Other Parts of the AS 2243 series are as follows:

- Part 1: *General*
- Part 2: *Chemical aspects*
- Part 4: *Ionizing radiations*
- Part 5: *Non-ionizing radiations*
- Part 6: *Mechanical aspects*
- Part 7: *Electrical aspects*
- Part 8: *Fume cupboards*
- Part 9: *Recirculating fume cabinets*

The term 'normative' referred to in the Appendices of this Standard indicates that the Appendix is an integral part of the Standard and that all requirements therein must be met. The term 'informative' indicates that the Appendices are for information only.

AS 2243.3/Amdt 1/1993-02-15

STANDARDS AUSTRALIA

Amendment No. 1

to

AS 2243.3—1991

Safety in laboratories

Part 3: Microbiology

REVISED TEXT

The 1991 edition of AS 2243.3 is amended as follows; the amendment should be inserted in the appropriate place.

SUMMARY: This Amendment applies to Clause 3.8.2.

Published on 15 February 1993.

© Copyright — STANDARDS AUSTRALIA

Users of Standards are reminded that copyright subsists in all Standards Australia publications and software. Except where the Copyright Act allows and except where provided for below no publications or software produced by Standards Australia may be reproduced, stored in a retrieval system in any form or transmitted by any means without prior permission in writing from Standards Australia. Permission may be conditional on an appropriate royalty payment. Requests for permission and information on commercial software royalties should be directed to the head office of Standards Australia.

Standards Australia will permit up to 10 percent of the technical content pages of a Standard to be copied for use exclusively in-house by purchasers of the Standard without payment of a royalty or advice to Standards Australia.

Standards Australia will also permit the inclusion of its copyright material in computer software programs for no royalty payment provided such programs are used exclusively in-house by the creators of the programs.

Care should be taken to ensure that material used is from the current edition of the Standard and that it is updated whenever the Standard is amended or revised. The number and date of the Standard should therefore be clearly identified.

The use of material in print form or in computer software programs to be used commercially, with or without payment, or in commercial contracts is subject to the payment of a royalty. This policy may be varied by Standards Australia at any time.

CONTENTS

	<i>Page</i>
FOREWORD	5
SECTION 1 SCOPE AND GENERAL	
1.1 SCOPE	6
1.2 REFERENCED DOCUMENTS	6
1.3 DEFINITIONS	6
SECTION 2 ORGANIZATION AND RESPONSIBILITY	
2.1 RESPONSIBILITY	8
2.2 HEALTH MANAGEMENT OF STAFF	8
2.3 PROVISION OF SERVICES	8
SECTION 3 GENERAL LABORATORY PRACTICES AND PROCEDURES	
3.1 GENERAL	10
3.2 APPAREL	10
3.3 WORK PRACTICES	10
3.4 TREATMENT OF LABORATORY SPILLS	11
3.5 GENERAL APPARATUS	12
3.6 FREEZE-DRYING OF CULTURES	13
3.7 USE OF AUTOCLAVES	13
3.8 WASTE DISPOSAL	14
3.9 TRANSPORT OF INFECTIOUS MATERIAL	15
3.10 ACCIDENT TREATMENT	15
SECTION 4 LABORATORY CLEANING	
4.1 GENERAL	17
4.2 CLEANING PERSONNEL	17
4.3 BENCHES AND WORK SURFACES	17
4.4 WALLS AND SHELVES	17
4.5 FLOOR CLEANING	17
SECTION 5 DEGREE OF HAZARD FROM MICROORGANISMS	
5.1 GENERAL	18
5.2 CLASSIFICATION OF INFECTIVE MICROORGANISMS BY RISK GROUP	18
5.3 RISK-GROUPING OF MICROORGANISMS BY TYPE	18
SECTION 6 TYPES OF LABORATORY	
6.1 CLASSIFICATION OF LABORATORIES	25
6.2 LABORATORIES FOR RECOMBINANT DNA AND GENETIC MANIPULATION	25
SECTION 7 WORK AREAS	
7.1 WARNING SIGNS	26
7.2 SEGREGATION OF SPECIAL AREAS	26
7.3 BENCHES AND SEATING HEIGHTS	26
7.4 VERMIN	26
7.5 VENTILATION AND AIR CONDITIONERS	26
SECTION 8 SPECIAL SAFETY EQUIPMENT	
8.1 PERSONNEL PROTECTION	27
8.2 FIRST AID KIT	27
8.3 BIOLOGICAL SAFETY CABINETS	27
8.4 CYTOTOXIC DRUG SAFETY CABINETS	27

	<i>Page</i>
SECTION 9 ANIMALS AND ANIMAL HOUSES	
9.1 GENERAL	28
9.2 SPECIAL AREAS	28
9.3 CAGES AND RACKS	28
9.4 CROSS-INFECTION	28
9.5 PERSONAL PROTECTION	28
9.6 DERMAL ENTRY OF INFECTION	28
9.7 ALLERGIES	28
9.8 DISSECTION AND POST-MORTEM EXAMINATIONS	28
9.9 STERILIZATION AND WASTE DISPOSAL	28
APPENDICES	
A TOXIC SUBSTANCES USED IN THE MICROBIOLOGY LABORATORY	29
B DISINFECTANTS AND ANTISEPTICS	30
C MICROBIOLOGICAL ACCIDENT REPORT	33
D TYPICAL MICROBIOLOGICAL INCIDENT REPORT	34
E CONTAINMENT LEVELS	35
F CONTAINMENT LEVEL FOR MANIPULATED DNA	39
G BIOLOGICAL HAZARD WARNING SYMBOL	40
H REFERENCES AND RELATED DOCUMENTS	41

FOREWORD

Safety in all laboratories is an individual, as well as a management responsibility. It is management's responsibility to provide protective equipment, a policy relating to safe work practices within a laboratory and to promote the institution of those practices. It is the laboratory staff's responsibility to carry out the safe work practices and to use protective equipment to minimize injury, not only to themselves, but also to their colleagues. Staff training must be directed toward making safety an attitude of mind and an integral part of all laboratory procedures, so that a constant, purposeful control of the laboratory environment will result. Accidents such as spillages are an obvious hazard, but the production of aerosols is a less obvious hazard that can be a serious source of contamination. In addition to the many problems commonly encountered in chemical laboratories, microbiological laboratories pose the following specific problems:

- (a) Infection of laboratory staff, and of the general public or of animals by dissemination of pathogens outside the laboratory.
- (b) Cross-contamination.
- (c) Contamination with adventitious microorganisms.

The safest procedure is to regard all microorganisms as potential pathogens and to treat them accordingly.

STANDARDS AUSTRALIA

Australian Standard
Safety in laboratories

Part 3: Microbiology

SECTION 1 SCOPE AND GENERAL

1.1 SCOPE This Standard sets out requirements, responsibilities and general guidelines relating to safety in laboratories where microorganisms are handled. It is intended for laboratories, including animal houses, where microbiological work such as research, teaching, diagnosis, quality control and regulatory analysis, e.g. of foodstuffs, water and effluents, pharmaceuticals and cosmetics, is undertaken. This Standard should be read in conjunction with AS 2243.1.

1.2 REFERENCED DOCUMENTS The following documents are referred to in this Standard:

AS

1132 Methods of test for air filters for use in air conditioning and general ventilation
1132.5 Method 5: Determination of arrestance efficiency, average arrestance efficiency, dust holding capacity, and dust holding capacity per unit of effective face area for test dusts Nos 1, 2 and 3

1319 Safety signs for the occupational environment

1324 Air filters for use in air conditioning and general ventilation

1336 Recommended practices for eye protection in the industrial environment

1337 Eye protectors for industrial applications

1410 Sterilizers—Steam—Pre-vacuum

1807 Classrooms, workstations and safety cabinets—Methods of test

1807.6 Method 6: Determination of integrity of terminally mounted HEPA filter installations

1885 Code of practice for recording and measuring work injury experience

2243 Safety in laboratories

2243.1 Part 1: General

2243.2 Part 2: Chemical aspects

2243.8 Part 8: Fume cupboards

2243.9 Part 9: Recirculating fume cabinets

2252 Biological safety cabinets

2252.1 Part 1: Biological safety cabinets (Class I) for personnel protection

2252.2 Part 2: Laminar flow biological safety cabinets (Class II) for personnel and product protection

2508 Safe storage and handling information cards for hazardous materials

2567 Cytotoxic drug safety cabinets

2639 Cytotoxic drug cabinets—Installation and use

2647 Biological safety cabinets—Installation and use

2982 Laboratory construction

Worksafe Australia

(NOHSC) exposure standards for atmospheric contaminants in the occupational environment (held in SSC)

ISO

3864 Safety colours and safety signs

ANSI

Z35.5 Biological hazard symbol

BS

502 Specification for safety requirements for laboratory centrifuges

1.3 DEFINITIONS For the purpose of this Standard, the definitions below apply.

1.3.1 Aerosol—suspension in air of finely dispersed solids or liquids.

1.3.2 Antiseptic—substance capable of destroying or preventing growth of microorganisms under prescribed conditions of use, and specifically for application to living tissues.

1.3.3 Aseptic technique—the exercise of special procedures for maintaining—

(a) the sterility of equipment, media, and other materials;

(b) the purity of cultures, by eliminating adventitious contamination; and

(c) protection for the operator and environment.