

Australian Standard[®]

Safety in laboratories

Part 6: Mechanical aspects

This Australian Standard was prepared by Committee CH/26, Safety in laboratories. It was approved on behalf of the Council of Standards Australia on 4 January 1990 and published on 7 May 1990.

The following interests are represented on Committee CH/26:

Australian Chemical Industry Council
Australian Government Analytical Laboratories
Australian Institute of Petroleum
National Association of Testing Authority, Australia

Additional interests participating in preparation of Standard:

Australian Nuclear Science and Technology Organization
Commonwealth Scientific and Industrial Research Organization
Department of Agriculture and Rural Affairs, Victoria
Department of Defence
Department of Science
University of Melbourne

Review of Australian Standards. To keep abreast of progress in industry, Australian Standards are subject to periodic review, and are kept up to date by the issue of amendments or new editions as necessary. It is important therefore that Standards users ensure that they are in possession of the latest edition, and any amendments thereto.

Full details of all Australian Standards and related publications will be found in the Standards Australia Catalogue of Publications; this information is supplemented each month by the magazine 'The Australian Standard' which subscribing members receive, and which gives details of new publications, new editions and amendments, and of withdrawn Standards.

Suggestions for improvements to Australian Standards, addressed to the head office of Standards Australia, are welcomed. Notification of any inaccuracy or ambiguity found in an Australian Standard should be made without delay in order that the matter may be investigated and appropriate action taken.

AS 2243.6—1990

Australian Standard[®]

Safety in laboratories

Part 6: Mechanical aspects

First published as AS 2243.6—1980.
Second edition—1990

PUBLISHED BY STANDARDS AUSTRALIA
(STANDARDS ASSOCIATION OF AUSTRALIA)
1 THE CRESCENT, HOMEBUSH, NSW 2140

ISBN 0 7262 6150 5

PREFACE

This Standard was prepared by the Standards Australia Committee on Safety in Laboratories under the direction of the Chemical Standards Board to supersede AS 2243.6—1980.

The revision was undertaken in keeping with the Standards Australia policy of revising Standards at least every five years. This edition has been expanded to include detailed information on the safeguarding of machinery and on the dangers from machinery.

Particular sections have also been expanded to include information on ventilation, principles of control of hazardous substances and operations, inventory of chemical substances, plasma cutting, cutting and welding of plastics materials, radiofrequency dielectric plastic welders, solvent welding of plastics, cryogenic liquids and hearing protection. In the preparation of this Standard cognizance was taken of the following documents:

- (a) Victorian Universities and Victoria Institute of Colleges, 'Code of Safety in Workshops'.
- (b) Commonwealth Scientific and Industrial Research Organization, 'Safety Handbook'
- (c) The University of Leeds, 'Safety Handbook'.
- (d) The University of New England, 'Safety Manual'.
- (e) The University of Warwick, 'Safety in Laboratories'.
- (f) Committee on Occupational Safety and Health in Commonwealth Government Employment, 'Draft Code of Practice: Laboratories (Physical and Metallurgical)'.
- (g) Handbook of Compressed Gases, Compressed Gas Association Inc, Reinforced, 2nd Edition, 1981.
- (h) AS 1219, Power presses—Safety requirements.
- (i) AS 1470, Health and safety at work—Principles and practices
- (j) AS 1473, Code of practice for the guarding and safe use of woodworking machinery
- (k) AS 1485, Safety and health in workrooms of educational establishments
- (l) AS 1893, Code of practice for the guarding and safe use of metal and paper cutting guillotines

This Standard is intended for use in conjunction with other Standards of the AS 2243 series and is applicable to all laboratory situations. Other parts of AS 2243 are as follows:

Part 1: General

Part 2: Chemical aspects

Part 3: Microbiology

Part 4: Ionizing radiations

Part 5: Non-ionizing radiations

Part 7: Electrical aspects

Part 8: Fume cupboards.

A further part of AS 2243 dealing with recirculating fume cabinets is in the course of preparation.

© Copyright — STANDARDS AUSTRALIA

Users of Standards are reminded that copyright subsists in all Standards Australia publications and software. Except where the Copyright Act allows and except where provided for below no publications or software produced by Standards Australia may be reproduced, stored in a retrieval system in any form or transmitted by any means without prior permission in writing from Standards Australia. Permission may be conditional on an appropriate royalty payment. Requests for permission and information on commercial software royalties should be directed to the head office of Standards Australia.

Standards Australia will permit up to 10 percent of the technical content pages of a Standard to be copied for use exclusively in-house by purchasers of the Standard without payment of a royalty or advice to Standards Australia.

Standards Australia will also permit the inclusion of its copyright material in computer software programs for no royalty payment provided such programs are used exclusively in-house by the creators of the programs.

Care should be taken to ensure that material used is from the current edition of the Standard and that it is updated whenever the Standard is amended or revised. The number and date of the Standard should therefore be clearly identified.

The use of material in print form or in computer software programs to be used commercially, with or without payment, or in commercial contracts is subject to the payment of a royalty. This policy may be varied by Standards Australia at any time.

CONTENTS

	<i>Page</i>
1. SCOPE AND GENERAL	
1.1 SCOPE	4
1.2 REFERENCED DOCUMENTS	4
1.3 MINIMIZATION OF COMMON SAFETY HAZARDS	4
1.4 PLANNING FOR SAFETY	5
1.5 FIRE, EMERGENCY AND RESCUE PROCEDURES	5
SECTION 2. EQUIPMENT IN GENERAL USE	
2.1 MACHINERY	6
2.2 FIXED POWER TOOLS	6
2.3 PORTABLE POWER TOOLS	9
2.4 HAND TOOLS	10
2.5 CENTRIFUGES	11
2.6 HEATING EQUIPMENT	11
2.7 ELECTRONIC APPARATUS	11
2.8 HIGH PRESSURE EQUIPMENT	12
2.9 VACUUM APPARATUS	12
2.10 TUBING AND PIPING	13
2.11 MISCELLANEOUS EQUIPMENT AND PROCESSES	13
SECTION 3. COMPRESSED GAS CYLINDERS	
3.1 IDENTIFICATION OF GAS CYLINDERS	15
3.2 HANDLING, USE AND STORAGE	15
3.3 PRECAUTIONS FOR SPECIFIC GASES	16
SECTION 4. HAZARDOUS SUBSTANCES AND PROCESSES	
4.1 GENERAL REQUIREMENTS	18
4.2 SOLDERING AND BRAZING	18
4.3 WELDING AND CUTTING	18
4.4 DEGREASING	19
4.5 CUTTING AND GRINDING OILS	19
4.6 GLASS WORKING	19
4.7 CRYOGENIC SUBSTANCES	19
APPENDICES	
A LIST OF REFERENCED DOCUMENTS	21
B GUIDE TO THE CHEMICAL RESISTANCE OF GLOVE MATERIALS	23

STANDARDS AUSTRALIA

**Australian Standard
Safety in laboratories****Part 6: Mechanical aspects**

1 SCOPE AND GENERAL

1.1 SCOPE. This Standard sets out general principles and requirements of safe working practice relevant to mechanical operations performed in laboratories and describes the nature of the potential mechanical hazards. This Standard also covers safe working practices relevant to the use of compressed gas cylinders and cryogenic substances.

1.2 REFERENCED DOCUMENTS. A list of documents referenced in this Standard is given in Appendix A.

1.3 MINIMIZATION OF COMMON SAFETY HAZARDS.

1.3.1 General. Mechanical safety hazards are present in the use of pressure and vacuum apparatus, heating and cooling apparatus, hand and power tools, in degreasing operations, and many other aspects of daily laboratory work. Particular danger is present when gaining access to moving or otherwise hazardous parts (which are safe-guarded in normal industrial practice) of machinery to perform measurements or adjustments. Potential safety hazards associated with mechanical apparatus should be identified by displaying the appropriate warning symbols in accordance with AS 1216.1 and AS 1319.

1.3.2 Workshop training. Many laboratories have small mechanical workshops for use by people not necessarily skilled in the use of machine tools. It is essential that such people be made aware of the operating characteristics, basic safe working practices and emergency procedures relevant to the tools involved, and in particular the need for wearing appropriate protective equipment.

1.3.3 Use of machines. Mechanical apparatus shall be used only within the range of operating conditions for which it was designed. Machines with rotating components are particularly susceptible to stress because, in such apparatus, the centrifugal stresses increase as the the speed of the rotor increases.

Movement of machinery parts consists basically of rotary, sliding or reciprocating motion which, individually or in combination, can produce cutting, shearing and crushing injuries. Rotating parts are also capable of inflicting injury by personal entanglement. Danger from machinery can be minimized by the operator wearing suitable clothing, and by fitting suitable guards, physical barriers and electrical interlocks to the machine to protect both the operator and passing traffic.

An obvious function of a machine guard is to keep the operator's body, fingers, clothing and arms away from the danger points, without impeding the intended operation of the machine, or obstructing vision. This function can usually be met by intelligent design.

Another function, which may be less obvious, is to keep hazardous objects from striking the operator. A grinding wheel guard is an example of this; a suitable guard should be of an appropriate shape and of adequate strength to contain any potential hazard.

A guard can serve a further function in preventing the fitting of an attachment likely to fail. In the case of the grinder this could be an over-size wheel, which would be more prone to disintegration. This aspect of guard functioning also applies to interlocks, where the machine cannot be started or operated unless the guard is correctly in position.

1.3.4 Temporary apparatus. Where the construction of temporary apparatus, such as supports for small loads, or small vessels for pressure or vacuum, is necessary, it is essential that apparatus design and manufacture be carried out by competent and experienced personnel.

1.3.5 Personal care. Persons wearing loose-fitting clothing, ties, jewellery including rings, and long hair (including beards) not confined close to the head by net or otherwise, shall not work with moving machinery. Sleeves of overalls or working-shirts should either be rolled up above the elbow, or be buttoned tightly at the wrist.

1.3.6 Use of personal protective equipment. Appropriate personal protective equipment for eyes, ears, face, hands, feet and respiratory system should always be available to laboratory personnel. The equipment should be comfortable to wear, and should be selected, used and maintained in accordance with SAA HB9.