

2, Part 1 - 1981

DUP
Under Revision PR 92158

see also Seminar Papers

Superseded by AS 2252.1-1994

AS 2252, Part 1—1981
UDC 614.44:615.97:542.2

↑

615.917
=

Australian Standard 2252, Part 1—1981

BIOLOGICAL SAFETY CABINETS Part 1—BIOLOGICAL SAFETY CABINETS (CLASS I) FOR PERSONNEL PROTECTION



STANDARDS ASSOCIATION OF AUSTRALIA
Incorporated by Royal Charter



THE FOLLOWING SCIENTIFIC, INDUSTRIAL AND GOVERNMENTAL ORGANIZATIONS and departments were officially represented on the committee entrusted with the preparation of this standard:

Australian Institute of Refrigeration, Air Conditioning and Heating
Australian Medical Association
Commonwealth Serum Laboratories
Confederation of Australian Industry
CSIRO, Division of Animal Health
Department of Housing and Construction
Department of Defence
Department of Public Works, N.S.W.
Firms and consultants specializing in equipment and design for controlled environments
Health Commission of New South Wales
Health Commission of Victoria
National Association of Testing Authorities, Australia
National Biological Standards Laboratory
National Council of Chemical and Pharmaceutical Industries
Royal Australian Institute of Architects

This standard, prepared by Committee MS/28, Controlled Environment, was approved on behalf of the Council of the Standards Association of Australia on 17 March 1981, and was published on 1 June 1981.

To keep abreast of progress in industry, Australian standards are subject to continuous review and are kept up-to-date by the issue of amendments or new editions as necessary. It is important therefore that standards users ensure that their standards are up-to-date. Full details of all SAA publications will be found in the Annual List of Australian Standards; these details are supplemented by listings in the SAA monthly journal 'The Australian Standard'. Information on the Annual List and 'The Australian Standard' may be obtained from any sales office of the Association, where details are also available of the current status of individual standards. Suggestions for improvements to published standards, addressed to the head office of the Association, are welcomed.

AUSTRALIAN STANDARD

BIOLOGICAL SAFETY CABINETS

Part 1

**BIOLOGICAL SAFETY
CABINETS (CLASS I)
FOR PERSONNEL
PROTECTION**

AS 2252, Part 1—1981

First published	1979
Second edition	1981

**PUBLISHED BY THE STANDARDS ASSOCIATION OF AUSTRALIA
STANDARDS HOUSE, 80 ARTHUR ST, NORTH SYDNEY, N.S.W.**

ISBN 0 7262 2211 9



2 JUN 1981

PREFACE

This edition of this standard was prepared by the Association's Committee on Controlled Environment. It is Part 1 of a 3-part standard for biological safety cabinets, and is a revision of the 1979 edition.

The other Parts of the standard are—

- Part 2 Laminar Flow Biological Safety Cabinets (Class II) for Personnel and Product Protection
- Part 3 Gas-tight Biological Safety Cabinets (Class III) for Personnel Protection Against Agents of Extreme Hazard

The issue in 1972 of a public review draft DR 72168 for cleanrooms and work-stations drew a comment from the Commonwealth Serum Laboratories which pointed to the need for work to be done under aseptic conditions and the need to contain potentially dangerous materials, and to the fact that the proposed standards did not adequately cater for these situations.

A need has also been established for standards applying to cabinets designed to minimize the inherent risks to personnel working with hazardous biological agents. There is an increasing awareness of the hazards of exposure to infectious agents that present an actual or potential risk to people handling them.

This edition aligns this Part with Part 2 in respect of style and format, and introduces an objective, quantitative method of test for indicating containment at the work access opening in place of the subjective, qualitative method for directional airflow.

In the preparation of this standard, reference was made to the booklet published by the United States Department of Health, Education and Welfare, Public Health Service entitled 'Classification of Etiological Agents on the Basis of Hazard'. Although written for American conditions, this booklet describes the basis for such classification of agents, and lists bacterial, fungal, parasitic, viral, rickettsial and chlamydial agents in five classes. The least hazardous agents are in Class 1 and those requiring the greatest restrictions are in Class 4. Class 5 contains agents which are specifically excluded from the U.S. by law. The list of organisms prohibited in Australia differs somewhat from the American list. Details can be obtained from the Quarantine Division of the Department of Health, Canberra.

Appendix A, Purchasing Guidelines, provides a basis for contractual matters.

This standard may require reference to the following standards:

- AS 1132 Methods of Test for Air Filters for Use in Air Conditioning and General Ventilation
 .5—Determination of Arrestance Efficiency, Average Arrestance Efficiency, Dust-holding Capacity and Dust-holding Capacity per Unit of Effective Face Area for Test Dusts Nos 1, 2 and 3
- AS 1216 Classification, Hazard Identification and Information Systems for Dangerous Goods
 Part 1—Classification and Class Labels for Dangerous Goods
- AS 1217 Methods of Measurement of Airborne Sound Emitted by Machines
- AS 1224 Air Filters for Use in Air Conditioning and General Ventilation
- AS 1677 SAA Refrigeration Code
- AS 1680 Code of Practice for Interior Lighting and the Visual Environment
- AS 1807 Methods of Test for Cleanrooms, Work-stations and Their Accessories
 .1 —Air Velocity and Uniformity of Clean Work-stations
 .2 —Air Velocity Under Loaded Filter Conditions of Clean Work-stations
 .5 —Induced Air Leakage
 .6 —Final Filter Installation Integrity
 .15—Light Intensity
 .18—Vibration in Work-stations
- AS C100 Approval and Test Specification for Definitions and General Requirements for Electrical Materials and Equipment
- AS Guide for Installation and Use of Biological Safety Cabinets*.

*In course of preparation.

CONTENTS

	<i>Page</i>
FOREWORD	4
SPECIFICATION	
1 Scope	6
2 General Requirements	6
3 Construction Requirements	6
4 Performance Requirements	9
5 Marking	9
APPENDICES	
A Purchasing Guidelines	10
B Procedures for Location, Inspection and Testing	11
C Method for Determining Gastightness of Outer Shell	12
D Method for Determining Inward Air Velocity	13
E Method for Determining Intensity of Radiation from Germicidal Ultraviolet Lamps	13

© Copyright — STANDARDS ASSOCIATION OF AUSTRALIA 1981

Users of standards are reminded that copyright subsists in all SAA publications. No part of this publication may be reproduced, stored in a retrieval system in any form or transmitted by any means without prior permission in writing of the Standards Association of Australia.

STANDARDS ASSOCIATION OF AUSTRALIA

Australian Standard
for
BIOLOGICAL SAFETY CABINETS

PART 1—BIOLOGICAL SAFETY CABINETS (CLASS I) FOR PERSONNEL PROTECTION

FOREWORD

In order to adequately guard against inherent risks to personnel, products and experiments associated with hazardous biological agents, consideration is necessary in respect of the type of cabinet containment required and of the competence of operators. This, in turn, calls for definition of the degree of hazard, and the following terms, which relate directly to those used by the United States Public Health Service (PHS), have been adopted:

- *No or minimal hazard*—the risk level of agents and/or operations of no or minimal danger to personnel, animals, or plants under ordinary conditions of handling (PHS Class 1).
 - *Ordinary or potential hazard*—the risk level associated with agents which produce disease in man, animals, or plants and which can be contained by normal microbiological techniques. The level of competence for personnel should be that expected of staff in a hospital diagnostic microbiological laboratory (PHS Class 2).
 - *Special hazard*—the risk level associated with agents which are highly infectious or toxic for man, animals, or plants with the production of dangerous disease. Also included are agents with genetic alterations and those which may have a synergistic effect with other materials. Appropriate containment measures are required. The level of competence for personnel should be at least that expected in staff of a hospital diagnostic microbiological laboratory. In addition personnel must have had proper training in the handling of dangerous agents (PHS Class 3).
 - *Extreme hazard*—the risk level associated with agents which are extremely dangerous for man, animals, or plants, or cause serious epidemic disease. They may have various dangerous combinations of the following characteristics:
 - (a) Low infective doses.
 - (b) High pathogenicity.
 - (c) Potential to spread outside the laboratory.
 - (d) Concentration.
 - (e) Genetic alteration or genetic recombination that significantly increases pathogenicity.
- Stringent containment measures are required. Personnel must have a high level of competence in microbiology and must have had special training in the handling of dangerous agents (PHS Class 4).

To cope with risks posed by hazardous biological agents it has been necessary to design cabinets which fall into three separate classes. Each class has a specific application according to the hazard and as to whether protection is required for product and experiment in addition to personnel.

For recombinant DNA containment, the use of a biological safety cabinet (Class II) is appropriate. The Australian Academy of Science has established a Committee on Recombinant DNA Molecules and application for advice on containment conditions for specific experimental proposals should be submitted to the Executive Secretary, Australian Academy of Science, Canberra.