

Australian Standard[®]

**Measurement of fluid flow in closed
conduits**

**Part 6.1. Volumetric methods—
By mass**

[ISO title: Measurement of liquid flow in closed conduits—Weighing
method]

This Australian Standard was prepared by Committee CE/24, Measurement of Water Flow in Open Channels and Closed Conduits. It was approved on behalf of the Council of Standards Australia on 3 August 1993 and published on 20 December 1993.

The following interests are represented on Committee CE/24:

Association of Consulting Engineers of Australia
Department of Water Resources, New South Wales
Engineering and Water Supply Department, South Australia
Forestry Commission of New South Wales
Institute of Instrumentation and Control, Australia
Melbourne Water
Monash University
Snowy Mountains Engineering Corporation
University of New South Wales
University of Queensland
Water Authority of Western Australia
Water Board, Sydney— Illawarra— Blue Mountains
Water Resources Commission, Queensland

Review of Australian Standards. To keep abreast of progress in industry, Australian Standards are subject to periodic review and are kept up to date by the issue of amendments or new editions as necessary. It is important therefore that Standards users ensure that they are in possession of the latest edition, and any amendments thereto.

Full details of all Australian Standards and related publications will be found in the Standards Australia Catalogue of Publications; this information is supplemented each month by the magazine 'The Australian Standard', which subscribing members receive, and which gives details of new publications, new editions and amendments, and of withdrawn Standards.

Suggestions for improvements to Australian Standards, addressed to the head office of Standards Australia, are welcomed. Notification of any inaccuracy or ambiguity found in an Australian Standard should be made without delay in order that the matter may be investigated and appropriate action taken.

Australian Standard®

**Measurement of fluid flow in closed
conduits**

**Part 6.1: Volumetric methods—
By mass**

First published as AS 2360.6.1—1993.

PUBLISHED BY STANDARDS AUSTRALIA
(STANDARDS ASSOCIATION OF AUSTRALIA)
1 THE CRESCENT, HOMEBUSH, NSW 2140

ISBN 0 7262 8497 1

PREFACE

This Standard was prepared by the Standards Australia Committee on Measurement of Water Flow in Open Channels and Closed Conduits. It is identical with and has been reproduced from ISO 4185:1980, *Measurement of liquid flow in closed conduits—Weighing method*.

Under arrangements made between Standards Australia and the international Standards bodies, ISO and IEC, as well as certain other Standards organizations, users of this Australian Standard are advised of the following:

- (a) Copyright is vested in Standards Australia.
- (b) The number of this Standard is not reproduced on each page; its identity is shown only on the cover and title pages.
- (c) There may be occasional dual language sections, but English is always one of the languages reproduced.
- (d) Where any cross-references to page numbers appear within the text, these relate to page numbering in the international Standard and are to be disregarded.

This Standard is one of a series, to be published progressively, which deals with methods of measurement of fluid flow in closed conduits. The following Parts were published concurrently with this Part:

AS

- | | |
|----------|--|
| 2360 | Measurement of fluid flow in closed circuits |
| 2360.0 | Part 0: Vocabulary and symbols |
| 2360.1.1 | Part 1.1: Pressure differential methods—Measurement using orifice plates, nozzles or Venturi tubes—Conduits with diameters from 50 mm to 1200 mm |
| 2360.1.2 | Part 1.2: Pressure differential method—Measurement using orifice plates or nozzles—Conduits with diameters less than 50 mm |
| 2360.1.3 | Part 1.3: Pressure differential methods—Measurement using orifice plates, nozzles or Venturi tubes—Guide to the use of methods specified in Parts 1.1 and 1.2 |
| 2360.1.4 | Part 1.4: Pressure differential methods—Measurement using orifice plates, nozzles or Venturi tubes—Guide to the effect of departure from the conditions specified in Part 1.1 |
| 2360.1.5 | Part 1.5: Pressure differential methods—Measurement using orifice plates, nozzles or Venturi tubes—Pulsating flow, in particular sinusoidal or square wave intermittent periodic-time fluctuations |
| 2360.6.1 | Part 6.1: Volumetric methods—By mass (<i>this Standard</i>) |
| 2360.6.2 | Part 6.2: Volumetric methods—By volume |
| 2360.7.1 | Part 7.1: Assessment of uncertainty in the calibration and use of flow measurement devices—Linear calibration relationships |
| 2360.7.2 | Part 7.2: Assessment of uncertainty in the calibration and use of flow measurement devices—Non-linear calibration relationships |

At the date of publication of this Part the following Parts, with the numbers of the parent international Standards in parenthesis, had not been published:

Pressure differential methods—Measurement using orifice plates, nozzles or Venturi tubes—Connections for pressure signal transmissions between primary and secondary elements (ISO 2186)

Pitot static tube methods—Measurement of velocity at a point of the cross-section of a conduit (ISO 7145)

Pitot static tube methods—Measurement using Pitot-static tubes (ISO 3966)

Pitot static tube methods—Measurement in swirling or asymmetric flow conditions using ISO 3966 or ISO 3354 (ISO 7194)

Current meters method—Measurement of clean water in full conduits and under regular flow conditions using current meters (ISO 3354)

Non-radioactive tracer methods—Review of alternative methods (ISO 2975.1)

Non-radioactive tracer methods—Measurement using constant rate injection (ISO 2975.2)

Non-radioactive tracer methods—Measurement using transit time (ISO 2975.6)

Weighing methods—Verification of static type (ISO 9368.1)

Weighing methods—Verification of dynamic type (ISO 9368.2, not published)

When published, the details for the above unpublished Australian Standards will be listed in the Catalogue of Australian Standards and Other Products.

For the purposes of this Australian Standard, the ISO text should be modified as follows:

- (i) Wherever the words 'International Standard' appear, referring to this Standard, they should be read as 'Australian Standard'.
- (ii) Substitute a full stop (.) for a comma (,) as a decimal marker.
- (iii) The references to other publications should be replaced by references to Australian Standards as follows:

<i>Reference to International Standard</i>	<i>Australian Standard</i>
ISO	AS
4006 Measurement of fluid flow in closed conduits—Vocabulary and symbols	2360 Measurement of fluid flow in closed conduits
	2360.0 Part 0: Vocabulary and symbols
5168 Measurement of fluid flow—Estimation of uncertainty of a flow-rate measurement	3778 Measurement of water flow in open channels
	3778.2.4 Part 2.4: General—Estimation of uncertainty of a flow-rate measurement

© Copyright — STANDARDS AUSTRALIA

Users of Standards are reminded that copyright subsists in all Standards Australia publications and software. Except where the Copyright Act allows and except where provided for below no publications or software produced by Standards Australia may be reproduced, stored in a retrieval system in any form or transmitted by any means without prior permission in writing from Standards Australia. Permission may be conditional on an appropriate royalty payment. Requests for permission and information on commercial software royalties should be directed to the head office of Standards Australia.

Standards Australia will permit up to 10 percent of the technical content pages of a Standard to be copied for use exclusively in-house by purchasers of the Standard without payment of a royalty or advice to Standards Australia.

Standards Australia will also permit the inclusion of its copyright material in computer software programs for no royalty payment provided such programs are used exclusively in-house by the creators of the programs.

Care should be taken to ensure that material used is from the current edition of the Standard and that it is updated whenever the Standard is amended or revised. The number and date of the Standard should therefore be clearly identified.

The use of material in print form or in computer software programs to be used commercially, with or without payment, or in commercial contracts is subject to the payment of a royalty. This policy may be varied by Standards Australia at any time.

CONTENTS

	<i>Page</i>
1 General	1
1.1 Scope and field of application	1
1.2 References	1
1.3 Definitions	1
1.4 Units	1
1.5 Notation	2
1.6 Certification	2
2 Principle	2
2.1 Statement of the principle	2
2.2 Accuracy of the method	7
3 Apparatus	7
3.1 Diverter	7
3.2 Time-measuring apparatus	8
3.3 Weighing tank	9
3.4 Weighing machine	9
3.5 Auxiliary measurements	9
4 Procedure	10
4.1 Static weighing method	10
4.2 Dynamic weighing method	10
4.3 Common provisions	10
5 Calculation of flow-rate	10
5.1 Calculation of mass flow-rate	10
5.2 Calculation of volume flow-rate	10
6 Calculation of the overall uncertainty of the measurement of the flow-rate	10
6.1 Presentation of results	10
6.2 Sources of error	11
6.3 Calculation of uncertainty in flow-rate measurement	14
Annexes	
A Corrections on the measurement of filling time	16
B Density of pure water	18
C Definition of terms and procedures used in error analysis	19
Student's <i>t</i> -distribution	21

Measurement of fluid flow in closed conduits

Part 6.1: Volumetric methods—By mass

1 General

1.1 Scope and field of application

This International Standard specifies a method of liquid flow-rate measurement in closed conduits by measuring the mass of liquid delivered into a weighing tank in a known time interval. It deals in particular with the measuring apparatus, the procedure, the method for calculating the flow-rate and the uncertainties associated with the measurement.

The method described may be applied to any liquid provided that its vapour pressure is such that any escape of liquid from the weighing tank by vaporization is not sufficient to affect the required measurement accuracy. Closed weighing tanks and their application to the flow measurement of liquids of high vapour pressure are not considered in this International Standard.

This International Standard does not cover the cases of corrosive or toxic liquids.

Theoretically, there is no limit to the application of this method which is used generally in fixed laboratory installations only. However, for economic reasons, usual hydraulic laboratories using this method can produce flow-rates of 1.5 m³/s or less.

Owing to its high potential accuracy, this method is often used as a primary method for calibration of other methods or devices for mass flow-rate measurement or volume flow-rate measurement provided that the density of the liquid is known accurately.

It must be ensured that the pipeline is running full with no air or vapour pockets present in the measuring section.

1.2 References

ISO 1006, *Measurement of fluid flow in closed conduits — Vocabulary and symbols.*

ISO 5168, *Measurement of fluid flow — Estimation of uncertainty of a flow-rate measurement.*

OIML, Recommendations Nos. 1, 2, 3, 20, 28, 33.

1.3 Definitions

Only terms which are used in a special sense or the meaning of which merits restatement are defined below.

1.3.1 static weighing : The method in which the net mass of liquid collected is deduced from tare and gross weighings made respectively before and after the liquid has been delivered for a measured time interval into the weighing tank.

1.3.2 dynamic weighing : The method in which the net mass of liquid collected is deduced from weighings made while fluid flow is being delivered into the weighing tank. (A diverter is not required with this method.)

1.3.3 diverter : A device which diverts the flow either to the weighing tank or to its by-pass without changing the flow-rate during the measurement interval.

1.3.4 flow stabilizer : A structure forming part of the measuring system, ensuring a stable flow-rate in the conduit being supplied with liquid; for example, a constant level head tank, the level of liquid in which is controlled by a weir of sufficient length.

1.3.5 buoyancy correction : The correction to be made to the readings of a weighing machine to take account of the difference between the upward thrust exerted by the atmosphere, on the liquid being weighed and on the reference weights used during the calibration of the weighing machine.

1.4 Units

The units used in this International Standard are the SI units, metre, kilogram, and second; the degree Celsius is used for convenience instead of the kelvin.