

Australian Standard[®]

**Single-use (sterile) infusion sets
for general medical use**

This Australian Standard was prepared by Committee HT/6, Transfusion Equipment for Medical Use. It was approved on behalf of the Council of Standards Australia on 6 February 1990 and published on 15 October 1990.

The following interests are represented on Committee HT/6:

Australian Chamber of Commerce
Australian Medical Devices and Diagnostics Association
Australian Red Cross Society
Australian Society of Anaesthetists
Department of Administrative Services, N.S.W.
Department of Community Services and Health
Department of Health, N.S.W.
Department of Health, Qld.
Health Department, W.A.
Royal Australasian College of Surgeons
Royal College of Pathologists of Australasia
South Australian Health Commission

Review of Australian Standards. To keep abreast of progress in industry, Australian Standards are subject to periodic review and are kept up to date by the issue of amendments or new editions as necessary. It is important therefore that Standards users ensure that they are in possession of the latest edition, and any amendments thereto.

Full details of all Australian Standards and related publications will be found in the Standards Australia Catalogue of Publications; this information is supplemented each month by the magazine 'The Australian Standard' which subscribing members receive, and which gives details of new publications, new editions and amendments, and of withdrawn Standards.

Suggestions for improvements to Australian Standards, addressed to the head office of Standards Australia, are welcomed. Notification of any inaccuracy or ambiguity found in an Australian Standard should be made without delay in order that the matter may be investigated and appropriate action taken.

Australian Standard[®]

**Single-use (sterile) infusion sets
for general medical use**

First published as AS 2385—1980.
Second edition 1990.
Incorporating:
Amdt 1—1993

PREFACE

This Standard was prepared by the Standards Australia Committee on Transfusion Equipment for Medical Use, under the direction of the Health Technology Standards Board, to supersede AS 2385—1980.

Unlike the previous edition, tests for reducing (oxidizable) matter, metal ions, non-volatile residue, absorbance, haemolytic effects, intracutaneous reactivity and alkylene oxide gas residues are included. The test for toxicity has been replaced by a test for cytotoxicity, the pyrogen test has been altered to allow the use of the *Limulus amoebocyte lysate* (bacterial endotoxins) test, and the requirements for the closure piercing device have been altered. This edition also offers a choice of drip rates.

© Copyright — STANDARDS AUSTRALIA

Users of Standards are reminded that copyright subsists in all Standards Australia publications and software. Except where the Copyright Act allows and except where provided for below no publications or software produced by Standards Australia may be reproduced, stored in a retrieval system in any form or transmitted by any means without prior permission in writing from Standards Australia. Permission may be conditional on an appropriate royalty payment. Requests for permission and information on commercial software royalties should be directed to the head office of Standards Australia.

Standards Australia will permit up to 10 percent of the technical content pages of a Standard to be copied for use exclusively in-house by purchasers of the Standard without payment of a royalty or advice to Standards Australia.

Standards Australia will also permit the inclusion of its copyright material in computer software programs for no royalty payment provided such programs are used exclusively in-house by the creators of the programs.

Care should be taken to ensure that material used is from the current edition of the Standard and that it is updated whenever the Standard is amended or revised. The number and date of the Standard should therefore be clearly identified.

The use of material in print form or in computer software programs to be used commercially, with or without payment, or in commercial contracts is subject to the payment of a royalty. This policy may be varied by Standards Australia at any time.

CONTENTS

	<i>Page</i>
FOREWORD	4
1 SCOPE	5
2 REFERENCED DOCUMENTS	5
3 DEFINITIONS	5
4 MATERIALS	6
5 SPECIFIC REQUIREMENTS FOR BLOOD SET	6
6 SPECIFIC REQUIREMENTS FOR SOLUTION SET	8
7 DELIVERY RATE	10
8 FREEDOM FROM LEAKAGE OF AIR	10
9 CYTOTOXICITY	10
10 PYROGENICITY	10
11 STERILITY	10
12 HAEMOLYTIC EFFECTS	10
13 INTRACUTANEOUS REACTIVITY	10
14 PARTICULATE CONTAMINATION	10
15 REDUCING (OXIDIZABLE) MATTER	10
16 EXTRACTABLE TRACE METALS	10
17 NON-VOLATILE RESIDUE	10
18 ABSORBANCE	10
19 ALKYLENE OXIDE GAS RESIDUES	10
20 END PROTECTORS	11
21 MARKING	11
APPENDICES	
A METHOD OF TEST FOR RESISTANCE TO CORROSION OF THE CLOSURE-PIERCING DEVICE	12
B METHOD FOR TESTING CLOSURE PIERCING	13
C METHOD OF TEST FOR LEAKAGE OF AIR UNDER INCREASED PRESSURE ..	14
D METHOD OF TEST FOR LEAKAGE OF AIR UNDER REDUCED PRESSURE ...	15
E METHOD FOR TESTING DELIVERY RATE	17
F METHOD OF TEST FOR CYTOTOXICITY	18
G METHOD OF TEST FOR PYROGENS (BACTERIAL ENDOTOXINS)	22
H METHOD OF TEST FOR HAEMOLYTIC EFFECTS	24
I METHOD OF TEST FOR PARTICULATE CONTAMINATION	25
J METHOD OF TEST FOR REDUCING (OXIDIZABLE) MATTER	27
K METHOD OF TEST FOR EXTRACTABLE TRACE METALS	28
L METHOD OF TEST FOR NON-VOLATILE RESIDUE	29
M METHOD OF TEST FOR ABSORBANCE	30
N METHOD OF TEST FOR ALKYLENE (ETHYLENE OR PROPYLENE) OXIDE GAS RESIDUES	31

FOREWORD

Two types of infusion sets are included in this Standard, one suitable for the administration of blood, blood derivatives and intravenous solutions, and the other for the administration of intravenous solutions only. The set for the administration of blood and blood products requires a piercing device as specified in this Standard to fit the outlet port of a plastics blood container. A simpler and cheaper piercing device would suffice for the set for administering intravenous solutions. Further, the filter for the blood and the solution sets have the same performance characteristics, but the filter for the solution set may be of a much smaller area than that for the blood set.

Wherever possible, stress has been laid on performance requirements rather than dimensional requirements so as to allow as much variation in design as possible without inhibiting innovation. For example, the Standard allows for sets with separate drip chambers and filter chambers and for sets with combined drip and filter chambers.

In regard to the measurement of particulate contamination in giving sets, preliminary investigations have indicated that accuracy of estimation of the number of particles is dependent on a number of experimental variables. Consequently the particulate contamination test described in Appendix I should be considered only as an interim test pending further investigations.

STANDARDS AUSTRALIA

Australian Standard

Single-use (sterile) infusion sets for general medical use

1 SCOPE This Standard specifies requirements for single-use sterile intravenous infusion sets intended for general medical use. It covers sets suitable for blood, blood derivatives and other intravenous fluids in general, and also sets suitable only for use with fluids containing no solid phase. The sets are intended to be restricted to one patient and not to be re-used.

The Standard does not apply to certain types of sets for paediatric use or other sets such as microdrip sets, pump sets and sets fitted with a central venous pressure monitoring device. Requirements for devices for intravenous insertion—whether catheter, needle or cannula—are also not included.

NOTE: The infusion sets covered in this Standard may not be suitable for the administration of platelet concentrates.

2 REFERENCED DOCUMENTS The following documents are referred to in this Standard:

AS

- 1094 Single-use syringes (sterile) for general medical use
- 1444 Wrought alloy steels—Standard and Hardenability (H) Series
- 1600 Medical equipment—Conical fittings with a 6% (Luer) taper for syringes, needles and certain other medical equipment
 - 1600.1 Part 1: General requirements
 - 1946 Hypodermic equipment—Single-use needles (sterile) for general medical use
 - 2134 Recommended practice for chemical analysis by atomic absorption spectrometry
 - 2134.1 Part 1: Flame atomic absorption spectrometry
 - 2134.2 Part 2: Graphite furnace spectrometry
 - 2193 Methods for calibration and grading of force-measuring systems of testing machines
 - 2506 Wrought alloy steels—En series
- CK 19 Code of recommended practice for the chemical analysis of materials by ultraviolet visible spectrophotometry

BS

- 5736 Evaluation of medical devices for biological hazards
 - Part 4: Method of test for intracutaneous reactivity of extracts from medical devices
 - British Pharmacopoeia (BP), 1988, Vol 2, Appendix XIVK, pA 183
 - United States Pharmacopoeia (USP), XII Monograph 85
 - Proposed Standard for particulate matter in large volume injections, Therapeutic Goods Act 1972
 - Therapeutic Goods Order 11, Standard for sterile therapeutic goods
 - Guideline on validation of the *Limulus amoebocyte* test as an end-product endotoxin for human and parenteral drugs, biological products and medical devices (FDA), U.S.A.

3 DEFINITIONS For the purpose of this Standard, the definitions below apply.

- 3.1 Set**—a single-use sterile intravenous infusion set, including any air-inlet assembly and intravenous injection device provided with the set, but excluding the end protectors.
- 3.2 Blood set**—a set suitable for use with blood, blood derivatives and intravenous fluids in general.
- 3.3 Solution set**—a set suitable only for intravenous fluids containing no solid phase.
- 3.4 Closure-piercing device**—component for perforating the closure of a container of an intravenous fluid and connecting the set to the container.
- 3.5 Drip chamber**—chamber, other than a drip-filter chamber, containing a drip tube and constructed so as to enable the rate of flow of fluid through the set to be observed.
- 3.6 Filter chamber**—chamber, other than a drip-filter chamber, containing a blood or fluid filter.
- 3.7 Drip-filter chamber**—chamber containing a drip tube and a blood or fluid filter, and constructed so as to enable the rate of flow of fluid through the set to be observed.
- 3.8 Drip tube**—tube protruding into a drip chamber or drip-filter chamber at the inlet end.
- 3.9 Blood filter**—porous material placed across a chamber of a blood set to remove clots or aggregates from blood.
- 3.10 Fluid filter**—porous material placed across the liquid pathway of a solution set as a safeguard against the inadvertent use of blood with the set.