

Australian Standard[®]

**Refractories and refractory materials—
Chemical analysis**

**Part 6: Refractories, refractory mortars
and silicate materials—Determination of
major and minor elements—Wavelength
dispersive X-ray fluorescence
spectrometry using lithium borate
fusion**



This Australian Standard® was prepared by Committee MN-007, Refractories and Refractory Materials. It was approved on behalf of the Council of Standards Australia on 22 December 2006.

This Standard was published on 23 February 2007.

The following are represented on Committee MN-007:

- Australian Ceramic Society
 - Australasian Institute of Mining and Metallurgy
 - Bureau of Steel Manufacturers of Australia
 - CSIRO—Manufacturing and Materials Technology
 - Institute of Refractories Engineers
 - Refractories Manufacturers Association of Australia
 - University of New South Wales
-

This Standard was issued in draft form for comment as D 106251.

Standards Australia wishes to acknowledge the participation of the expert individuals that contributed to the development of this Standard through their representation on the Committee and through the public comment period.

Keeping Standards up-to-date

Australian Standards® are living documents that reflect progress in science, technology and systems. To maintain their currency, all Standards are periodically reviewed, and new editions are published. Between editions, amendments may be issued.

Standards may also be withdrawn. It is important that readers assure themselves they are using a current Standard, which should include any amendments that may have been published since the Standard was published.

Detailed information about Australian Standards, drafts, amendments and new projects can be found by visiting www.standards.org.au

Standards Australia welcomes suggestions for improvements, and encourages readers to notify us immediately of any apparent inaccuracies or ambiguities. Contact us via email at mail@standards.org.au, or write to Standards Australia, GPO Box 476, Sydney, NSW 2001.

STANDARDS AUSTRALIA

RECONFIRMATION

OF

AS 2503.6—2007

**Refractories and refractory materials—Chemical analysis
Part 6: Refractories, refractory mortars and silicate materials—Determination of
major and minor elements—Wavelength dispersive X-ray fluorescence
spectrometry using lithium borate fusion**

RECONFIRMATION NOTICE

Technical Committee MN-007 has reviewed the content of this publication and in accordance with Standards Australia procedures for reconfirmation, it has been determined that the publication is still valid and does not require change.

Certain documents referenced in the publication may have been amended since the original date of publication. Users are advised to ensure that they are using the latest versions of such documents as appropriate, unless advised otherwise in this Reconfirmation Notice.

Approved for reconfirmation in accordance with Standards Australia procedures for reconfirmation on 19 August 2015.

The following are represented on Technical Committee MN-007:

Australian Ceramic Society
Bureau of Steel Manufacturers of Australia
Cement Industry Federation
CSIRO
Institute of Refractories Engineers
Refractories Manufacturers Association of Australia
The University of New South Wales

NOTES

Currently in preview, click buy full vers.

Australian Standard[®]

**Refractories and refractory materials—
Chemical analysis**

**Part 6: Refractories, refractory mortars
and silicate materials—Determination of
major and minor elements—Wavelength
dispersive X-ray fluorescence
spectrometry using lithium borate
fusion**

First published as AS 2503.6—2007.
Reissued incorporating Amendment No. 1 (May 2011).

COPYRIGHT

© Standards Australia Limited

All rights are reserved. No part of this work may be reproduced or copied in any form or by any means, electronic or mechanical, including photocopying, without the written permission of the publisher, unless otherwise permitted under the Copyright Act 1968.

Published by SAI Global Limited under licence from Standards Australia Limited, GPO Box 476, Sydney, NSW 2001, Australia

ISBN 0 7337 8058 X

PREFACE

This Standard was prepared by the Standards Australia Committee MN-007, Refractories and Refractory Materials, to provide an XRF method for the analysis of refractories, refractory mortars and silicate materials.

This Standard incorporates Amendment No. 1 (May 2011). The changes required by the Amendment are indicated in the text by a marginal bar and amendment number against the clause, note, table, figure or part thereof affected.

The objective of this Standard is to provide those responsible for the analysis of refractories, refractory mortars and silicate materials with standardized procedures that will ensure the integrity of the analysis.

The terms 'normative' and 'informative' have been used in this Standard to define the application of the appendix to which they apply. A 'normative' appendix is an integral part of a Standard, whereas an 'informative' appendix is only for information and guidance.

CONTENTS

	<i>Page</i>
1 SCOPE.....	4
2 REFERENCED DOCUMENTS.....	4
3 PRINCIPLE	4
4 SAFETY	5
5 REAGENTS AND MATERIALS	5
6 APPARATUS	7
7 SAMPLING AND SAMPLES	8
8 PROCEDURE.....	8
9 PRECISION TESTS	16
10 STATISTICAL TOLERANCES WITH CERTIFIED REFERENCE MATERIALS (CRMs).....	16
APPENDICES	
A LIST OF MATERIAL TYPES AND COMPONENT RANGES FOR LISTED MATERIALS	18
B THE USE AND CARE OF PLATINUM AND PLATINUM ALLOY.....	22

STANDARDS AUSTRALIA

Australian Standard

Refractories and refractory materials—Chemical analysis

Part 6: Refractories, refractory mortars and silicate materials—Determination of major and minor elements—Wavelength dispersive X-ray fluorescence spectrometry using lithium borate fusion

1 SCOPE

This Standard sets out a wavelength dispersive X-ray fluorescence spectrometry method for the analysis of refractories and refractory mortars, including their feedstocks, using a spectrometer and current XRF software. Silicate materials can also be analysed using this method. This method is applicable to the elements and concentrations ranges in Appendix A (Tables A1–A8).

2 REFERENCED DOCUMENTS

The following documents are referred to in this Standard:

AS	
2243	Safety in laboratories (series)
2563	Wavelength dispersive X-ray fluorescence spectrometers—Determination of precision
2879	Alumina
2879.7	Part 7: Determination of trace elements—Wavelength dispersive X-ray fluorescence spectrometric method
4392	Heavy mineral sands—Analysis by wavelength dispersive X-ray fluorescence spectrometry
4392.2	Part 2: Zircon materials
4433	Guide to sampling of particulate materials
4433.1	Part 1: Sampling procedures
4433.2	Part 2: Preparation of samples
ISO	
12677	Chemical analysis of refractory products by XRF—Fused cast bead method

3 PRINCIPLE

X-ray fluorescence measurements are made on a glass disc prepared by fusing and casting the test sample into a lithium borate disc.

Calibration is primarily carried out using synthetic standards prepared from pure chemicals using a two-point regression. Reference materials are used in addition to the synthetic standards. Matrix corrections are employed to ensure linearity throughout the concentration range of the calibration. Reference materials are used to verify the calibration.

Intensity measurements are corrected for spectrometer drift via a monitor fused glass disc.

NOTE: It is assumed that the analyst is using current XRF software for calibration and instrument control. (See Section 8.)