

Australian Standard[®]

**Determination of flash and fire points—
Cleveland open cup method**

STANDARDS
Australia



This Australian Standard® was prepared by Committee CH-009, Safe Handling of Chemicals. It was approved on behalf of the Council of Standards Australia on 30 June 2006. This Standard was published on 24 July 2006.

The following are represented on Committee CH-009:

- Australasian Fire Authorities Council
 - Australian Paint Manufacturers' Federation
 - AVCARE
 - Consumers' Federation of Australia
 - Department for Administrative and Information Services, SA
 - Department of Conservation and Environment, NSW
 - Department of Consumer and Employment Protection, WA
 - Department of Emergency Services, QLD
 - New Zealand Chemical Industry Council
 - Plastics and Chemicals Industries Association, Australia
 - Trans NZ
 - Victorian WorkCover Authority
-

This Standard was issued in draft form for comment as DR 06210.

Standards Australia wishes to acknowledge the participation of the expert individuals that contributed to the development of this Standard through their representation on the Committee and through public comment received.

Keeping Standards up-to-date

Australian Standards® are living documents that reflect progress in science, technology and systems. To maintain their currency, all Standards are periodically reviewed, and new editions are published. Between editions, amendments may be issued.

Standards may also be withdrawn. It is important that readers assure themselves they are using a current Standard, which should include any amendments that may have been published since the Standard was published.

Detailed information about Australian Standards, drafts, amendments and new projects can be found by visiting www.standards.org.au

Standards Australia welcomes suggestions for improvements, and encourages readers to notify us immediately of any apparent inaccuracies or ambiguities. Contact us via email at mail@standards.org.au, or write to Standards Australia, GPO Box 476, Sydney, NSW 2001.

Australian Standard[®]

**Determination of flash and fire points—
Cleveland open cup method**

First published as AS 2585—2006.

COPYRIGHT

© Standards Australia

All rights are reserved. No part of this work may be reproduced or copied in any form or by any means, electronic or mechanical, including photocopying, without the written permission of the publisher.

Published by Standards Australia, GPO Box 476, Sydney, NSW 2001, Australia

ISBN 0 7337 7635 3

PREFACE

This Standard was prepared by Standards Australia/Standards New Zealand Committee CH-009, Safe Handling of Chemicals. It is identical with, and has been reproduced from ISO 2592:2000, *Determination of flash and fire points—Cleveland open cup method*.

The objective of this Standard is to provide a method for the determination of open-cup flash and fire points using the Cleveland (open cup) apparatus.

As this Standard is reproduced from an International Standard, the following modifications apply:

- (a) Its number does not appear on each page of text and its identity is shown on the cover and title page.
- (b) In the source text, 'this International Standard' should read 'this Australian Standard'.
- (c) A full point substitutes for a comma when referring to a decimal marker.

References to International Standards should be replaced by references to equivalent Australian Standards as follows:

<i>Reference to International Standard</i>		<i>Reference to Australian Standard</i>	
ISO		AS	
2719	Petroleum products and lubricants— Determination of flash point— Pensky-Martens closed cup method	2106	Methods for the determination of the flash point of flammable liquids (closed cup)
		2106-2	Part 2: Determination of flash point—Pensky-Martens closed cup method
Guide 33	Uses of certified reference materials	AS 19	Reference materials
		19.33	Guide 33: Uses of certified reference materials
Guide 35	Certification of reference materials— General and statistical principles	19.35	Guide 35: Certification of reference materials—General and statistical principles

Other International Standards referenced in the source document have not been adopted as Australian Standards.

The terms 'normative' and 'informative' have been used in this Standard to define the application of the annex to which they apply. A 'normative' annex is an integral part of a Standard, whereas an 'informative' appendix is only for information and guidance.

CONTENTS

	<i>Page</i>
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 Principle	2
5 Reagents and materials	2
6 Apparatus	2
7 Preparation of apparatus	2
8 Sampling	3
9 Sample handling	3
10 Procedure for determining flash point	4
11 Procedure for determining fire point	5
12 Calculation	5
13 Expression of results	5
14 Precision	5
15 Test report	6
Annex A (normative) Cleveland open cup apparatus	7
Annex B (normative) Thermometer specification	10
Annex C (informative) Verification of apparatus	11
Bibliography	14

Currently in preview, click buy full version

AUSTRALIAN STANDARD

Determination of flash and fire points — Cleveland open cup method

WARNING — The use of this International Standard may involve hazardous materials, operations and equipment. This International Standard does not purport to address all of the safety problems associated with its use. It is the responsibility of the user of this International Standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.

1 Scope

This International Standard specifies a procedure for the determination of flash and fire points of petroleum products using the Cleveland open cup apparatus. It is applicable to petroleum products having an open cup flash point above 79 °C, except fuel oils, which are most commonly tested by the closed cup procedure described in ISO 2719 [1].

NOTE Flash point and fire point are indications of the ability of a substance to form a flammable mixture with air under controlled conditions, and then to support combustion. They are only two of a number of properties that may contribute towards the assessment of overall flammability and combustibility of a material.

2 Normative references

The following normative documents contain provisions which, through reference in this text, constitute provisions of this International Standard. For dated references, subsequent amendments to, or revisions of, any of these publications do not apply. However, parties to agreements based on this International Standard are encouraged to investigate the possibility of applying the most recent editions of the normative documents indicated below. For undated references, the latest edition of the normative document referred to applies. Members of ISO and IEC maintain registers of currently valid International Standards.

ISO 3170:1988, *Petroleum liquids — Manual sampling*.

ISO 3171:1988, *Petroleum liquids — Automatic pipeline sampling*.

3 Terms and definitions

For the purposes of this International Standard, the following terms and definitions apply.

3.1

flash point

lowest temperature of the test portion, corrected to a barometric pressure of 101,3 kPa, at which application of a test flame causes the vapour of the test portion to ignite and the flame to propagate across the surface of the liquid, under the specified conditions of test

3.2

fire point

lowest temperature of the test portion, corrected to a barometric pressure of 101,3 kPa, at which application of a test flame causes the vapour of the test portion to ignite and sustain burning for a minimum of 5 s under the specified conditions of test