

# Australian Standard 2647—1983

---

## BIOLOGICAL SAFETY CABINETS—INSTALLATION AND USE

AS 2647—1994  
Biological safety cabinets—  
installation and use  
(In Professional Packages 17F,  
34D) 27pp G  
Sets out recommended practices  
for the installation, operation,  
maintenance, decontamination  
and inspection of Class I and  
Class II biological safety cabinets.  
Supersedes AS 2647—1983; DR 92160.  
Publication date 1994-04-11.



**STANDARDS ASSOCIATION OF AUSTRALIA**  
*Incorporated by Royal Charter*

This Australian standard was prepared by Committee MS/28, Controlled Environment. It was approved on behalf of the Council of the Standards Association of Australia on 7 July 1983 and published on 5 September 1983.

---

The following interests are represented on Committee MS/28:

Australian Institute of Refrigeration, Air Conditioning and Heating  
Australian Medical Association  
Commonwealth Serum Laboratories  
Confederation of Australian Industry  
CSIRO, Australian National Animal Health Laboratory  
Department of Housing and Construction  
Department of Defence  
Firms and Consultants specializing in equipment and design for controlled environments  
Health Commission of Victoria  
Health Department of New South Wales  
National Association of Testing Authorities Australia  
National Biological Standards Laboratory  
National Council of Chemical and Pharmaceutical Industries  
Public Works Department, N.S.W.  
Royal Australian Institute of Architects

---

To keep abreast of progress in industry, Australian standards are subject to continuous review and are kept up-to-date by the issue of amendments or new editions as necessary. It is important therefore that standards users ensure that their standards are up-to-date. Full details of all SAA publications will be found in the Annual List of Australian Standards; these details are supplemented by listings in the SAA monthly journal 'The Australian Standard'. Information on the Annual List and 'The Australian Standard' may be obtained from any sales office of the Association, where details are also available of the current status of individual standards. Suggestions for improvements to published standards, addressed to the head office of the Association, are welcomed.

---

*This standard was issued in draft form for comment as DR 81274.*

**AUSTRALIAN STANDARD**

**BIOLOGICAL SAFETY  
CABINETS—INSTALLATION  
AND USE**

**AS 2617—1983**

First published .....1983

**PUBLISHED BY THE STANDARDS ASSOCIATION OF AUSTRALIA  
STANDARDS HOUSE, 80 ARTHUR STREET, NORTH SYDNEY, N.S.W.**

ISBN 0 7262 3098 7

## PREFACE

This standard was prepared by the Association's Committee on Controlled Environment to satisfy the need for a guide to the two classes of biological safety cabinets specified in Parts 1 and 2 of AS 2252, Biological Safety Cabinets.

The issue in 1972 of a public review draft DR 72168 for cleanrooms and workstations drew a comment from the Commonwealth Serum Laboratories which pointed to the need for work to be done under aseptic conditions and the need to contain potentially dangerous materials, and to the fact that the standards then proposed did not adequately cater for these situations.

A need has been established for standards applying to cabinets designed to minimize the immediate risks from aerosols to personnel working with hazardous biological agents. There is an increasing awareness of the hazards of exposure to infectious agents that present an actual or potential risk to people handling them. However, it must be appreciated that biological safety cabinets act only as the primary protective barrier and secondary and tertiary barriers are necessary to establish a safe laboratory environment. A quaternary barrier may be necessary to ensure a safe community environment.

In the preparation of this standard, reference was made to the booklet published by the United States Department of Health, Education and Welfare, Public Health Service entitled 'Classification of Etiological Agents on the Basis of Hazard'. Although written for American conditions, this booklet describes the basis for such classification of agents, and lists bacterial, fungal, parasitic, viral, rickettsial and chlamydial agents in five classes. The least hazardous agents are in Class 1 and those requiring the greatest restrictions are in Class 4. Class 5 contains agents which are specifically excluded from the U.S. by law. The list of organisms prohibited in Australia differs somewhat from the American list. Details can be obtained from the Quarantine Division of the Department of Health, Canberra.

©Copyright — STANDARDS ASSOCIATION OF AUSTRALIA 1983

Users of standards are reminded that copyright subsists in all SAA publications. No part of this publication may be reproduced, stored in a retrieval system in any form or transmitted by any means without prior permission in writing of the Standards Association of Australia.

## CONTENTS

	<i>Page</i>
FOREWORD .....	4
<b>SECTION 1: SCOPE AND GENERAL</b>	
1.1 Scope .....	5
1.2 Referenced Documents .....	5
1.3 Selection and Class of Cabinet .....	5
<b>SECTION 2: INSTALLATION, OPERATION, DECONTAMINATION AND INSPECTION</b>	
2.1 Quaternary Barrier .....	7
2.2 Tertiary Barrier .....	7
2.3 Secondary Barrier .....	7
2.4 Primary Barrier .....	7
<b>APPENDICES</b>	
A Consideration for Exhaust Arrangements .....	10
B Suggested Procedures for Cleaning up Following Spillage in a Class I Cabinet or a Class II Cabinet .....	13
C Decontamination of Cabinets .....	14
Annex 1 to Appendix C. Safety Information, Advice and Instruction for Personnel Involved in Formaldehyde Decontamination .....	17
D Monitoring Effectiveness of Decontamination .....	19

## STANDARDS ASSOCIATION OF AUSTRALIA

**Australian Standard**  
**for**  
**BIOLOGICAL SAFETY CABINETS—INSTALLATION AND USE**

**FOREWORD**

This standard applies to the two classes of cabinets specified in Parts 1 and 2 of AS 2252. Classes I and II provide equal protection for personnel against agents of ordinary, potential or special hazard. Class III provides protection to personnel against agents of extreme or suspected extreme hazard and presents a special case not covered by this standard.

The choice of cabinet depends on the nature of the agent(s) being handled. A list of organisms of special hazard which are commonly found in Australia is given in Table 1.2.

The choice of cabinet introduces the question of containment and there are three degrees of physical containment provided by the following barriers:

- (a) *Primary barriers*, which provide immediate containment at the source. Examples are biological safety cabinets, centrifuge enclosures and ventilated animal enclosures.
- (b) *Secondary barriers*, which provide secondary containment in the event of failure of the primary barrier. An example of a secondary barrier is the room (and its support facilities) within which the primary barrier is located.
- (c) *Tertiary barriers*, which are the building(s) in which the secondary barrier(s) are housed.
- (d) *Quaternary barriers*, which are the geographical sites for buildings constituting the tertiary barriers. An example of a quaternary barrier is an island.

Compliance with this standard should achieve the following:

- (i) Protection, by the primary barrier, of personnel directly engaged in handling infectious materials.
- (ii) Protection of supportive staff, animals and facilities in adjacent work areas and in the surrounding community.
- (iii) For a Class II cabinet protection of the product.

The Australian concept of self-contained cabinets is in contrast to that of some other countries, where the cabinet exhaust air is ducted into ventilating, airconditioning or exhaust systems which discharge to atmosphere introducing the possibility of uncontrollable variations in cabinet airflow velocity and pressure (see Appendix A).