

# Australian Standard 2696—1984

---

## MEDICAL EQUIPMENT— SINGLE-USE URETHRAL CATHETERS (STERILE) FOR GENERAL MEDICAL USE

[Title allocated by Defence Cataloguing Authority:  
CATHETER, URETHRAL (Sterile, Single-use)]

---



**PUBLISHED BY THE STANDARDS ASSOCIATION OF AUSTRALIA  
STANDARDS HOUSE, 80 ARTHUR ST, NORTH SYDNEY, N.S.W.**

*Incorporated by Royal Charter*



This Australian standard was prepared by Committee MD/1, Hypodermic and Other Equipment for General Medical Use. It was approved on behalf of the Council of the Standards Association of Australia on 11 January 1984 and published on 4 June 1984.

---

The following interests are represented on Committee MD/1:

- Australian Chamber of Commerce
- Australian Dental Standards Laboratory
- Australian Medical Association
- Australian Medical Devices and Diagnostics Association
- Confederation of Australian Industry
- Department of Defence
- Department of Veterans Affairs
- Department of Health
- Hospitals and Hospital Associations
- New South Wales Government Stores Department
- Pharmaceutical Association of Australia

Representatives of the following interest also participated in the drafting of this standard:

- Urological Society of Australasia

---

*Review of Australian Standards* - To keep abreast of progress in industry, Australian standards are subject to periodic review and are kept up-to-date by the issue of amendments or new editions as necessary. It is important therefore that standards users ensure that they are in possession of the latest edition, and any amendments thereto.

Full details of all SAA publications will be found in the Catalogue of Australian Standards; this information is supplemented each month by SAA's journal 'The Australian Standard', which subscribing members receive, and which gives details of new publications, new editions and amendments, and of withdrawn standards.

Suggestions for improvements to Australian standards, addressed to the head office of the Association, are welcome. Notification of any inaccuracy or ambiguity found in an Australian standard should be made without delay in order that the matter may be investigated and appropriate action taken.

First published ..... 1984

*This standard was issued in draft form for comment as DR 78119.*

ISBN 0 7262 3340 4

## PREFACE

This standard was prepared by the Association's Committee on Hypodermic and Other Equipment for General Medical Use, under the direction of the Medical Materials and Equipment Standards Board.

In the preparation of the standard, account was taken of BS 1695, Catheters, and its draft revision.

The standard applies only to certain types of urinary catheters which are introduced through the urethra, and which are considered to be in widespread general use and have been carefully selected from the large variety currently available. Although other types of catheters may still be used, adherence to the types covered by the standard would simplify the inventories of manufacturers and hospitals, yet still provide an adequate range for most purposes.

The committee was aware of the fact that there are problems associated with long term use, compatibility and rigidity of the catheter.

The diagrams illustrate a typical catheter for the purpose of defining the relevant dimensions and terms but the illustrated design of the catheter does not form part of the standard unless specifically stated as such.

Some facilities required for the testing of materials for compliance with this standard are available at the Australian Dental Standards Laboratory, 240 Langridge Street, Abbotsford, Victoria, 3067.

An illustrated document describing the various types of catheter referred to in this standard is in process of preparation by the Urological Society of Australia, and it is envisaged that this will be published in 1984.

## CONTENTS

	<i>Page</i>
<b>SPECIFICATION</b>	
1 Scope .....	4
2 Referenced Documents .....	4
3 Classification .....	4
4 Definitions .....	4
5 Tolerance on Size .....	4
6 Size Range .....	4
7 Materials .....	4
8 Design and Construction—General Requirements .....	5
9 Specific Requirements According to Type .....	5
10 Sterilization .....	6
11 Sterility .....	6
12 Packaging .....	6
13 Labelling .....	6
<b>APPENDICES</b>	
A Method of Testing Catheters for Toxicity .....	13
B Simulated In-use Test .....	16
C Method for Determining Flow Rate of Urethral Catheter .....	17
D Method for Determining Balloon Performance of Catheter .....	19
E Method of Testing Catheters for Sterility .....	20
F Assessment of Compliance .....	23

©Copyright — STANDARDS ASSOCIATION OF AUSTRALIA 1984

Users of standards are reminded that copyright subsists in all SAA publications. No part of this publication may be reproduced, stored in a retrieval system in any form or transmitted by any means without prior permission in writing of the Standards Association of Australia.



## STANDARDS ASSOCIATION OF AUSTRALIA

## Australian Standard

for

**MEDICAL EQUIPMENT—SINGLE-USE URETHRAL CATHETERS (STERILE)  
FOR GENERAL MEDICAL USE**

**1 SCOPE.** This standard specifies requirements for commercially available sterile, single-use urethral catheters intended to be introduced through the urethra into the urinary bladder.

**NOTES:**

1. Materials to be used for the construction of sterile urethral catheters for single-use are not specified in detail as their selection will depend, to some extent, upon the design, process of manufacture and method of sterilization employed by the individual manufacturers.
2. Advisory information on sampling and assessing compliance with this standard is given in Appendix F.

**2 REFERENCED DOCUMENTS.** The following documents are referred to in this standard:

- AS 1094 Single-use Syringes (Sterile) for General Medical Use
- AS 1600 Conical Fittings with 6 percent (Luer) Taper for Hypodermic and Other Surgical Equipment\*
- AS 2070 Plastics Materials for Food Contact Use
- BS 3574 Recommendations for the Storage of Vulcanized Rubber  
United States Pharmacopoeia XIX.

**3 CLASSIFICATION.** For the purpose of this standard, catheters are classified into two types in accordance with the following:

- (a) *Type I—Urethral catheters without balloon.* Type I classification comprises the following sub-types:
  - (i) Nelaton, Jaques, or other 'ordinary' basically cylindrical urethral catheters.
  - (ii) Whistle tip urethral catheters.
  - (iii) Tiemann urethral catheters.
  - (iv) Gibbon urethral catheters.
- (b) *Type II—Urethral catheters with self-retaining balloon.* Type II classification comprises the following sub-types:
  - (i) Foley type urethral catheters (a Nelaton 'ordinary' basically cylindrical urethral catheter with balloon).
  - (ii) Whistle tip urethral catheters with balloon.
  - (iii) Tiemann urethral catheters with balloon.
  - (iv) Gibbon urethral catheters with balloon.

**4 DEFINITIONS.** For the purpose of this standard, the following definitions apply:

- 4.1 Urethral catheter**—a device for the drainage of fluids from the urinary bladder and urethra.
- 4.2 French gauge** (also known as Charriere gauge and commonly abbreviated as F, Fr or Ch)—a system indicating the external circumference of the catheter, in millimetres.

**4.3 Tip end**—the end of a urethral catheter which is furnished with one or more 'eyes' communicating with the catheter lumen.

**4.4 Eye**—the aperture at the tip end of the catheter.

**4.5 Shaft**—the main length of tubing excluding the tip end.

**4.6 Free end**—the end of the catheter opposite to the tip end.

**4.7 Lumen**—the space enclosed by the (internal) wall of the catheter.

**4.8 Size**—the nominal external circumference of the shaft, expressed in millimetres (and commonly referred to as French or Charriere gauge (symbols F, Fr or Ch)).

**4.9 Unit**—the catheter.

**4.10 Unit pack**—a pack containing a single unit.

**4.11 Multiple pack**—a pack containing a number of unit packs.

**4.12 Store pack**—a pack containing one or more multiple packs.

**4.13 Balloon**—an integral portion of the catheter designed for inflation and deflation, for use as a retention device.

**4.14 Nominal capacity of balloon**—the minimum balloon capacity at which the catheter is designed to function satisfactorily.

**4.15 Rated capacity (maximum) of balloon**—the nominal capacity of the balloon or the nominal capacity plus an additional volume designated by the manufacturer.

**5 TOLERANCE ON SIZE.** When measured 75 mm from the tip end, the external circumference of the catheter shall be in accordance with Table 1, appropriate to the nominal French size and within the tolerances specified therein.

NOTE: For catheters with balloons, the size measurement should not be made over or in close proximity to the balloon.

**6 SIZE RANGE.** The size range of catheters shall be in accordance with Table 1.

**7 MATERIALS.**

**7.1 General.** The catheter shall be made of a suitable plastics material or of natural or synthetic rubber or elastomers, or of a combination of components made from any of these materials, and shall not become soft and sticky or harden excessively when used or during storage prior to use.

\*In course of preparation.