

Australian Standard™

**Workplace atmospheres—Method for  
sampling and gravimetric determination  
of respirable dust**

This Australian Standard was prepared by Committee CH-031, Methods for Examination of Workplace Atmospheres. It was approved on behalf of the Council of Standards Australia on 22 January 2004 and published on 20 February 2004.

---

The following are represented on Committee CH-031:

Australian Aluminium Council  
Australian Chamber of Commerce and Industry  
Australian Institute of Occupational Hygienists  
Australian Mines and Metals Association  
Bureau of Steel Manufacturers of Australia  
Clean Air Society of Australia & New Zealand  
Coal Services  
Commonwealth Department of Health and Ageing  
Department for Administrative and Information Services, SA  
Department of Consumer & Employment Protection, WorkSafe Division, WA  
Department of Mineral Resources, NSW  
National Association of Testing Authorities Australia  
WorkCover New South Wales

---

### **Keeping Standards up-to-date**

Standards are living documents which reflect progress in science, technology and systems. To maintain their currency, all Standards are periodically reviewed, and new editions are published. Between editions, amendments may be issued. Standards may also be withdrawn. It is important that readers assure themselves they are using a current Standard, which should include any amendments which may have been published since the Standard was purchased.

Detailed information about Standards can be found by visiting the Standards Web Shop at [www.standards.com.au](http://www.standards.com.au) and looking up the relevant Standard in the on-line catalogue.

Alternatively, the printed Catalogue provides information current at 1 January each year, and the monthly magazine, *The Global Standard*, has a full listing of revisions and amendments published each month.

Australian Standards™ and other products and services developed by Standards Australia are published and distributed under contract by SAI Global, which operates the Standards Web Shop.

We also welcome suggestions for improvement in our Standards, and especially encourage readers to notify us immediately of any apparent inaccuracies or ambiguities. Contact us via email at [mail@standards.org.au](mailto:mail@standards.org.au), or write to the Chief Executive, Standards Australia International Ltd, GPO Box 5420, Sydney, NSW 2001.

---

*This Standard was issued in draft form for comment as DR 02321.*

Australian Standard™

**Workplace atmospheres—Method for  
sampling and gravimetric determination  
of respirable dust**

Revised and reissued as AS 2985—1987.  
Second edition 2004.

**COPYRIGHT**

© Standards Australia International

All rights are reserved. No part of this work may be reproduced or copied in any form or by any means, electronic or mechanical, including photocopying, without the written permission of the publisher.

Published by Standards Australia International Ltd  
GPO Box 5420, Sydney, NSW 2001, Australia

ISBN 0 7337 5755 3

## PREFACE

This Standard was prepared by the Joint Standards Australia/Standards New Zealand Committee CH-031, Methods for the Examination of Workplace Atmospheres to supersede AS 2985—1987, *Workplace atmospheres—Method for sampling and gravimetric determination of respirable dust*. After consultation with stakeholders in both countries, Standards Australia and Standards New Zealand decided to develop this Standard as an Australian Standard rather than an Australian/New Zealand Standard.

Since the publication of the previous edition, the International Standards Organization has superseded its technical report with ISO 7708:1995, *Air quality—Particle size fraction definitions for health-related sampling*. The objective of this revision is to align this Standard more closely with international practices in the field of respirable workplace dust measurement.

Sampling devices other than those previously specified have now become available and are included in the Standard.

The Committee is aware of the role that errors, especially in weighing, play in the overall uncertainty of the measurements and has attempted to address the issue in this revision, by reference to a new International Standard, ISO 15767:2003, *Workplace atmospheres—Controlling and characterizing errors in weighing collected aerosols*.

The terms ‘normative’ and ‘informative’ have been used in this Standard to define the application of the appendix to which they apply. A ‘normative’ appendix is an integral part of a Standard, whereas an ‘informative’ appendix is only for information and guidance.

## CONTENTS

	<i>Page</i>
FOREWORD.....	4
1 SCOPE.....	5
2 OBJECTIVE .....	5
3 REFERENCED DOCUMENTS .....	5
4 DEFINITIONS .....	5
5 PRINCIPLE .....	7
6 APPARATUS .....	7
7 FLOW RATE DETERMINATION .....	10
8 PROCEDURE .....	10
9 CALCULATIONS .....	12
10 EXPRESSION OF UNCERTAINTY IN MEASUREMENT.....	13
11 DETECTION LIMIT.....	13
12 REPORTING OF RESULTS.....	13
APPENDICES	
A BIBLIOGRAPHY.....	14
B CALIBRATION OF SECONDARY FLOW METER.....	16

## FOREWORD

Most airborne industrial dusts contain particles of a wide range of sizes. The behaviour, deposition and fate of any particle after entry into the human respiratory system and the response that it elicits depends on the nature and size of the particle.

Occupational hygiene practice commonly differentiates between two size fractions of airborne dust, namely respirable and inhalable dust. Where particles may have toxic effects if absorbed in the nasopharyngeal (nose and throat) region or may have toxic effects if ingested after deposition in this region, it is appropriate to measure the mass concentration of inhalable particles in the atmosphere. It may also be apt to measure this size fraction for particles that exhibit no specific toxic effects, namely 'particulates/dusts not otherwise classified.' AS 3640, *Workplace atmospheres—Method for sampling and gravimetric determination of inhalable dust* should be referred to for determining inhalable particles in workplace atmospheres.

Respirable particles can be measured when the nature of these particles is such that they exhibit toxic effects primarily when deposited in the alveolar region (deepest reserve) of the lungs. This usually applies to toxic insoluble particles that accumulate in the lungs such as crystalline silica, coal dust and cadmium oxide fume. This Standard sets down the method for determining the mass concentration of these respirable sized particles in workplace atmospheres.

## STANDARDS AUSTRALIA

### Australian Standard

## Workplace atmospheres—Method for sampling and gravimetric determination of respirable dust

### 1 SCOPE

This Standard sets out a method for the collection and gravimetric determination of respirable dust in workplace atmospheres. This method does not consider the measurement of 'inhalable' dust which is covered in AS 3640.

### 2 OBJECTIVE

The objective of this Standard is to provide a method to assess personal exposure to respirable dust by sampling in a worker's breathing zone.

Whilst the method only allows for personal sampling, it can also be used to assist in controlling the occupational environment by means of static sampling, i.e. samples taken at a fixed location. However, static samples are not to be used to evaluate health risks unless a specific situation or circumstances indicates otherwise.

### 3 REFERENCED DOCUMENTS

The following documents are referred to in this Standard:

#### AS

- 2162 Verification and use of volumetric apparatus
- 2162.1 Part 1: General—Volumetric glassware
- 2380 Electrical equipment for explosive atmospheres —Explosion-protection techniques
- 2380.7 Part 7: Intrinsic safety
- 2430 Classification of hazardous areas
- 2430.1 Part 1: Explosive gas atmospheres
- 3640 Workplace atmospheres—Method for sampling and gravimetric determination of inhalable dust

#### AS/NZS

- 61241 Electrical apparatus for use in the presence of combustible dust
- 61241.3 Part 3: Classification of areas where combustible dusts are or may be present

#### ISO

- 708 Air quality—Particle size fraction definitions for health-related sampling
- 5707 Workplace atmospheres—Controlling and characterizing errors in weighing collected aerosols

MORRIS, Edwin C. and FEN, Kitty M.K. *The Calibration of Weights and Balances* Monograph 4: Technology Transfer Series, National Measurement Laboratory, CSIRO 2001.