

Australian Standard<sup>®</sup>

---

**APPROVAL AND TEST  
SPECIFICATION—  
ELECTROMEDICAL EQUIPMENT—  
GENERAL REQUIREMENTS**

---

[Title Allocated by Reference Cataloguing Authority:  
MEDICAL AND DENTAL EQUIPMENT, ELECTRICAL  
(Approval and Test Specification, General Safety  
Requirements) . . . NSC 65 GP]

This Australian standard was prepared by Committee EL/18/1, Safety Requirements of Electromedical Equipment under the direction of Committee EL/18, Electromedical Equipment. It was approved on behalf of the Council of the Standards Association of Australia on 28 July 1986 and published on 5 September 1986.

---

The following interests are represented on Committee EL/18/1:

Australian Association of Physical Medicine and Rehabilitation Medicine  
Australian Federation for Medical and Biological Engineering  
Australian Medical Devices & Diagnostics Association  
Australian Private Hospitals Association  
Australian Society of Anaesthetists  
Australian and New Zealand Intensive Care Society  
Confederation of Australian Industry  
Department of Defence  
Department of Health  
Department of Health, N.S.W.  
Department of Health, Qld  
Department of Hospital and Allied Services, V.A.  
Department of Public Works, N.S.W.  
Department of Veterans' Affairs  
Diagnostic Imaging and Medical Electronics Association of Australia  
Institution of Biomedical Engineers (Australia)  
Queensland Electricity Commission  
Royal Australasian College of Physicians  
Royal Australasian College of Surgeons  
Royal Australian Nursing Federation  
South Australian Health Commission  
State Electricity Commission of Victoria

---

**Review of Australian Standards.** To keep abreast of progress in industry, Australian Standards are subject to periodic review and are kept up to date by the issue of amendments or new editions as necessary. It is important therefore that Standards users ensure that they are in possession of the latest edition, and any amendments thereto.

Full details of all Australian Standards and related publications will be found in the Standards Australia Catalogue of Publications; this information is supplemented each month by the magazine 'The Australian Standard', which subscribing members receive, and which gives details of new publications, new editions and amendments, and of withdrawn Standards.

Suggestions for improvements to Australian Standards, addressed to the head office of Standards Australia, are welcomed. Notification of any inaccuracy or ambiguity found in an Australian Standard should be made without delay in order that the matter may be investigated and appropriate action taken.

Australian Standard<sup>®</sup>

---

**APPROVAL AND TEST  
SPECIFICATION—  
ELECTROMEDICAL EQUIPMENT—  
GENERAL REQUIREMENTS**

---

First published (as AS C200) . . . . .	1970
AS 3200 first published . . . . .	1978
Second edition . . . . .	1980
Third edition . . . . .	1986

Incorporating:  
Amdt 1—1990

PUBLISHED BY STANDARDS AUSTRALIA  
(STANDARDS ASSOCIATION OF AUSTRALIA)  
1 THE CRESCENT, HOMEBUSH, NSW 2140

ISBN 0 7262 4350 7

## PREFACE

This edition of this standard was prepared by the Association's Committee on Safety Requirements of Electromedical Equipment to supersede AS 3200—1980.

This amendment to the Standard is intended to—

- correct several requirements identified as necessary during application of the 1986 edition;
- remove critical conflicts with the requirements subsequently adopted in the 1990 edition of this Standard (likely to be numbered AS 3200.1, 1990);
- provide for the Standard, as modified by this amendment, to continue to be available for application as an alternative to the 1990 edition, for a period of two years after the date of publication of this amendment (after which time the Standard will be withdrawn).

This standard incorporates general requirements for the safety of electromedical equipment (including associated accessories and materials) and is the 'parent' of a series of approval and test specifications covering requirements for a wide range of electromedical equipment. These general requirements are intended to be read in conjunction with AS 3100, Definitions and General Requirements for Electrical Materials and Equipment, which prescribes the safety requirements for materials and equipment used in electrical installations complying with AS 3000, SAA Wiring Rules.

Specific requirements for individual items of electromedical equipment are contained in the 'particular' specifications which indicate whether or not the requirements and test methods stated in this general specification apply to the particular item of electromedical equipment and whether they are to be supplemented, modified or replaced. Under this system, the particular specification converts the general specification into the specification for the appliance or group of appliances covered by the particular specification.

The purpose of the series of specifications is to outline conditions which must be met to secure approval for the sale and use of electromedical and electrodentary equipment in Australia. Only safety matters and conditions closely allied thereto are covered. In some instances, however, these are more stringent than for most electrical appliances because of the exacting additional requirements necessary to ensure the safety of the patient, and because of the environmental conditions (e.g. high humidity, hazardous locations) in which some equipment may be used.

In every State of Australia legislation has been enacted which requires that electrical equipment of a prescribed class or type shall not be marketed unless approved by the relevant Statutory Electricity Authority. In general, the basis of approval is the series of approval and test specifications referred to above.

For equipment which is not prescribed (and electromedical equipment is currently not prescribed), statutory and supply authorities may require proof that it complies with the relevant approval and test specification(s) or some requirements thereof. For this reason, voluntary examination schemes are operated by the statutory authorities.

Irrespective of the need for approval of equipment covered by these specifications, the Association emphasizes that the specifications establish minimum standards of safety for electromedical equipment. For this reason alone there is a responsibility on each manufacturer or importer of electromedical equipment to ensure that such equipment complies with these specifications in all respects.

The standard does not cover tests to be applied as acceptance tests or routine maintenance tests on equipment, such tests being specified in AS 3551.

This edition incorporates the following technical changes:

- (a) The nomenclature for identifying patient-circuits has been changed to:
  - 'Cardiac protected'—'Type CF' . . . . . (equivalent to old 'Class A'.)
  - 'Body protected'—'Type BF' . . . . . (equivalent to old 'Class B'.)
  - 'Unprotected'—'Type B' . . . . . (equivalent to old 'Class Z'.)

NOTE: This nomenclature aligns with AS 3003\* and IEC 601.1† and compares with earlier classifications specified in previous editions of this standard (i.e. Type CF equivalent to Class A; Type BF equivalent to Class B; Type B equivalent to Class Z). In order to avoid confusion and provide necessary guidance to users of this edition of the standard, explanatory footnotes have been included on all pages referring to the new classification scheme for equipment and patient-circuits.

\* AS 3003 Electrical Installations—Patient Treatment Areas of Hospitals and Medical and Dental Practices.

† IEC 601.1 Medical Electrical Equipment, Part 1: General Requirements.

- (b) Increased earth leakage current for Class I equipment when tested under the fault condition of having either supply conductor interrupted.
- (c) Increased earth leakage current for Class I mobile X-ray equipment.
- (d) Safety requirements for ceiling-supported equipment now permit devices other than a safety brake.
- (e) The requirements for insulation of patient-circuits from earth have been clarified. This has allowed the complex explanation of investigation techniques for patient-circuit isolation/insulation to be deleted.
- (f) The permissible temperature limits for accessible parts of equipment have been specified separately, depending on whether the equipment is likely to contact the patient during normal operation.
- (g) The requirement for multi-pole overheating protective devices has been reduced for equipment where a single fault will not result in a hazard, e.g. double insulated components, and aligns with the philosophy of IEC 601.1\*.
- (h) Transformers in electromedical equipment are required to comply with AS 3200†.
- (j) The illustration of the frequency response characteristic of the measuring device in Appendix D, Fig D1, has been corrected.

This standard requires reference to the following:

(a) *Standards*

- AS 1169 Minimizing of Combustion Hazards Arising from the Medical Use of Flammable Anaesthetic Agents
- AS 1200 SAA Boiler Code
- AS 1319 Rules for the Design and Use of Safety Signs for the Occupational Environment
- AS 2030 SAA Gas Cylinders Code
- AS 2034 Electrical Equipment for Explosive Atmospheres—Flameproof Electric Lighting Fittings
- AS 2398 Fixed Diagnostic X-ray Equipment—Design, Construction and Installation—Safety Requirements
- AS 2500 Safe Use of Electricity in Patient Care
- AS 3000 Rules for the Electrical Equipment of Buildings, Structures and Premises (SAA Wiring Rules)
- AS 3003 Electrical Installations—Patient Treatment Areas of Hospitals and Medical and Dental Practices
- IEC 601.1 Medical Electrical Equipment, Part 1: General Requirements

(b) *Approval and Test Specifications*

- AS 3100 Definitions and General Requirements for Electrical Materials and Equipment
- AS 3112 Plugs and Plug Sockets
- AS 3116 Elastomer Insulated Electric Cables and Flexible Cables for Working Voltages of 0.6/1 kV
- AS 3142 Electric Water Heaters
- AS 3147 PVC Insulated Electric Cables and Flexible Cables for Working Voltages of 0.6/1 kV
- AS 3151 Thermostats and Energy Regulators
- AS 3191 Electric Flexible Cords
- AS 3201.2 Electrically-heated Incubators for Babies‡
- AS 3205 Dental and Mobile Diagnostic X-ray Equipment
- AS 3203 Electrocardiographs
- AS 3208 Transformers in Electromedical Equipment
- AS 3300 General Requirement for Household and Similar Electrical Appliances
- AS 3551 Acceptance and In Service Testing of Electromedical Equipment.

---

\* IEC 601.1 Medical Electrical Equipment, Part 1: General Requirements.

† AS 3208 Approved and Test Specification—Transformers in Electromedical Equipment.

‡ Under revision.



## SECTION 6. PROTECTION AGAINST HEAT AND FIRE

6.1	SURFACE TEMPERATURE .....	27
6.2	TEMPERATURE CONTROL DEVICES .....	27
6.3	HEATING DEVICES .....	28
6.4	EQUIPMENT CONTAINING OXYGEN-ENRICHED ATMOSPHERES .....	28

## SECTION 7. EQUIPMENT MARKING AND ACCOMPANYING DOCUMENTATION

7.1	MARKING ON EQUIPMENT .....	29
7.2	ACCOMPANYING DOCUMENTATION .....	29

## APPENDICES

A	TESTS TO BE APPLIED FOR TYPE APPROVAL .....	31
B	SYMBOLS .....	32
C	PLUGS AND SOCKETS USED FOR RETICULATING SUPPLY TO FREE-STANDING EQUIPMENT .....	33
D	MEASURING INSTRUMENT (MEASURING DEVICE) .....	34

## © Copyright — STANDARDS AUSTRALIA

Users of Standards are reminded that copyright subsists in all Standards Australia publications and software. Except where the Copyright Act allows and except where provided for below no publications or software produced by Standards Australia may be reproduced, stored in a retrieval system in any form or transmitted by any means without prior permission in writing from Standards Australia. Permission may be conditional on an appropriate royalty payment. Requests for permission and information on commercial software royalties should be directed to the head office of Standards Australia.

Standards Australia will permit up to 10 percent of the technical content pages of a Standard to be copied for use exclusively in-house by purchasers of the Standard without payment of a royalty or advice to Standards Australia.

Standards Australia will also permit the inclusion of its copyright material in computer software programs for no royalty payment provided such programs are used exclusively in-house by the creators of the programs.

Care should be taken to ensure that material used is from the current edition of the Standard and that it is updated whenever the Standard is amended or revised. The number and date of the Standard should therefore be clearly identified.

The use of material in print form or in computer software programs to be used commercially, with or without payment, or in commercial contracts is subject to the payment of a royalty. This policy may be varied by Standards Australia at any time.

## FOREWORD

The safe use of electromedical equipment can only be assured if—

- (a) the equipment is safe, and is appropriate for the procedure;
- (b) the environment in which the procedure is to be carried out is wired in accordance with an appropriate code; and
- (c) the procedure is carried out under an appropriate set of rules that links the equipment with the type of procedure and the type of wiring in the area.

Compared with other equipment, electromedical equipment warrants special care being taken in its use, for the following reasons:

- (i) The patient may not react normally to electric shock or heat because he may be unconscious, anaesthetized, or fastened to the examination or treatment equipment.
- (ii) The normal high contact resistance between the skin and an electrical conductor is considerably reduced when electrodes are applied to the surface of the body with electrode paste; even the contact resistance resulting from casual encounter with other conductors is significantly reduced in many electromedical procedures by the presence of copious amounts of other electrolytes, such as blood and saline. Furthermore, the application of electrodes or electrically-conducting transducers to internal tissues of the body also produces very low contact resistance.
- (iii) Many electromedical procedures require the introduction of an insulated conductor into direct contact with heart muscle. Ventricular fibrillation may be induced by minute currents well below the threshold of feeling if such currents flow through a small cross-section of heart muscle. This is termed 'micro-electrocution'.
- (iv) Electromedical equipment may be used to support or to replace vital organ functions, either temporarily or permanently; it may also be used in emergencies. Failure of the equipment or the supply may result in hazard to the patient.
- (v) Care must be taken to minimize the risk of fire or explosion arising from the use of flammable medical agents.
- (vi) Some electromedical procedures entail delivery to the patient of various forms of energy (i.e. electric current, heat, ionizing radiation, etc) in excess of the levels normally considered safe.

This specification is confined to minimum standards for the safety of equipment. Requirements for the safety of the environment in which electromedical procedures are undertaken are specified in AS 1003, and AS 2500 covers the safe application and use of equipment. In particular the flowchart in AS 2500 provides a summary of the safe application and use of the various classifications of electromedical equipment.

STANDARDS ASSOCIATION OF AUSTRALIA

---

**Australian Standard**

**APPROVAL AND TEST SPECIFICATION—**  
**ELECTROMEDICAL EQUIPMENT—GENERAL REQUIREMENTS**

---

SECTION 1. SCOPE AND APPLICATION

**1.1 SCOPE.** This specification sets out the general safety requirements for electromedical equipment which are additional to those in AS 3100 and additional to those in such other SAA approval and test specifications as are applicable to the equipment and component parts thereof.

**1.2 APPLICATION.** This specification shall be read in conjunction with AS 3100 and with any other relevant approval and test specification.

As and when particular approval and test specifications dealing with specific features of design, construction and testing of any particular class or type of electromedical equipment or parts thereof are issued, they shall supersede those general requirements of this specification which are specifically dealt with in the particular specification.

NOTE: Approvals authorities may at their discretion accept a manufacturer's declaration or an endorsed test report in respect to the testing of the whole machine or any part thereof.