

**SUPERSEDED BY:** AS/NZS 3200.2.18:1997

AS 3200.2.18—1992  
IEC 601-2-18:1990

Australian Standard®

---

**Approval and test specification—  
Medical electrical equipment**

**Part 2.18: Particular requirements  
for safety— Endoscopic equipment**

---

[IEC title: Medical electrical equipment—  
Part 2: Particular requirements for the safety of endoscopic equipment]



This Australian Standard was prepared by Committee HT/16, High Frequency Surgical and Therapy Equipment. It was approved on behalf of the Council of Standards Australia on 29 April 1992 and published on 13 July 1992.

---

The following interests are represented on Committee HT/16:

Australian Dental Association  
Australian Physiotherapy Association  
Department of Health, Housing and Community Services  
Department of Health, New South Wales  
Institute of Biomedical Engineering (Australia)  
Medical Industry Association of Australia  
South Australian Health Commission  
Urological Society of Australasia

---

*Review of Australian Standards* — To keep abreast of progress in industry, Australian Standards are subject to periodic review and are kept up to date by the issue of amendments or new editions as necessary. It is important therefore that standards users ensure that they are in possession of the latest edition, and any amendments thereto.

Full details of all Australian Standards and related publications will be found in the Standards Australia Catalogue of Publications; this information is supplemented each month by the magazine 'The Australian Standard', which subscribing members receive, and which gives details of new publications, new editions and amendments, and of withdrawn Standards.

Suggestions for improvements to Australian Standards, addressed to the head office of Standards Australia, are welcomed. Notification of any inaccuracy or ambiguity found in an Australian Standard should be made without delay in order that the matter may be investigated and appropriate action taken.

© Copyright — STANDARDS AUSTRALIA

Users of Standards are reminded that copyright subsists in all Standards Australia publications and software. Except where the Copyright Act allows and except where provided for below no publications or software produced by Standards Australia may be reproduced, stored in a retrieval system in any form or transmitted by any means without prior permission in writing from Standards Australia. Permission may be conditional on an appropriate royalty payment. Requests for permission and information on commercial software royalties should be directed to the head office of Standards Australia.

Standards Australia will permit up to 10 percent of the technical content pages of a Standard to be copied for use exclusively in-house by purchasers of the Standard without payment of a royalty or advice to Standards Australia.

Standards Australia will also permit the inclusion of its copyright material in computer software programs for no royalty payment provided such programs are used exclusively in-house by the creators of the programs.

Care should be taken to ensure that material used is from the current edition of the Standard and that it is updated whenever the Standard is amended or revised. The number and date of the Standard should therefore be clearly identified.

The use of material in print form or in computer software programs to be used commercially, with or without payment, or in commercial contracts is subject to the payment of a royalty. This policy may be varied by Standards Australia at any time.

Australian Standard®

---

**Approval and test specification—  
Medical electrical equipment**

**Part 2.18: Particular requirements  
for safety—Endoscopic equipment**

---

First published as AS 3200.2.18—1992.

## PREFACE

This Standard was prepared on behalf of the Standards Australia Multi-technics Board. It is technically equivalent to IEC 601-2-18—*Medical electrical equipment, Part 2: Particular requirements for the safety of endoscopic equipment*.

The Standard is one of a series of Approval and Test Specifications issued by Standards Australia for individual items of medical electrical equipment. It is supplementary to AS 3200.1—1990, *Approval and test specification—Medical electrical equipment, Part 1: General requirements for safety*. Only safety matters and closely aligned conditions are covered by this series. In some instances, however, these are more stringent than for most electrical appliances because of the additional requirements necessary to ensure the safety of the patient and because of the environmental conditions, e.g. high humidity and hazardous locations, in which some equipment is likely to be used.

The International Standard IEC 601-2-18 modifies and supplements the corresponding clauses of IEC 601-1 (1977), *Safety of medical electrical equipment, Part 1: General requirements*, first edition. This has since been revised and published as IEC 601-1 (1988), *Medical electrical equipment, Part 1: General requirements for safety*, second edition. The second edition has been adopted as the Australian Standard AS 3200.1—1990, *Approval and test specification—Medical electrical equipment—Part 1: General requirements for safety*, hereinafter referred to as the General Standard.

In order that IEC 601-2-18 and the General Standard are compatible the following editorial amendments have been made to IEC 601-2-18:

- (a) The section and clause titles have been reworded.
- (b) The following clauses have been renumbered:
  - 6.1 p): Items 'a)' and 'b)' to dashes '—'
  - 14.6: 'Amendment' to become 'Addition aa')
  - 19.3: Item d) to aa)
  - 20.4: Item 'a)' to item 'aa)'
  - Item 'c)' to item 'bb)' and
  - Item 'd)' to item 'cc)'
  - 42.3: Item 'a)' to item 'aa)'.
- (c) Clauses 5.2 and 17 aa) d) have been reworded.
- (d) The requirements of Clauses 5.1, 14.3, 14.4, 14.5 and 20.4 have been incorporated in the General Standard, therefore these clauses have been deleted.

The clauses of this Particular Standard supplement or modify the corresponding clauses in AS 3200.1—1990. As stated in Clause 3 of AS 3200.1—1990, the requirements of a Particular Standard take priority, where appropriate, over those of AS 3200.1—1990. Where the reference in the text of this Standard indicates an 'amendment', 'addition' or 'replacement' of the relevant requirements, tests or explanatory notes of AS 3200.1—1990, these changes are made to the relevant text which then becomes part of the Standard.

Subclauses and figures which are additional to the General Standard are numbered starting from 101; additional appendices are indicated by a sequence of double capitals e.g., AA), BB), ... and additional items are denoted by a (aa), bb), ... sequence.

In some parts of the series Australian requirements differ from the IEC specifications. To accommodate Australian conditions and well-accepted safety practices, a marginal bar is placed alongside the IEC text, and Appendix ZZ details the Australian specifications and rationale for the deviations.

In the text of this Standard, the following print types are used:

- Requirements, compliance with which can be tested and definitions: . . . . in roman type.
- Explanations, advice, introductions, general statements, exceptions and references . . . . . in smaller roman type.
- Headings of subclauses and test specifications . . . . . *in italic type*.
- Terms used throughout the Standard, which have been defined in Clause 2 and which are also in the index . . . . . IN SMALL CAPITALS.
- An asterisk is placed before each clause for which rationale is included in Appendix AA.

This Standard requires reference to International Standards which are detailed in Appendix LL.

Under arrangements made between Australia and the International Standards bodies, ISO and IEC, as well as certain other Standards organizations, users of this Australian Standard are advised that copyright is vested in Standards Australia.

## CONTENTS

Clause	<i>Page</i>
<b>SECTION ONE - GENERAL</b>	
1. Scope and object .....	4
2. Terminology and definitions .....	4
3. General requirements .....	5
4. General requirements for tests .....	5
5. Classification .....	5
6. Identification, marking and documents .....	5
7. Power output .....	5
<b>SECTION TWO - ENVIRONMENTAL CONDITIONS</b>	
<b>SECTION THREE - PROTECTION AGAINST ELECTRIC SHOCK HAZARD</b>	
13. General .....	7
14. Requirements related to classification .....	7
15. Limitation of voltage and/or energy .....	7
16. ENCLOSURES and PROTECTIVE COVERS .....	8
17. Separation .....	8
18. Protective earthing, functional earthing and potential equalisation .....	9
19. Continuous LEAKAGE CURRENTS and PATIENT AUXILIARY CURRENTS .....	9
20. Dielectric strength .....	9
<b>SECTION FOUR - PROTECTION AGAINST MECHANICAL HAZARDS</b>	
21. Mechanical strength .....	10
<b>SECTION FIVE - PROTECTION AGAINST HAZARDS FROM UNWANTED OR EXCESSIVE RADIATION</b>	
<b>SECTION SIX - PROTECTION AGAINST HAZARDS OF IGNITION OF FLAMMABLE ANESTHETIC MIXTURES</b>	
<b>SECTION SEVEN - PROTECTION AGAINST EXCESSIVE TEMPERATURES AND OTHER SAFETY HAZARDS</b>	
42. Excessive temperatures .....	11
<b>SECTION EIGHT - ACCURACY OF OPERATING DATA AND PROTECTION AGAINST HAZARDOUS OUTPUT</b>	
<b>SECTION NINE - ABNORMAL OPERATION AND FAULT CONDITIONS; ENVIRONMENTAL TESTS</b>	
<b>SECTION TEN - CONSTRUCTIONAL REQUIREMENTS</b>	
FIGURE 101 - Measurement of the high frequency leakage current through the eyepiece ....	12
FIGURE 102 - Measurement of PATIENT LEAKAGE CURRENT .....	12
APPENDIX AA - Rationale .....	13
APPENDIX LL - References—Publications mentioned in this Standard .....	16
APPENDIX ZZ - Amendments to IEC 601-2-18 for application in Australia .....	16

## STANDARDS AUSTRALIA

## Australian Standard

## Approval and test specification—Medical electrical equipment

## Part 2.18: Particular requirements for safety—Endoscopic equipment

## SECTION ONE — GENERAL

## 1. Scope and object

This clause of the General Standard applies except as follows:

## 1.1 Scope

*Addition:*

This Particular Standard specifies requirements for the safety of ENDOSCOPIC EQUIPMENT and integrated instrumentation used for medical diagnosis and for treatment in body cavities.

Requirements for particular applications of ENDOSCOPES are subdivided and identified as follows:

- a) ENDOSCOPES for direct visual observation, using visible light as the transmitting means;
- b) ENDOSCOPES integrated with GLOWCAUTERY DEVICE or THERMOCAUTERY DEVICE;
- c) ENDOSCOPES for ELECTROSURGERY;
- d) ENDOSCOPES integrated with ELECTROHYDRAULIC LITHOTRIPSE;
- \*e) other specialized ENDOSCOPES.

## 1.2 Object

*Replacement:*

The object of this Particular Standard is the assurance of the safety of the entirety of the ENDOSCOPIC EQUIPMENT together with its ACCESSORIES.

The object is also to ensure component parts of ENDOSCOPIC EQUIPMENT to be tested and certified separately and individually.

*Additional sub-clause:*

## 1.101 References to International Standards

This Particular Standard makes reference to IEC Publications 417G, 536, 601-1, 601-2-2 and 878.

## 2. Terminology and definitions

This clause of the General Standard applies except as follows:

## 2.1 EQUIPMENT parts, auxiliaries and ACCESSORIES

*Additional definitions:*

## 2.1.101 ENDOSCOPIC EQUIPMENT

As MEDICAL ELECTRICAL EQUIPMENT, the combination of an ENDOSCOPE, its ACCESSORIES and its supply unit (see Sub-clause 6.1 p) first dash).

NOTE. — The electrically heated elements of cryoprobes are regarded as ENDOSCOPIC EQUIPMENT.