

Australian Standard<sup>®</sup>

---

**APPROVAL AND TEST  
SPECIFICATION—**

**DENTAL AND MOBILE  
DIAGNOSTIC X-RAY EQUIPMENT**

---

[Title allocated by Defence Cataloguing Authority:  
X-RAY APPARATUS, DENTAL AND MOBILE MEDICAL  
(APPROVAL AND TEST SPECIFICATION)]

This Australian standard was prepared by Committee EL/18/8, Mobile and Dental X-ray Machines. It was approved on behalf of the Council of The Standards Association of Australia on 22 March 1985 and published on 7 June 1985.

---

The following interests are represented on Committee EL/18/8:

- Australasian College of Physical Scientists in Medicine
- Australian Dental Association Incorporated
- Australian Institute of Radiography
- Australian Medical Devices and Diagnostic Association
- Confederation of Australian Industry
- Department of Defence
- Department of Health
- Department of Health, Queensland
- Department of Public Works, New South Wales
- Department of Veterans Affairs
- Diagnostic Imaging and Medical Electronics Association of Australia
- Health Department of Western Australia
- Royal Australasian College of Radiologists
- Statutory Electricity Authorities
- United Dental Hospital of Sydney
- University of Sydney

---

**Review of Australian Standard.** To keep abreast of progress in industry, Australian Standards are subject to periodic review and are kept up to date by the issue of amendments or new editions as necessary. It is important therefore that Standards users ensure that they are in possession of the latest edition, and any amendments thereto.

Full details of all Australian Standards and related publications will be found in the Standards Australia Catalogue of Publications, this information is supplemented each month by the magazine 'The Australian Standard', which subscribing members receive, and which gives details of new publications, new editions and amendments, and of withdrawn Standards.

Suggestions for improvements to Australian Standards, addressed to the head office of Standards Australia, are welcomed. Notification of any inaccuracy or ambiguity found in an Australian Standard should be made without delay in order that the matter may be investigated and appropriate action taken.

---

*This standard was issued in draft form for comment as DR 83210.*

First published (as AS 3201, Part 5) . . . . .	1977
AS 3205 first published . . . . .	1985

## PREFACE

This standard was prepared by the Association's Committee on Dental and Mobile X-ray Machines. It supersedes AS 3201, Part 5—1977, Dental and Mobile Medical X-ray Machines. This standard incorporates requirements for the safety of dental and mobile diagnostic X-ray equipment and is one of a series of approval and test specifications issued by the Association. These approval and test specifications for individual items of electromedical equipment are supplementary to the parent approval and test specification for all electromedical equipment, i.e. AS 3200, Approval and Test Specification for Electromedical Equipment—General Requirements.

The purpose of the series is to specify requirements with which compliance may be necessary to gain approval for the sale and use of electromedical equipment in Australia. Only safety matters and conditions closely allied thereto are covered. In some instances, however, these are more stringent than for most electrical appliances because of the exacting additional requirements necessary to ensure the safety of the patient, and because of the environmental conditions, e.g. high humidity, hazardous location, in which X-ray equipment is liable to be used.

The clauses of this particular standard supplement or modify the corresponding clauses in AS 3200. As stated in Clause 1.2, the requirements of a particular standard take priority, where appropriate, over those of AS 3200. Where the reference in the text of this standard indicates an 'addition' or 'replacement' of the relevant requirements, tests or explanation notes of AS 3200, these changes are made to the relevant text which then becomes part of the standard. Where no changes are necessary the words 'this Section of AS 3200 applies' are used.

The revision of AS 3201, Part 5 was primarily undertaken to accommodate new and modified requirements applicable to dental and mobile diagnostic X-ray equipment and to align the content with the following standards:

AS 2398 Fixed Diagnostic X-ray Equipment—Design Construction and Installation—Safety Requirements

AS XXXX Diagnostic X-ray Facilities—Safe Practices\*

Significant changes from AS 3201, Part 5 include the following:

- (a) The application of the new edition to—
  - (i) dental panoramic X-ray equipment having an intra-oral X-ray source;
  - (ii) dental panoramic X-ray equipment for tomography; and
  - (iii) cephalometric X-ray equipment.
- (b) New requirements for battery-powered mobile X-ray equipment.
- (c) Additional safety requirements for capacitor-discharge X-ray equipment.
- (d) Expanded requirements for protection against the hazards of radiation.

During the preparation of this standard reference was made to documentation published by the following organizations:

International Electrotechnical Commission (IEC)

International Commission on Radiological Protection (ICRP)

National Health and Medical Research Council (NHMRC)

Acknowledgement is made of the assistance received from these sources.

---

\* In course of preparation

## CONTENTS

	<i>Page</i>
SPECIFICATION	
1.1	Scope . . . . . 5
1.2	Application . . . . . 5
1.3	Referenced Documents . . . . . 5
2.2	General Definitions . . . . . 6
3.1	Design and Construction . . . . . 8
3.1.100	Mechanical Requirements . . . . . 8
3.3	Mechanical Stability (Other Than Fixed Equipment) . . . . . 9
3.6	Controls, Instruments and Indicators . . . . . 9
3.6.100	Indication of Exposure Factors for Radiography . . . . . 9
3.7	Enclosure and Location of Batteries . . . . . 11
3.9	Suspended Masses . . . . . 11
4.6.2	Supply Switch . . . . . 11
4.14	Transformers . . . . . 12
5	Protection Against Electric Shock Hazards . . . . . 13
6	Protection Against Heat and Fire . . . . . 14
7	Equipment Marking and Accompanying Documents . . . . . 15
7.1	Marking on Equipment . . . . . 15
7.1.100	Protective Devices . . . . . 15
8	Protection Against Radiation . . . . . 16
8.1	X-ray Tube, X-ray Tube Housing, X-ray Tube Assembly and Collimator Devices . . . . . 16
8.2	Leakage Radiation from Capacitor-discharge Equipment . . . . . 16
8.3	Beam Quality . . . . . 16
8.4	Field Limitations of Dental X-ray Equipment . . . . . 16
8.5	Mobile or Portable Diagnostic X-ray Equipment . . . . . 17
APPENDIX A.	TEST TO BE APPLIED FOR TYPE APPROVAL . . . . . 20

## © Copyright — STANDARDS AUSTRALIA

Users of Standards are reminded that copyright subsists in all Standards Australia publications and software. Except where the Copyright Act allows and except where provided for below no publications or software produced by Standards Australia may be reproduced, stored in a retrieval system in any form or transmitted by any means without prior permission in writing from Standards Australia. Permission may be conditional on an appropriate royalty payment. Requests for permission and information on commercial software royalties should be directed to the head office of Standards Australia.

Standards Australia will permit up to 10 percent of the technical content pages of a Standard to be copied for use exclusively in-house by purchasers of the Standard without payment of a royalty or advice to Standards Australia.

Standards Australia will also permit the inclusion of its copyright material in computer software programs for no royalty payment provided such programs are used exclusively in-house by the creators of the programs.

Care should be taken to ensure that material used is from the current edition of the Standard and that it is updated whenever the Standard is amended or revised. The number and date of the Standard should therefore be clearly identified.

The use of material in print form or in computer software programs to be used commercially, with or without payment, or in commercial contracts is subject to the payment of a royalty. This policy may be varied by Standards Australia at any time.

## STANDARDS ASSOCIATION OF AUSTRALIA

## Australian Standard

APPROVAL AND TEST SPECIFICATION—  
DENTAL AND MOBILE DIAGNOSTIC X-RAY EQUIPMENTSECTION 1. SCOPE, APPLICATION AND REFERENCED  
DOCUMENTS

This Section of AS 3200 is applicable, except as follows:

**ADDITION**

**1.1 SCOPE.** Add the following:

This specification sets out the particular safety requirements for dental and mobile diagnostic X-ray equipment as defined in Clause 2.2.

The specification also applies to—

- (a) X-ray equipment for dental panoramic radiography with an intra-oral X-ray source;
- (b) X-ray equipment for dental panoramic tomography; and
- (c) cephalometric X-ray equipment.

It also specifies the measures to be taken for the provision of protection additional to that incorporated in the X-ray equipment against the hazards of ionizing radiation.

**ADDITION**

**1.2 APPLICATION.** Add the following:

**1.2.1 Compliance with AS 3100, and AS 3200.** This specification shall be read in conjunction with AS 3100 and AS 3200, and the appropriate provisions of these standards shall apply to the construction of dental and mobile diagnostic X-ray equipment and to the installation and safeguarding of parts which normally carry current.

**1.2.2 Special requirements.**

**1.2.2.1 Special requirements of this standard.** X-ray equipment covered by this specification shall be deemed to comply with this specification only if it complies with all the requirements of this specification and passes the tests specified herein.

**1.2.2.2 Requirements of other standards.** Electrical accessories and components incorporated in dental and mobile diagnostic X-ray equipment which are dependent upon for safety shall also comply with the relevant requirements of individual Approval and Test Specifications unless these requirements are varied herein.

**1.2.2.3 Protection against radiation.** Where other Australian standards are issued, which deal with specific requirements for the safety of radiological staff, the patient, and the public, these standards shall supersede the protection requirements of this standard where the requirements are more stringent.

**1.2.2.4 Requirements for hazardous locations.** X-ray equipment suitable for use in areas where flammable medical agents are used including hospitals, medical rooms in industrial premises, dental surgeries, and the like, shall comply with the relevant requirements of AS 1169.

**ADDITION**

**1.3 REFERENCED DOCUMENTS.** The following documents are referred to in this standard:

**1.3.1 Standards**

- AS 1169 Minimizing of Combustion Hazards Arising from the Medical Use of Flammable Anaesthetic Agents
- AS 1556 Steel Wire Ropes (Other than for Mining Purposes)
- AS 2398 Fixed Diagnostic X-ray Equipment—Design, Construction and Installation—Safety Requirements
- AS MB1 SAA Steel Wire Rope Manual
- AS XXXX Diagnostic X-ray Facilities—Safe Practices\*

**1.3.2 Approval and Test Specifications:**

- AS 3100 Definitions and General Requirements for Electrical Materials and Equipment
- AS 3126 Extra-low Voltage Transformers
- AS 3193 Transformer-type Battery Chargers
- AS 3200 Electromedical Equipment—General Requirements
- AS 3208 Transformers in Electromedical Equipment

\* In course of preparation.