

Australian Standard[®]

Waters

**Part 4: Determination of solids—
Gravimetric method**

This Australian Standard was prepared by Committee CH/22, Methods for Examination of Waters. It was approved on behalf of the Council of Standards Australia on 6 July 1990 and published on 10 December 1990.

The following interests are represented on Committee CH/22:

Australian Government Analytical Laboratories
Australian Mining Industry Council
Board of Works, Melbourne
Confederation of Australian Industry
Department of Administrative Services, Australian Construction Services
Department of Conservation, Forests and Lands
Department of Health, N.S.W.
Department of Local Government, Qld
Engineering and Water Supply Department, S.A.
Environment Protection Authority, Vic.
Government Chemical Laboratory, Qld
National Association of Testing Authorities, Australia
Royal Australian Chemical Institute
Rural Water Commission, Vic.
University of Melbourne
University of Sydney
Water Board, Sydney

Review of Australian Standards. To keep abreast of progress in industry, Australian Standards are subject to periodic review and are kept up to date by the issue of amendments or new editions as necessary. It is important therefore that Standards users ensure that they are in possession of the latest edition, and any amendments thereto.

Full details of all Australian Standards and related publications will be found in the Standards Australia Catalogue of Publications; this information is supplemented each month by the magazine 'The Australian Standard', which subscribing members receive, and which gives details of new publications, new editions and amendments, and of withdrawn Standards.

Suggestions for improvements to Australian Standards, addressed to the head office of Standards Australia, are welcomed. Notification of any inaccuracy or ambiguity found in an Australian Standard should be made without delay in order that the matter may be investigated and appropriate action taken.

This Standard was issued in draft form for comment as DR 87151.

Australian Standard[®]

Waters

**Part 4: Determination of solids—
Gravimetric method**

First published as AS 3550.4—1990.

PUBLISHED BY STANDARDS AUSTRALIA
(STANDARDS ASSOCIATION OF AUSTRALIA)
1 THE CRESCENT, HOMEBUSH, NSW 2140

ISBN 0 7262 6483 0

PREFACE

This Standard was prepared by the Standards Australia Committee on the Examination of Waters under the direction of the Chemical Standards Board, as Part 4 in the AS 3550 series of Standards. Other parts in the series are as follows:

- Part 1 *Determination of dissolved sulfide—Spectrophotometric method*
- Part 2 *Determination of carbon dioxide—Alkalimetric titration method*
- Part 3 *Determination of alkalinity—Acidimetric titration method*
- Part 5 *Determination of gross alpha and gross beta activities*
- Part 6 *Determination of filtrable calcium—Flame atomic absorption spectrometric method.*

CONTENTS

	<i>Page</i>
1 SCOPE	3
2 REFERENCED DOCUMENTS	3
3 DEFINITIONS	3
4 PRINCIPLE	3
5 REAGENTS	3
6 APPARATUS	3
7 SAMPLING AND SAMPLES	4
8 PROCEDURE	4
9 CALCULATION AND EXPRESSION OF RESULTS	5
10 PRECISION	6
11 TEST REPORT	6
APPENDIX A ALTERNATIVE METHOD FOR THE DETERMINATION OF SUSPENDED SOLIDS	7

© Copyright — STANDARDS AUSTRALIA

Users of Standards are reminded that copyright subsists in all Standards Australia publications and software. Except where the Copyright Act allows and except where provided for below no publications or software produced by Standards Australia may be reproduced, stored in a retrieval system in any form or transmitted by any means without prior permission in writing from Standards Australia. Permission may be conditional on an appropriate royalty payment. Requests for permission and information on commercial software royalties should be directed to the head office of Standards Australia.

Standards Australia will permit up to 10 percent of the technical content pages of a Standard to be copied for use exclusively in-house by purchasers of the Standard without payment of a royalty or advice to Standards Australia.

Standards Australia will also permit the inclusion of its copyright material in computer software programs for no royalty payment provided such programs are used exclusively in-house by the creators of the programs.

Care should be taken to ensure that material used is from the current edition of the Standard and that it is updated whenever the Standard is amended or revised. The number and date of the Standard should therefore be clearly identified.

The use of material in print form or in computer software programs to be used commercially, with or without payment, or in commercial contracts is subject to the payment of a royalty. This policy may be varied by Standards Australia at any time.

STANDARDS AUSTRALIA

Australian Standard
Waters

Part 4: Determination of solids—Gravimetric method

1 SCOPE This Standard sets out methods for the gravimetric determination of total, fixed, volatile dissolved and suspended solids in waters. These methods are directly applicable to potable, waste and saline waters having concentrations of dissolved and suspended solids up to 40 000 mg/L. Dilution of the filtered sample may be used to extend the limit for dissolved solids. This method may underestimate the amount of suspended solids if their nature alters during the procedure, due to solubility, colloidal stability or changes in physical integrity.

NOTES:

- 1 A method for the determination of suspended solids in samples which are difficult to filter is given in Appendix 1.
- 2 The determination of low concentrations of volatile solids in the presence of high concentrations of fixed solids may be subject to considerable error.

2 REFERENCED DOCUMENTS The following documents are referred to in this Standard:

AS

- 2031 Selection of containers and preservation of water samples for chemical and microbiological analysis
- 2031.1 Part 1: Chemical
- 2162 Code of practice for the use of volumetric glassware
- 2163 Graduated measuring cylinders
- 2850 Chemical analysis — Interlaboratory test programs—For determining precision of analytical method(s)—Guide to the planning and conduct

3 DEFINITIONS

3.1 Total solids—the material, both suspended and dissolved, remaining after evaporation of a sample of water and its subsequent drying at $105 \pm 2^\circ\text{C}$, or the sum of suspended and dissolved solids.

3.2 Fixed total solids—the total solids remaining after heating at $550 \pm 50^\circ\text{C}$.

3.3 Volatile total solids—the total solids that are volatilized by heating at $550 \pm 50^\circ\text{C}$.

3.4 Total dissolved solids—the solids obtained by evaporating the filtrate passing a glass fibre filter with a nominal pore size of $1.2 \mu\text{m}$, and drying at $105 \pm 2^\circ\text{C}$.

NOTE: 'Total dissolved solids' is sometimes referred to as 'the filtrable residue'.

3.5 Fixed dissolved solids—the dissolved solids remaining after heating at $550 \pm 50^\circ\text{C}$.

3.6 Volatile dissolved solids—the dissolved solids that are volatilized by heating at $550 \pm 50^\circ\text{C}$.

3.7 Total suspended solids—the solids retained on a glass fibre filter with a nominal pore size of $1.2 \mu\text{m}$, after drying at $105 \pm 2^\circ\text{C}$.

NOTE: 'Total suspended solids' is sometimes referred to as the 'non-filtrable residue'.

3.8 Fixed suspended solids—the suspended solids remaining after heating at $550 \pm 50^\circ\text{C}$.

3.9 Volatile suspended solids—the suspended solids that are volatilized by heating at $550 \pm 50^\circ\text{C}$.

4 PRINCIPLE The various solids as defined in Clause 3 are determined, either gravimetrically or by calculation, as shown in Figures 1 and 2.

5 REAGENTS Unless otherwise specified, distilled water or water of equivalent purity shall be used.

6 APPARATUS

6.1 Glassware—volumetric glassware complying with the relevant Australian Standards and for use in accordance with AS 2162. Measuring cylinders shall comply with AS 2163.

6.2 Evaporating dishes—manufactured from platinum, porcelain or silica.

6.3 Vacuum filtration apparatus

NOTE: Pressure apparatus may be used.

6.4 Filters—Glass fibre filter discs of approximately 2.2 cm diameter and with a nominal pore size of $1.2 \mu\text{m}$.

NOTE: Whatman GF/c or equivalent have been found suitable.