

# Australian Standard®

## Methods for sampling and analysis of ambient air

### Method 9.9: Determination of suspended particulate matter—PM<sub>10</sub> low volume sampler—Gravimetric method

AS 3580.9.9:2017

#### PREFACE

This Standard was prepared by the Joint Standards Australia/Standards New Zealand Committee EV-007, Methods for Examination of Air, to supersede AS/NZS 3580.9.9:2006.

After consultation with stakeholders in both countries, Standards Australia and Standards New Zealand decided to develop this Standard as an Australian Standard rather than an Australian/New Zealand Standard.

The objective of this Standard is to provide regulatory and testing bodies with a standard method for determining suspended particulate matter with an equivalent aerodynamic diameter of less than 10 µm utilizing a low volume sampler and size selective inlet.

This Standard deals with the determination of suspended particulate matter with an equivalent aerodynamic diameter (EAD) of less than approximately 10µm (PM<sub>10</sub>). This is one in a series of Standards for the determination of particulate matter in ambient air.

The procedure described in this Standard involves batch sampling and the gravimetric determination of PM<sub>10</sub>, and is based on the United States Code of Federal Regulations, Title 40, Chapter 1, Part 50 Appendix J, *Reference method for the determination of particulate matter as PM<sub>10</sub> in the atmosphere*.

The term ‘normative’ has been used in this Standard to define the application of the appendix to which it applies. A ‘normative’ appendix is an integral part of a Standard.

#### FOREWORD

Suspended particulate matter measured by this method includes particles with an equivalent aerodynamic diameter (EAD) of less than 10 µm, as passed by a size selective inlet (PM<sub>10</sub>). Particles with EAD of 10 µm and less are classified as respirable and hence may affect health. Particles larger than 10 µm generally have nuisance and aesthetic impacts only. PM<sub>10</sub> emission sources include industrial processes, fuel combustion, burning of vegetation, incineration and natural causes such as wind blown dust and salt laden air.

## METHOD

### 1 SCOPE

This Standard sets out a gravimetric method for determination of PM<sub>10</sub> in ambient air utilizing low volume sequential and non-sequential samplers equipped with size selective inlets. The method provides a measure of mean concentration of PM<sub>10</sub> in units of micrograms per cubic metre ( $\mu\text{g}/\text{m}^3$ ), normally sampled over a 24 h period.

This method may be used to collect particle samples for subsequent physical or chemical analysis.

### 2 REFERENCED DOCUMENTS

The following documents are referred to in this Standard:

#### AS ISO/IEC

17025 General requirements for the competence of testing and calibration laboratories

#### AS/NZS

3580 Methods for sampling and analysis of ambient air

3580.1.1 Part 1.1: Guide to siting air monitoring equipment

3760 In-service safety inspection and testing of electrical equipment

#### ISO/IEC

Guide 98 Uncertainty of measurement

Guide 98-3 Part 3: Guide to the expression of uncertainty in measurement (GUM:1995)

#### EN

12341 Ambient air—Standard gravimetric measurement method for the determination of the PM<sub>10</sub> or PM<sub>2.5</sub> mass concentration of suspended particulate matter.

#### ASTM

E563 Standard practice for preparation and use of an ice point bath as a reference temperature

#### US EPA

US Code of Federal Regulations—Environment Protection Agency 40 CFR, Chapter I, Part 50, Appendix J

US Code of Federal Regulations—Environment Protection Agency 40 CFR, Chapter I, Part 53

### 3 DEFINITIONS

For the purposes of this Standard the definitions below apply.

#### 3.1 Equivalent aerodynamic diameter (EAD)

The diameter of a spherical particle of unit density ( $1 \text{ g}/\text{cm}^3$ ) which exhibits the same aerodynamic behaviour as the particle in question.

#### 3.2 Field blank filter

A filter not used for sampling, which has been taken into the field. The filter is handled as a sample filter, including placing it in the sampler filter holder, and removing it, except the sampler is not in operation during this time. A field blank filter is used to assess the effect of field conditions, transport conditions and field operator technique on the weighing procedure.