

Australian Standard[®]

Sphygmomanometers

[Title allocated by Defence Cataloguing Authority:
SPHYGMOMANOMETERS (Manual or Electronic) NSC 6515]

This Australian Standard was prepared by Committee MD/12, Sphygmomanometers. It was approved on behalf of the Council of Standards Australia on 17 August 1988 and published on 20 March 1989.

The following interests are represented on Committee MD/12:

Australian Medical Association
Australian Medical Devices and Diagnostics Association
Commonwealth and State Departments of Health
Department of Veterans' Affairs
Government Supply Department, N.S.W.
Institution of Biomedical Engineering (Australia)
Institute of Hospital Engineering, Australia
National Heart Foundation of Australia
National Standards Commission
Royal Australasian College of Physicians
Royal Australian College of General Practitioners
Royal Australian Nursing Federation

Additional interests participating in the preparation of this Standard:

Manufacturers
Hospital scientists and engineers

Review of Australian Standards. To keep abreast of progress in industry, Australian Standards are subject to periodic review and are kept up to date by the issue of amendments or new editions as necessary. It is important therefore that Standards users ensure that they are in possession of the latest edition, and any amendments, etc.

Full details of all Australian Standards and related publications will be found in the Standards Australia Catalogue of Publications; this information is supplemented each month by the magazine 'The Australian Standard', which subscribing members receive, and which gives details of new publications, new editions and amendments, and of withdrawn Standards.

Suggestions for improvements to Australian Standards, addressed to the head office of Standards Australia, are welcomed. Notification of any inaccuracy or ambiguity found in an Australian Standard should be made without delay in order that the matter may be investigated and appropriate action taken.

This Standard was issued in draft form for comment as DR 86181.

Australian Standard[®]

Sphygmomanometers

[Title allocated by Defence Cataloguing Authority:
SPHYGMOMANOMETERS (Manual or Electronic) NSC 6515]

First published as AS 3655—1989.

PUBLISHED BY STANDARDS AUSTRALIA
(STANDARDS ASSOCIATION OF AUSTRALIA)
1 THE CRESCENT, HOMEBUSH, NSW 2140

ISBN 0 7262 5471 1

PREFACE

This Standard was prepared by the Association's Committee on Sphygmomanometers, under the direction of the Medical Materials and Equipment Standards Board, in response to a request by the Commonwealth Department of Community Services and Health.

In the preparation of this Standard account as been taken of ANSI/AAMI SP9—1985, *Non-automated sphygmomanometers*, and ANSI/AAMI SP10—1987, *Electronic or automated sphygmomanometers*, and this is acknowledged.

© Copyright — STANDARDS AUSTRALIA

Users of Standards are reminded that copyright subsists in all Standards Australia publications and software. Except where the Copyright Act allows and except where provided for below no publications or software produced by Standards Australia may be reproduced, stored in a retrieval system in any form or transmitted by any means without prior permission in writing from Standards Australia. Permission may be conditional on an appropriate royalty payment. Requests for permission and information on commercial software royalties should be directed to the head office of Standards Australia.

Standards Australia will permit up to 10 percent of the technical content pages of a Standard to be copied for use exclusively in-house by purchasers of the Standard without payment of a royalty or advice to Standards Australia.

Standards Australia will also permit the inclusion of its copyright material in computer software programs for no royalty payment provided such programs are used exclusively in-house by the creators of the programs.

Care should be taken to ensure that material used is from the current edition of the Standard and that it is updated whenever the Standard is amended or revised. The number and date of the Standard should therefore be clearly identified.

The use of material in print form or in computer software programs to be used commercially, with or without payment, or in commercial contracts is subject to the payment of a royalty. This policy may be varied by Standards Australia at any time.

CONTENTS

	<i>Page</i>
FOREWORD	4
1 SCOPE	5
2 APPLICATION	5
3 REFERENCED DOCUMENTS	5
4 DEFINITIONS	5
5 DESCRIPTION OF SPHYGMOMANOMETER	5
6 GENERAL REQUIREMENTS	5
7 SPECIFIC REQUIREMENTS	7
8 ELECTRICALLY POWERED DEVICES	7
9 REQUIREMENTS FOR DEVICES WITH AUTOMATED SYSTOLIC AND DIASTOLIC PRESSURE DETECTION	8
10 REQUIREMENTS FOR SEMI-AUTOMATIC DEVICES	9
11 REQUIREMENTS FOR SPECIFIC TYPES OF DEVICES	9
12 PRODUCT LITERATURE	9
13 MARKING	10
APPENDICES	
A TEST FOR DURABILITY OF INFLATION SYSTEM	11
B TEST FOR STATIC AND DYNAMIC ACCURACY OF PRESSURE SENSING DEVICE	12
C TEST FOR AIR LEAKAGE	14
D TEST FOR EFFECTS OF OVERPRESSURE	15
E TEST FOR PRESSURE RELEASE RATE OF MANUALLY OPERATED VALVES	16
F TEST FOR VENTING	17
G TEST FOR PRESSURE RELEASE RATE OF SELF-BLEEDING VALVES	18
H CYCLIC TESTING OF PRESSURE RELEASE VALVE	19
I INTEGRITY OF INFLATABLE BLADDER AND CUFF	20
J TEST FOR MERCURY SPILLAGE	20
K TEST FOR TUBE DIAMETER	21
L TEST FOR INFLATION SURFACE	21
M METHOD FOR TESTING SPHYGMOMANOMETER (SYSTEM) ACCURACY BY THE INTRA-ARTERIAL COMPARISON METHOD ..	22
N METHOD FOR TESTING SPHYGMOMANOMETER (SYSTEM) ACCURACY BY THE INDIRECT COMPARISON METHOD	27
O BLOOD PRESSURE UNITS	29
COMMENTARY	
RATIONALE FOR SPHYGMOMANOMETER ACCURACY/PER- FORMANCE TESTING REQUIREMENTS	31

FOREWORD

Sphygmomanometers are among the most frequently used clinical measuring devices and are extensively used in monitoring blood pressure and the treatment of blood pressure abnormalities. Additionally, they influence life and death decisions in critical care. Increasingly, sphygmomanometers make use of microcomputers and other novel features and are being utilised by patients themselves at home and in public places with little expert supervision.

This Standard aims to assure that sphygmomanometers are capable of a maintained high order of accuracy in the hands of the average trained observer.

A basic sub-division into manual, semi-automated and automated devices is used in this Standard, and features which appear to cover virtually all types of instrument in current use are dealt with. New features appearing with advancing technology could be readily accommodated because of the modular layout.

It is acknowledged that in clinical medicine blood pressure is usually measured in mmHg, but to encourage the use of SI units (kPa) both units are referred to in this Standard. The Standard requires that blood pressure measuring instruments be marked in 'mmHg', and in 'kPa'. In order to minimize the chance of error, the Standard requires that the units should be inscribed in different colours with equal prominence given to both. For digital displays the two units need not be displayed simultaneously, but they should be clearly differentiated.

While the Standard makes provision for all the commonly used sphygmomanometers, it is possible that the accelerating pace of electronic development will lead to new instruments with features not addressed by this Standard.

STANDARDS AUSTRALIA

Australian Standard

Sphygmomanometers

1 SCOPE. This Standard specifies requirements for manual, semi-automated and automated devices for indirect measurement of arterial blood pressure.

The Standard applies to liquid manometer, mechanical gauge, oscillotonometer or electro-manometer sphygmomanometers used in conjunction with a stethoscope or other methods for the detection and display of pulsations, flow or sounds in connection with the measurement, display or recording of blood pressure by electronic means, regardless of whether they have an automatic inflation system.

Safety aspects of the instruments and their servicing and maintenance are also dealt with.

The Standard does not apply to invasive devices such as those used for the direct measurement of blood pressure, nor does it apply to the non-invasive determination of blood pressure during exercise.

NOTE: Specialized instruments, such as the population of new ambulatory blood pressure recorders, the London School of Hygiene (LSH), or the random zero machines (RZM) and devices based on the vascular unloading principle combined with plethysmography may require additional consideration, as these research-based instruments are not completely covered by this Standard.

2 APPLICATION. The requirements of this Standard may be read in conjunction with, but do not take precedence over, any Statutory Regulations that may apply in any area.

3 REFERENCED DOCUMENTS. The following documents are referred to in this Standard:

AS

2901 Medical devices—Characteristics of audible and visible alarm signals

3200 Electromedical equipment—General requirements

4 DEFINITIONS. For the purpose of this Standard, the definitions below apply:

4.1 Sphygmomanometer—an instrument for the indirect (non-invasive) measurement of arterial blood pressure.

NOTE: Such an instrument would typically include a manometer or other pressure sensing device, inflation system and occluding cuff with bladder that is used in conjunction with a method of detection of Korotkoff sounds, pulsation or arterial wall movement.

4.2 Cuff—that component of the sphygmomanometer which in operation is applied to a limb of the subject whose arterial blood pressure is to be measured. The cuff consists of an inflatable section which may be integral to it or be a separate bladder enclosed by an elastic sleeve with fastening mechanism.

NOTE: The definition for the term 'Cuff', refers generally and inclusively to that part of the device which is applied to a limb. This meaning can be adopted effectively by also describing the component parts of the 'cuff', e.g. the bladder and sleeve. 'Bladder' and 'sleeve' do not require definition since these terms are self descriptive of the cuff components to which they are applied, particularly when the adjectives 'inflatable' and 'inelastic' are used.

4.3 Manual device—an aneroid, mercury gravity or electronic transducer/display sphygmomanometer used in conjunction with a stethoscope or with other manual methods for detecting Korotkoff sounds, arterial wall movement, pulsation or blood flow.

4.4 Semi-automated device—a sphygmomanometer with an in-built method of detection of Korotkoff sounds, arterial wall movement, pulsation, or blood flow which may alert the user, typically by sound or visual indication when to take a reading of its pressure display. The device may have either an automatic cuff inflation and deflation system or an automatic detection system.

4.5 Fully automated device—a sphygmomanometer with an in-built method of detection of Korotkoff sounds, arterial wall movement, pulsation, or blood flow and an automatic inflation and deflation system which, once activated, measures and displays blood pressure without further user intervention.

4.6 Audible alarm—any alert tone generation intended to attract attention.

5 DESCRIPTION OF SPHYGMOMANOMETER.

The sphygmomanometer may contain some or all of the following components:

- (a) Manometer (tube, reservoir, trap, scale plate).
- (b) Valves.
- (c) Cuff.
- (d) Tubing.
- (e) Inflating and deflating system.
- (f) Transducer(s).
- (g) Signal processing.
- (h) Display.
- (i) Hard copy device.
- (j) Alarm system.

6 GENERAL REQUIREMENTS.

6.1 Environmental performance and stability.

6.1.1 Operating conditions. The performance of the device, when set up in its normal operating configuration, shall be such that it complies with Clause 6.2.4, after stabilization for 2 h and whilst maintained at each of the following conditions:

- (a) A temperature of not less than 34°C with a relative humidity of not less than 85 percent.
- (b) Atmospheric pressure in the range 85 kPa to 100 kPa.

NOTE: This is considered equivalent to an altitude of -50 m to 1500 m, relative to sea level.