

Australian Standard™

Measurement of water flow in open channels

**Part 3.3: Velocity-area methods—
Measurement by slope-area method**

[ISO title: Liquid flow measurement in open channels—Slope-area method]



This Australian Standard was prepared by Committee CE-024, Measurement of Water Flow in Open Channels and Closed Conduits. It was approved on behalf of the Council of Standards Australia on 25 September 2000 and published on 12 March 2001.

The following interests are represented on Committee CE-024:

Australian Water and Wastewater Association
Department of Natural Resources, Qld
Institute of Instrumentation and Control Australia
Department of Land and Water Conservation, New South Wales
Department of Public Works and Services, New South Wales
South Australian Water Corporation
Sydney Water Corporation
University of New South Wales
University of Adelaide
University of Technology, Sydney

Keeping Standards up-to-date

Standards are living documents which reflect progress in science, technology and systems. To maintain their currency, all Standards are periodically reviewed, and new editions are published. Between editions, amendments may be issued. Standards may also be withdrawn. It is important that readers assure themselves they are using a current Standard, which should include any amendments which may have been published since the Standard was purchased.

Detailed information about Standards can be found by visiting the Standards Australia website at www.standards.com.au and looking up the relevant Standard in the on-line catalogue.

Alternatively, the printed Catalogue provides information current at 1 January each year, and the monthly magazine, *The Australian Standard*, has a full listing of revisions and amendments published each month.

We also welcome suggestions for improvement in our Standards, and especially encourage readers to notify us immediately of any apparent inaccuracies or ambiguities. Contact us via email at mail@standards.com.au, or write to the Chief Executive, Standards Australia International Ltd, GPO Box 5420, Sydney, NSW 2001.

Australian Standard™

Measurement of water flow in open channels

**Part 3.3: Velocity-area methods—
Measurement by slope-area method**

Originated as AS 3778.3.3—1990.
Second edition 2001.

COPYRIGHT

© Standards Australia International

All rights are reserved. No part of this work may be reproduced or copied in any form or by any means, electronic or mechanical, including photocopying, without the written permission of the publisher.

Published by Standards Australia International Ltd
GPO Box 5420, Sydney, NSW 2001, Australia

ISBN 0 7337 3647 5

PREFACE

This Standard was prepared by the Standards Australia Committee, CE-024, Measurement of Water Flow in Open Channels and Closed Conduits.

The Standard is identical to and is reproduced from ISO 1070:1992, *Liquid flow measurement in open channels—Slope area method*.

This Standard is Part 3.3 of AS 3778, *Measurement of water flow in open channels*, which is published in parts as follows:

AS		
3778		Measurement of water flow in open channels
3778.1	Part 1:	Vocabulary and symbols
3778.2	Part 2:	General
3778.2.1	Part 2.1:	Guidelines for the selection of methods of measurement
3778.2.2	Part 2.2:	Establishment and operation of a gauging station
3778.2.3	Part 2.3:	Determination of the stage-discharge relation
3778.2.4	Part 2.4:	Estimation of uncertainty of a flow-rate measurement
3778.2.5	Part 2.5:	Guidelines for the selection of flow gauging structures
3778.3	Part 3:	Velocity-area method
3778.3.1	Part 3.1:	Measurement by current meters and floats
3778.3.2	Part 3.2:	Measurement by moving boat method
3778.3.3	Part 3.3:	Measurement by slope-area method (this Standard)
3778.3.4	Part 3.4:	Collection and processing of data for determination of errors in measurement
3778.3.5	Part 3.5:	Investigation of total error
3778.3.6	Part 3.6:	Measurement of flow in tidal channels
3778.3.7	Part 3.7:	Measurement by ultrasonic (acoustic) method
3778.3.8	Part 3.8:	Electromagnetic method using a full-channel-width coil
3778.4	Part 4:	Measurement using flow gauging structures
3778.4.1	Part 4.1:	Thin-plate weirs
3778.4.2	Part 4.2:	Rectangular broad-crested weirs
3778.4.3	Part 4.3:	Round-nose horizontal broad-crested weirs
3778.4.4	Part 4.4:	V-shaped broad-crested weirs
3778.4.5	Part 4.5:	Triangular profile weirs
3778.4.6	Part 4.6:	Flat-V weirs
3778.4.7	Part 4.7:	Rectangular, trapezoidal and U-shaped flumes
3778.4.8	Part 4.8:	Trapezoidal profile weirs
3778.4.9	Part 4.9:	Parshall and Saniiri flumes
3778.4.10	Part 4.10:	End-depth method for estimation of flow in rectangular channels with a free overfall
3778.4.11	Part 4.11:	End-depth method for estimation of flow in rectangular channels with a free overfall (approximate method)
3778.5	Part 5:	Dilution method
3778.5.1	Part 5.1:	Constant-rate injection method for the measurement of steady flow
3778.5.2	Part 5.2:	Integration method for the measurement of steady flow
3778.6	Part 6:	Measuring devices, instruments and equipment
3778.6.1	Part 6.1:	Rotating element current-meters
3778.6.2	Part 6.2:	Direct depth sounding and suspension equipment
3778.6.3	Part 6.3:	Calibration of rotating element current meters in straight open tanks
3778.6.4	Part 6.4:	Echo sounders for water depth measurements
3778.6.5	Part 6.5:	Water level measuring devices
3778.6.6	Part 6.6:	Cableway system for stream gauging
3778.6.7	Part 6.7:	Ultrasonic (acoustic) velocity meters
3778.6.8	Part 6.8:	Position fixing equipment for hydrometric boats

Under arrangements made between Standards Australia and the international Standards bodies, ISO and IEC, as well as certain other Standards organisations, users of this Australian Standard are advised of the following:

- (a) Copyright is vested in Standards Australia.
- (b) The number of this Standard is not reproduced on each page, its identity is shown only on the cover and title pages.
- (c) In the source text 'this standard' should read 'this Australian Standard'.
- (d) Where the word 'fluid' appears, it should read 'water'.
- (e) A full point should be substituted for a comma when referring to a decimal marker.

References to International Standards should be replaced by references to equivalent Australian Standards, as follows:

<i>Reference to International Standard</i>		<i>Australian Standard</i>	
ISO		AS	
772	Liquid flow measurement in open channels—Vocabulary and symbols	3778	Measurement of water flow in open channels
1100	Liquid flow measurement in open channels	3778.1	Part 1: Vocabulary and symbols
1100-1	Part 1: Establishment and operation of a gauging station	3778.2.2	Part 2.2: General—Establishment and operation
1100-2	Part 2: Determination of the stage-discharge relation	3778.2.3	Part 2.3: General—Determination of the stage-discharge relation
748	Liquid flow measurement in open channels—Velocity area methods	3778.3.1	Part 3.1: Velocity area methods—Measurement by current-meters and floats

Currently in preview, click buy full version.

AUSTRALIAN STANDARD

Measurement of water flow in open channels

Part 3.3:

Velocity-area methods—Measurement by slope-area method

1 Scope

This International Standard specifies a method of determining liquid flow in open channels from observations of the surface slope and cross-sectional area of the channel. It is suitable for use under somewhat special conditions when direct measurement of discharge by more accurate methods, such as the velocity-area method, is not possible.

The slope-area method can be used with reasonable accuracy in open channels having stable boundaries, bed and sides (e.g. rock or very cohesive clay), in lined channels and in channels with relatively coarse material. It may also be used in alluvial channels, including channels with overbank flow or non-uniform channel cross-sections, but in these cases the method is subject to large uncertainties owing to the selection of the rugosity coefficient (such as Manning's coefficient n or Chezy's coefficient C).

Generally the method may be used to determine discharge

- at the time of determining gauge heights from a series of gauges;
- for a peak flow that left marks on a series of gauges or where peak stages were recorded by a series of gauges;
- for a peak flow that left high-water marks along the stream banks.

This method is not suitable for use in very large channels, channels with very flat surface slopes and high sediment load or channels having significant curvature.

Although the accuracy of the results given by the slope-area method is less than that of the results given by the velocity-area method, the slope-area method is sometimes the only method that can be used for determining the extreme high-stage end of rating curves in cases where the magnitude of

floods is such that other methods of measuring discharge cannot be used.

2 Normative references

The following standards contain provisions which, through reference in this Standard, constitute provisions of this International Standard. At the time of publication, the editions indicated were valid. All standards are subject to revision, and parties to agreements based on this International Standard are encouraged to investigate the possibility of applying the most recent editions of the standards indicated below. Members of IEC and ISO maintain registers of currently valid International Standards.

ISO 772:1988, *Liquid flow measurement in open channels — Vocabulary and symbols*.

ISO 1100-2:1982, *Liquid flow measurement in open channels — Part 2: Determination of the stage-discharge relation*.

ISO 4373:1979, *Measurement of liquid flow in open channels — Water level measuring devices*.

ISO 5168:1978, *Measurement of fluid flow — Estimation of uncertainty of a flow-rate measurement*.

3 Definitions

For the purposes of this International Standard, the definitions given in ISO 772 apply.

4 Principle of the method of measurement

A measuring reach is chosen for which the mean area of the stream or river cross-section is determined and the surface slope of the flowing water in that reach is measured. The mean velocity is then established by using known empirical formulae which relate the velocity to the hydraulic mean depth, and the surface slope is corrected for the kinetic energy of the flowing water and the charac-