

Australian Standard™

**Laboratory glassware—Boiling flasks  
with interchangeable conical ground  
glass joints**

**STANDARDS**  
Australia



This Australian Standard was prepared by Committee CH-001, Laboratory Glassware and Related Apparatus. It was approved on behalf of the Council of Standards Australia on 20 October 2005.

This Standard was published on 23 December 2005.

---

The following are represented on Committee CH-001:

National Association of Testing Authorities  
National Measurement Institute  
Royal Australian Chemical Institute  
Royal College of Pathologists of Australasia  
Science Industry Australia  
University of New South Wales  
University of Sydney

---

#### **Keeping Standards up-to-date**

Standards are living documents which reflect progress in science, technology and systems. To maintain their currency, all Standards are periodically reviewed, and new editions are published. Between editions, amendments may be issued. Standards may also be withdrawn. It is important that readers assure themselves they are using a current Standard, which should include any amendments which may have been published since the Standard was purchased.

Detailed information about Standards can be found by visiting the Standards Web Shop at [www.standards.com.au](http://www.standards.com.au) and looking up the relevant Standard in the on-line catalogue.

Alternatively, the printed Catalogue provides information current at 1 January each year, and the monthly magazine, *The Global Standard*, has a full listing of revisions and amendments published each month.

Australian Standards™ and other products and services developed by Standards Australia are published and distributed under contract by SAI Global, which operates the Standards Web Shop.

We also welcome suggestions for improvement in our Standards, and especially encourage readers to notify us immediately of any apparent inaccuracies or ambiguities. Contact us via email at [mail@standards.org.au](mailto:mail@standards.org.au), or write to the Chief Executive, Standards Australia, GPO Box 476, Sydney, NSW 2001.

---

*This Standard was issued in draft form for comment as DR 05067.*

Australian Standard™

**Laboratory glassware—Rolling flasks  
with interchangeable conical ground  
glass joints**

First published as AS 3812—2005.

**COPYRIGHT**

© Standards Australia

All rights are reserved. No part of this work may be reproduced or copied in any form or by any means, electronic or mechanical, including photocopying, without the written permission of the publisher.

Published by Standards Australia, GPO Box 476, Sydney, NSW 2001, Australia

ISBN 0 7337 7068 1

## PREFACE

This Standard was prepared by the Standards Australia Committee CH-001, Laboratory Glassware and Related Apparatus. This Standard is identical with, and has been reproduced from ISO 4797:2004, *Laboratory glassware—Boiling flasks with conical ground joints*.

The objective of this Standard is to ensure that boiling flasks with conical ground joints meet the requirement of users for laboratory purposes.

As this Standard is reproduced from an International Standard, the following applies:

- (a) Its number does not appear on each page of text and its identity is shown only on the cover and title page.
- (b) In the source text 'this International Standard' should read 'this Australian Standard'.
- (c) A full point should be substituted for a comma when referring to a decimal marker.
- (d) Substitute 'L' for 'l' and 'mL' for 'ml' whenever they appear.
- (e) The following note should be added to the Scope:

**NOTE:** Good laboratory practice demands that the smaller volume conical flasks only be used as storage or reaction vessels where physical stability is required, and not for heating the contents. It is strongly recommended that conical flasks of less than 50 mL capacity NOT be used for boiling liquids, since the shape of the flask is such that, should bumping or superheating of the liquid occur during heating, the contents may be/are likely to be violently expelled.

Reference to International Standards should be replaced by reference to Australian Standards

<i>Reference to International Standard</i>	<i>Australian Standard</i>
ISO	AS
383 Laboratory glassware—Interchangeable conical ground joints	2409 Laboratory glassware—Interchangeable conical ground glass joints

Other ISO documents listed as normative references in Clause 2 have not been adopted as Australian Standards.

This Australian Standard adopts the Series 2 dimensions as given in Tables 1, 2 and 3.