

AS 4362.2—1996

IEC 1180-2:1994

Australian Standard<sup>®</sup>

---

**High-voltage test techniques  
for low-voltage equipment**

**Part 2: Test equipment**

---

This Australian Standard was prepared by Committee EL/7, Power Switchgear. It was approved on behalf of the Council of Standards Australia on 22 November 1995 and published on 5 February 1996.

---

The following interests are represented on Committee EL/7:

- Australian British Chamber of Commerce
  - Australian Chamber of Commerce and Industry
  - Australian Electrical and Electronic Manufacturers Association
  - Electricity Supply Association of Australia
  - Institution of Engineers, Australia
  - Railways of Australia
  - Testing Interests, Australia
  - WorkCover Authority of N.S.W.
- 

**Review of Australian Standards.** To keep abreast of progress in industry, Australian Standards are subject to periodic review and are kept up to date by the issue of amendments or new editions as necessary. It is important therefore that Standards users ensure that they are in possession of the latest edition, and any amendments thereto.

Full details of all Australian Standards and related publications will be found in the Standards Australia Catalogue of Publications; this information is supplemented each month by the magazine 'The Australian Standard', which subscribing members receive, and which gives details of new publications, new editions and amendments, and of withdrawn Standards.

Suggestions for improvements to Australian Standards, addressed to the head office of Standards Australia, are welcomed. Notification of any inaccuracy or ambiguity found in an Australian Standard should be made without delay in order that the matter may be investigated and appropriate action taken.

---

This Standard was issued in draft form for comment as DR 95222.

Australian Standard<sup>®</sup>

---

**High-voltage test techniques  
for low-voltage equipment**

**Part 2: Test equipment**

---

## PREFACE

This Standard was prepared by the Joint Standards Australia/Standards New Zealand Committee EL/7 on Power Switchgear. It is identical with and has been reproduced from IEC 1180-2:1994, *High-voltage test techniques for low-voltage equipment, Part 2: Test equipment*.

It is the result of a consensus among representatives on the Joint Committee to produce it as an Australian Standard.

This Standard is Part 2 of AS 4362, *High-voltage test techniques for low-voltage equipment*, which is published in two Parts as follows:

Part 1: Definitions, test and procedure requirements

Part 2: Test equipment

The purpose of this Standard is to cover test equipment used for high-voltage testing of low-voltage equipment.

The term 'informative' has been used in this Standard to define the application of the annex to which it applies. An 'informative' annex is only for information and guidance.

As this Standard is reproduced from an international Standard, the following applies:

- Its number is shown only on the cover and title page, while the international Standard number appears only on the cover.
- In the source text, 'this international Standard' should read 'this Australian Standard'.
- A full point substitutes for a comma when referring to a decimal marker.

References to International Standards should be replaced by equivalent Australian Standards, as follows:

<i>References to International Standard</i>		<i>Australian Standard</i>	
IEC		AS	
60	High-voltage test techniques	1931	High voltage testing techniques
60-1	Part 1: General definitions and test requirements	1931.1	General definitions, test requirements, test procedures and measuring devices
60-2	Part 2: Measuring systems	1931.2	Application guide for measuring devices
68	Environmental testing	1099	Basic environmental testing procedures for electrotechnology
68-1	Part 1: General and guidance	1099.1	Part 1: General
790	Oscilloscope and peak voltmeters for impulse tests	—	
1083	Digital recorders for measurements in high-voltage impulse tests		
1083-1	Part 1: Requirements for digital recorders	—	
1180	High-voltage test techniques for low-voltage equipment		
1180-1	Part 1: Definitions, test and procedure requirements	—	

## CONTENTS

	<i>Page</i>
1 Scope .....	1
2 Normative references .....	1
3 Definitions .....	2
4 General conditions for verification of test equipment .....	2
4.1 Atmospheric conditions .....	2
4.2 Connections and arrangement .....	2
4.3 Influence of the load .....	3
4.4 Comparison procedure .....	3
4.5 Frequency of verifications .....	3
5 Verification of characteristics of direct voltage generators .....	3
5.1 Ripple factor .....	3
5.2 Permitted measuring error .....	3
5.3 Voltage regulation .....	3
6 Verification of characteristics of alternating voltage generators .....	3
6.1 Voltage wave-form .....	3
6.2 Permitted measuring error .....	4
6.3 Minimum test voltage .....	4
7 Verification of characteristics of impulse voltage generators .....	4
7.1 Voltage wave-form .....	4
7.2 Permitted measuring error .....	4
7.3 Conventional output impedance .....	4
8 Verification of characteristics of impulse current generators .....	5
8.1 Current wave-form .....	5
8.2 Permitted measuring error .....	5
9 Verification of characteristics of hybrid impulse voltage generators .....	5
9.1 Wave-form and permitted measuring error .....	5
9.2 Virtual impedance .....	5
10 Requirements for reference measuring systems .....	6
10.1 Direct voltage .....	6
10.2 Alternating voltage .....	6
10.3 Lightning impulse voltage .....	6
10.4 Impulse current .....	6
10.5 Comparative measurement .....	6
Annex A – Bibliography .....	7

First published as AS 4362.2— 1996.

© Copyright — STANDARDS AUSTRALIA

Users of Standards are reminded that copyright subsists in all Standards Australia publications and software. Except where the Copyright Act allows and except where provided for below no publications or software produced by Standards Australia may be reproduced, stored in a retrieval system in any form or transmitted by any means without prior permission in writing from Standards Australia. Permission may be conditional on an appropriate royalty payment. Requests for permission and information on commercial software royalties should be directed to the head office of Standards Australia.

Standards Australia will permit up to 10 percent of the technical content pages of a Standard to be copied for use exclusively in-house by purchasers of the Standard without payment of a royalty or advice to Standards Australia.

Standards Australia will also permit the inclusion of its copyright material in computer software programs for no royalty payment provided such programs are used exclusively in-house by the creators of the programs.

Care should be taken to ensure that material used is from the current edition of the Standard and that it is updated whenever the Standard is amended or revised. The number and date of the Standard should therefore be clearly identified.

The use of material in print form or in computer software programs to be used commercially, with or without payment, or in commercial contracts is subject to the payment of a royalty. This policy may be varied by Standards Australia at any time.

## AUSTRALIAN STANDARD

**HIGH-VOLTAGE TEST TECHNIQUES FOR  
LOW-VOLTAGE EQUIPMENT –****Part 2: Test equipment****1 Scope**

This part of IEC 1180 is applicable to the test equipment used for dielectric tests on low-voltage equipment. It covers tests with direct, alternating or impulse voltage, impulse current, and tests with a combination of impulse voltage and impulse current. Termination procedures necessary for ensuring that the dielectric tests comply with the voltage, or current, requirements stated in part 1 of this standard in shape and magnitude are stated.

The test equipment comprises a voltage and/or current generator and a measuring system. This standard covers test equipment in which the measuring system is protected against external interference and coupling by appropriate screening, for example a continuous conducting shield. Therefore, simple comparison tests are sufficient to ensure valid results.

Test equipment having measuring systems composed of non-screened components and/or connected by long leads is not covered in this standard. In this case guidance can be obtained from IEC 60-2 keeping in mind the less stringent requirements of this standard.

**2 Normative references**

The following normative documents contain provisions which, through reference in this text, constitute provisions of this part of IEC 1180. At the time of publication of this standard, the editions indicated were valid. All normative documents are subject to revision, and parties to agreements based on this part of IEC 1180 are encouraged to investigate the possibility of applying the most recent editions of the normative documents indicated below. Members of IEC and ISO maintain registers of currently valid International Standards.

IEC 68-1:1975, *Environmental testing — Part 1: General and guidance*

IEC 790:1984, *Oscilloscopes and peak voltmeters for impulse test*

IEC 1083-1:1991, *Digital recorders for measurements in high-voltage impulse tests – Part 1: Requirements for digital recorders*

IEC 1180-1:1992, *High-voltage test techniques for low-voltage equipment – Part 1: Definitions, test and procedure requirements*

NOTE – The requirements of IEC 790 and IEC 1083-1 may be reduced because the uncertainty limits of this part of 1180 are less stringent than those in IEC 60-1, for example,  $\pm 5\%$  for peak value ( $\pm 3\%$  in IEC 60-1).