



**Single-use face masks for use in health care**

**STANDARDS**  
Australia

Currently in preview, click buy full version

This Australian Standard® was prepared by Committee HE-013, Surgical Apparel. It was approved on behalf of the Council of Standards Australia on 11 November 2015. This Standard was published on 23 December 2015.

---

The following are represented on Committee HE-013:

- Australasian College for Infection Prevention and Control
  - Australian Chamber of Commerce and Industry
  - Australian College of Operating Room Nurses
  - Australian Dental Association
  - Australian Nursing and Midwifery Federation
  - Medical Technology Association of Australia
  - NSW Health
  - Queensland Health
  - Royal College of Pathologists of Australasia
  - Testing Interests, Australia
  - Therapeutic Goods Administration
- 

This Standard was issued in draft form for comment as DR AS 4381:2015.

Standards Australia wishes to acknowledge the participation of the expert individuals that contributed to the development of this Standard through their representation on the Committee and through the public comment period.

---

#### **Keeping Standards up-to-date**

Australian Standards® are living documents that reflect progress in science, technology and systems. To maintain their currency, all Standards are periodically reviewed, and new editions are published. Between editions, amendments may be issued.

Standards may also be withdrawn. It is important that readers assure themselves they are using a current Standard, which should include any amendments that may have been published since the Standard was published.

Detailed information about Australian Standards, drafts, amendments and new projects can be found by visiting [www.standards.org.au](http://www.standards.org.au)

Standards Australia welcomes suggestions for improvements, and encourages readers to notify us immediately of any apparent inaccuracies or ambiguities. Contact us via email at [mail@standards.org.au](mailto:mail@standards.org.au), or write to Standards Australia, GPO Box 476, Sydney, NSW 2001.

---

Australian Standard<sup>®</sup>

**Single-use face masks for use in health care**

Original as AS 4381—1996.  
Second edition 2002.  
Third edition 2015.

**COPYRIGHT**

© Standards Australia Limited

All rights are reserved. No part of this work may be reproduced or copied in any form or by any means, electronic or mechanical, including photocopying, without the written permission of the publisher, unless otherwise permitted under the Copyright Act 1968.

Published by SAI Global Limited under licence from Standards Australia Limited, GPO Box 476, Sydney, NSW 2001, Australia

ISBN 978 1 76035 363 6

## PREFACE

This Standard was prepared by the Standards Australia Committee HE-013, Surgical Apparel, to supersede AS 4381—2002, *Single-use face masks for use in health care*.

The objective of this revision is to update and harmonize the testing of single-use surgical masks with relevant American Standards (ASTM) and European Norms (EN) as well as to provide a new classification system for the selection of a single-use face mask with a suitable barrier rating for various applications in health care.

Single-use face masks have specific, limited applications and are designed for use in procedures that do not require respiratory protection from the airborne transmission pathway. It is therefore incumbent on users to ensure that appropriate face masks are worn only for their intended purpose, i.e. to provide barrier protection to minimize mucous membrane exposure to infectious microbial droplets.

AS/NZS 1715, *Selection, use and maintenance of respiratory protective equipment* provides advice on the selection of respirators for use when respiratory protection from airborne hazards is required.

This edition differs from the 2002 Standard in the following principal areas:

- (a) There are three defined barrier level ratings for single-use masks each for use in certain defined procedures.
- (b) All testing required for the masks has been aligned with the equivalent ASTM and/or EN test methods. The Particle Filtration Efficiency (PFE) test is no longer required—this created some confusion regarding the type of protection provided and did not improve the relevant performance characteristics of the mask for use in health care.

## CONTENTS

	<i>Page</i>
1 SCOPE.....	4
2 REFERENCED DOCUMENTS .....	4
3 DEFINITIONS .....	4
4 MATERIALS AND CONSTRUCTION.....	-
5 PHYSICAL REQUIREMENTS .....	7
6 PACKAGING .....	7
7 STORAGE AND DISPOSAL .....	7
8 LABELLING .....	7

## STANDARDS AUSTRALIA

**Australian Standard**  
**Single-use face masks for use in health care****1 SCOPE**

This Standard sets out requirements for single-use face masks which are used in health care where it is necessary to keep cross contamination between the health care worker and the patient to a minimum. The masks are intended for use in surgical, medical and dental procedures. They form part of the personal protective equipment (PPE) used to minimize mucous membrane exposure to infectious microbial droplets.

NOTE: For respiratory protection from airborne infectious agents, e.g. tuberculosis (TB) and surgical plume, refer to AS/NZS 1715.

**2 REFERENCED DOCUMENTS**

The following documents are referred to in this Standard:

AS/NZS 1715	Selection, use and maintenance of respiratory protective equipment
AS ISO 13485	Medical devices—Quality management systems—Requirements for regulatory purposes
ISO 22609	Clothing for protection against infectious agents—Medical face masks—Test method for resistance against penetration by synthetic blood (fixed volume, horizontally projected)
ASTM F2101-14	Standard Test Method for Evaluating the Bacterial Filtration Efficiency (BFE) of Medical Face Mask Materials, Using a Biological Aerosol of <i>Staphylococcus aureus</i>
F1862/ F1862M-13	Standard Test Method for Resistance of Medical Face Masks to Penetration by Synthetic Blood (Horizontal Projection of Fixed Volume at a Known Velocity)
EN 14683	Medical face masks—Requirements and test methods

**3 DEFINITIONS**

For the purpose of this Standard, the following definitions apply.

**3.1 Bacterial filtration efficiency**

The efficiency of the mask as a barrier to the passage of aerosolized bacteria.

**3.2 Differential pressure ( $\Delta P$ ) (breathability)**

Air permeability of the mask, measured by determining the pressure differential across the mask.

**3.3 Facial fit**

Ability of the mask to adequately cover the nose and mouth area of the wearer, so as to prevent the inhalation and exhalation of deleterious matter.