

Australian Standard™

**Analysis of soils**

**Part 4: Determination of metals in aqua  
regia extracts of soil by inductively  
coupled plasma-atomic emission  
spectrometry**

This Australian Standard was prepared by Committee EV/9, Sampling and Analysis of Soils and Biota. It was approved on behalf of the Council of Standards Australia on 15 October 1999 and published on 5 December 1999.

---

The following interests are represented on Committee EV/9:

Agriculture Victoria  
Association of Consulting Engineers Australia  
Australian Collaborative Land Evaluation Program (ACLEP)  
Australian Government Analytical Laboratories  
Australian Institute of Environmental Health  
Australian Institute of Medical Scientists  
Australian Society of Soil Science Incorporated  
CSIRO Land and Water  
CSIRO Mathematical and Information Sciences  
Department of Mineral Resources, N.S.W.  
Department of Natural Resources, Qld  
Environment Institute of Australia  
Environment Protection Authority of N.S.W.  
Environment Protection Authority of Victoria  
Minerals Council of Australia  
National Association of Testing Authorities Australia  
Queensland Health Scientific Services  
Royal Australian Chemical Institute  
University of Sydney

---

#### **Keeping Standards up-to-date**

Standards are living documents which reflect progress in science, technology and systems. To maintain their currency, all Standards are periodically reviewed, and new editions are published. Between editions, amendments may be issued. Standards may also be withdrawn. It is important that readers assure themselves they are using a current Standard, which should include any amendments which may have been published since the Standard was purchased.

Detailed information about Standards can be found by visiting the Standards Australia website at [www.standards.com.au](http://www.standards.com.au) and looking up the relevant Standard in the on-line catalogue.

Alternatively, the printed Catalogue provides information current at 1 January each year, and the monthly magazine, *The Australian Standard*, has a full listing of revisions and amendments published each month.

We also welcome suggestions for the improvement in our Standards, and especially encourage readers to notify us immediately of any apparent inaccuracies or ambiguities. Contact us via email at [mail@standards.com.au](mailto:mail@standards.com.au), or write to the Chief Executive, Standards Australia International Ltd, PO Box 1055, Strathfield, NSW 2135.

---

Australian Standard™

## Analysis of soils

### Part 4: Determination of metals in aqua regia extracts of soil by inductively coupled plasma-atomic emission spectrometry

First published as AS 4479.4—1999.

#### **COPYRIGHT**

© Standards Australia International

All rights are reserved. No part of this work may be reproduced or copied in any form or by any means, electronic or mechanical, including photocopying, without the written permission of the publisher.

Published by Standards Australia International Ltd  
PO Box 1055, Strathfield, NSW 2135, Australia

ISBN 0 7337 3035 3

## PREFACE

This Standard was prepared by the Joint Standards Australia/Standards New Zealand Committee EV/9, Sampling and Analysis of Soils and Biota. This Standard is the result of a consensus among the representatives of the Joint Committee to produce it as an Australian Standard.

This Standard is to provide an inductively coupled plasma atomic emission spectrometric method for the determination of metals in aqua regia extracts of soil. It should be read in conjunction with AS 3641.2, *Recommended practice for atomic emission spectrometric analysis, Part 2: Inductively coupled plasma excitation*.

This Standard is Part 4 of a series comprising:

AS

4479 Analysis of soils

4479.1 Part 1: Pretreatment of potentially contaminated soil samples for heavy metal and metalloid analysis

4479.2 Part 2: Extraction of heavy metals and metalloids from soil by aqua regia—Hotplate digestion method

4479.3 Part 3: Determination of metals in aqua regia extracts of soil by flame atomic absorption spectrometry

4479.4 Part 4: Determination of metals in aqua regia extracts of soil by inductively coupled plasma-atomic emission spectrometry

It has been assumed that executing the provisions of this document is restricted to appropriately qualified chemists, familiar with the use of atomic emission spectrometers. This Standard calls for the use of procedures that may be hazardous or injurious to health, if adequate provisions are not taken.

---

## CONTENTS

	<i>Page</i>
1 SCOPE .....	3
2 REFERENCED DOCUMENTS .....	3
3 DEFINITIONS .....	3
4 PRINCIPLE .....	4
5 APPARATUS .....	4
6 REAGENTS .....	4
7 PROCEDURE .....	6
8 CALCULATION OF RESULTS .....	7
9 QUALITY CONTROL .....	7
10 PRECISION .....	8
11 TEST REPORT .....	8

## STANDARDS AUSTRALIA

## Australian Standard

## Analysis of soils

## Part 4: Determination of metals in aqua regia extracts of soil by inductively coupled plasma-atomic emission spectrometry

**1 SCOPE** This Standard specifies an inductively coupled plasma-atomic emission spectrometric (ICPAES) method for the determination of one or more metals in aqua regia extracts of soils or related materials. Working solution concentration range is dependent upon instrument and operating conditions.

NOTE: This is an intricate procedure and should be performed only by an analyst familiar with the chemistry of the technique and the operational procedures of the instrument used.

**2 REFERENCED DOCUMENTS** The following documents are referred to in this Standard:

## AS

- 2162 Verification and use of volumetric apparatus
- 2162.1 Part 1: General—Volumetric glassware
- 2162.2 Part 2: Guide to the use of piston-operated volumetric apparatus (POVA)
- 2164 Laboratory glassware—One-mark volumetric flasks
- 2166 One-mark pipettes
- 2167 Graduated straight pipettes
- 2850 Chemical analysis—Interlaboratory test programs—For determining precision of analytical method(s)—Guide to the planning and conduct
- 3641 Recommended practice for atomic emission spectrometric analysis
- 3641.1 Part 1: Principles and techniques
- 3641.2 Part 2: Inductively coupled plasma excitation
- 4479 Analysis of soils
- 4479.2 Part 2: Extraction of heavy metals and metalloids from soil by aqua regia—Hotplate digestion method

## AS/NZS

- 2243 Safety in laboratories
- 2243.1 Part 1: General
- 2243.2 Part 2: Chemical aspects

## ISO

- 659 Water for analytical laboratory use—Specification and test methods

**3 DEFINITIONS** For the purpose of this Standard, the definitions below apply.

- 3.1 Accuracy**—the closeness of agreement between the test result and the true value.
- 3.2 Precision**—the closeness of agreement between replicate test results obtained by applying the procedure under prescribed conditions.
- 3.3 Repeatability conditions**—conditions where mutually independent test results are obtained on identical test material with the same method in the same laboratory by the same operator using the same equipment and within short intervals of time.