

AS 4549.1—1999
ISO 13321:1996
Reconfirmed 2016

Australian Standard™

Particle size analysis

**Part 1: Photon correlation
spectroscopy**

This Australian Standard was prepared by Committee CH/32, Particle Size Analysis. It was approved on behalf of the Council of Standards Australia on 1 March 1999 and published on 5 June 1999.

The following interests are represented on Committee CH/32:

Australian Nuclear Science and Technology Organisation
Australian Pre-Mixed Concrete Association
CSIRO Land and Water
The Royal Australian Chemical Institute
Scientific Suppliers Association of Australia
University of South Australia
Queensland University Technology

Review of Australian Standards. To keep abreast of progress in industry, Australian Standards are subject to periodic review and are kept up to date by the issue of amendments or new editions as necessary. It is important therefore that Standards users ensure that they are in possession of the latest edition, and any amendments thereto.

Full details of all Australian Standards and related publications will be found in the Standards Australia Catalogue of Publications. This information is supplemented each month by the magazine 'The Australian Standard', which subscribing members receive, and which gives details of new publications, new editions and amendments, and of withdrawn Standards.

Suggestions for improvements to Australian Standards, addressed to the head office of Standards Australia, are welcomed. Notification of any inaccuracy or ambiguity found in an Australian Standard should be made without delay in order that the matter may be investigated and appropriate action taken.

This Standard was issued in draft form for comment as DR 98550.

STANDARDS AUSTRALIA

RECONFIRMATION

OF

AS 4549.1—1999

Particle size analysis

Part 1: Photon correlation spectroscopy

RECONFIRMATION NOTICE

Technical Committee NT-001 has reviewed the content of this publication and in accordance with Standards Australia procedures for reconfirmation, it has been determined that the publication is still valid and does not require change.

Certain documents referenced in the publication may have been amended since the original date of publication. Users are advised to ensure that they are using the latest versions of such documents as appropriate, unless advised otherwise in this Reconfirmation Notice.

Approved for reconfirmation in accordance with Standards Australia procedures for reconfirmation on 9 September 2016.

The following are represented on Technical Committee NT-001:

Accord Australasia
Australasian College of Toxicology & Risk Assessment
Australian Industry Group
Australian Solar Council
Consumers Federation of Australia
CSIRO
Engineers Australia
National Association of Testing Authorities Australia
National Industrial Chemical Notification and Assessment Scheme
National Measurement Institute
Safe Work Australia
Safety Institute of Australia
University of Adelaide

NOTES

Currently in preview, click buy full vers.

Australian Standard™

Particle size analysis

**Part 1: Photon correlation
spectroscopy**

First published as AS 4549.1—1999.

PREFACE

This Standard was prepared by the Standards Australia Committee CH/32, Particle Size Analysis. This Standard is identical with and has been reproduced from ISO 13321:1996, *Particle size analysis—Photon correlation spectroscopy*.

The objective of this Standard is to describe the application of photon correlation spectroscopy (PCS) to the measurement of an average particle size and a measure of the broadness of the size distribution of particles dispersed in liquids.

As this Standard is reproduced from an International Standard, the following applies:

- (a) Its number appears on the cover and title page while the International Standard number appears only on the cover.
- (b) In the source text, 'this International Standard' should read 'this Australian Standard'.
- (c) A full point substitutes for a comma when referring to a decimal marker.
- (d) Substitute 'mL' for 'ml' wherever it appears.

The terms 'normative' and 'informative' have been used in this Standard to define the application of the annex to which they apply. A 'normative' annex is an integral part of a Standard, whereas an 'informative' annex is only for information and guidance.

© Copyright – STANDARDS AUSTRALIA

Users of Standards are reminded that copyright subsists in all Standards Australia publications and software. Except where the Copyright Act allows and except where provided for below no publications or software produced by Standards Australia may be reproduced, stored in a retrieval system in any form or transmitted by any means without prior permission in writing from Standards Australia. Permission may be conditional on an appropriate royalty payment. Requests for permission and information on commercial software royalties should be directed to the head office of Standards Australia.

Standards Australia will permit up to 10 percent of the technical content pages of a Standard to be copied for use exclusively in-house by purchasers of the Standard without payment of a royalty or advice to Standards Australia.

Standards Australia will also permit the inclusion of its copyright material in computer software programs for no royalty payment provided such programs are used exclusively in-house by the creators of the programs.

Care should be taken to ensure that material used is from the current edition of the Standard and that it is updated whenever the Standard is amended or revised. The number and date of the Standard should therefore be clearly identified.

The use of material in print form or in computer software programs to be used commercially, with or without payment, or in commercial contracts is subject to the payment of a royalty. This policy may be varied by Standards Australia at any time.

CONTENTS

	<i>Page</i>
Introduction	iv
1 Scope	1
2 Definitions	1
3 Symbols	1
4 Principle	2
5 Apparatus	2
6 Preliminary procedures	2
7 Measurement procedure	3
8 Calibration and validation	4
9 Repeatability	4
10 Test report	4
Annexes	
A Computation of the average particle diameter and polydispersity index	6
B Recommended test report form	8
C Theoretical background	10
D Typical PCS instruments and recommended specifications	13
E Recommendations for sample preparation	15
F Bibliography	20

INTRODUCTION

Particle sizing in the submicrometer size range is nowadays performed on a routine basis using photon correlation spectroscopy (PCS). The success of the technique is mainly based on the facts that it provides estimates of average particle size in measuring times of a few minutes and that user-friendly commercial equipment is available. Nevertheless, proper use of the instrument and interpretation of the results require certain precautions. Therefore there is a need for an International Standard for the determination of particle size by photon correlation spectroscopy, in order to provide a methodology that allows the users to obtain good interlaboratory agreement on accuracy and reproducibility.

Although PCS allows the determination of particle size distribution, this International Standard is limited to the description of size distribution by means of only two parameters: an average size and a polydispersity index, as obtained by so-called cumulants analysis (see Annex A). This does not exclude more detailed information about particle size distributions being obtained. However, the reproducibility and reliability of the method of calculation for full distributions is, at the present state of the art, not good enough to include in an International Standard. Again, this does not exclude determination of acceptable size distributions in particular applications.

This International Standard recommends measurements at a single scattering angle of 90° using a He-Ne laser light source with a wavelength *in vacuo* of 632.8 nm. Since solid state laser sources operating at other wavelengths have become available and may be used in future instruments, this International Standard already includes recommendations for such instruments. Although the procedure given is limited to a single angle measurement, with some instruments additional measurements and valuable additional information can be obtained at other scattering angles or by simultaneous analysis of measurements performed at different angles.

This International Standard uses isotropic spherically shaped particles within a test procedure. Measurement of nonspherical and/or nonisotropic particles can be made by this technique where the size of such particles is reported by a spherical equivalence.

A list of suitable references for further reading is given in Annex F.

Particle size analysis — Photon correlation spectroscopy

1 Scope

This International Standard describes the application of photon correlation spectroscopy (PCS) to the measurement of an average particle size and a measure of the broadness of the size distribution of particles dispersed in liquids. It is applicable to particle sizes ranging from a few nanometres to about 1 μm , or to the onset of sedimentation. In the data analysis procedure (see annexes A and C) it is assumed that the particles are isotropic and spherically shaped.

NOTE 1 — The technique is also known or referred to under other names, e.g. quasi-elastic light scattering (QELS) and dynamic light scattering (DLS).

2 Definitions

For the purposes of this International Standard, the following definitions apply.

2.1 average particle diameter, x_{PC} : Harmonic intensity-averaged particle diameter, as determined by equation (C.10) of annex C.

It is expressed in nanometres (10⁻⁹ m).

2.2 polydispersity index, PI: Dimensionless measure of the broadness of the size distribution, as determined by equation (C.9) of annex C.

2.3 scattering volume, V : Section of the incident laser beam viewed by the collecting or detection optics.

Typical order of magnitude is 10⁻⁶ cm³.

3 Symbols

B value of the intercept of the intensity autocorrelation function [see equation (C.6) of annex C];

B_{max} maximum value of the intercept for a given setting of the detection optics;

c concentration of particulate material, in moles per litre;

$G_2(\tau)$ intensity autocorrelation function;

n refractive index of the dispersion medium;

N_V number of particles in scattering volume V ;

η viscosity of the dispersion medium;

Γ decay rate;

λ_0 laser wavelength *in vacuo* (632,8 nm for He-Ne laser);

ϕ particle volume fraction;

ρ particle density;

θ scattering angle;

μ_2 second cumulant.

4 Principle

A monochromatic and coherent laser light beam illuminates a representative sample for particle size analysis, dispersed at a suitable concentration in a liquid. The light scattered by the particles at an angle (typically 90°) is recorded by a detector whose output is fed to a correlator. The decay of the autocorrelation function of the scattered intensity is interpreted in terms of average particle size and polydispersity index by the so-called cumulants method.

Annex C provides some theoretical background for particle sizing by PCS.