

Australian Standard[®]

Meters for non-urban water supply

**Part 6: Installation and commissioning
of open channel meters**

STANDARDS
Australia



This Australian Standard® was prepared by Committee CE-024, Measurement of Water Flow in Open Channels and Closed Conduits. It was approved on behalf of the Council of Standards Australia on 9 January 2013.
This Standard was published on 21 February 2013.

The following are represented on Committee CE-024:

- Australian Industry Group
 - Australian Water Association
 - Department of Environment and Resource Management, Qld
 - Institute of Instrumentation, Control and Automation Australia
 - Irrigation Australia
 - National Farmers Federation
 - National Irrigators Council
 - National Measurement Institute
 - University of South Australia
 - Water Services Association of Australia
-

This Standard was issued in draft form for comment as Draft AS 4747.6.

Standards Australia wishes to acknowledge the participation of the expert individuals that contributed to the development of this Standard through their representation on the Committee and through the public comment period.

Keeping Standards up-to-date

Australian Standards® are living documents that reflect progress in science, technology and systems. To maintain their currency, all Standards are periodically reviewed, and new editions are published. Between editions, amendments may be issued.

Standards may also be withdrawn. It is important that readers assure themselves they are using a current Standard, which should include any amendments that may have been published since the Standard was published.

Detailed information about Australian Standards, drafts, amendments and new projects can be found by visiting www.standards.org.au

Standards Australia welcomes suggestions for improvements, and encourages readers to notify us immediately of any apparent inaccuracies or ambiguities. Contact us via email at mail@standards.org.au, or write to Standards Australia, GPO Box 476, Sydney, NSW 2001.

Australian Standard[®]

Meters for non-urban water supply

**Part 6: Installation and Commissioning
of open channel meters**

First published as ATS 4747.6—2008.
Revised and redesignated AS 4747.6—2013.

COPYRIGHT

© Standards Australia Limited

All rights are reserved. No part of this work may be reproduced or copied in any form or by any means, electronic or mechanical, including photocopying, without the written permission of the publisher, unless otherwise permitted under the Copyright Act 1968.

Published by SAI Global Limited under licence from Standards Australia Limited, GPO Box 476, Sydney, NSW 2001, Australia

ISBN 978 1 74342 356 1

PREFACE

This Standard was prepared by the Standards Australia Committee CE-024, Measurement of Water Flow in Open Channels and Closed Conduits, to supersede ATS 4747.6—2008. *Meters for non-urban water supply Part 6: Installation and commissioning of open channel meters.*

The objective of the 4747 series of Standards is to provide requirements for the installation and commissioning of irrigation and non-urban water meters to meet the requirements of the National Water Initiative Clause 88.

NMI M 11-1, *Meters Intended for the Metering of Water in Open Channels and Partially Filled Pipes, Part 1: Metrological and Technical Requirements*, specifies the metrological and technical requirements for pattern approval and verification of water meters for compliance with the National Measurement Regulations.

This Standard includes essential requirements that are up to date with current progress in the water meter industry technology.

This document is part of a series of Standards covering the metering of non-urban water supply, as follows:

AS

4747	Meters for non-urban water supply
4747.1	Part 1: Glossary of terms
4747.2	Part 2: Technical requirements for closed conduit meters fully charged
4747.3	Part 3: Technical requirements for open channel meters
4747.5	Part 5: Installation and commissioning of closed conduit meters fully charged
4747.6	Part 6: Installation and commissioning of open channel meters (this Standard)
4747.8	Part 8: In-service compliance for non-urban water meters

The above documents were originally released as Australian Technical Specifications for a period of over two years. Following review and consideration of industry feedback, they have been revised and released as Standards.

Significant changes to the series include the following:

- (a) The method of handling meters that are beyond the limit of what can be practicably pattern-approved has been revised, which applies to AS 4747.2 and AS 4747.3.
- (b) Overlap with the National Measurement Institute (NMI) documents has also been removed, and the NMI documents are to be read in conjunction with the above Standards.

The term 'informative' has been used in this Standard to define the application of the appendix to which it applies. An 'informative' appendix is for information and guidance only.

CONTENTS

	<i>Page</i>
SECTION 1 SCOPE AND GENERAL	
1.1 SCOPE.....	4
1.2 APPLICATION	4
1.3 NORMATIVE REFERENCES	5
1.4 DEFINITIONS.....	5
SECTION 2 GENERAL CRITERIA FOR THE DESIGN OF WATER METER INSTALLATIONS	
2.1 MATERIALS.....	6
2.2 METER DESIGN	6
2.3 INSTALLATION DESIGN—GENERAL	6
2.4 INFORMATION TO BE PROVIDED BY THE MANUFACTURER	6
2.5 ELECTRICALLY POWERED METERS	6
SECTION 3 PARTICULAR INSTALLATION REQUIREMENTS	
3.1 LOCATION.....	7
3.2 INSTALLATION	7
3.3 REMOTE-MOUNTED INDICATING AND OTHER REMOTE-MOUNTED DEVICES	8
3.4 SECURITY OF OPERATION.....	8
3.5 PROVISION FOR IN SITU VERIFICATION BY VOLUMETRIC MEANS	8
SECTION 4 PARTICULAR HYDRAULIC REQUIREMENTS	
4.1 HYDRAULIC DISTURBANCES.....	9
4.2 ELIMINATING HYDRAULIC DISTURBANCES	9
SECTION 5 COMMISSIONING	
5.1 GENERAL.....	10
5.2 METERS THAT DO NOT REQUIRE CONFIGURATION AFTER INSTALLATION	10
5.3 METERS REQUIRING CONFIGURATION AFTER INSTALLATION.....	10
5.4 COMMISSIONING.....	11
APPENDIX A GUIDELINES FOR MANAGING FLOW DISTURBANCE	12
BIBLIOGRAPHY.....	14

STANDARDS AUSTRALIA

Australian Standard
Meters for non-urban water supply

Part 6: Installation and commissioning of open channel meters

SECTION 1 SCOPE AND GENERAL

1.1 SCOPE

This Standard sets out requirements for the installation and commissioning of meters complying with NMI M11-1 and AS 4747.3, for use in non-urban applications in open channel and part-filled pipe applications. The Standard principally covers metering for irrigation water.

Two types of meter installations are covered in this Standard, as follows:

- (a) A 'self-contained unit'—a meter that defines its own geometry, is able to be pattern-approved as a complete water meter, calibrated upon assembly and supplied by a single supplier.
- (b) A 'modular metering system'—a device or group of instruments (which may be supplied as a pattern-approved metering system) that does not define the installation geometry and requires configuration after installation.

NOTES:

- 1 The requirements for these types of meters to demonstrate compliance with this Standard, and hence with the National Framework are set out and illustrated in Section 7 of AS 4747.3.
- 2 Meters with undefined or variable channel are not covered by this Standard.

In this Standard, a meter installation is considered to include—

- (i) the meter, its indicating device and all ancillary equipment; and
- (ii) the upstream and downstream pipe or emplacement that may have a material effect on the meter unit.

The Standard covers meters that operate under the following conditions:

- (A) Open channels.
- (B) Working water temperature range of 0.1°C to 30°C. The maximum admissible temperature is 50°C, which is the limiting condition for the meter.
- (C) Ambient temperature range of –5°C to 55°C.
- (D) Ambient humidity range of 0% to 100% except for remote indicating devices where the range is 0% to 93%.

1.2 APPLICATION

This Standard is intended to be used to meet the requirements of the Australian Water Initiative for metering non-urban water. It shall be read in conjunction with NMI M11-1 and NMI M11-2.

Meters that do not comply with the performance requirements of Section 7 of AS 4747.3 shall comply with this Standard.