

Australian Standard[®]

Transfusion sets for single use

STANDARDS
Australia



This Australian Standard® was prepared by Committee HE-021, Transfusion Equipment. It was approved on behalf of the Council of Standards Australia on 23 June 2006. This Standard was published on 4 October 2006.

The following are represented on Committee HE-021:

- Australian Chamber of Commerce
 - Australian College of Operating Room Nurses
 - Australian Red Cross
 - College of Emergency Nursing Australasia
 - Commonwealth Department of Health and Ageing
 - The Medical Industry Association of Australia
-

This Standard was issued in draft form for comment as DR 6114.

Standards Australia wishes to acknowledge the participation of the expert individuals that contributed to the development of this Standard through their representation on the Committee and through public comment period.

Keeping Standards up-to-date

Australian Standards® are living documents that reflect progress in science, technology and systems. To maintain their currency, all Standards are periodically reviewed, and new editions are published. Between editions, amendments may be issued.

Standards may also be withdrawn. It is important that readers assure themselves they are using the current Standard, which should include any amendments that may have been published since the Standard was published.

Detailed information about Australian Standards, drafts, amendments and new projects can be found by visiting www.standards.org.au

Standards Australia welcomes suggestions for improvements, and encourages readers to notify us immediately of any apparent inaccuracies or ambiguities. Contact us via email at mail@standards.org.au, or write to Standards Australia, GPO Box 476, Sydney, NSW 2001.

Australian Standard[®]

Transfusion sets for single use

Originally a part of AS 2385—1980.
Previous edition part of AS 2385—1990.
Revised in part and redesignated as AS 4885—2006.

COPYRIGHT

© Standards Australia

All rights are reserved. No part of this work may be reproduced or copied in any form or by any means, electronic or mechanical, including photocopying, without the written permission of the publisher.

Published by Standards Australia, GPO Box 476, Sydney, NSW 2001, Australia

ISBN 0 7337 7692 2

PREFACE

This Standard was prepared by the Standards Australia Committee HE-021, Transfusion Equipment, to supersede, in part, AS 2385—1990, *Single-use (sterile) infusion sets for general medical use*.

This Standard is identical with, and has been reproduced from ISO 1135-4:2004, *Transfusion equipment for medical use—Part 4: Transfusion sets for single use*.

The objective of this Standard is to specify requirements for single use transfusion sets for medical use in order to ensure their compatibility with containers for blood and blood components as well as with intravenous equipment. Also, to provide requirements for air inlet devices for use with rigid containers for blood and blood components and guidance on specifications relating to the quality and performance of materials used in infusion sets and to present designations for transfusion set components.

The term ‘normative’ has been used in this Standard to define the application of the annex to which it applies. A ‘normative’ annex is an integral part of a Standard.

As this Standard is reproduced from an international standard, the following applies:

- (a) Its number appears on the cover and title page while the international standard number appears only on the cover.
- (b) In the source text ‘this part of ISO 1135’ should read ‘this Australian Standard’.
- (c) A full point substitutes for a comma when referring to a decimal marker.

References to International Standards should be replaced by references to Australian or Australian/New Zealand Standards, as follows:

<i>Reference to International Standard</i>		<i>Australian Standard</i>	
ISO		AS	
594	Conical fittings with a 6% (Luer) taper for syringes, needles and certain other medical equipment	1600	Medical equipment—Conical fittings with a 6% (Luer) taper for syringes, needles and certain other medical equipment
594-1	Part 1: General requirements	1600.1	Part 1: General requirements
594-2	Part 2: Lock fittings	1600.2	Part 2: Lock fittings
7864	Sterile hypodermic needles for single use	1946	Sterile hypodermic needles for single use

Only international references that have been adopted as Australian Standards have been listed.

CONTENTS

	<i>Page</i>
1 Scope	1
2 Normative references	1
3 General requirements	1
3.1 Nomenclature for components of the transfusion set	1
3.2 Maintenance of sterility	3
3.3 Designation	3
4 Materials	4
5 Physical requirements	4
5.1 Particulate contamination	4
5.2 Leakage	4
5.3 Tensile strength	4
5.4 Closure-piercing device	4
5.5 Air-inlet device	5
5.6 Tubing	5
5.7 Filter for blood and blood components	5
5.8 Drip chamber and drip tube	5
5.9 Flow regulator	5
5.10 Flow rate of blood and blood components	5
5.11 Injection site	6
5.12 Male conical fitting	6
5.13 Protective caps	6
6 Chemical requirements	6
6.1 Reducing (oxidizable) matter	6
6.2 Metal ions	6
6.3 Titration acidity or alkalinity	6
6.4 Residue on evaporation	6
6.5 UV absorption of extract solution	6
7 Biological requirements	7
7.1 General	7
7.2 Sterility	7
7.3 Pyrogenicity	7
7.4 Haemolysis	7
7.5 Toxicity	7
8 Labelling	7
8.1 Unit container	7
8.2 Shelf or multi-unit container	8
9 Packaging	8
Annex A (normative) Physical tests	9
Annex B (normative) Chemical tests	13

	<i>Page</i>
Annex C (normative) Biological tests	15
Bibliography	16

Currently in preview, click buy full version

AUSTRALIAN STANDARD

Transfusion sets for single use

1 Scope

This part of ISO 1135 specifies requirements for single-use transfusion sets for medical use in order to ensure their compatibility with containers for blood and blood components as well as with intravenous equipment.

This part of ISO 1135 also specifies requirements for air-inlet devices for use with rigid containers for blood and blood components.

Secondary aims of this part of ISO 1135 are to provide guidance on specifications relating to the quality and performance of materials used in transfusion sets and to present designations for transfusion set components.

In some countries, the national pharmacopoeia or other national regulations are legally binding and take precedence over this part of ISO 1135.

2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 594-1:1986, *Conical fittings with a 6 % (Luer) taper for syringes, needles and certain other medical equipment — Part 1: General requirements*

ISO 594-2:1998, *Conical fittings with 6 % (Luer) taper for syringes, needles and certain other medical equipment — Part 2: Lock fittings*

ISO 3696:1987, *Water for analytical laboratory use — Specification and test methods*

ISO 7864:1993, *Sterile hypodermic needles for single use*

ISO 14644-1:1999, *Cleanrooms and associated controlled environments — Part 1: Classification of air cleanliness*

ISO 15223, *Medical devices — Symbols to be used with medical device labels, labelling and information to be supplied*

3 General requirements

3.1 Nomenclature for components of the transfusion set

The nomenclature for components of transfusion sets is given in Figure 1. An air-inlet device as shown in Figure 2 is required for use with rigid containers for blood and blood components.