

Australian Standard[®]

Hearing aids

**Part 7: Measurement of the
performance characteristics of hearing
aids for production, supply and delivery
quality assurance purposes**

STANDARDS
Australia



This Australian Standard® was prepared by Committee AV-003, Acoustics—Human Effects. It was approved on behalf of the Council of Standards Australia on 3 November 2006. This Standard was published on 20 February 2007.

The following are represented on Committee AV-003:

- Association of Accredited Certification Bodies
 - Association of Australian Acoustical Consultants
 - Association of Consulting Engineers Australia
 - Audiological Society of Australia
 - Australasian Faculty of Occupational Medicine
 - Australian Acoustical Society
 - Australian Chamber of Commerce and Industry
 - Australian Council of Trade Unions
 - Australian Industry Group
 - Department of Consumer & Employment Protection, WorkSafe Division, WA
 - Department of Labour, NZ
 - National Acoustic Laboratories
 - New South Wales Nurses, Association
 - New Zealand Audiological Society
 - Royal Institute of Naval Architects
 - Safety Institute of Australia
 - The Australian Society of Otolaryngological Head and Neck Surgery
 - Victorian WorkCover Authority
 - WorkCover New South Wales
-

This Standard was issued in draft form for comment as DR 06472.

Standards Australia wishes to acknowledge the participation of the expert individuals that contributed to the development of this Standard through their representation on the Committee and through public comment period.

Keeping Standards up-to-date

Australian Standards® are living documents that reflect progress in science, technology and systems. To maintain their currency, all Standards are periodically reviewed, and new editions are published. Between editions, amendments may be issued.

Standards may also be withdrawn. It is important that readers assure themselves they are using a current Standard, which should include any amendments that may have been published since the Standard was published.

Detailed information about Australian Standards, drafts, amendments and new projects can be found by visiting **www.standards.org.au**

Standards Australia welcomes suggestions for improvements, and encourages readers to notify us immediately of any apparent inaccuracies or ambiguities. Contact us via email at **mail@standards.org.au**, or write to Standards Australia, GPO Box 476, Sydney, NSW 2001.

STANDARDS AUSTRALIA

RECONFIRMATION

OF

AS 60118.7—2007

Hearing aids

**Part 7: Measurement of the performance characteristics of hearing aids for
production, supply and delivery quality assurance purposes**

RECONFIRMATION NOTICE

Technical Committee AV-003 has reviewed the content of this publication and in accordance with Standards Australia procedures for reconfirmation, it has been determined that the publication is still valid and does not require change.

Certain documents referenced in the publication may have been amended since the original date of publication. Users are advised to ensure that they are using the latest versions of such documents as appropriate, unless advised otherwise in this Reconfirmation Notice.

Approved for reconfirmation in accordance with Standards Australia procedures for reconfirmation on 22 August 2019.

The following are represented on Technical Committee AV-003:

Accident Compensation Corporation (New Zealand)
Association of Australasian Acoustical Consultants
Australian Acoustical Society
Australian Chamber of Commerce and Industry
Australian Council of Trade Unions
Australian Hearing
Department of Defence (Australian Government)
Engineers Australia
Ministry of Health (New Zealand)
New South Wales Nurses Association
New Zealand Audiological Society
Worksafe Division, Department of Commerce, Western Australia
WorkSafe Victoria

NOTES

Currently in preview, click buy full vers.

Australian Standard[®]

Hearing aids

Part 7: Measurement of the performance characteristics of hearing aids for production, supply and delivery quality assurance purposes

Originally as part of AS 1088—1971.
Previous edition AS 1087.7—1987.
Revised and redesignated as AS 60118.7—2007.

COPYRIGHT

© Standards Australia

All rights are reserved. No part of this work may be reproduced or copied in any form or by any means, electronic or mechanical, including photocopying, without the written permission of the publisher.

Published by Standards Australia GPO Box 476, Sydney, NSW 2001, Australia

ISBN 0 7337 8039 3

PREFACE

This Standard was prepared by the Joint Standards Australia/Standards New Zealand Committee AV-003, Acoustics—Human Effects, to supersede AS 1088.7—1987, *Hearing aids—Measurement of the performance characteristics of hearing aids for quality inspection for delivery purposes*.

After consultation with stakeholders in both countries, Standards Australia and Standards New Zealand decided to develop this Standard as an Australian Standard rather than an Australian/New Zealand Standard.

The objective of this Standard is to give recommendations for the measurement of the performance characteristics of air-conduction hearing aids of a particular model for the purposes of comparing measured properties with those specified by the manufacturer.

This Standard is identical with, and has been reproduced from IEC 60118-7, Ed. 2.0 (2005), *Electroacoustics—Hearing aids—Part 7: Measurement of the performance characteristics of hearing aids for production, supply and delivery quality assurance purposes*.

As this Standard is reproduced from an International Standard, the following applies:

- (a) Its number does not appear on each page of text and its identity is shown only on the cover and title page.
- (b) In the source text ‘this part of IEC 60118’ should read ‘this Australian Standard’.
- (c) A full point should be substituted for a comma when referring to a decimal marker.

CONTENTS

	<i>Page</i>
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 General conditions	4
4.1 Acoustic test method	4
4.2 Reporting of data	4
5 Nominal characteristics and tolerances	4
6 Test box and test equipment	5
6.1 General	5
6.2 Unwanted stimuli in the test box	5
6.3 Sound source	5
6.4 Acoustic coupler	5
6.5 Measurement system for the measurement of sound pressure level and harmonic distortion in the acoustic coupler	5
7 Test conditions	6
7.1 General	6
7.2 Control of the sound field	6
7.3 Normal operating conditions for a hearing aid	8
8 Measurements, specifications and tolerances	10
8.1 Frequency response curves	10
8.2 Output sound pressure level frequency response curve for an input sound pressure level of 90 dB (OSPL90 frequency response curve)	10
8.3 Full-on acoustic gain response curve	11
8.4 Basic frequency response curve at reference test gain setting	11
8.5 Battery current	13
8.6 Total harmonic distortion	13
8.7 Equivalent input noise	13
8.8 Additional measurements for hearing aids having induction pick-up coil	14
8.9 Additional measurements applying to AGC hearing aids	14
9 Maximum permitted expanded uncertainty of measurements	15
Bibliography	17

Currently in preview, click buy full vers.

STANDARDS AUSTRALIA

Australian Standard**Hearing aids—Part 7: Measurement of the performance characteristics of hearing aids for production, supply and delivery quality assurance purposes**

1 Scope

This part of IEC 60118 gives recommendations for the measurement of the performance characteristics of air-conduction hearing aids of a particular model for production, supply and delivery quality assurance purposes. The manufacturer will normally assign nominal values.

This standard does not relate to mechanical or environmental tests. It should not be used as the basis for the exchange of information about hearing aid characteristics in general, nor is it intended to be used as a predictor for real-ear performance.

NOTE Terms such as "manufacturer" and "purchaser" are used in this standard. These terms may be understood, however, to refer to the supplier and recipient respectively in any arrangement for the supply of hearing aids in which the use of this standard is called for.

Though the number of measurements covered by this standard is limited, it is not intended that all measurements described herein shall be made in every case.

This second edition now specifies performance requirements. Conformance to the specifications in this standard is demonstrated only when the result of a measurement, extended by the actual expanded uncertainty of measurement of the testing laboratory, lies fully within the tolerances specified in this standard extended by the values for U_{\max} given in Table 4.

In case of custom-made in-the-ear instruments, the data supplied by the manufacturer applies only to the particular hearing aid being tested.

2 Normative references

The following reference documents are indispensable for the application of this document. For dated references only, the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

IEC 60318-5, *Electroacoustics – Simulators of human head and ear – Part 5: 2 cm³ coupler for the measurement of hearing aids and earphones coupled to the ear by means of ear inserts*¹⁾

3 Terms and definitions

For the purpose of this document, the following terms and definitions apply:

¹⁾ To be published. IEC 60318-5 is a revision of IEC 60126:1973, *IEC reference coupler for the measurement of hearing aids using earphones coupled to the ear by means of ear inserts*.