



Medical electrical equipment

**Part 2.10: Particular requirements
for the basic safety and essential
performance of nerve and muscle
stimulators (IEC 60601-2-10:2016
(ED. 2.1), MOD)**

STANDARDS
Australia



AS 60601.2.10:2018

This Australian Standard® was prepared by Joint Technical Committee HE-003, Medical Electrical Equipment. It was approved on behalf of the Council of Standards Australia on 11 September 2018.

This Standard was published on 27 September 2018.

The following are represented on Committee HE-003:

- Australasian College of Physical Scientists and Engineers in Medicine
- Australian and New Zealand College of Anaesthetists
- Australian Chamber of Commerce and Industry
- Australian College of Perioperative Nurses
- Australian Radiation Protection and Nuclear Safety Agency
- Australian Society of Anaesthetists
- Certification Body Australia (Certification Interests Australia)
- College of Biomedical Engineering, Engineers Australia
- Department of Defence (Australian Government)
- Engineers Australia
- Medical Technology Association of Australia
- Therapeutic Goods Administration (Australian Government)

This Standard was issued in draft form for comment as DR AS/NZS 60601.2.10:2018.

Keeping Standards up-to-date

Ensure you have the latest versions of our publications and keep up-to-date about Amendments, Rulings, Withdrawals, and new projects by visiting:

www.standards.org.au

www.saiglobal.com (sales and distribution)

ISBN 978 1 76072 181 7



Medical electrical equipment

Part 2.10: Particular requirements for the basic safety and essential performance of nerve and muscle stimulators (IEC 60601-2-10:2016 (ED. 2.1), MOD)

First published as AS 60601.2.10:2018.

COPYRIGHT

© IEC 2018 — All rights reserved
© Standards Australia Limited 2018

All rights are reserved. No part of this work may be reproduced or copied in any form or by any means, electronic or mechanical, including photocopying, without the written permission of the publisher, unless otherwise permitted under the Copyright Act 1968 (Cth).

Published by SAI Global Pty Limited under licence from Standards Australia Limited, GPO Box 476, Sydney, NSW 2001.

Preface

This Standard was prepared by the Australian members of the Joint Standards Australia/Standards New Zealand Committee HE-003, Medical Electrical Equipment.

After consultation with stakeholders in both countries, Standards Australia and Standards New Zealand decided to develop this Standard as an Australian Standard rather than an Australian/New Zealand Standard.

The objective of this Standard is to specify the requirements for the safety and essential performance of nerve and muscle stimulators, for use in the practice of physical medicine. This includes transcutaneous electrical nerve stimulators (TENS) and electrical muscle stimulators (EMS).

This Standard amends and supplements the general Standard IEC 60601-1, *Medical electrical equipment — Part 1: General requirements for basic safety and essential performance*.

This Standard is an adoption with national modifications, and has been reproduced from, IEC 60601-2-10:2016 (ED. 2.1), *Medical electrical equipment — Part 2-10: Particular requirements for the basic safety and essential performance of nerve and muscle stimulators*. Appendix Z7 lists the variations for the application of this Standard in Australia.

As this document has been reproduced from an International Standard, the following applies:

- (a) In the source text “this particular Standard” should read “this Australian Standard”.
- (b) A full point substitutes for a comma when referring to a decimal marker.

Australian or Australian/New Zealand Standards that are identical adoptions of international normative references may be used interchangeably. Refer to the online catalogue for information on specific Standards.

The terms “normative” and “informative” are used in Standards to define the application of the appendices or annexes to which they apply. A “normative” appendix or annex is an integral part of a Standard, whereas an “informative” appendix or annex is only for information and guidance.

NOTES

Currently in preview, click buy full version

CONTENTS

FOREWORD.....	3
INTRODUCTION.....	6
201.1 Scope, object and related standards.....	7
201.2 Normative references.....	8
201.3 Terms and definitions.....	9
201.4 General requirements.....	9
201.5 General requirements for testing of ME EQUIPMENT.....	10
201.6 Classification of ME EQUIPMENT and ME SYSTEMS.....	10
201.7 ME EQUIPMENT identification, marking and documents.....	10
201.8 Protection against electrical HAZARDS from ME EQUIPMENT.....	12
201.9 Protection against MECHANICAL HAZARDS of ME EQUIPMENT and ME SYSTEMS.....	12
201.10 Protection against unwanted and excessive radiation HAZARDS.....	12
201.11 Protection against excessive temperatures and other HAZARDS.....	12
201.12 Accuracy of controls and instruments and protection against hazardous outputs.....	12
201.13 HAZARDOUS SITUATIONS and fault conditions for ME EQUIPMENT.....	14
201.14 PROGRAMMABLE ELECTRICAL MEDICAL SYSTEMS (PEMS).....	14
201.15 Construction of ME EQUIPMENT.....	14
201.16 ME SYSTEMS.....	14
201.17 Electromagnetic compatibility of ME EQUIPMENT and ME SYSTEMS.....	14
202 Electromagnetic disturbances – requirements and tests.....	15
Annexes.....	16
Annex C (informative) Guide to marking and labelling requirements for ME EQUIPMENT and ME SYSTEMS.....	17
Annex AA (informative) Particular guidance and rationale.....	18
Index of defined terms used in this particular standard.....	21
Figure 202.101 – Testing layout.....	16
Table 201.101 – Pulse frequency versus applied current limits.....	14
Table 201.C.101 – Marking on the outside of STIMULATORS or their parts.....	17

INTERNATIONAL ELECTROTECHNICAL COMMISSION

MEDICAL ELECTRICAL EQUIPMENT –

**Part 2-10: Particular requirements for the basic safety
and essential performance of nerve and muscle stimulators**

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as far as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

DISCLAIMER

This Consolidated version is not an official IEC Standard and has been prepared for user convenience. Only the current versions of the standard and its amendment(s) are to be considered the official documents.

This Consolidated version of IEC 60601-2-10 bears the edition number 2.1. It consists of the second edition (2012-06) [documents 62D/1003/FDIS and 62D/1015/RVD] and its amendment 1 (2016-04) [documents 62D/1332/FDIS and 62D/1352/RVD]. The technical content is identical to the base edition and its amendment.

This Final version does not show where the technical content is modified by amendment 1. A separate Redline version with all changes highlighted is available in this publication.

International standard IEC 60601-2-10 has been prepared by IEC subcommittee 62D: Electromedical equipment of IEC technical committee 62: Electrical equipment in medical practice.

This second edition cancels and replaces the first edition, published in 1987 and its Amendment 1 (2001). This edition constitutes a technical revision and is aligned with IEC 60601-1.

This publication has been drafted in accordance with the ISO/IEC Directives, Part 2.

In this standard, the following print types are used:

- Requirements and definitions: roman type.
- *Test specifications: italic type.*
- Informative material appearing outside of tables, such as notes, examples and references: smaller type. Normative text of tables is also in a smaller type.
- TERMS DEFINED IN CLAUSE 3 OF THE GENERAL STANDARD, IN THIS PARTICULAR STANDARD OR AS NOTED: SMALL CAPITALS.

In referring to the structure of this standard, the term

- “clause” means one of the seventeen numbered divisions within the table of contents, inclusive of all subdivisions (e.g. Clause 7 includes subclauses 7.1, 7.2, etc.);
- “subclause” means a numbered subdivision of a clause (e.g. 7.1, 7.2 and 7.2.1 are all subclauses of Clause 7).

References to clauses within this standard are preceded by the term “Clause” followed by the clause number. References to subclauses within this particular standard are by number only.

In this standard, the conjunctive “or” is used as an “inclusive or” so a statement is true if any combination of the conditions is true.

The verbal forms used in this standard conform to usage described in Annex H of the ISO/IEC Directives, Part 2. For the purposes of this standard, the auxiliary verb:

- “shall” means that compliance with a requirement or a test is mandatory for compliance with this standard;
- “should” means that compliance with a requirement or a test is recommended but is not mandatory for compliance with this standard;
- “may” is used to describe a permissible way to achieve compliance with a requirement or test.

An asterisk (*) as the first character of a title or at the beginning of a paragraph or table title indicates that there is guidance or rationale related to that item in Annex AA.

A list of all parts of the IEC 60601 series, published under the general title *Medical electrical equipment*, can be found on the IEC website.

The committee has decided that the contents of the base publication and its amendment will remain unchanged until the stability date indicated on the IEC web site under "<http://webstore.iec.ch>" in the data related to the specific publication. At this date, the publication will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

Currently in preview, click buy full version

INTRODUCTION

The minimum safety requirements specified in this particular standard are considered to provide for a practical degree of safety in the operation of nerve and muscle stimulators.

This particular standard amends and supplements IEC 60601-1, *Medical electrical equipment – Part 1: General requirements for safety and essential performance* hereinafter referred to as the General Standard (see 201.1.4).

The requirements are followed by specifications for the relevant tests.

A "Particular guidance and rationale" section giving some explanatory notes, where appropriate, about the more important requirements is included in Annex AA.

Clauses or subclauses for which there are explanatory notes in Annex AA are marked with an asterisk (*).

It is considered that a knowledge of the reasons for these requirements will not only facilitate the proper application of the standard but will, in due course, expedite any revision necessitated by changes in clinical practice or as a result of developments in technology. However, this annex does not form part of the requirements of this standard.

MEDICAL ELECTRICAL EQUIPMENT –

Part 2-10: Particular requirements for the basic safety and essential performance of nerve and muscle stimulators

201.1 Scope, object and related standards

Clause 1 of the general standard¹⁾ applies, except as follows:

201.1.1 * Scope

Replacement:

This International Standard specifies the requirements for the safety of nerve and muscle STIMULATORS, defined in subclause 201.3.204, for use in the practice of physical medicine, hereinafter referred to as ME EQUIPMENT. This includes transcutaneous electrical nerve STIMULATORS (TENS) and electrical muscle STIMULATORS (EMS).

NOTE A muscle STIMULATOR may also be known as a neuromuscular STIMULATOR.

The following ME EQUIPMENT is excluded:

- ME EQUIPMENT intended to be implanted or to be connected to implanted electrodes;
- ME EQUIPMENT intended for the stimulation of the brain (e.g. electroconvulsive therapy ME EQUIPMENT);
- ME EQUIPMENT intended for neurological research;
- external cardiac pacemakers (see IEC 60601-2-31);
- ME EQUIPMENT intended for averaged evoked potential diagnosis (see IEC 60601-2-40);
- ME EQUIPMENT intended for electromyography (see IEC 60601-2-40);
- ME EQUIPMENT intended for cardiac defibrillation (see IEC 60601-2-4).

201.1.2 Object

Replacement:

The object of this particular standard is to establish particular BASIC SAFETY and ESSENTIAL PERFORMANCE requirements for nerve and muscle STIMULATORS as defined in 201.3.204.

201.1.3 Collateral standards

Addition:

This particular standard refers to those applicable collateral standards that are listed in Clause 2 of the general standard and Clause 201.2 of this particular standard.

IEC 60601-1-2 applies as modified in Clause 202. IEC 60601-1-3 does not apply. All other published collateral standards in the IEC 60601-1 series apply as published.

¹⁾ The general standard is IEC 60601-1, *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance*