



Medical electrical equipment

**Part 1.8: General requirements for basic safety and essential performance—
Collateral Standard: General requirements, tests and guidance for alarm systems in medical electrical equipment and medical electrical systems**

STANDARDS
Australia

Currently in preview. Click buy full version.

This Australian Standard® was prepared by Committee HE-003, Medical Electrical Equipment. It was approved on behalf of the Council of Standards Australia on 10 May 2017. This Standard was published on 8 June 2017.

The following are represented on Committee HE-003:

- Australasian College of Physical Scientists and Engineers in Medicine
 - Australian and New Zealand College of Anaesthetists
 - Australian Chamber of Commerce and Industry
 - Australian Radiation Protection and Nuclear Safety Agency
 - Australian Society of Anaesthetists
 - Certification Body of Australia (Certification Interests Australia)
 - College of Biomedical Engineering Engineers Australia
 - Department of Defence (Australian Government)
 - Engineers Australia
 - Therapeutic Goods Administration
-

This Standard was issued in draft form for comment as **DIS AS IEC 60601.1.8:2017**.

Standards Australia wishes to acknowledge the participation of the expert individuals that contributed to the development of this Standard through their representation on the Committee and through the public comment period.

Keeping Standards up-to-date

Australian Standards® are living documents that reflect progress in science, technology and systems. To remain in their currency, all Standards are periodically reviewed, and new editions are published. Between editions, amendments may be issued.

Standards may also be withdrawn. It is important that readers assure themselves they are using a current Standard, which should include any amendments that may have been published since the Standard was published.

Detailed information about Australian Standards, drafts, amendments and new projects can be found by visiting www.standards.org.au

Standards Australia welcomes suggestions for improvements, and encourages readers to notify us immediately of any apparent inaccuracies or ambiguities. Contact us via email at mail@standards.org.au, or write to Standards Australia, GPO Box 476, Sydney, NSW 2001.

Australian Standard®

Medical electrical equipment

**Part 1.8: General requirements for basic safety and essential performance—
Collateral Standard. General requirements, tests and guidance for alarm systems in medical electrical equipment and medical electrical systems**

Originally as AS/NZS 3200.1.8:2005.
Revised in Australia and redesignated as AS IEC 60601.1.8:2017.

COPYRIGHT

© Standards Australia Limited

All rights are reserved. No part of this work may be reproduced or copied in any form or by any means, electronic or mechanical, including photocopying, without the written permission of the publisher, unless otherwise permitted under the Copyright Act 1968.

Published by SAI Global Limited under licence from Standards Australia Limited, GPO Box 476, Sydney, NSW 2001, Australia

ISBN 978 1 76035 807 5

PREFACE

This Standard was prepared by the Standards Australia Committee HE-003, Medical Electrical Equipment, to supersede AS/NZS 3200.1.8:2005, *Medical electrical equipment, Part 1.8: General requirements for safety—Collateral Standard: General requirements, tests and guidance for alarm systems in medical electrical equipment and medical electrical systems*.

The objective of this Standard is to specify basic safety and essential performance requirements and tests for alarm systems in medical equipment (ME) and ME systems and to provide guidance for their application. This is accomplished by defining alarm categories (priorities) by degree of urgency, consistent alarm signals and consistent control states and their marking for all alarm systems.

This Standard does not specify:

- (a) whether any particular ME equipment or ME system is required to be provided with alarm systems;
- (b) the particular circumstances which initiate an alarm condition;
- (c) the allocation of priorities to a particular alarm condition; or
- (d) the means of generating alarm signals.

This Standard is identical with, and has been reproduced from IEC 60601-1-8:2006+AMD1:2012 CSV (ED. 2.1), *Medical electrical equipment, Part 1-8: General requirements for basic safety and essential performance—Collateral Standard: General requirements, tests and guidance for alarm systems in medical electrical equipment and medical electrical systems*.

As this Standard is reproduced from an International Standard, the following applies:

- (i) In the source text 'this International Standard' should read 'this Australian Standard'.
- (ii) A full point substitutes for a comma when referring to a decimal marker.

None of the normative references in the source document have been adopted as Australian or Australian/New Zealand Standards.

CONTENTS

FOREWORD.....	i
INTRODUCTION TO THE AMENDMENT	i
INTRODUCTION.....	i
1 * Scope, object and related standards	9
1.1 Scope.....	9
1.2 Object	9
1.3 Related standards	9
2 Normative references	10
3 Terms and definitions	10
4 General requirements	15
5 ME EQUIPMENT identification marking and documents	15
5.1 Indicator lights and controls.....	15
5.2 ACCOMPANYING DOCUMENTS	15
6 ALARM SYSTEMS.....	15
6.1 ALARM CONDITION	15
6.2 * Disclosures for INTELLIGENT ALARM SYSTEM	16
6.3 Generation of ALARM SIGNALS	17
6.4 * Disclosure of delays.....	24
6.5 ALARM PRESETS.....	24
6.6 ALARM LIMIT	27
6.7 * ALARM SYSTEM security.....	28
6.8 * ALARM SIGNAL inactivation state	28
6.9 * ALARM RESET	31
6.10 * NON-LATCHING and LATCHING ALARM SIGNALS	31
6.11 * DISTRIBUTED ALARM SYSTEM	32
6.12 * ALARM-CONDITION SYSTEM logging	33
Annex A (informative) General guidance and rationale.....	35
Annex B (informative) Guide to marking and labelling requirements for ME EQUIPMENT and ME SYSTEMS.....	72
Annex C (normative) Symbols on marking.....	75
ANNEX D (informative) Guidance for auditory ALARM SIGNALS	81
ANNEX E (informative) Verbal ALARM SIGNALS.....	83
ANNEX F (normative) * Reserved melodies for ALARM SIGNALS.....	85
Bibliography.....	86
Index of defined terms used in this collateral standard	88
Figure 1 – Illustration of temporal characteristics of auditory ALARM SIGNALS	21
Figure A.1 – Graphical representation of components of ALARM SYSTEM delay	53

Table 1 – ALARM CONDITION priorities Determination of ALARM CONDITIONS and assignment of priorities	16
Table 2 – Characteristics of alarm indicator lights	18
Table 3 – * Characteristics of the BURST of auditory ALARM SIGNALS	20
Table 4 – * Characteristics of the PULSE of auditory ALARM SIGNALS.....	20
Table 5 – ALARM SIGNAL inactivation states.....	30
Table A.1 – Reference interpretation of Table F.1	71
Table A.2 – Reference interpretation of Table F.2.....	71
Table B.1 – Cross-reference of marking.....	72
Table B.2 – Cross-reference of ACCOMPANYING DOCUMENTS	73
Table B.3 – Cross-reference of instructions for use.....	73
Table B.4 – Cross-reference of technical description	74
Table C.1 – Graphical symbols for ALARM SYSTEMS	75
Table C.2 – Alternative ALARM SYSTEM related markings.....	80
Table D.1 – Attributes of perceived urgency.....	81
Table F.1 – * Equipment encoded auditory ALARM SIGNALS categorized by ALARM CONDITION and priority complying with Table 3 and Table 4	85
Table F.2 – * Auditory LOW PRIORITY ALARM SIGNAL complying with Table 3 and Table 4	85

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

This consolidated version of IEC 60601-1-8 consists of the second edition (2006) [documents 62A/519/CDV and 62A/537A/RVC] and its amendment 1 (2012) [documents 62A/824/FDIS and 62A/837/RVD]. It bears the edition number 2.1.

The technical content is therefore identical to the base edition and its amendment and has been prepared for user convenience. A vertical line in the margin shows where the base publication has been modified by amendment 1. Additions and deletions are displayed in red, with deletions being struck through.

International standard IEC 60601-1-8 has been prepared by IEC subcommittee 62A: Common aspects of electrical equipment used in medical practice of IEC technical committee 62: Electrical equipment in medical practice, and ISO subcommittee SC 3: Lung ventilators and related devices of ISO technical committee 121: Anaesthetic and respiratory equipment.

It is published as double logo standard.

IEC 60601-1-8 constitutes a collateral standard to IEC 60601-1: *Medical electrical equipment – Part 1: General requirements for safety and essential performance* hereafter referred to as the general standard.

This edition of IEC 60601-1-8 was revised to structurally align it with the 2005 edition of IEC 60601-1 and to implement the decision of IEC Subcommittee 62 A that the clause numbering structure of collateral standards written to IEC 60601-1:2005 would adhere to the form specified in ISO/IEC Directives, Part 2:2004. The principle technical changes are in clause 4, which now recognizes that there is a general requirement for a risk management process in IEC 60601-1:2005.

This publication has been drafted in accordance with the ISO/IEC Directives, Part 2.

In the 60601 series of publications, collateral standards specify general requirements for safety applicable to:

- a subgroup of MEDICAL ELECTRICAL EQUIPMENT (e.g. radiological equipment); or
- a specific characteristic of all MEDICAL ELECTRICAL EQUIPMENT, not fully addressed in the general standard (e.g. ALARM SYSTEMS).

In this collateral standard, the following print types are used:

- Requirements and definitions: roman type.
- *Test specifications: italic type. In addition, in Annex A text in italics indicates guidance that describes means to achieve the safety objectives of this collateral standard.*
- Informative material appearing outside of tables, such as notes, examples and references: in smaller type. Normative text of tables is also in a smaller type.
- TERMS DEFINED IN CLAUSE 3 OF THE GENERAL STANDARD, IN THIS COLLATERAL STANDARD OR AS NOTED: SMALL CAPITALS.

In referring to the structure of this standard, the term

- “clause” means one of the seventeen numbered divisions within the table of contents, inclusive of all subdivisions (e.g. Clause 6 includes Subclauses 6.1, 6.2, etc.);
- “subclause” means a numbered subdivision of a clause (e.g. 6.1, 6.2 and 6.3.1 are all subclauses of Clause 6).

References to clauses within this standard are preceded by the term “Clause” followed by the clause number. References to subclauses within this standard are by number only.

In this standard, the conjunctive “or” is used as an “inclusive or” so a statement is true if any combination of the conditions is true.

The verbal forms used in this standard conform to usage described in Annex H of the ISO/IEC Directives, Part 2. For the purposes of this standard, the auxiliary verb:

- “shall” means that compliance with a requirement or a test is mandatory for compliance with this standard;
- “should” means that compliance with a requirement or a test is recommended but is not mandatory for compliance with this standard;
- “may” is used to describe a permissible way to achieve compliance with a requirement or test.

Clauses, subclauses and definitions for which a rationale is provided in informative Annex A are marked with an asterisk (*).

A list of all parts of the IEC 60601 series, under the general title: *Medical electrical equipment*, can be found on the IEC website.

The committee has decided that the contents of the base publication and its amendments will remain unchanged until the stability date indicated on the IEC web site under "<http://webstore.iec.ch>" in the data related to the specific publication. At this date, the publication will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

NOTE The attention of National Committees is drawn to the fact that equipment manufacturers and testing organizations may need a transitional period following publication of a new, amended or revised IEC or ISO publication in which to make products in accordance with the new requirements and equip themselves for conducting new or revised tests. It is the recommendation of the committee that the content of this publication be adopted for mandatory implementation nationally not earlier than 3 years from the date of publication.

IMPORTANT – The “colour inside” logo on the cover page of this publication indicates that it contains colours which are considered to be useful for the correct understanding of its contents. Users should therefore print this publication using a colour printer.

INTRODUCTION TO THE AMENDMENT

The second edition of IEC 60601-1-8 was published in 2006. Since its publication, an issue has been identified with respect to pulse and burst testing. In addition, issues have been raised by IEC/62D/MT 22, *Electromedical diagnostic and patient monitoring equipment*, during implementation of alarm system requirements in particular standards within their scope of work.

At the Brussels meeting, IEC/SC 62A accepted a proposal, based on ISO/TC 121/SC 3 Resolution Orebro 6, to develop the 1st amendment to IEC 60601-1-8:2006 to address the issues identified above. IEC/SC 62A – ISO/TC 121/SC 3 Joint Working Group 2, *Alarms*, was reactivated as a maintenance team to develop this amendment.

INTRODUCTION

MEDICAL ELECTRICAL EQUIPMENT and MEDICAL ELECTRICAL SYSTEMS are increasingly used in medical practice. ALARM SIGNALS are frequently used to indicate unsatisfactory physiological PATIENT states, unsatisfactory functional states of the MEDICAL ELECTRICAL EQUIPMENT or MEDICAL ELECTRICAL SYSTEM or to warn the OPERATOR of HAZARD to the PATIENT or OPERATOR due to the MEDICAL ELECTRICAL EQUIPMENT or MEDICAL ELECTRICAL SYSTEM. INFORMATION SIGNALS convey information that is independent of an ALARM CONDITION.

Surveys of healthcare personnel have indicated significant discontent with ALARM SIGNALS. Problems include difficulty in identifying the source of an ALARM SIGNAL, loud and distracting ALARM SIGNALS, and the high incidence of FALSE POSITIVE or NEGATIVE ALARM CONDITIONS [16]¹⁾. Surveys of MANUFACTURERS of medical monitors demonstrated a wide variety of DEFAULT ALARM PRESETS. The leading reason for disabling ALARM SIGNALS is the large number of ALARM SIGNALS associated with FALSE POSITIVE ALARM CONDITIONS. See also bibliography.

Safety of PATIENTS depends on the ability of the OPERATOR to correctly discern the characteristics of ALARM SIGNALS. USABILITY is an important element in the design of ALARM SIGNALS that are readily discernible without being unnecessarily distracting or disturbing. This approach is intended to rationalize the current situation, to reduce confusion by limiting proliferation of ALARM SIGNALS and their control states, and to minimize distraction for other people. This collateral standard was developed with contributions from clinicians, engineers and applied psychologists.

The terminology requirements, general recommendations and guidance of this collateral standard are intended to be useful for MANUFACTURERS of MEDICAL ELECTRICAL EQUIPMENT and MEDICAL ELECTRICAL SYSTEMS and for technical committees responsible for particular standards.

The effectiveness of any ALARM SYSTEM depends critically on its implementation by the RESPONSIBLE ORGANIZATION. It is important that the RESPONSIBLE ORGANIZATION configure the ALARM SYSTEM so that an OPERATOR is not able to compromise it.

À
À
À
À
À
À
À

1) Figures in brackets refer to the bibliography.

AUSTRALIAN STANDARD

Medical electrical equipment

Part 1.8:

General requirements for basic safety and essential performance—
Collateral Standard: General requirements, tests and guidance for alarm systems in medical electrical equipment and medical electrical systems

1 * Scope, object and related standards**1.1 Scope**

This International Standard applies to the BASIC SAFETY and ESSENTIAL PERFORMANCE of MEDICAL ELECTRICAL EQUIPMENT and MEDICAL ELECTRICAL SYSTEMS, hereafter referred to as ME EQUIPMENT and ME SYSTEMS.

This collateral standard specifies requirements for ALARM SYSTEMS and ALARM SIGNALS in ME EQUIPMENT and ME SYSTEMS.

It also provides guidance for the application of ALARM SYSTEMS.

1.2 Object

The object of this collateral standard is to specify BASIC SAFETY and ESSENTIAL PERFORMANCE requirements and tests for ALARM SYSTEMS in ME EQUIPMENT and ME SYSTEMS and to provide guidance for their application. This is accomplished by defining alarm categories (priorities) by degree of urgency, consistent ALARM SIGNALS and consistent control states and their marking for all ALARM SYSTEMS.

This collateral standard does not specify:

- whether any particular ME EQUIPMENT or ME SYSTEM is required to be provided with ALARM SYSTEMS;
- the particular circumstances which initiate an ALARM CONDITION;
- the allocation of priorities to a particular ALARM CONDITION; or
- the means of generating ALARM SIGNALS.

1.3 Related standards**1.3.1 IEC 60601-1**

For ME EQUIPMENT and ME SYSTEMS, this collateral standard complements IEC 60601-1.

When referring to IEC 60601-1 or to this collateral standard, either individually or in combination, the following conventions are used:

- "the general standard" designates IEC 60601-1 alone (latest edition including any amendments);
- "this collateral standard" designates IEC 60601-1-8 alone;
- "this standard" designates the combination of the general standard and this collateral standard.